PROCEEDINGS OF SPIE

SPIDigitalLibrary.org/conference-proceedings-of-spie

In-orbit characterization of the ISRF

Thierry Garnier, Hervé Bénard, Stéphanie Béhar-Lafenêtre, Ralph Snel, Marcela Pelica Páscoa, et al.

Thierry Garnier, Hervé Bénard, Stéphanie Béhar-Lafenêtre, Ralph Snel, Marcela Pelica Páscoa, Niels Dijkhuizen, Bart Speet, Geert Slegtenhorst, Luca Maresi, Arnaud Pasquet, Mathijs Arts, "In-orbit characterization of the ISRF," Proc. SPIE 11852, International Conference on Space Optics — ICSO 2020, 118523L (11 June 2021); doi: 10.1117/12.2599611

SPIE.

Event: International Conference on Space Optics — ICSO 2021, 2021, Online Only

International Conference on Space Optics—ICSO 2020

Virtual Conference

30 March–2 April 2021

Edited by Bruno Cugny, Zoran Sodnik, and Nikos Karafolas

In-orbit characterization of the ISRF

In-orbit characterisation of the ISRF

Thierry Garnier, Hervé Bénard, Stéphanie Béhar-Lafenêtre, Thales Alenia Space France
Ralph Snel, Marcela Pelica Páscoa, Niels Dijkhuizen, Bart Speet, Geert Slegtenhorst, TNO
Luca Maresi, Arnaud Pasquet, ESA
Mathijs Arts, ATG Europe BV

ABSTRACT

Recent Earth Observation instruments require a highly accurate knowledge of their Instrument Spectral Response Function (ISRF). This translates into lengthy and costly characterisation programs during the Assembly, Integration, and Test phase of the instrument. In addition, potential changes of ISRF after launch suggests to use an on-board equipment dedicated to this accurate characterisation in flight. This topic has been studied by Thales Alenia Space and TNO in the frame of a TRP study funded by ESA during years 2019 and 2020.

This paper first recalls the potential application cases and the related target performances of the study. Then we identify conceptual solutions, such as operating tuneable laser diodes, Fabry-Perot interferometer, microresonators. Next, we describe the principle of the selected solution: a Fourier Transform Spectrometer (FTS), used as a calibration light source. Indeed, a single FTS can cover a very large spectral range with an extremely high spectral accuracy. But FTS are known as complex and sensitive devices: we have designed a simplified solution, well adapted to space applications.

In a second step, we describe the flight hardware designed by Thales Alenia Space, targeting optimised manufacturing and implementation. The main performances, being spectral resolution and signal to noise ratio (SNR), are discussed. Breadboard activities also took place in TNO in order to validate the main hypotheses. Innovative data processing has been tested, that only makes use of interferogram data, while maintaining a high accuracy. Description of the breadboard as well as the first test results are presented.

Keywords: ISRF, FTS, Earth Observation, CO2 Instrument

1. INTRODUCTION

Spectrometers for atmospheric chemistry and climate applications, such as TROPOMI, Sentinel 5 or the future Copernicus missions for CO2 monitoring, require an extremely accurate calibration of the Instrument Spectral Response Function (ISRF).

The ISRF, relating the measured spectra with the monochromatic Top-Of-Atmosphere (TOA) radiance, needs to be known across the entire spectral range and Field-of-View (FoV) of the instruments. Several measurements are necessary for the calibration of the ISRF as a function of the wavelength and the field of view. Furthermore, an accurate calibration of the ISRF has to be performed in a thermal vacuum chamber at the operating temperature. Typically, such measurements are performed in the frame of extensive on-ground calibration campaigns lasting several months and involving the complete instrument. Therefore, the calibration of the ISRF of high-performance spectrometer is one of the cost and schedule drivers.

In addition, the ISRF may be slightly modified due to the settling of the optics caused by launch vibrations and shocks and due to the degradation of the instrument throughout its lifetime. Thermo-mechanical deformations can modify the distances between optical elements and lead to a shift of the focal distances, which translate into changes of the ISRF shape. Therefore, an in-flight calibration of the ISRF shape accounting for launch effects, short-term modifications and long-term degradation would be beneficial for maintaining high instrument performance during the mission lifetime. Previous space-borne imaging spectrometers for atmospheric applications (e.g. SCIAMACHY, GOME, OMI) have mainly relied on thorough on-ground calibration with only very limited ISRF monitoring in orbit, combined with mechanical stability of the instrument. This in turn lead to bulky and heavy mechanical designs for the optical bench and mounts, with corresponding impact on mass and volume, which are also regarded as cost driving. A reliable on-board technique for monitoring of the spectral response, ensuring ISRF knowledge compliant with the mission requirement,

could potentially relax the mechanical stability needs and therefore reduce the overall complexity and cost of the instrument.

A method to perform accurate in-orbit calibration of the ISRF is therefore expected to reduce cost and delivery time of future space-borne imaging spectrometers, as well as improving their performance stability by periodically re-calibrating the ISRF in-orbit. A first step in this direction is the inclusion of tuneable laser sources in flight calibration units of recent (TROPOMI) [1] and future (Sentinel-5) imaging spectrometers. However, these are restricted to a very narrow wavelength band defined by the tuneable range (typically < 2 nm) and are not expected to yield the accuracies required over the entire spectral range and FoV. The main objective of these limited devices is the monitoring of potential degradations, rather than a full-scale calibration. The objective of this activity is to explore possibilities for on-board ISRF calibration over the full spectral range and FoV, beyond the current capabilities for degradation monitoring.

After review of the requirements set forth for several instruments, namely OMI/EOS, TROPOMI, Sentinel-5/UVNS, FLEX, CHIME, Copernicus CO2 Monitoring, the requirements for the ISRF calibration system have been enriched and completed. Here are the most relevant ones.

Table 1: ISRF calibration requirements

ISRF-1	The ISRF shall be known with an accuracy of 1% of its maximum over the spectral range Δ where the ISRF is greater than 1% of its maximum. Weight: mandatory.
ISRF-2	The ISRF shall be known with at least 20 sampling points per FWHM over the spectral range Δ .
ISRF-3	The shape of the ISRF shall be sufficiently well known to calculate the FWHM with an accuracy of 1%.
ISRF-4	The shape of the ISRF shall be sufficiently well known to calculate the barycentre with an accuracy of 0.05 SSI (Spectral Sampling Interval).
ISRF-5	The following reference instrument shall be used: <ul style="list-style-type: none"> • spectral range 270 to 2400 nm • spectral resolution 0.5 nm • spectral sampling interval 0.2 nm • spatial dimensions with 500 resolution elements in one dimension, 1 resolution element in the other • the instrument has a radiance and an irradiance mode
ISRF-6	The ISRF shall be known for all pixels that can see direct light (possibly by interpolation).
ISRF-7	The ISRF shall be known with sufficient accuracy to perform acceptance tests on the instrument, a representative subset of pixels is acceptable.
ISRF-8	It shall be possible to measure the orbital variation of the ISRF.

2. POTENTIAL SOLUTIONS TRADE-OFF AND SELECTION

2.1 Overview of potential solutions and pre-selection

Quite a large number of solutions have been identified in the initial proposal. However, most of them have been discarded at the beginning of the study, based on simple practical considerations reminded below.

The use of natural light sources available in space (Earth, Sun, Moon, planets or stars) have the following drawbacks:

- They cannot be used on-ground to validate the device or to perform a first characterisation of the instrument using the in-flight device
 - They may require special manoeuvres to be observed (e.g. for the Moon)
 - They are usually available during short slots only, during which the incidence angle is continuously changing
 - They exhibit a limited number of spectral features (e.g. Fraunhofer lines, atmospheric absorption lines), of fixed wavelength and spectral width. As such, they usually cannot be used to calibrate all spectral pixels of the instrument.
- Therefore, those natural sources have been discarded for an extensive ISRF calibration.

Wide-band or largely tuneable laser and laser-based sources, such as OPO and super-continuum, have also been discarded, as their flight operation often proved delicate and the flight experiences are actually few. Spectral lamps have also been discarded. Although some have been successfully used in flight, they offer a limited number of spectral lines, of fixed wavelength and spectral width, and as such are not suited to a measurement over a large spectral range.

Once the image of a monochromatic line is projected onto the detector, it must be sampled with an adequate step. Several methods can be envisaged for that:

- Relative shift of the detector with respect to the image plane across the spectral direction
- Use of a slanted slit: several contiguous spatial pixels are used to sample the Instrument Line Shape (ILS)
- Over-sample the ILS thanks to pixels smaller than the ISRF FWHM
- Scan of the calibration source wavelength.

The relative shift requires some scan mechanism inside the spectrometer, between the entrance slit and the detector: this is considered as a risk for the performances and requires to be taken into account very early in the instrument design. Similarly, the slanted slit shows some advantages regarding the ISRF measurements but cannot be used for normal observation mode and therefore requires some mechanism to switch between the nominal and the slanted slit. Here again, this is considered too much intrusive in the instrument design. Over-sampling of the ISRF by the detector itself has not been retained here, since the number of available spectral pixel is often limited by the detector technology. This could be a solution for low or medium resolution instruments, but these ones usually do not require an accurate knowledge of the ISRF.

Therefore, the only solution that have been considered for the trade-off is the scan of the calibration source wavelength. This is ensured by the calibration device itself and does not rely on any particularity of the instrument. This is also in line with the conclusion about the light source presented above.

2.2 Final trade-off and selection

Finally, 4 solutions have been considered for the trade-off:

2.2.1 Tuneable Laser diodes (LD)

The use of a tuneable LD for ISRF calibration is straightforward. Taking benefit from the single mode operation of Distributed FeedBack (DFB) LD, and of the sensitivity of the emitted wavelength to the diode temperature and/or supply current, narrow emission lines can be produced, narrower than the resolution of the instrument to be tested. The sweep of the emitted wavelength over a range larger than the ISRF FWHM directly allows the retrieval of the ISRF. This is already implemented in some instruments such as TROPOMI and is planned for Sentinel 5/UVNS and CO2I instruments.

Figure 1: Example performances of a Laser Diode. Left: Typical room temperature cw spectrum.
Right: Typical tuning of a Laser Diode by current and temperature.

2.2.2 Micro-resonators

The micro-resonator-based frequency combs are devices that are still in development. While they are addressed by a number of laboratory researches and related papers, they are not commercially available yet [2].

The main parts of such resonators are illustrated below. A laser diode generates a CW laser line, usually close to 1550nm. This line is further amplified, using an Erbium-Doped Fibre Amplifier, then propagates through one or several micro-cavities in which non-linear Kerr effect occurs and generates several lines by four-wave mixing. Thanks to the resonant properties of the cavity, the output of the micro-cavities is a frequency comb that covers quite a large spectral range. Frequency generation over a full octave have been demonstrated in laboratory [3][4].

Figure 2: Primary building blocks of a micro-resonator

While materials and technology are available for micro-resonator devices in the SWIR range, this is no longer true for VIS wavelengths[5][6], not to mention UV, for which usual materials are not suited (strongly normal Group Velocity Dispersion GVD), and resonator Q-factors degraded by the micro roughness of the cavity.

The delivered power demonstrated in laboratory may reach almost 1mW per comb line for an input power of the order of 10W. This quite high power, combined with the small footprint of hybrid photonics technology, make micro-resonators very attractive for space applications such as in-flight ISRF calibration. However, they have not been further addressed in this study, because still a bit far from maturity.

Figure 3: Example of photonic hybrid package including micro-cavities: here a laser tuneable by steps

2.2.3 Fabry-Perot Etalons (FP)

Fabry-Perot etalons are well known devices, however they are rarely used in space-borne instrument. Tuneable Fabry-Perot etalon are commercially available, usually based on PZT for adjustment of the plates spacing, and on capacitive measurements for the closed-loop control of this spacing.

Because the spectral resolution of the FP and its Free Spectral Range FSR depends upon the spacing of the etalon plates and on their reflectivity, 1 single device cannot cover the whole spectral range required. Moreover, practical considerations such as polishing quality limit their performances: they are not suited to high resolution in the UV/VIS range and would exhibit a resolution close to the instrument to be characterised in SWIR range. Finally, the output flux is low.

2.2.4 Fourier-Transform Spectrometer (FTS)

The FTS can be considered as a way to modulate its input light beam with a frequency that is strictly related to the wavelength. As such, it may provide for a convenient way to measure the ISRF, benefiting from the advantage of measuring all channels at once, and also from the number of measurements that ensures a high SNR in the spectral domain. This technique has already been reported in the literature for the AIRS instrument [7], although in a different spectral range (from 3.7 to 15.4 μm), and in the frame of ground testing.

FTS are however compatible with space environment: this has been demonstrated in flight by IASI instrument, and the IRS instrument is currently under test, to be flown on third generation of Meteosat satellites.

The next figure illustrates the way a spectrometer ISRF can be retrieved from measurement of the FTS interferogram.

Figure 4: Principle of ISRF retrieval using an FTS as a light source

The drawback of the FTS is the need to accurately know the OPD, usually through a dedicated measurement path, its size, and the volume of data produced.

2.2.5 Trade-off conclusion

The advantages and limitations of the 4 solutions are briefly summarised in the table 2 below.

Table 2: Comparison of the 4 pre-selected solutions

	PROS	CONS
Tuneable Laser Diode (LD)	Flight proven High output power Fibre-coupled Small size device	Limited tuning range (0.2nm/K) Wavelength primarily dictated by available materials: only discrete wavelengths available, no device available in UV range and short VIS wavelengths,
Micro-resonator	Fair output power Simultaneous measurements (frequency comb) Fibre-coupled Small size device	Only available about 1.5 μ m Not mature yet
Fabry-Perot Etalon (FP)	Simultaneous measurements (frequency comb)	Delicate H/W Several devices required Insufficient spectral resolution Low output power
Fourier-Transform Spectrometer	Already demonstrated Simultaneous and wide-band measurements High SNR on retrieved ISRF	Medium to low signal OPD knowledge required Data volume H/W volume

At this stage of the study, the FTS is selected as the best solution: it can cover the whole spectral range, it can deliver sufficient signal, it is mature enough. The challenge is now to devise a simple and robust concept that is suitable for accommodation aboard a space instrument.

3. DESIGN OF THE CALIBRATION SYSTEM

3.1 Application case (CO2I/CO2M)

While the trade-off reported above has led to the selection of a conceptual solution, the best way to go further in the detailed study is to define a realistic application case, to which the ISRF calibration device will be dedicated.

Finally, the CO2I instrument has been selected as the application case. Not only is this instrument fully relevant because of its very demanding ISRF knowledge in flight, but also it is well known in Thales Alenia Space who is now in charge of its definition study.

The CO2M mission aims at monitoring anthropogenic CO2 emissions, relying on simultaneous accurate CO2 and NO2 observations with high spatial and temporal resolution at global scale, thanks to a constellation of identical satellites in sun-synchronous orbit (altitude 735 km, LTDN 11:30).

To fulfil the mission's objectives, each satellite will include the following capabilities:

- Perform measurements in NIR, VIS, SWIR-1, SWIR-2 channels and determine total column concentrations for atmospheric CO2 and CH4 as well as atmospheric NO2 content for accurate tracing of anthropogenic CO2 plumes with the CO2I instrument (Thales Alenia Space in France)
- Perform measurements of aerosols to better account for scattering effects using the Multi-Angle Polarimeter (MAP) instrument (Thales Alenia Space in UK)
- Accurately filter data for cloud-contaminated cases thanks to the Cloud Imager (CLIM) (OIP in Belgium).

Observations run continuously for Sun zenith angles up to 80 degrees. A Sun calibration is performed over the North pole, and other kinds of calibration during eclipse.

Figure 5: The CO2M payload carries 3 distinct instruments

The CO2I cross-track Spatial Sampling Angle SSA is 2.85mrad, and the integration time is 0.29s.

As mentioned, the CO2I instrument operates in 4 channels specified below, which are therefore to be addressed by the ISRF calibration device in replacement of the requirement ISRF-7 defined above.

Table 3: CO2I instrument performances

	VIS	NIR	SWIR-1	SWIR-2
Spectral range (nm)	405 - 490	747 - 773	1590 - 1675	1990 - 2095
Spectral resolution (nm)	0.6	0.12	0.3	0.35
Spectral sampling (nm)	0.2	0.04	0.1	0.12
ACT Spatial Angular Sampling (mrad)	2.85	2.85	2.85	2.85

3.2 Design drivers

One important part of FTS is often the sub-system devoted to the accurate measurement of the OPD, which usually requires a dedicated measurement path in the interferometer and its related electronics. This is why an innovative technique has been considered, to get rid of this OPD measurement specific part and to derive the OPD from the interferogram measurement themselves, taking advantage of the different spectral channels available. This has been the major objective of the breadboarding activities described below.

A second driver relates to the retroreflectors used on each arm of the interferometer. The simplest device is a flat mirror, but then the guiding accuracy of the drive mechanism is severely constrained. The use of cube-corners (CC) allows for a relaxation of these constraints, and we have selected solid-glass CC with a relatively small diameter of 25mm, or 1 inch. They can be quite easily procured, and the small dimension and mass again contributes to relaxing the requirements on the mechanism and of the overall system. The counterpart is a beam expander to be inserted between the FTS output and the instrument, in order to recover a diameter sufficient to illuminate the entire entrance pupil of the instrument, and an increase of the beam divergence within the interferometer, which adversely affects its spectral resolution.

Another important topic is the light level at the FTS output. In the basic concept that has first been naively analysed, a diffuser plate was placed between the FTS and the instrument entrance pupil, embedded in the rotating drum of the instrument Flight Calibration Unit (FCU), thus allowing to illuminate the entire instrument FoV at once. However, the signal level revealed too low in that configuration. To solve this issue, it has been decided instead to sequentially characterise only a few spatial samples of the instrument, through a direct, collimated beam, reflected by a flat mirror located in lieu of the diffusing plate. This way, the light level at FTS output is more than sufficient, and even excessive for some channels. It has the secondary advantage that a low power lamp is sufficient, and therefore the thermal control is simplified.

Finally, the operational scenario must also be considered. Let's take the channel SWIR 1 as an example. Numerical simulations have demonstrated that a maximum OPD of 10 mm was sufficient. Applying the usual sampling criterion, the OPD step should be smaller than half the measured wavelength, that is roughly $0.8\mu\text{m}$. The number of samples to be acquired then reaches 12500 for a single-sided interferogram. Let's assume that measurement is performed in 1 integration period, while the OPD step is performed within another integration step. Then the overall measurement duration amounts to $12500 \times 2 \times 0.29 = 7250\text{s}$, that is approximately 2 hours. Calibration would therefore require more than 1 orbit for completion, thus precluding normal observation, or it would be limited to eclipse periods and therefore requires to be split between several orbits. This is not in favour of accuracy and does not allow to meet requirement ISRF-8. Therefore, taking advantage of the narrow spectral channels of the instrument, and of the a priori knowledge of their spectral positions, interferograms must be undersampled such that a full band characterisation takes less than 40 minutes, that is approximately less than 3200 acquired samples per channel.

3.3 Calibration system design

3.3.1 Overview

The main part of the calibration device is shown on the next figure. It is composed of an optical bench that supports:

- The interferometer, that is the heart of the system
- The light source assembly
- The beam expander and the scan mirror, that shape the output beam and direct it towards the instrument.

Figure 6: The optical bench supporting the light source, the interferometer, and the beam coupling at its output. The right-hand view also shows the FCU drum containing the field mirrors that reflect the FTS beam toward the instrument

This ensemble still needs to be completed by a set of field mirrors located in the instrument FCU, and by electronic boxes required to drive the 2 mechanisms and the lamps, assumed to be hosted in the instrument itself. The overall dimensions of the equipped bench are modest. Its mass is estimated as less than 8kg, and it fits well within the instrument as shown by the figure 7.

Figure 7 - Centre: dimensions of the calibration device are reasonable. Left: Footprint of the optical bench compared to that of the payload. Right: the calibration device is close to an aperture of the side panel, to benefit from a radiative area.

3.3.2 Interferometer

The interferometer is classically built around a Silica beam splitter BS and its Silica compensating plate CP, and features 2 arms with 30° incidence angles, which ensures smaller optical parts and less polarisation. The largest optical dimension of the BS is 30mm, and its thickness 5mm. One face of the BS is coated with a metallic coating (e.g. chromium) that ensures a reflection and a transmission value of at least 20% for the useful spectral bands. The other face of the BS, as well as the 2 faces of the CP, are left uncoated, with Fresnel losses close to 4% at each interface. An anti-reflection coating ARC, optimised for the 4 CO2I spectral bands, could have also been used, however the gain is considered low. The use of a CP is somewhat mandatory to balance the glass thickness crossed by the 2 arms, and therefore reduce the chromatic OPD between the 2 arms. Ideally, the thickness of the 2 plates should be similar to within a few tens of microns.

The solid-glass CC are also made of fused Silica. Their overall diameter is 25mm for a clear aperture of 22mm. Their front face can be left uncoated, with a reflection of about 4%, or be AR-coated to reduce this a bit. The rear faces are left uncoated, so that total internal reflection occurs. As for the BS, the thickness of the 2 CC should be similar in order to reduce the chromatic OPD between the 2 arms. Since the wavefront error difference between the 2 arms of the interferometer is likely to reduce the contrast of the light modulation at its output, the CC should be selected with a low WFE in order to minimise their difference. Beam deviation is often used as a proxy for the CC's WFE: a beam deviation of 1.3 arcsec is convenient regarding the short wavelengths of VIS band.

Figure 8: Top view of the interferometer. BS and CP are close together and share the same mount

The mobile CC is mounted on a linear actuator capable of a 20mm stroke and of step smaller than $0.5\mu\text{m}$. The selected mechanism is an existing linear actuator, that meets these requirements and is already qualified for spatial environments similar to the ones of CO2M.

3.3.3 Light source

The lamp selected for the calibration device is a low-power, Quartz-Tungsten-Halogen lamp (Halostar Starlite). It is rated for a 10W electrical power, and a colour temperature of 2800K. Similar lamps have been already qualified for flight aboard EnMap instrument [8]. The light flux from the lamp is collected by an off-axis, aspherical mirror, and focused onto the input of a fibre bundle. The latter allows for a flexible accommodation, and also for a cold redundancy by using a Y-shaped bundle. The output of the bundle is placed at the object focus of an off-axis parabola that collimates the beam before it enters the interferometer.

With this configuration, where the instrument basically looks at the lamp through the calibration device and its transmission, it happens the light level at the interferometer output is fair for the VIS channel, but far too high for IR channels. The collecting mirrors are therefore coated with a specific 'cold filter' that attenuates the IR wavelengths and balances the spectrum.

Figure 9: Left: the reflection factor of the cold filter as designed. Right: Radiance obtained at instrument input with this cold filter are close to the target.

3.3.4 Beam coupling

The beam output by the interferometer is first expanded by a beam expander composed of 2 off-axis, confocal parabolas. Then it is reflected by a rotating mirror (scan mirror) towards the so-called field mirror, located within the instrument FCU, that again reflects it towards the instrument entrance pupil with the right angle. Since the volume is limited inside the FCU, only 3 such mirrors can be implemented in the case of CO2I, meaning that only 3 spatial samples can be characterised in flight. This however is considered sufficient to detect a variation, and possibly to derive the ISRF of other spatial samples through some interpolating scheme.

Figure 10: Field mirrors re-direct the beam from the calibrator to the instrument. Only 3 such mirrors can be accommodated within the CO2M FCU.

The rotary actuator that has been selected to drive the scan mirror is an existing, space qualified model which offers convenient performances such as small angular steps of approximately 20 arcsec.

3.3.5 Thermal control

Despite its low power, the heat produced by the lamp needs to be dissipated by a radiator towards the cold space and should not affect the stability of the optical bench and of the interferometer. To that purpose, the lamps and the collecting mirrors are decoupled from the optical bench using insulating washers, and are enclosed in a cavity, decoupled from the instrument and the ISRF calibration device by MLI. This cavity is coupled to a 0.025m^2 radiator located on the CO2M – Y panel, covered by SSM on its external side and black coated on its internal side. An active thermal control is implemented on the radiator to ensure a stable environment to the lamp during all the mission. The radiator temperature is then stabilised to -10°C , thanks to an active thermal control power of about 4W, while the temperature of the lamp remains stabilised at about 300°C .

3.3.6 Optical bench

The optical bench is made from a 50mm thick honeycomb CFRP panel with 1.6mm skin. A finite element model analysis has demonstrated a first resonant frequency above 190Hz. The model has also been used to check the thermo-elastic behaviour of the calibration bench, and especially of the interferometer.

3.4 Calibration system expected performances

3.4.1 ISRF barycentre (ISRF-5)

The error on the retrieved ISRF barycentre is essentially dominated by the error on the wavenumber scaling, hence by the linear error on the OPD step. This in turn depends on the following contributors:

- Knowledge of the mobile corner cube displacement step
- Residual chromatism between the 2 arms of the interferometer
- Accurate alignment of optical axis with the mobile CC stroke axis
- Angular spatial extent of the measured instrument sample, which induces self-apodisation and spectral shift

The usual process is to rely upon a laser beam passing through the interferometer, and the knowledge of its wavelength. This in turn introduces some complexity in the design, such as accommodation of the laser source, its receiver, the metrology optical path, the electronics, etc.

For sake of simplicity, one of the design drivers was to not use such metrology. Instead, the baseline is to design a robust and stable system, that can be calibrated on-ground, then remains stable throughout the launch and the operational life. According to the requirement ISRF-5, the required relative accuracy on the retrieved wavelength amounts to $0.05 \times \text{SSI} / \lambda$, that is as low as $2\text{E-}6$. The relative accuracy required for the knowledge of the OPD should therefore be of the order of $1\text{E-}6$, and the stability should be even better, of the order of $1\text{E-}7$ over the duration of 1 calibration, that is roughly 40 minutes. This is considered achievable through a careful thermal control of the OPD drive mechanism and of the interferometer as a whole.

Another solution is to accurately derive the OPD from the interferogram data, taking benefit from the numerous adjacent channels. Investigating this way is one of the goals assigned to the breadboard activities reported below.

3.4.2 ISRF FWHM (ISRF-3)

The determination of ISRF FWHM is not affected by errors on wavelength scaling. However, it is affected by some of the errors that also affects the shape of the ISRF. The most critical contribution appears to be the residual chromatism between the 2 arms of the interferometer, and its variation over the wavelength range of the ISRF. This is why this chromatism must be reduced to the maximum extent by a careful match of the BS and CP thicknesses, and of the 2 CC thicknesses. For shorter wavelengths of VIS band, a match to within $20\mu\text{m}$ is required.

OPD harmonic errors may also have a significant impact, depending on their period. The strategy is here to characterise these errors on ground, then to account for them during the data processing and the ISRF retrieval. As they are linked to the CC drive mechanism and especially to its reduction stage, they are not supposed to evolve in flight.

Non-linearity of the instrument detection chain is also a potential contributor, much similar in its effect to the harmonic OPD error. There is no mean for the calibration system to reduce this: the best we can do is to characterise it, then to use this characterisation to further correct all the acquisition data.

3.4.3 ISRF Shape (ISRF-1)

The shape of the retrieved ISRF may be affected by several contributions, some of them also affecting the performances discussed above. Meeting the requirement of 1% accuracy requires tight to very tight tolerances.

One important contribution comes from the self-apodisation, which happens when the size of the detector becomes non negligible with respect to the interferometer field angle, or when the detector is no longer centred in this field. The figure below illustrates the loss of contrast for a detector of large dimension compared to the interference pattern.

Figure 11. (left) Fringe pattern generated by a small OPD: the irradiance on the detector is quite uniform. (right): Fringe pattern generated by a large OPD: the various rings observed scramble each other.

Without any particular processing, each point in the pattern will return a frequency which is slightly shifted with respect to the one retrieved at the centre. The neat effect for a large and off-centred detector is then an overall spectral shift and a broadening of the retrieved spectrum (loss of resolution). Most of this effect can be modelled and corrected by an adequate processing of the data, however this assumes the alignment between the interferometer and the instrument to be calibrated is well-known, and stable when in flight. The figure below plots the error resulting from a misalignment between the interferometer and the NIR channel of the CO2M instrument: the alignment must always be known to better than 0.5mrad, i.e. 1/5th of a CO2M spatial sample.

Figure 12. The retrieved ISRF is very sensitive to the alignment between interferometer and instrument

The contributions to ISRF-shape error are summarised in the table below.

Table 4. Error budget on ISRF shape

Contribution	Tolerance	Impact on ISRF shape
Chromatism and induced ZPD error	BS and CP thickness equal to less than 20µm CC thickness equal to less than 40µm	0.22%
OPD linear error	Linear error corrected to less than 1E-7	< 0.4%
OPD harmonic errors	Characterised on-ground	< 0.1%
MCC alignment	1 arcmin	negligible
FTS/Instrument alignment and self-apodisation	0.5 arcmin	0.5%
Light source instability	1.5% caused by temperature variations. Supply regulation assumed negligible.	0.35%
SNR		negligible
Total error (Linear / RSS)		1.57% / 0.77%

As a conclusion, meeting the requirement of 1% accuracy requires tight tolerances, but looks as an achievable target.

4. PROCESSING BREADBOARD

4.1 Breadboard objectives and description

Two difficulties are evidenced in the ISRF measurement concept presented above:

1. OPD errors play a crucial role in the performance of the system, and essentially its accuracy. One solution is to calibrate the OPD errors on-ground and to rely on its stability when in flight: this is the way the ISRF calibration system presented above has been designed. There is however another possibility, that is to process the interferogram data in such a way to retrieve the accurate OPD information from them.
2. The number of samples to be acquired in an interferogram is usually very large. But in order to keep the measurement duration in a reasonable range, not larger than one half of an orbit, the number of samples must be limited to approximately 3200. This requires to sub-sample the interferograms while keeping a sufficient accuracy on the retrieved ISRF.

The goal of the breadboarding activities is to demonstrate the solutions to the two issues listed above: retrieval of accurate OPD data from the interferogram; retrieval of an accurate ISRF from a sub-sampled interferogram.

The in-orbit design of the breadboard was generated with the Copernicus CO2M mission as a target application for this study. For this breadboard activity the following spectral bands on this mission were selected: Visible channel [405-490] nm and Near Infrared channel [747-773] nm. For this reason, the test instruments chosen for the breadboard were: Tropolite spectrometer within [320-500] nm and Hamamatsu spectrometer respectively [200-800] nm, covering different spectral ranges and different spectral resolution.

The simplified lay-out to fulfil all the In-Flight ISRF breadboarding goals is presented in Figure 13 as being the design of the breadboard and its OGSE for this study.

Figure 13. Setup overview of the In-flight ISRF breadboarding

A Bruker Vertex 80v vacuum FTS (Fourier Transform Spectrometer) will be used as an interferometric spectrally filtered light source (a QTH lamp will be used covering VIS and NIR ranges from 400-2500 nm) to illuminate the integrating sphere. The forward output homogenised light goes into the Tropolite push-broom spectrometer (320 - 500 nm) telescope, illuminating a limited number of spatial pixels and goes as well into the fibre-coupled Hamamatsu mini-spectrometer (200 - 800 nm spectral range).

The FTS instrument uses a HeNe laser (which cannot be switched off) as metrology for controlling its OPD and for active mirror alignment. For reaching the OPD knowledge goal, a different setup was used with a simple Michelson interferometer without active metrology resulting in large ($> 10 \mu\text{m}$) OPD deviations from the expected linear behaviour. In addition to the white light source, a HeNe laser was used simultaneously as light source, allowing verification of the OPD determination.

4.2 Derivation of OPD from measured interferogram data

The interferogram was measured by collecting spectra obtained by the spectrometer under test at each of the OPD positions imposed by the interferometer. At zero path difference of the interferometer the spectrometer is expected to see the unmodified spectrum of the incoming light (e.g. an incandescent lamp continuous spectrum), affected by the spectral transmissivity features of the interferometer and spectrometer, and the spectral quantum efficiency of the detector in the spectrometer. At non-zero OPD, a spectral modulation on the spectrum is expected, with faster modulation as the OPD increases. For a non-ideal interferometer, wavelength dependence of the ZPD position is expected (due to e.g. dispersion of glass used for the beam splitter and corner cubes, in combination with non-perfect alignment). This will result in a spectrally modulated spectrum even at “zero” OPD position.

The frequency and phase of the spectral modulations on the observed spectrum are directly coupled to the OPD.

Provided the input spectrum is known, as well as the ISRF of all pixels of the spectrometer, a very accurate determination of the a posteriori effective OPD can be derived from the spectrum, see for an example Figure 14.

Figure 14. OPD fit (blue line) to the measured spectrum (black asterisks) at OPD=2004.6 nm

Figure 15 shows an example of an interferogram obtained with a simple (unbalanced) beam splitter, a translation stage with irregular motion, and a commercially off-the-shelf Hamamatsu mini spectrometer. A fibre-fed incandescent lamp was used for illumination. The interferogram is for this image corrected for the average spectrum, resulting in oscillations around 0 (medium grey in the image).

Figure 15. Interferogram data of a simple interferometer with a Hamamatsu spectrometer. The horizontal axis shows the OPD dimension, the vertical axis shows the spectrometer pixel dimension, with short wavelength pixels towards the bottom

The measured interferogram was modelled using the FROID method (Fourier Reconstruction of Optical Interferometer Data [9].

The model uses as input the following parameters:

- A model wavelength grid
- For each of the detector pixels modelled:
 - a spectrum at the model wavelength grid (normalised to unity)
 - an intensity offset for the interferogram
 - an intensity scaling term for the interferogram
 - a Zero Path Difference shift
 - A linear slope with OPD
 - A quadratic term with OPD
- Dispersion of the interferometer at the model wavelength grid (responsible for spectral dependence of the Zero Path Difference)

- For each of the measured spectra, an OPD position of the interferometer at the time that spectrum was recorded
- A non-linearity term for the spectrometer detector

This way, the model can describe the interferogram as measured, including all relevant aspects affecting the data. In order to speed up (initial) processing, subsets in pixel- and / or OPD sample dimensions can be selected.

In brief, a best fit solution of the model to the measurements is determined by iteratively improving previous model values for the spectra observed by each pixel and the OPD position at each spectrometer read-out. This yields ZPD values which may be different for each spectrometer pixel, which are subsequently described by an appropriately smooth and continuous function for the (wavelength dependent) dispersion of the interferometer, minimising the residual ZPD values needed for each spectrometer pixel.

The OPD values derived this way were compared with the expected interferogram caused by the HeNe laser. Assuming a wavelength for this laser of 632.816 nm a systematic scaling factor of the OPD was found of about $1+2.3\text{e-}6$. Since no constraint was built in the fitting routine, a non-unity scaling factor for the OPD and wavelength axis is expected. Removing the effects of the OPD scaling, Figure 16 shows the differences between the OPD as expected from the HeNe laser interferogram and that extracted from the white light part of the interferogram. The large peak around ZPD is the result of mixing of white light and HeNe laser interferograms and can be disregarded. The discontinuity around $-370\text{ }\mu\text{m}$ and $+370\text{ }\mu\text{m}$ was observed from the fit residuals and is the result of incomplete convergence of the fit routine. Convergence is very slow at OPD values where very little fringe contrast is present in the interferogram. The OPD differences are clearly not randomly distributed and show somewhat periodic structures at various scales. E.g. in the OPD range between $110\text{ }\mu\text{m}$ and $370\text{ }\mu\text{m}$ (at an OPD scale of $260\text{ }\mu\text{m}$ and smaller), the standard deviation of the differences is 2.6 nm RMS, but after removal of a 20th order polynomial (resulting in OPD scales of $13\text{ }\mu\text{m}$ and smaller), the RMS difference drops to 1 nm.

Figure 16. Difference between retrieved OPD and HeNe laser line fit, after scaling the OPD

Figure 17. Averaged fit residuals of the measurement - model, as function of measurement intensity. The model did not take into account non-linearity, which shows up here as a systematic fit residual

Figure 17 shows the average fit residuals as function of measured intensity. Overall, the fit residuals are small and close to zero, but above 14000 BU a systematic deviation occurs that increases towards higher signal levels. This can be ascribed to spectrometer detector non-linearity (0.35% at 20000 BU) and is one of the spin-offs of this analysis, in addition to the ISRFs of all the spectrometer pixels.

4.3 Interferogram subsampling

In-flight measurement time of the ISRF is limited by the frame rate of a well-exposed spectrum times the number of measurements required to derive the ISRF. In practice, it is desirable to measure the ISRF in a fraction of an orbit (lasting typically about 1.5 hours for a Low Earth Orbit), thus the requirement to perform the relevant ISRF measurements within 3200 samples. However, a full scan at the required resolution and wavelength range requires considerably more samples. By taking larger OPD steps while maintaining the full OPD scan range, the resolution of the retrieved ISRF is maintained but due to the larger steps it is aliased to longer wavelengths. Without prior information, it is not possible to distinguish the original wavelength from the aliased wavelength. However, a coarse wavelength calibration of the spectrometer pixels (e.g. design values) is sufficient to provide the correct information to extract the ISRF at the correct wavelength.

A full OPD scan at 10000 samples was performed, from which the ISRF was extracted for a sample pixel at wavelength 773 nm (Figure 18, black curve). The same data set was decimated by selecting one in every 4 OPD samples, causing aliasing. The FROID method was provided with prior information on the approximate wavelength of the pixel, after which the ISRF was extracted for the sub-sampled data set (Figure 18, green curve). The absolute difference between the two ISRFs is shown as the dashed black curve, as well as the 1% goal value for the ISRF knowledge.

Since the number of points used in the full scan is 4 times higher than that of the sub-sampled scan, the signal to noise ratio in the ISRF is expected to be higher in the full scan, which explains part of the difference between the two curves. The RMS difference between the full and aliased ISRF is 0.3% in this example. For pixels at other wavelengths the difference ranges between 0.10 and 0.9 %.

Figure 18. Full scan (black) and aliased (green) ISRF of pixel 1845 at 773 nm, as well as the absolute difference between the two ISRFs (dashed curve) and the 1% goal (dashed horizontal line)

5. CONCLUSIONS

The new generation of atmosphere-sensing instruments requires a very accurate knowledge of the ISRF when in flight. Solutions have been investigated for an on-board ISRF-calibration system, and the best solution identified appears to be a Fourier Transform Spectrometer.

Advantages of this solution are a large spectral coverage from deep blue to SWIR wavelength, an excellent spectral resolution, a good immunity to noise.

The FTS however usually suffers from a low sensitivity, a high complexity and a huge volume of data produced.

We have devised a simple and robust system that should rule out these difficulties:

- A direct illumination provides for a high signal and a high SNR
- The a priori knowledge of the spectral response of the instrument allows to significantly undersample the interferograms hence to reduce the volume of data
- A sophisticated analysis software, in which the interferogram is simulated using a priori knowledge and actual OPD is determined through a best fit approach, allows to get rid of the OPD metrology system that is usually the main source of complexity in FTS hardware. The OPD accuracy evidenced in breadboard measurements demonstrated an improvement of OPD knowledge from $>10\ \mu\text{m}$ to the nm level.

A preliminary design has been established, and first runs of the data processing have been done on experimentally acquired data. All this sounds promising, and we look forward to go on with an elegant breadboard of the hardware, which would also be used to get experimental data on which to exercise our analysis software.

ACKNOWLEDGEMENT

This work has been realised in the frame of the ESA contract 4000126068/18/NL/AR: Concepts for the in-orbit characterisation of the Instrument Spectral Response Function.

REFERENCES

- [1] New method to determine the instrument spectral response function, applied to TROPOMI-SWIR, R. van Hees et al., Atmospheric Measurements Techniques Discussions, December 2017
- [2] Compact optical frequency comb source based on a DFB butt-coupled to a silicon nitride microring, S. Boust et al., 2019 International Topical Meeting on Microwave Photonics (MWP), Ottawa, ON, Canada, 2019, pp. 1-4. DOI: 10.1109/MWP.2019.8892102
- [3] Octave Spanning Frequency Comb on a Chip, P. Del'Haye et al., Conference on Lasers and Electro-Optics, OSA, 2010
- [4] Octave Spanning Frequency Comb generation in a silicon nitride chip, Y. Okawachi et al., Optics Letters Vol 36 No. 17, 2011, 3398-3400
- [5] Frequency comb generation in the green using silicon nitride microresonators, L. Wang et al., Laser Photonics Rev. 10, No.4, 631-638 (2016), DOI 10.1002/lpor.201600006
- [6] Visible Kerr comb generation in a High-Q silica microdisk resonator with a large wedge angle, J. Ma et al., Photonics Research Vol 7 No. 5, 2019, 573-578
- [7] Spectral test and calibration of the atmospheric infrared sounder, M.H. Weiler et al., Proc. SPIE 4483 (2002), 44-52, doi: 10.1117/12.453471
- [8] An on-board calibration assembly (OBCA) on the ENMAP satellite, L. Wilkens et al., ICSO 2016, Proc. SPIE Vol 10562, doi: 10.1117/12.2296123
- [9] Initial results from the USNO dispersed fourier transform spectrograph. Hajian, A. R., et al. 2007, The Astrophysical Journal. [astro-ph/0611495] Initial Results from the USNO Dispersed Fourier Transform Spectrograph (arxiv.org)].