

Spectral Computed Tomography

Spectral Computed Tomography

**Björn Heismann
Bernhard Schmidt
Thomas Flohr**

**SPIE
PRESS**

Bellingham, Washington USA

Library of Congress Cataloging-in-Publication Data

Heismann, Björn J.

Spectral computed tomography / Björn J. Heismann, Bernhard T. Schmidt,
Thomas Flohr.

p. ; cm.

Includes bibliographical references and index.

ISBN 978-0-8194-9257-9

I. Schmidt, Bernhard T. II. Flohr, T. (Thomas), 1960- III. SPIE
(Society) IV. Title.

[DNLM: 1. Radiography, Dual-Energy Scanned Projection. 2.
Tomography, X-Ray Computed. 3. Absorptiometry, Photon. 4.
Contrast Media--diagnostic use. 5. Radiographic Image Interpretation,
Computer-Assisted. WN 150]

616.07'572--dc23

2012028603

Published by

SPIE

P.O. Box 10

Bellingham, Washington 98227-0010 USA

Phone: +1 360.676.3290

Fax: +1 360.647.1445

Email: Books@spie.org

Web: <http://spie.org>

Copyright © 2012 Society of Photo-Optical Instrumentation Engineers
(SPIE)

All rights reserved. No part of this publication may be reproduced or
distributed in any form or by any means without written permission of
the publisher.

The content of this book reflects the work and thought of the author(s).
Every effort has been made to publish reliable and accurate information
herein, but the publisher is not responsible for the validity of the informa-
tion or for any outcomes resulting from reliance thereon.

Printed in the United States of America.

First printing



Contents

Introduction	vii
Chapter 1 Clinical Motivation for Spectral CT	1
References	4
Chapter 2 Physics of Spectral CT Measurements	5
2.1 X-Ray Source	5
2.2 Object	5
2.3 Detector	9
2.3.1 Detector responsivity details	12
2.4 Spectral Weighting	15
2.4.1 The linear Radon approximation	17
2.5 Measurement Results	19
References	20
Chapter 3 Spectral CT Algorithms	25
3.1 Basis Material Decomposition	25
3.1.1 Projection-based basis material decomposition	25
3.1.2 Image-based basis material decomposition	29
3.1.3 Practical implications of the basis material decomposition	30
3.2 Density and Atomic Number Reconstruction	34
3.2.1 Theory	35
3.2.2 Analytical approximation	36
3.2.3 Numerical solution	37
3.2.4 Fractional atomic numbers	39
3.2.5 Algorithm flow chart	41
3.2.6 Basic imaging properties	41
3.2.7 Quantitative measurement of chemical solutions .	43
3.2.8 Quantitative reconstruction of body fluids	44

3.3	Limits of Spectral CT Algorithms.....	49
3.3.1	Measurement errors	49
3.3.2	Algorithmic information transfer	50
3.4	Comparison of Spectral CT Algorithms.....	51
	References	52
Chapter 4 Techniques to Acquire Spectral CT Data		55
4.1	Use of Different X-Ray Spectra	55
4.1.1	Slow kV switching	55
4.1.2	Rapid kV switching.....	59
4.1.3	Dual-source CT	62
4.2	Use of Energy-Resolving Detectors	73
4.2.1	Dual-layer detectors.....	73
4.2.2	Photon-counting detectors	78
	References	87
Chapter 5 Clinical Applications		93
5.1	Raw-Data-based Applications	94
5.1.1	Basic principle	94
5.1.2	Medical applications of raw-data-based approaches.....	95
5.1.2.1	Bone densitometry.....	96
5.1.2.2	Monoenergetic images	96
5.1.2.3	K-edge imaging	100
5.2	Image-Data-based Applications	101
5.2.1	Simple techniques: Image mixing, optimized contrast, dual-energy index, and monoenergetic... 102	
5.2.2	Medical applications for image-based two-material decomposition.....	106
5.2.3	Medical applications for image-based three-material decomposition.....	108
5.3	Conclusions.....	112
	References	113
Index		117

Introduction

The first computed tomography (CT) system was built by Godfrey Hounsfield in 1971. A few years earlier, his co-inventor Allan McLeod Cormack had used the Radon transform and its inverse to theoretically describe a radiological x-ray scanning machine and image reconstruction method. For their research, Cormack and Hounsfield received the 1979 Nobel Prize in Physiology or Medicine.

Currently, CT is a widely used x-ray scanning technique. In its prominent use as a medical imaging device, CT serves as a workhorse in many clinical settings throughout the world. It provides answers to urgent diagnostic tasks such as oncology tumor staging, acute stroke analysis, or radiation therapy planning. Moreover, CT systems are also used in the quality analysis of industrial products or for security screening of luggage at airports.

Spectral CT was introduced in 1975 as an improvement in measurement technology. CT devices were enabled to gain information on the energy-dependent (i.e., spectral) attenuation properties of the object. Different technical realizations were employed. The straightforward solution is the so-called dual-kVp technique. Two CT scans with different x-ray tube acceleration voltages are performed. The two resulting data sets contain information on the spectral x-ray attenuation characteristics of the object. These spectral data can be used to obtain additional information on the object.

The corresponding spectral CT algorithms have two fundamentally different targets. First, two patient tissue types, such as bone and iodine-contrast-media-filled blood, can produce the same range of attenuation gray values in a CT image. In order to differentiate between the two materials, spectral CT data can be weighted for an optimum contrast-to-noise ratio. This task is readily solved for two specific tissues types. The general solution is achieved when the spectral measurement channels are summed with the inverse cube of their mean detected energy as scaling weights.

Alternatively, spectral CT algorithms can provide absolute and quantitative information on the scanned object, e.g., its chemical composition. It is important to understand how spectral CT differs from

standard single-energy CT scanning. With the latter, the resulting images consist of water-normalized effective x-ray attenuation coefficients given in Hounsfield unit (HU) numbers. A value of -1000 HU represents air, and a value of 0 HU indicates water. In medical imaging, the absolute HU numbers are only used in a limited number of diagnostic fields, e.g., calcium scoring in coronary heart disease assessments. This is due to the dependency of the HU numbers on the measurement system properties such as the x-ray tube spectrum, the CT reconstruction process, and the detector response function. In comparison, quantitative spectral CT algorithms aim to provide absolute material parameters such as concentrations, densities, or atomic numbers. These parameters can be compared to ground truth values in terms of their accuracy and precision. As a natural limit, x-ray measurement noise and systematic errors challenge the quantitative characteristics of the output data.

In 1976, Alvarez and Macovski (see Ref. 1 in Chapter 3) proposed the first and most successful approach to date, the basis-material decomposition (BMD) algorithm. This algorithm transfers the well-known principal component analysis to x-ray physics. As a result, images of basis-material coefficients (e.g., the distribution of bone and soft tissue in a patient) are obtained. The BMD algorithm has been used as a standard method in scientific research for the last few decades.

However, several practical problems were observed. The technical limitations of early dual-kVp CT scanners in terms of temporal resolution and measurement precision resulted in movement artifacts and limited accuracy of the algorithmic input data. Moreover, the algorithms themselves were affected by noise amplifications and model mismatch that led to a limited overall precision and accuracy. For a combination of these reasons, the method did not enter clinical routine CT imaging at that stage.

In 2006, dual-source CT (DSCT) was introduced as a CT measurement technique in which a second x-ray tube and detector pair was integrated into the CT system. DSCT allows for simultaneous data acquisition at two different x-ray tube voltages. The spectral separation between the two measurement channels can be improved by additional tube prefiltration. In the years following the introduction of DSCT, kVp switching of the x-ray tube and detector-based prototype spectral CT systems were developed as alternative system approaches. These rapid advances in spectral measurement technologies have driven a renewed interest in spectral CT algorithms.

To achieve a structured approach to the fundamentals and clinical applications of spectral CT, this book is divided as follows: In the first chapter, we consider the main clinical motivations for spectral CT applications. In Chapter 2, the measurement properties of spectral CT systems are described. Chapter 3 provides an overview of the current

state of research on spectral CT algorithms. Based on this overview, we evaluate the technical realization of spectral CT systems in Chapter 4. Device approaches such as DSCT, kVp switching, and energy-resolving detectors are compared. Finally, Chapter 5 summarizes various algorithms for spectral CT reconstructions and spectral CT image postprocessing, and links these algorithms to clinical use cases.

Björn Heismann
Bernhard Schmidt
Thomas Flohr
October 2012