

# Index

---

- “active video area”, 10
- alpha ( $\alpha$ ) angles 1 and 2, 26
- aluminum backing, 9
- aperture grill, 14
- Aquadag, 9
- area brightness, 40
- ASCII terminals, 99
  
- background raster, 101
- bandwidth, 53
- bar pattern, 74
- barium oxide, 29
  - style, 13
- barium–lead glass, 8
- beam landing angle, 65
- bipotential, 23
- black pedestal, 84
- blanking, 101
- Braun, Karl Ferdinand, 1
- break-through voltage, 68
- Briggs Test Pattern 4, 71, 75
  
- cathode, 13, 21
  - aging, 29
  - drift, 32
  - life, 30
- color CRT mask, 14
- commercial off the shelf (COTS), 83
- composite sync, 99
- composite video, 99
- contrast modulation, 47
- control grids, 21
- coulombs, 41
- Crookes, Sir William, 1
- crosshatch, 72
- cross talk, 102
- current density, 33
  
- DAC’s conversion rate, 86
- deltoid, 21
- Department of Health and Human Services (DHHS), 11
- depth of modulation, 56
- DICOM Part 3.14 Grayscale Standard Display Function (GSDF), 25
- dielectric properties, 9
- digital visual interface (DVI), 88
- dispenser, 29
  - types, 13
- drive fraction, 24
- dynamic focus, 34
  
- electron backscatter, 16
- electron cloud, 21
- Electronic Industries Association (EIA), 37
- extent, 72
  
- faceplate, 5
- full width, half maximum (FWHM), 47
- funnel, 5
  
- G<sub>1</sub> control grid, 22
- G<sub>2</sub> grid, first anode, 22
- gamma, 24
- Gaussian-shaped pixel, 47
- Geisler discharge tube, 1
- geometric distortion, 103
- ghosting, 101
- glass, dielectric material, 8
- grayscale steps, 74
  
- heater, 21
- high-frequency stability, 73
- horizontal pixel modulation, 52
  
- iconoscope, 2
- interlaced scan, 103
- International Standards Organization (ISO)
  - exposure rate limit curves, 11
- Invar alloy, 14
  
- jail bars, 102
- JEDEC, 37
  
- Kasal, 17
- Kelly Chart of Color Designations for Lights, the 1931 CIE system, 38
- kinescope, 2
  
- lead-free barium–strontium, 6
- line pairing, 103
- lower gun structure, 22
- luminance nonuniformity, 63
  
- mask, 14
- Maxwellian, 22
- moiré, 15
  
- National Information Display Laboratory (NIDL), 49
- neck tubing, 5
- Nyquist frequency, 52, 75

- pattern burn, 42
- phosphor efficacy, 38
- phosphor screen weight, 45
- pixel profile, 53
- point of focus, 34
- potash-soda lead glass, 6
- progressive scan, 103
  
- RAMDAC, 50
- raster addressability ratio (RAR), 49
- raster modulation, 68
- ringing, 101
- rolloff, 53
  
- scandium oxide, 29
- screen burn, 42
- screen weight, 17
- sequential scan, 103
- shadow mask, 14
- slot mask, 14
- smearing, 101
- Society of Motion Picture and Television Engineers (SMPTE), 71
- space charge, 21
- spatial noise, 44
- spot cutoff voltage, 22
  
- static focus, 34
- sweep circuit, 64
- sync on green, 99
  
- thermal quenching, 39
- thermionic, 13
- theta ( $\theta$ ) angle, 26
- thin screen, 45
- transistor-to-transistor logic (TTL), 100
- triode, 22, 24
  
- unipotential, 23
- upper gun structure, 23
- upper gun structure (lens), 28
  
- variable gain amplifier, 68
- vertical linearity, 51
- video card, 83
- Video Electronics Standards Association (VESA), 71
  
- Worldwide Type Designation System (WTDS), 37
  
- Zworykin, Vladimir, 2



**Kenneth D. Compton** received his BS degree from Moravian College in 1966 and has been with Clinton Electronics for 16 years, working with display manufacturer's who use cathode ray tube technology in commercial and medical imaging applications. Over the last nine years, Mr. Compton has focused on medical imaging and the development of medical displays. His on going involvement with various medical research organizations has brought the knowledge of the end user into the design phase of a display system to enhance the quality of softcopy. Mr. Compton is an active member of SPIE and SCAR and participates in panel discussions on display related questions. He has published more than 15 papers on imaging and display technology; his speaking experience includes symposium and formal training environments at HealthTech, SPIE Medical Imaging, and IWDM2000, in addition to internal training for Clinton Electronics and its distributors. Mr. Compton has served as part of the working group on Digital Mammography sponsored by the Office of Women's Health, and has been an industry liaison, consultant, and author for the AAPM Task Group 18.