High-speed spectral domain polarization-sensitive optical coherence tomography using a single camera and an optical switch at 1.3 μm

Sang-Won Lee,^{a,b} Hyun-Woo Jeong,^{b,c} and Beop-Min Kim^{b,*}

^aKorea University, Research Institute of Health Sciences, Jeongneung 3-dong, Seongbuk-gu, Seoul, 136-703 Korea

^bKorea University, Department of Biomedical Engineering, Jeongneung 3-dong, Seongbuk-gu, Seoul, 136-703 Korea

^cYonsei University, Department of Biomedical Engineering, 234 Meaji-ri, Heungup-myeon, Wonju-si, Gangwon-do, 220-710 Korea

Abstract. We propose high-speed spectral domain polarization-sensitive optical coherence tomography (SD-PS-OCT) using a single camera and a 1×2 optical switch at the 1.3- μm region. The PS-low coherence interferometer used in the system is constructed using free-space optics. The reflected horizontal and vertical polarization light rays are delivered via an optical switch to a single spectrometer by turns. Therefore, our system costs less to build than those that use dual spectrometers, and the processes of timing and triggering are simpler from the viewpoints of both hardware and software. Our SD-PS-OCT has a sensitivity of 101.5 dB, an axial resolution of 8.2 μm , and an acquisition speed of 23,496 A-scans per second. We obtain the intensity, phase retardation, and fast axis orientation images of a rat tail tendon *ex vivo*.

© 2010 Society of Photo-Optical Instrumentation Engineers.
[DOI: 10.1117/1.3277668]

Keywords: polarization-sensitive optical coherence tomography; single camera; medical biological imaging.

Paper 09222LR received Jun. 11, 2009; revised manuscript received Sep. 23, 2009; accepted for publication Nov. 9, 2009; published online Jan. 11, 2010.

Polarization-sensitive optical coherence tomography (PS-OCT) is an optical imaging technique that can provide not only cross sectional images with a high resolution approaching that of conventional histology, but also polarization information such as Stokes parameters, Jones and Mueller matrices, phase retardation, fast axis orientation, and the degree of polarization.¹⁻⁴ A number of medical applications using PS-OCT have been reported, such as dermatology, ophthalmology, and gynecology.⁵⁻⁷

Recent developments in PS-OCT have focused on frequency domain implementations using either a spectrometer [spectral-domain OCT (SD-OCT)] or a frequency-swept laser source [swept-source OCT (SS-OCT)] because of the sensi-

tivity and imaging speed advantages over conventional time-domain OCT.⁸⁻¹⁰ Earlier versions of the SD-PS-OCT system required two spectrometers to detect two orthogonal states of light split by a polarization beamsplitter (PBS).¹¹ However, this approach has disadvantages such as additional costs, larger-sized apparatuses, and hardware- and software-related difficulties with regard to trigger time control. To overcome these problems, several groups have proposed using a single camera in the 800-nm region.¹²⁻¹⁴ Baumann et al.¹² and Cense et al.¹³ demonstrated that two orthogonal states of light could be detected, side by side, using a single camera with 2048 pixels, half of which detected the horizontal component, while the other half detected the vertical component. They used dispersive grating at different incident angles for the two components (or a specialized Wollaston prism). This is a relatively low-cost configuration, where the trigger control is simplified. However, if the two orthogonally polarized light rays reach the detector array at slightly different angles, the two spectra may not match exactly, which results in distorted spectra and differences of the full bandwidth associated with the rays.^{12,14} Fan et al.¹⁴ also presented a single-spectrometer SD-PS-OCT configuration based on the full range of complex domain method for separating the vertical and horizontal polarization beam components. They used whole pixels of the camera for each polarization to avoid the spectral distortion between the two orthogonal components. To separate the two orthogonal components, they used different optical path-length differences for the vertically and horizontally polarized light. Complex conjugate mirror images were eliminated using a constant frequency modulation of the reference mirror. However, this method requires long calculation time because of the need for several postprocessing steps such as averaging, interpolation, full range complex image processing, and image shifting after data acquisition.¹⁴

In this work, we demonstrate a simple design for enabling the application of SD-PS-OCT to the 1.3- μm region using a single line-scan InGaAs camera with 1024 pixels and an optical switch. We obtained the intensity, phase retardation, and fast axis orientation images of a rat tail tendon *ex vivo*.

A schematic of our system is shown in Fig. 1(a). The polarization-sensitive low coherence interferometer used in the system is based on the original free-space PS-OCT system.¹ A superluminescent diode (SLD) at 1310 nm (DenseLight Semiconductors Incorporated, Singapore) with a full-width-half-maximum ($\Delta\lambda$) of 170 nm, which corresponds to a theoretical axial resolution of 4.4 μm in air, and a power of 25 mW, was used. Light from the SLD passes through the polarization controller (PC1) and PBS1 to produce horizontally polarized light with maximum optical power. The polarized light is split into the sample and reference arms. In the reference arm, light passes through a quarter-wave plate (QWP1) oriented at 22.5 deg with respect to the horizontal axis, so that after double pass, the reflected light from the stationary mirror is in a 45 deg linear polarized state. In the sample arm, after passing through the QWP2 oriented at 45 deg to the horizontal plane, the light is circularly polarized when incident on the sample. The circularly polarized light reflected from the 2-D galvanometer passes through the telescope constructed with two convex lenses and

*Tel: 82-2-940-2883; Fax: 82-2-941-2883; E-mail: bmk515@korea.ac.kr

Fig. 1 (a) Schematic of our PS-OCT system using a single spectrometer and a 1×2 optical switch. (b) Timing diagrams to synchronize the line-scan camera with the optical switch. The synchronized line trigger (red) for the line-scan camera with the driving signal for the optical switch (blue). (Color online only.)

an objective lens (NA=0.26, Mitutoyo, Japan) with a long working distance of 30.5 mm, as shown in the upper right corner of Fig. 1(a). This telescope is used to maintain a flat image plane while scanning. The lateral resolution of our system is approximately $14 \mu\text{m}$. Recombined light from the reference and sample arms is split into two orthogonally polarized states by PBS2. The horizontal and vertical polarization light rays are then delivered to the 2×1 optical switch (Boston Applied Technologies, Incorporated, Woburn, Massachusetts), which transmits each polarized component toward a single spectrometer by turns. The optical switch is made of electro-optic material; it is a solid-state switch operating in free-space architecture without any moving parts. The insertion loss is minimal and an ultrahigh switching frequency can be achieved ($< 1 \text{ MHz}$). The spectrometer is composed of a transmission diffraction grating (1145 lines/mm, Wa-

Fig. 2 The vertical (black) and horizontal (red) plane spectra after being resampled into the k -domain by a method of the linear interpolation. (Color online only.)

satch Photonics, Logan, Utah), achromatic doublet lens ($f=70 \text{ mm}$, Thorlabs Incorporated, Newton, New Jersey), and InGaAs line-scan camera (Goodrich Corporation). The camera has 1024 pixels, a resolution of 14 bits, and a maximum line rate of 46.992 klines/s.

To synchronize the camera with the optical switch and the 2-D galvanometer, an analog output board with four channels (PCI-6711, National Instruments, Austin, Texas) was used. Figure 1(b) shows that: 1. a TTL signal of 46.992 kHz (red) is required for line trigger of the camera, and 2. a TTL signal of 23.496 kHz (blue) is needed to drive the optical switch. When the channel for the optical switch generates 5 V, the vertically polarized light is transmitted to the spectrometer. On the other hand, when the channel for the optical switch generates 0 V, the spectrometer detects the horizontally polarized light. Therefore, an A-scan speed of our system is half the speed of the camera. The other two channels of the analog output board generate triangular signals to control the 2-D galvanometer. While B-scanning, the galvanometer was moved by a step so that the two orthogonal signals could be obtained from the same location.

After digitization with a frame grabber (PCIE-1427, National Instruments, Austin, Texas), the orthogonal datasets are rescaled, respectively, using the linear interpolation method, because the spectral domain data must be sampled as a function of the wavenumber (k). Figure 2 shows the raw spectra of the vertically polarized light (black) and the horizontally polarized light (red) after resampling into the k -domain when light from the sample arm is blocked. Both spectra with an exact same bandwidth can be detected. Therefore, it should be noted that no software corrections were required. We used PC2 and PC3 to maximize the spectrometer's efficiency and to correct the distorted spectra of two orthogonal light rays, because the efficiency of the grating and the camera depend to some extent on polarization states. After inverse Fourier transform, we can obtain the total reflected intensity, the phase retardation, and the fast orientation axis images by using equations in the previous paper.^{11,12}

The interval between image pixels in the axial direction of an OCT image should be smaller than half the theoretical axial resolution to satisfy the Nyquist sampling theorem.¹⁵ Therefore, the spectrum full bandwidth (λ_{full}) detected by the spectrometer should be defined by $\lambda_{\text{full}} \geq \pi \cdot \Delta\lambda / (2 \cdot \ln 2)$.¹⁵ Using our SLD, λ_{full} has to be as broad as 385 nm to achieve the theoretical axial resolution. However, a broader λ_{full} results in a low depth range, because a 1024-pixel camera is a state of the art product. Instead, we chose a λ_{full} value of

Fig. 3 Ex vivo PS-OCT images of the rat tail tendon obtained using Jones matrix formalism. (a) Intensity, (b) phase retardation with grayscale from 0 (black) to $\pi/2$ (white), and (c) fast axis orientation with grayscale from $-\pi/2$ (black) to $\pi/2$ (white).

Video 1 Intensity 3-D volume movie of the rat tail tendon. Each 3-D image has 512 (x axis)×512 (z axis)×512 (y axis) voxels (MPEG, 3.9 MB). [URL: <http://dx.doi.org/10.1117/1.3277668.1>].

214 nm, because a depth range of at least 2.0 mm is needed in most biological tissues. The axial resolution was measured to be 8.2 μm in air. The interval value between image pixels in the axial direction was 3.9 μm . This value is much greater than half the theoretical axial resolution, which is the main source of the broadening of the axial resolution. An InGaAs line-scan camera with a greater number of pixels can overcome this problem. The sensitivity of the system was approximately 101.5 dB for an optical path difference of close to zero when the exposure time was 16.96 μs . We could observe a decrease of approximately 10 dB within a depth range of 1.5 mm because of the depth-dependent sensitivity roll-off.¹⁰

To test the accuracy of the SD-PS-OCT system, we inserted a QWP into the sample arm between the objective lens and the sample mirror. When we rotated the QWP in steps of 10 deg, from 0 to 180 deg, we were able to measure the retardation with a constant value of 86.65 ± 1.04 deg, which shows that the SD-PS-OCT system is stable.

To demonstrate the system's performance in biological tissue, we obtained *ex vivo* images of a rat tail tendon. Figures 3(a), 3(b), and 3(c) show the intensity, phase retardation, and fast axis orientation images, respectively. The images are constructed with 512 A-lines separated by 5 μm in the lateral direction. We reconstructed 3-D volume images (Videos 1 and 2) from the intensity and phase retardation datasets using Amira 4.1 (Mercury Computer Systems, Chelmsford, MA).

In conclusion, we have developed a new PS-OCT system based on the spectral domain method at the 1.3- μm region.

Video 2 Phase retardation 3-D volume movie of the rat tail tendon. Each 3-D image has 512 (x axis)×512(z axis)×512 (y axis) voxels (MPEG, 3.3 MB). [URL: <http://dx.doi.org/10.1117/1.3277668.2>].

We used a high-speed InGaAs line-scan camera and an optical switch to detect the two orthogonal polarizations of light with a single spectrometer. Because we detected the vertically and horizontally polarized light by turns, the line rate of our SD-PS-OCT is half the speed (23.496 klines/s) of the line-scan camera. The axial resolution, depth range, and sensitivity of our system are 8.2 μm , 2.0 mm, and 101.5 dB, respectively. To demonstrate the system's performance in biological tissue, we obtained the intensity, phase retardation, and fast axis orientation images of a rat tail tendon.

Acknowledgments

This research was supported by a grant (2009K001280) from the Brain Research Center of the 21st Century Frontier Research Program and a grant (20090090250) from Basic Research Program through the National Research Foundation of Korea (NRF) funded by the Ministry of Education, Science, and Technology, Republic of Korea.

References

1. M. R. Hee, D. Huang, E. A. Swason, and J. G. Fujimoto, "Polarization sensitive low coherence reflectometer for birefringence characterization and ranging," *J. Opt. Soc. Am. B* **9**, 903–908 (1992).
2. S. Jiao and L. H. V. Wang, "Jones-matrix imaging of biological tissues with quadruple-channel optical coherence tomography," *J. Biomed. Opt.* **7**, 350–358 (2002).
3. S. Jiao, G. Yao, and L. H. V. Wang, "Depth-resolved two-dimensional Stokes vectors of backscattered light and Mueller matrices of biological tissue measured with optical coherence tomography," *Appl. Opt.* **39**, 6318–6324 (2000).
4. S. W. Lee, J. H. Kang, J. Y. Yoo, M. S. Kang, J. T. Oh, and B. M. Kim, "Quantification of scattering changes using polarization-sensitive optical coherence tomography," *J. Biomed. Opt.* **13**(5), 054032 (2008).
5. M. C. Pierce, J. Strasswimmer, B. H. Park, B. Cense, and J. F. de Boer, "Advances in optical coherence tomography imaging for dermatology," *J. Invest. Dermatol.* **123**, 458–463 (2004).
6. B. Cense, T. C. Chen, B. H. Park, M. C. Pierce, and J. F. de Boer, "In vivo depth-resolved birefringence measurements of the human retinal nerve fiber layer by polarization-sensitive optical coherence tomography," *Opt. Lett.* **27**, 1610–1612 (2002).
7. S. W. Lee et al., "Optical diagnosis of cervical intraepithelial neoplasia (CIN) using polarization-sensitive optical coherence tomography," *Opt. Express* **16**, 2709–2719 (2008).
8. Y. Yasuno, S. Makita, Y. Sutoh, M. Itoh, and T. Yatgai, "Birefringence imaging of human skin by polarization-sensitive spectral interferometric optical coherence tomography," *Opt. Lett.* **27**, 1803–1805 (2002).
9. B. H. Park et al., "Real-time fiber-based multi-functional spectral-domain optical coherence tomography at 1.3 μm ," *Opt. Express* **13**, 3931–3944 (2005).
10. J. Zhang, W. Jung, J. S. Nelson, and Z. Chen, "Full range polarization-sensitive Fourier domain optical coherence tomography," *Opt. Express* **12**, 6033–6039 (2004).
11. E. Gotzinger, M. Pircher, and C. K. Hitzenberger, "High speed spectral domain polarization sensitive optical coherence tomography of the human retina," *Opt. Express* **13**, 10217–10229 (2005).
12. B. Baumann, E. Gotzinger, M. Pircher, and C. K. Hitzenberger, "Single camera based spectral domain polarization sensitive optical coherence tomography," *Opt. Express* **15**, 1054–1063 (2007).
13. B. Cense, M. Mujat, T. C. Chen, B. H. Park, and J. F. de Boer, "Polarization-sensitive spectral-domain optical coherence tomography using a single line scan camera," *Opt. Express* **15**, 2421–2431 (2007).
14. C. M. Fan, Y. Wang, and R. K. K. Wang, "Spectral domain polarization sensitive optical coherence tomography achieved by single camera detection," *Opt. Express* **15**, 7950–7961 (2007).
15. M. Wojtkowski, V. J. Srinivasan, T. H. Ko, J. G. Fujimoto, A. Kowalczyk, and J. S. Duker, "Ultrahigh-resolution, high-speed, Fourier domain optical coherence tomography and methods for dispersion compensation," *Opt. Express* **12**, 2404–2422 (2004).