Journal of Biomedical Optics

SPIEDigitalLibrary.org/jbo

Multifocus confocal Raman microspectroscopy for rapid single-particle analysis

Lingbo Kong
Pengfei Zhang
Peter Setlow
Yong-qing Li

Multifocus confocal Raman microspectroscopy for rapid single-particle analysis

Lingbo Kong,^a Pengfei Zhang,^a Peter Setlow,^b and Yong-qing Li^a

^aEast Carolina University, Department of Physics, Greenville, North Carolina 27858-4353

^bUniversity of Connecticut Health Center, Department of Molecular, Microbial and Structural Biology, Farmington, Connecticut 06030-3305

Abstract. We have developed a multifocus confocal Raman microspectroscopy system that allows simultaneous analyses of ~ 80 individual biological or airborne microparticles based on a precise image-guided technique. Multiple individual particles adhered in random positions on a coverslip were illuminated by a multifocus excitation pattern formed by rapidly steering a single laser beam with a pair of galvo-mirrors, and their Raman scatterings were synchronously projected with another galvo-mirror to different rows of a CCD chip for parallel spectroscopic analyses. We show that this technique can be used to rapidly identify single airborne particles or bacteria collected on a slide and to monitor germination dynamics of multiple bacterial spores in real-time. © 2011 Society of Photo-Optical Instrumentation Engineers (SPIE). [DOI: 10.1117/1.3662456]

Keywords: Raman microspectroscopy; single-cell analysis.

Paper 115321R received Sep. 20, 2011; revised manuscript received Oct. 24, 2011; accepted for publication Oct. 27, 2011; published online Nov. 23, 2011.

Raman microspectroscopy combines Raman spectroscopy with optical microscopy¹ and has been widely applied in biomedicine, materials, and environmental science.^{1,2} It uses a focused laser beam as a single-point illumination on the sample by an objective, and the Raman-scattered signal from a tiny volume ($< 1 \mu\text{m}^3$) is collected in a confocal configuration to obtain high sensitivity and resolution.^{1,2} Raman microspectroscopy has been employed as a single-particle analyzer for noninvasive and label-free chemical analyses of single bacteria, viruses, mammalian cells, organelles, and nanoparticles.^{1,2} With the help of surface-enhancement, it allows ultrasensitive chemical analysis and single molecule detection such as crystal violet.³ The combination with optical tweezers allows real-time analysis of single living cells suspended in an aqueous medium.^{4,5} Recently, Raman microspectroscopy has been used to analyze airborne particles collected with the impactors,^{6,7} which is important in analyses related to human health, pollution monitoring, and global climate change. However, the weak nature of Raman scattering generally limits its use since it typically takes a few or

tens of seconds to obtain a Raman spectrum of an unknown particle, especially biological particles, and single-focus excitation can only analyze one particle within a Raman acquisition time. Consequently, this technique becomes time-consuming when the analysis of large numbers of single particles is desired, and the laser power shining on each particle is constrained, such as when monitoring physiological dynamics of multiple individual microbial cells in a population⁴ and when attempting to quantify individual airborne particles impacted at random positions on a cover slip.^{6,7}

In this paper, we report the development of a multifocus confocal Raman microspectroscopy system that allows parallel analysis of multiple individual particles in random positions on a cover slip based on a precise image-guided technique. The multifocus excitation technique has been used in laser scanning fluorescence microscopy,⁸ coherent anti-stokes Raman scattering microscopy,⁹ Raman imaging,¹⁰ and recently in multiple-trap laser tweezers Raman spectroscopy.⁴ Here, we show that this technique can be used to rapidly measure relative concentration of microparticles, to rapidly identify individual aerosol particles collected in indoor air and monitor biological dynamics of single bacterial spores during their germination. As a result, it allows parallel monitoring of dynamic process of 80 individual cells in 1 h, which otherwise would take at least 80 h to obtain the statistical data by using single-focus Raman microspectroscopy.

The experimental schematic is shown in Fig. 1(a). A laser beam at 780 nm was introduced into an inverted microscope (Nikon TiS) equipped with an objective (Plan Apo 60 $\times$, NA = 1.4) and a green-filtered halogen lamp was used for illumination.^{4,5} An imaging camera was used to record a bright-field or phase-contrast image which was analyzed by a MATLAB program to locate the centroid positions (x_i , y_i) of 80 particles in a field of view. These coordinates were used to drive a pair of galvo-mirrors GM₁ and GM₂ (Cambridge Technology, 6220H) to rapidly steer a single laser beam with step-function waveforms to form multifocus excitation on 80 individual particles. Another galvo-mirror GM₃ (Thorlabs, GVS001) in front of the spectrograph (Princeton Instruments, LS785) synchronously steered the backward Raman scattering from each focus onto different vertical positions of a multichannel CCD chip (Princeton Instruments, PIXIS 400BR) so that full Raman spectra from each focus can be collected simultaneously.¹⁰ The time that the laser focus illuminated each particle was equal within a cycle and much longer than the transition time for the laser beam to jump from one particle to another. The scanning frequency at which the laser focus was steered across each particle was high (~ 10 Hz) such that an effective time-averaged pattern illumination was created during the CCD acquisition time (1 to 10 s). A 50- μm confocal pinhole was used such that the Raman spectral image of each focus on the CCD chip occupied 2 to 3 pixels (20 μm /pixel). As a result, the CCD chip with 400 vertical pixels can be used to simultaneously collect 80 Raman spectra (5 pixels/spectrum). The measured lateral and axial resolutions of the confocal system are about 0.5 and 1.2 μm , respectively.

The advantages of using synchronous galvo-mirrors include: 1. the multifocus excitation pattern under the microscope is programmable and image-guided; and 2. the separation of Raman scattering from each focus on the CCD chip can be

Address all correspondence to: Yong-qing Li, East Carolina University, Department of Physics, 10th Street, Greenville, North Carolina 27858; Tel: 252-328-1858; Fax: 252-328-6314; E-mail: liy@ecu.edu.

Fig. 1 Experimental setup. (a) Schematic of multifocus confocal Raman microscopy. GM: galvo-mirrors; DM: dichroic mirrors; L: lens. (b) Phase-contrast image of a mixture of polystyrene beads and *B. megaterium* spores. Beads (numbered in green) and spores (numbered in red) have been identified by multifocus confocal Raman microscopy as shown in (c). (c) Individual Raman spectral images of 80 particles in (b) measured by the spectrograph's CCD chip with one exposure time. Typical Raman spectra of a polystyrene bead and a *B. megaterium* spore are shown at the top and bottom, respectively.

independently changed by the third galvo-mirror so that cross talk between different spectral channels can effectively be avoided.⁴ The difference between current work and multifoci-scan Raman imaging¹⁰ is that the 80 laser foci were only switched to each individual particle's position here, while in Raman imaging the laser foci are scanned across the entire area of the sample. This switching scheme effectively avoids useless Raman acquisition from the positions without particle occupation and thus increases the speed of multiple particle analysis.

Figure 1(b) shows a phase-contrast image of 80 particles of an unknown mixture of 1- μm polystyrene beads and dormant *Bacillus megaterium* spores. Apparently, these particles cannot be distinguished by their image sizes and shapes. Figure 1(c) shows the Raman spectral image of the 80 particles on the CCD chip with typical Raman spectra of a polystyrene bead (top spectrum) and a *B. megaterium* spore (bottom spectrum) recorded with an incident power of 1.0 mW per focus (with total power of 80 mW for the 80 foci) and a 5 s exposure time. From the single-particle Raman spectra, these particles can be easily distinguished, based on characteristic bands at 1001 cm^{-1} for polystyrene beads and 1017 cm^{-1} for *B. megaterium* spores. The latter band is due to the presence of ($\sim 10\%$ of dry wt) 1:1 chelate of Ca^{2+} with pyridine-2,6-dicarboxylic acid (dipicolinic acid) (CaDPA) in spores.⁴

Fig. 2 Relative concentration measurements of mixed *B. megaterium* and *B. subtilis* spores. Raman spectra of (a) *B. megaterium* and (b) *B. subtilis* spores. (c) Phase-contrast image of the mixed spores of *B. megaterium* (numbered in red) and *B. subtilis* (numbered in black). (d) Measured ratios of *B. megaterium* and *B. subtilis* spores versus the relative concentration of *B. megaterium* spores in the mixture. The red curve is the fitted line.

Due to the rapid identification capability noted above, it is possible to rapidly measure the relative percentage of specific microparticles in a mixture. As a demonstration, we mixed *B. megaterium* and *B. subtilis* spores with relative ratios of 1:3, 1:2, 1:1, 2:1, and 3:1, and $\sim 10 \mu\text{l}$ of each mixture was dried on a coverslip. For each mixture, we randomly measured 20 fields of view that each contained ~ 40 spores, as shown in Fig. 2(c), and ~ 800 particles were analyzed for each mixture. Figures 2(a) and 2(b) show Raman spectra of *B. megaterium* and *B. subtilis* spores, respectively. They can be easily distinguished because of the two carotenoid-specific bands at 1157 and 1520 cm^{-1} uniquely from *B. megaterium* spores.¹⁰ Therefore, the ratio of the numbers of *B. megaterium* and *B. subtilis* spores in a mixture can be determined for each condition simply by measuring the carotenoid-specific bands. Figure 2(d) shows the ratios as the function of the relative concentrations of the sample mixtures and a linear relationship was observed.

Multifocus confocal Raman microspectroscopy can also be used to rapidly identify airborne particles. Aerosol particles in a laboratory room were collected into deionized water (10 ml) with an AGI-30 impinger at a flow rate of 15 l/min for 30 min, as shown in Fig. 3(b). A small drop of the collected particles were deposited and dried on a quartz coverslip. Figure 3(c) shows a bright field image of 40 individual unknown particles. Raman spectra of these particles were simultaneously collected with a laser power of 2.5 mW per particle and an exposure time of 5 s. These spectra can be classified into five different types, as shown in Fig. 3(a). These particles were identified as i. carbon (with 1300 and 1600 cm^{-1} bands); ii. glassy silicates; iii. the mixed compounds of carbonates and silicates; iv. and v. silicate particles with and without fluorescence background.^{6,7}

The most effective application of multifocus confocal Raman microspectroscopy could be simultaneous monitoring of dynamic physiological processes of multiple living cells adhered at random positions on a coverslip. As an example, we monitored the germination of 20 *B. megaterium* spores using Raman microspectroscopy and phase-contrast microscopy. The spores were first heat-activated at 60 $^{\circ}\text{C}$ for 15 min in a water bath and were then adhered firmly on the surface of a quartz coverslip. The germination solution of 0.5 mM D-glucose with

Fig. 3 Identification of airborne particles in the atmosphere. (a) Raman spectra of five typical aerosol particles with particle numbers as labeled in (c). (b) Air sampler that collected the aerosol particles into water for spectroscopic analysis. (c) Selected aerosol particles that were recognized with a MATLAB program.

25 mM KPO₄ buffer (pH 7.4) was added at $t = 0$ min to initiate germination at 37 °C.^{4,5} The phase-contrast images were captured at a rate of 10 s per frame and the first frame was used to locate the positions of the 20 spores, as shown in Fig. 4(a). During the germination process of 60 min, the microscope stage may have slight movements in the horizontal direction, causing the displacement of spores from the laser foci. Therefore, a new frame of phase-contrast images was used to recalculate the coordinates of the 20 spores every 2 min to precisely guide the laser foci to the spores. The Raman spectra of these spores were acquired with an exposure time of 20 s and a laser power of 5 mW per focus. Figures 4(b) and 4(c) show the time-lapse changes in the CaDPA level (intensity of the 1017 cm⁻¹ band) and refractivity (intensity of the phase-contrast image) of seven germinating spores, respectively. The data showed that for each germinating spore, the completion of CaDPA release precisely corresponded to the end of the rapid fall in phase-contrast image

Fig. 4 Simultaneous monitoring of the germination of multiple individual *B. megaterium* spores. (a) Phase-contrast image of 20 *B. megaterium* spores before the addition of germinant. (b) Intensities of the CaDPA-specific band at 1017 cm⁻¹ and (c) intensities of phase-contrast images of seven germinating spores as a function of the germination time.

intensity and the rapid CaDPA release process normally took 1 to 2 min.⁵ Up to 80 spores can be monitored in parallel by using this multifocus Raman microspectroscopy.

In summary, we have developed a multifocus confocal Raman microspectroscopy system based on an image-guided synchronous galvo-mirror scheme, allowing ~80 individual particles in random positions to be analyzed simultaneously. We further demonstrated the utility of this analyzer for rapid identification of single bacteria and airborne particles collected on a cover slip, accurate concentration measurements, and the monitoring of physiological dynamics of multiple individual bacterial spores during their germination. This technique can further be integrated with differential inference-contrast and fluorescence microscopy,⁴ and may find broad applications in microbiology and environmental science.

Acknowledgments

This work was supported by a Multi-University Research Initiative (MURI) award through the U.S. Army Research Laboratory and the U.S. Army Research Office (PS/YQL) via contract Grant No. W911NF-09-1-0286 and by a grant from the Army Research Office (YQL/PS) via contract Grant No. W911NF-08-1-0431.

References

1. G. J. Puppels, F. F. M. de Mul, C. Otto, J. Greve, M. Robert-Nicoud, D. J. Arndt-Jovin, and T. M. Jovin, "Studying single living cells and chromosomes by confocal Raman microspectroscopy," *Nature* **347**(6290), 301–303 (1990).
2. G. Turrell and J. Corset, *Raman Microscopy: Developments and Applications*, Academic Press, London (1996).
3. K. Kneipp, Y. Wang, H. Kneipp, L. T. Perelman, I. Itzkan, R. Dasari, and M. S. Feld, "Single molecule detection using surface-enhanced Raman scattering (SERS)," *Phys. Rev. Lett.* **78**(9), 1667–1670 (1997).
4. L. B. Kong, P. F. Zhang, G. W. Wang, P. Setlow, and Y. Q. Li, "Characterization of bacterial spore germination using phase contrast microscopy, fluorescence microscopy, Raman spectroscopy and optical tweezers," *Nat. Protoc.* **6**(5), 625–639 (2011).
5. L. B. Kong, P. F. Zhang, P. Setlow, and Y. Q. Li, "Characterization of bacterial spore germination using integrated phase contrast microscopy, Raman spectroscopy and optical tweezers," *Anal. Chem.* **82**(9), 3840–3847 (2010).
6. N. P. Ileva, U. McKeon, R. Niessner, and U. Pöschl, "Raman microspectroscopic analysis of size-resolved atmospheric aerosol particle samples collected with an ELPI: soot, humic-like substances, and inorganic compounds," *Aerosol Sci. Technol.* **41**(7), 655–671 (2007).
7. S. Mertes, B. Dippel, and A. Schwarzenbock, "Quantification of graphitic carbon in atmospheric aerosol particles by Raman spectroscopy and first application for the determination of mass absorption efficiencies," *Aerosol Sci.* **35**(3), 347–361 (2004).
8. J. Bewersdorf, R. Pick, and S. W. Hell, "Multifocal multiphoton microscopy," *Opt. Lett.* **23**(9), 655–657 (1998).
9. T. Minamikawa, M. Hashimoto, K. Fujita, S. Kawata, and T. Araki, "Multi-focus excitation coherent anti-Stokes Raman scattering (CARS) microscopy and its applications for real-time imaging," *Opt. Express* **17**(12), 9526–9536 (2009).
10. L. B. Kong, P. F. Zhang, J. Yu, P. Setlow, and Y. Q. Li "Rapid confocal Raman imaging using a synchro multifoci-scan scheme for dynamic monitoring of single living cells," *Appl. Phys. Lett.* **98**(21), 213703 (2011).