Journal of Biomedical Optics

SPIEDigitalLibrary.org/jbo

Near-infrared light penetration profile in the rodent brain

Ammar Abdo
Ali Ersen
Mesut Sahin

Near-infrared light penetration profile in the rodent brain

Ammar Abdo, Ali Ersen, and Mesut Sahin

New Jersey Institute of Technology, Biomedical Engineering Department, New Jersey 07102

Abstract. Near-infrared (NIR) lasers find applications in neuro-medicine both for diagnostic and treatment purposes. Penetration depth and profile into neural tissue are critical parameters to be considered in these applications. Published data on the optical properties of rodent neural tissue are rare, despite the frequent use of rats as an animal model. The aim of this study was to measure the light intensity profile inside the rat brain using a direct method, while the medium is being illuminated by an NIR laser beam, and compare the results with *in vitro* measurements of transmittance in the rat brain slices. The intensity profile along the vertical axis had an exponential decline with multiple regions that could be approximated with different coefficients. The Monte Carlo method that was used to simulate light-tissue interactions and predict the scattering coefficient of brain tissue from the measurements suggested that more scattering occurred in deeper layers of the cortex. A single scattering coefficient of 125 cm^{-1} was estimated for cortical layers from 300 to $1500\text{ }\mu\text{m}$ and a gradually increasing value from 125 to 370 cm^{-1} for depths of 1500 to $3000\text{ }\mu\text{m}$. The deviations of *in vivo* results from the *in vitro* transmittance measurements, as well as the postmortem *in vivo* results from the alive measurements were significant. © The Authors. Published by SPIE under a Creative Commons Attribution 3.0 Unported License. Distribution or reproduction of this work in whole or in part requires full attribution of the original publication, including its DOI. [DOI: [10.1117/1.JBO.18.7.075001](https://doi.org/10.1117/1.JBO.18.7.075001)]

Keywords: neural tissue optics; light absorption and scattering; laser-tissue interactions; neural stimulation; Monte Carlo simulation.

Paper 130111R received Feb. 28, 2013; revised manuscript received May 28, 2013; accepted for publication Jun. 3, 2013; published online Jul. 5, 2013.

1 Introduction

In our laboratory, we are developing a wireless implantable device for neural microstimulation, where near-infrared (NIR) light is used for energy transfer to the microstimulators through neural tissue.¹⁻³ Microelectrode implants often fail either due to the chronic tissue response caused by the tethering forces of the wires or their breakage.^{4,5} A wireless neural stimulator can solve the problems associated with wire interconnects.

While the main objective of this study was to determine whether we could transmit such amounts of light into the depths of neural tissue that can generate sufficient electric currents for neural stimulation, our results can also provide practical insights for those who are investigating NIR light-neural tissue interactions. NIR lasers have been utilized for many diagnostic and therapeutic applications in medicine such as neural imaging⁶ and chronic pain treatment.⁷ Hematoma detection⁸ and treatments for stroke patients are new applications under investigation.⁹ An NIR light between 700 and 900 nm has minimal absorption and scattering,¹⁰ therefore light penetration into neural tissue is maximum at these wavelengths. Recent NIR applications considered for the clinic have produced a perception that NIR light can penetrate a few centimeters into the tissue through skin exposure where it can cause functional changes, whether due to the heat generated or other mechanisms.⁹ Although this may be correct in general, one should be cautious about how the penetration depth is defined. Photon density decreases steeply with depth in tissue and the remaining light intensity at the targeted depths may be too small to induce the desired effects. The intensity at the surface will always be higher relative

to the deeper regions whether it is muscle or neural tissue that is under investigation.

When light photons are sent into tissue, they may be reflected back at the surface of entry, absorbed or scattered as they propagate inside, or transmitted to the other side. In general, light propagation inside a uniform medium is defined by the absorption (μ_a), scattering (μ_s), and anisotropy (g) coefficients. The wavelength and the tissue type are the primary factors that determine these coefficients. The light intensity profile by depth is governed by Beer's law, a simple exponential decay, if multiple scatterings can be avoided. In tissue slices thicker than $\sim 100\text{ }\mu\text{m}$, however, the intensity profile deviates from an exponential where multiple scatterings take place before a photon is absorbed.

Both direct and indirect methods have been employed to measure the optical properties of biological tissues.¹¹ Direct methods are based on Beer's law where a thin piece of tissue is used, as in the collimated transmission technique. In this method, the unscattered light transmitted through the slab is detected on the other side for various tissue thicknesses and penetration depth is calculated by fitting an exponential curve to the measurements according to the Beer's law. The scattering coefficient can be predicted from experimental measurements of transmittance and reflectance together.¹² Indirect methods are based on a model of light transport. The integrating sphere method is a common technique where the absorption coefficient (μ_a) and the reduced scattering coefficient (μ_s) are deduced from measurements of diffuse reflectance and total transmittance. The coefficients are then predicted based on light transport models such as the Monte Carlo method.¹³

The Monte Carlo simulation is a powerful technique where probabilistic locations of a large number of photons are calculated as they propagate inside the tissue with known optical properties. Because scattering is the dominant phenomenon in the neural tissue at NIR wavelengths, it is crucial to know the

Address all correspondence to: Mesut Sahin, New Jersey Institute of Technology, Biomedical Engineering Department, Newark, New Jersey 07102. Tel: 973-596-5573; Fax: 973-596-5222; E-mail: sahin@njit.edu

scattering coefficient in order to predict the light penetration profile.

The rat nervous system is frequently used as an animal model in testing lasers for various applications.¹⁴ There are several reports on optical properties of human neural tissue also.^{15–17} However, the scattering and absorption coefficients predicted in these reports spread over a wide range of values.^{10,18} Thus, we set out to determine the light intensity profile with direct measurements of light intensity inside the brain of an intact animal model using an optical fiber. Our main objective was to narrow the reported range of scattering coefficients down and determine its value with a higher confidence in the rat brain. The results from the *in vivo* experiments were compared with the *in vitro* transmission measurements published in an abstract form earlier¹⁹ and Monte Carlo simulations were used to estimate the scattering coefficients.

2 Methods

2.1 In Vivo Experiments

Nine Sprague–Dawley rats (275 to 450 g) were used in this part. All experimental procedures were approved by the Animal Care Committee at Rutgers University, New Jersey. Under a single injection of urethane anesthesia (1.25 g/kg, IP), a 4 × 6 mm cranial opening was made immediately rostral to the bregma on the right side of the central fissure [Fig. 1(a)]. Dexamethasone (3 mg/kg, IM) was injected prior to surgery to reduce edema of the brain. Some protrusion of the brain pressing against the edge of the skull through the cranial opening was allowed to stabilize the tissue and minimize the motion artifact due to respiration and cardiac pulsation. A slight traction on the tail further reduced the brain motion in some cases. Connective tissue and any blood clot over the surface were cleaned without removing the dura mater. Surgical wax was applied to seal between the dura and the skull around the edges of the opening, and a shallow pool of normal saline (<1-mm high) was made to prevent dehydration. The rectal temperature was continuously monitored and maintained at 36°C using a temperature regulated heating pad underneath the animal. The end-tidal CO₂ was observed to stay within the normal range throughout the experiment, which usually lasted about 6 to 7 h including the surgery.

After positioning the head perfectly horizontally in the stereotaxic frame, a 26-g needle was slowly inserted from below at a point a few millimeters right of the trachea, until the tip of the

needle passed through a hole at the base of the skull and reached a few millimeters below the cortical surface in the center of the skull opening [Fig. 1(b)]. Care was taken to stay away from major blood vessels of the cortex. An optical fiber (Ø100 µm, graded index, Newport Inc., CA) was desheathed from its plastic insulation and cladding for about 5 cm of its tip. The fiber tip was polished flat by three different grain size sand papers in a decreasing order (3, 1, and 0.5 µm), attached to a micromanipulator facing up, and inserted through the lower end of the needle until its tip was leveled with the pial surface under the dura. The fiber tip was visible through the dura under 40× microscopic vision. A free-air, NIR laser beam (74 mW, 830 nm, DLS-500-830FS-100, StockerYale, Canada) with a circular footprint of 0.4 mm in diameter was aimed at the fiber tip from 13.5 cm above the cortex. The light power entering the polished end of the optical fiber was measured with a photodiode and a current amplifier is connected to the other end of the fiber. The laser activation pattern and data acquisition were controlled through MATLAB™.

A train of NIR pulses (10 ms pulse width, 25 pulses at 0.5-s intervals) was sent to the cortex while the light intensities were measured at depths from 0 to 2500 µm in steps of 50 or 100 µm from the dura. Pulsed laser was preferred over continuous exposure to avoid overheating the tissue. For each depth of the fiber tip position, the laser source was horizontally moved in a rostral or caudal direction in steps of 50 or 100 µm up to 1000 µm away from the fiber axis while repeating the light measurements. Thus, an intensity map was developed in a parasagittal plane for an area of 1000 × 2500 µm. Due to the vertical placement of the optical fiber, photons traveling from top to bottom were collected more efficiently than those traveling in other directions. Thus, it should be noted that the light intensity measurements were directionally biased in this study.

The laser intensity profile at a distance of 13.5 cm was verified using a commercial photodiode (G9842, Hamamatsu, GaAs PIN photodiode, Ø0.08 mm) as shown in Fig. 4(a) on top. The profile was Gaussian with a circular cross section. The radius was taken as the $\sqrt{2}\sigma$ value of the Gaussian intensity curve, which is the horizontal distance where the intensity profile drops down to 1/e of its peak in the center. This diameter was equal to that of a flat profile beam with the same peak power intensity and the same total power as the Gaussian profile laser that was used. The light power that was collected by the optical fiber just underneath the dura mater was measured as the calibration

Fig. 1 (a) Sketch of the cranial opening made rostral to the bregma on the right side of the central fissure. (b) The preparation used to measure light intensities in the rat brain.

value. All measurements inside the brain were normalized by this maximum light intensity level. If the fiber optic tip was broken or damaged during an experiment, it was prepared again by sand papering; then the calibration measurement was repeated and used as normalization value for further experiments. Measurements of the light power obtained by placing the fiber in air were compared with the subdural values to estimate the contribution of the dura matter to NIR light attenuation.

2.2 Data Processing

The average of 25 pulses in the laser train was taken as the representative light intensity measurement at the corresponding point in the brain. This averaging step substantially reduced the periodic variations due to cardiac and respiration artifacts. The trial was repeated when the animal had a sigh, which caused a small but significant mechanical displacement of the brain.

The exact peak position of the light intensity profile in the horizontal direction usually occurred between two adjacent measurement points, although we decreased the step size to $50\ \mu\text{m}$ near the center. Cubic spline interpretation was used to find the exact horizontal position of the maximum as well as all the other rostrocaudal points at $50\ \mu\text{m}$ steps from the new horizontal zero point.

2.3 In Vitro Experiments

The description of *in vitro* experiments can be found elsewhere¹⁹ and they are summarized, here, for clarity of this paper. Seven Sprague–Dawley rats were used for *in vitro* measurements. The brain was explanted quickly after sacrificing the animal and placed in sucrose cutting media (0.75 mmol KCl, 0.625 mmol NaH_2PO_4 , 6.5 mmol NaHCO_3 , 2.5 mmol glucose, 0.25 mmol MgCl_2 , 50.39 mmol sucrose, 0.5 mmol CaCl_2 , $\sim 0^\circ\text{C}$). Horizontal brain slices were cut using a Vibratome at various thicknesses from 300 to $1500\ \mu\text{m}$. The sliced samples were kept in the sucrose cutting media for less than 3 h until *in vitro* measurements were completed.

A commercial photodiode (G9842) was used to measure the transmitted light intensity through the brain slices. The photodiode was first secured at the bottom of a $9 \times 9\ \text{mm}$ plastic chamber such that the photodiode active area was even with the bottom surface. The same free-air laser source of the *in vivo* experiments was aimed at the surface of the tissue slab from above and aligned with the center of the photodiode.

The incident light power was measured first with no neural tissue present. The brain slices were then laid out on the bottom of the chamber in direct contact with the photodiode. The samples were kept moist with normal saline during measurements, which took maximally a few minutes. The horizontal extent of the tissue sample orthogonal to the incident light beam was assumed to be sufficiently large to neglect the boundary effects. The transmittance values were obtained for various thicknesses of the gray matter samples in the center of the laser beam.

2.4 Monte Carlo Simulations

A public domain software was utilized for the Monte Carlo simulations.²⁰ An infinitely narrow photon beam was vertically aimed at an infinitely wide tissue (Fig. 2). Each layer was characterized by the refractive index (n), the absorption coefficient (μ_a), the scattering coefficient (μ_s), the anisotropy factor (g), and the thickness of the layer (d). The simulation software keeps a

Fig. 2 The three-dimensional model geometry for Monte Carlo simulations. Tissue slice extends infinitely in the two-dimensional (2-D) horizontal plane.

score of the photon densities on a three-dimensional grid inside the tissue (fluence) and across the surfaces of the medium boundaries, i.e., reflectance at the entry surface and transmittance on the opposite side.

Neural tissue has an inhomogeneous structure with multiple regions of different scattering coefficients.¹⁸ From our measurement of light intensity in the cortex, it was also evident that multiple layers with varying slopes of the photon density curve existed in deep locations ($>1500\ \mu\text{m}$).

The scattering coefficient was estimated from the averaged NIR light intensity curve. The scattering coefficient was adjusted iteratively in the simulations to curve fit the light intensity values measured during the *in vivo* and *in vitro* experiments (T_{exp}) as a function of depth in the beam center. The transmittance for an infinitely narrow beam size (Monte Carlo modeling of light transport in multi-layered tissues component of software package) was computed first and then the results were convolved (convolution program component) with the Gaussian beam profile to simulate the experimental case. Finally, the scattering coefficient was incrementally adjusted while repeating the simulations until a minimum was found for the average percent error for all depths (or slice thicknesses for the *in vitro* data) between the simulated (T_{sim}) and experimental curves (T_{exp}).

Other parameters of simulation included the following: number of photons 1,000,000, refractory index for saline $n_s = 1.33$ and gray matter $n_g = 1.38^{10,18}$; absorption coefficient for biological tissue $\mu_a = 0.1\ \text{cm}^{-1}$ (Ref. 18); and anisotropy factor $g = 0.9^{10,18}$. All simulations were conducted on a computer with Intel 2.3 GHz processor and each run took less than 3 min.

3 Results

3.1 In Vivo Experiments

Photon density decreased exponentially with depth from the dura at the center of the laser beam (Fig. 3). Photon densities were $79 \pm 4.5\%$ at $100\ \mu\text{m}$ and $0.04 \pm 0.003\%$ (mean $\pm$ SE) at $3000\ \mu\text{m}$ into the cerebrum. In order to exclude the effect of the dura, the light intensities were plotted in percentages as a fraction of the subdural light intensity in the beam center. This was necessary to factor out the large variations we noticed in the dura thickness and the subdural light intensities between the animals. Light attenuation by the dura was calculated separately and found to be $58 \pm 4\%$ by comparing the subdural and in-air measurements. Results obtained from postmortem animals (red curve in Fig. 3) were significantly different

Fig. 3 The average NIR light intensity inside the brain as a function of depth at the center of the laser beam. (x) Average data from 16 penetrations in nine live animals and (■) data shortly after termination in six animals. (Δ) Average *in vitro* transmittance as a function of slice thickness ($N = 8$). The *in vivo* light intensities from 1500 to 3000 μm are reported in the inset on a larger vertical scale.

($p \leq 0.01$) from that of the alive photon densities between 100 and 1500 μm .

Figure 4(a) shows the two-dimensional (2-D) illumination profile in the parasagittal plane of the measurements. In the horizontal direction near the surface, the light intensity profile drops quickly, almost following a similar course to the laser intensity. In deeper locations, however, the light intensity profile becomes more spread horizontally [Fig. 4(b)]. The illumination contours extend a few times the laser diameter in the vertical direction and becomes wider at depths, resembling the shape of a bulb. A possible explanation for this shape is that the scattering increases in deeper regions of the brain.

3.2 In Vitro Experiments

The transmittance measurement for each slice thickness was obtained by averaging measurements from seven animals (green curve in Fig. 3). Specifically, the transmittance

values were $65.0 \pm 0.02\%$, $44.9 \pm 0.02\%$, $32.2 \pm 0.02\%$, $22.7 \pm 0.03\%$, $14.1 \pm 0.01\%$, $7.78 \pm 0.005\%$, $5.38 \pm 0.005\%$, and $3.02 \pm 0.004\%$ (mean $\pm$ SD) at thicknesses of 300, 400, 500, 600, 800, 1000, 1200, and 1500 μm respectively. These *in vitro* values were much larger than the light intensities obtained from live and postmortem animals (paired t -test, $p \leq 0.01$). Note that the *in vitro* experiments were normalized with the light intensity received by the photodiode without any tissue present, whereas *in vivo* measurements were normalized to the subdural light intensity. Using the Fresnel equation [$R = (n_1 - n_2)^2 / (n_1 + n_2)^2$, where $n_1 = 1$ for air and $n_2 = 1.38$ for neural], the reflection for a light beam entering the tissue perpendicularly was calculated as 2.55%, and neglected in the *in vitro* plot. This means the reflected light at the surface of entry was lumped into the attenuation by the tissue.

3.3 Monte Carlo Simulations

The scattering coefficient for *in vitro* samples was estimated by fitting simulated transmittance curves to the measured values and found to be 108 cm^{-1} .¹⁹ The average deviation from the experimental curve was 9.9% for the optimal fit. The mean error would approximately double if the scattering coefficient was set to 90 or 125 cm^{-1} on each side of the optimal value. Thus, the optimal value of μ_s (108 cm^{-1}) clearly produced the minimum error within this range. Simulations were repeated using two extreme values of absorption coefficient to investigate the effect of μ_a on transmittance.¹⁸ The average deviation from the *in vitro* transmittance curve was 9.4% at $\mu_a = 0.01$ and 1.0 cm^{-1} , which was smaller than the mean error at optimal μ_s . This indicated that μ_a had a minimal effect on the simulated transmittance. Therefore, we varied the scattering coefficient alone and fixed μ_a at a typical value of 0.1 cm^{-1} (Ref. 21) in further simulations.

The scattering coefficient was estimated also from *in vivo* data using the Monte Carlo method. This time, however, the curves appeared to have multiple regions with different exponentials. Figure 5 depicts the goodness of the fit using three different values of the scattering coefficient ($\mu_s = 125, 175$, or 370 cm^{-1}). Two main regions with different optical properties

Fig. 4 (a) Two-dimensional illumination inside the rat brain in a parasagittal plane. Results shown are the average data from four animals. Measurements were made in steps of 100 μm in the vertical and horizontal directions and then interpolated for intermediate points. The light intensities are normalized by the intensity measured subdurally in the beam center and expressed on a logarithmic scale. Measurements were made in half the plane since the data are vertically symmetrical. The bell shaped curve on the top shows the laser intensity profile at the dura surface. (b) Horizontal spread of the illumination profile at 2100 μm from dura.

Fig. 5 Photon density simulations using different scattering coefficients. The simulated values are fitted to the experimental group averages by adjusting the scattering coefficient. Experimental data (●, blue), simulated data using $\mu_s = 125 \text{ cm}^{-1}$ (×), $\mu_s = 175 \text{ cm}^{-1}$ (Δ), and $\mu_s = 370 \text{ cm}^{-1}$ (■). The data points from 1000 to 3000 μm are reported in the inset on a larger vertical scale.

existed as shown in the figure; first from the surface down to 1500 μm and the second from 1500 to 3000 μm . The photon intensity curve could be approximated with a single scattering coefficient (125 cm^{-1}) in the first region with an average error of 7.42% only. If the measurements at 100 to 400 μm were excluded, the average error would reduce to 3.81%. In the second region, multiple layers with different coefficients were present. The average deviations from the experimental curve were 77.5%, 36.5%, and 31.2% for scattering coefficient values of 125, 175, 370 cm^{-1} , respectively. Each scattering value was good only for a small section of the curve but not for the entire range. We concluded that the optimal scattering coefficient varied between 125 and 370 cm^{-1} at depths of 1500 to 3000 μm in the rat brain.

4 Discussion

In this study, we report direct measurements of photon density inside a rodent brain under a beam of NIR laser exposure. The results are compared to *in vitro* transmittance measurements from neural tissue slices and the scattering coefficients were predicted using Monte Carlo simulations. Overall, the intensity decreases exponentially by depth in the beam center, although multiple regions exist in the curve suggesting different scattering coefficients. These regions with different optical properties most probably correspond to the gray matter of the cortex and the underlying white matter of the brain.

The *in vitro* transmittance values were larger than the light intensities obtained *in vivo* from live and postmortem animals (Fig. 3). The discrepancies can be attributed to the differences between the *in vivo* experimental setup and the *in vitro* sample preparation techniques. Small amounts of blood residue on the dura surface and the mechanical displacement of the brain due to breathing and cardiac pulsation might have contributed to the variations observed in the *in vivo* measurements. Data from postmortem animals were slightly different from that of the live animals suggesting that the optical properties of neural tissue change when the arterial blood pressure diminishes.

The light intensity profile followed an exponential decline with an initial phase that could be simulated with a single scattering coefficient up to 1500 μm . Excluding the first few points lowered the estimation error. Outliers at the superficial layers

may be due to the contribution of the dura matter, which was not characterized in this study. A single scattering coefficient can probably represent the optical properties from 0 to 1500 μm where the brain cortex predominantly has cell bodies. The data suggest that the scattering coefficient begins to increase at the border of gray and white matter and continues to increase gradually within the white matter. It was reported that the gray matter inside the rat brain cortex extends from the surface down to 1 to 2 mm.¹⁸ Our results suggest that the border of gray and white matter is located at a depth of $\sim 1500 \mu\text{m}$.

The profile obtained from the *in vitro* samples had a simple exponential decrease as well, except at the beginning of the curve. Obtaining brain slices below 300 μm in thickness was not possible without compromising the intactness of the tissue samples.

The estimated scattering coefficients for the rat brain tissue were within the ranges reported in the literature. Johns et al. reported the reduced scattering coefficient value (μ_s) as 10 to 61 cm^{-1} (which corresponds to a scattering coefficient (μ_s) of 100 to 610 cm^{-1} for $g = 0.9$) in the rat brain tissue while the detector was moving from gray to white matter.¹⁸ They found the μ_s of the white matter to be a little higher (35 to 61 cm^{-1}). Scattering coefficients reported in the literature for human brain tissue also vary within a large range. Yaroslavsky et al. found the μ_s for the human gray matter to be around 70 cm^{-1} at 850 nm *in vitro*.¹⁵ Gottschalk et al. reported μ_s also for the human brain gray matter to be 364 and 267 cm^{-1} at 675 and 1064 nm, respectively.²² It is difficult to make a comparison even between the rat and human gray matter with these reported scattering coefficients differing between 4- and 6-folds within each species.

Sample preparation technique is very important in measurements of optical tissue properties *in vitro*. We believe that the data presented in this report are more realistic because the measurements were made in live animals. Roggan et al.²³ found that both shock and slow freezing procedures alter the optical properties of biological tissue. The freezing procedure changes the optical properties by damaging the tissue due to mechanical forces generated by ice crystals. Chan et al. observed that compressing the tissue between microscope slides also alters the optical properties,²⁴ and thus we abandoned the usage of microscope slides for the *in vitro* part of this study.

On the other hand, the *in vivo* techniques of this study may have other shortcomings. Insertion of a 100 μm fiber optic into the brain tissue might have caused breaking of microvessels and cellular damage. The interstitial liquid that was oozing slowly at the lower end of the inserted needle was clear, suggesting that microvessel bleeding was minimal. The track left behind as the fiber was being lowered in the cortex was collapsing in each trial, which could be verified under microscopic vision through the dura. Therefore, the integrity of the tissue structure was presumably not compromised by the optical fiber to the level that would affect the reproducibility of the measurements.

The light intensity profiles in a 2-D vertical plane led to some additional observations. The penetration in the direction of the beam is much larger than its spread in the radial direction. This forward biasing of the profile is mostly due to the high anisotropy factor (g) of neural tissue. Another contributing factor might be the directionally biased cone of reception of the optical fiber. The tip of the glass fiber, a flat circular surface, was positioned upward facing the NIR beam. The light photons coming towards the tip in the normal direction can enter the fiber with only a small percentage of them being reflected at the

interface. However, because glass has a higher refractory index than tissue, more and more photons reflect back into tissue with increasing incident angles from normal and all the photons reflect coming tangential to the surface (Fresnel equation). This drawback is difficult to overcome since optical sensors and light guides in general have a nonuniform cone of reception. This directional bias in photon collection might have caused the dark regions on each side of the 2-D plot near the dura, if the light photons are mostly traveling at smaller angles to the horizontal axis in these regions.

A glass surface, however, well mimics the case that we have with the optically activated microstimulators with a silicon nitride coating over the light receiving window.² The main objective of this study was to estimate the light energy that will be collected by the optical microstimulators at various depths of implantation inside the neural tissue. The dark regions in Fig. 4 also suggest that the energy transfer efficiency may be sensitive to horizontal movements of the laser beam with respect to the implanted device in the locations near the surface. On the other hand, experimental results in the rat spinal cord with implants at a depth of 2.35 mm showed that the light intensity needs to be increased only by a factor of four for 1000 μm horizontal displacement of the source (Fig. 9 in Ref. 2). This roughly agrees with the horizontal spread of the intensity plot in Fig. 4(b) taken at a depth of 2.1 mm. In conclusion, the light energy harvested by the implant should be much less sensitive to lateral movements of the source for deeper implant locations due to spreading of the beam in the horizontal direction.

Acknowledgments

This study was supported by NIBIB, National Institute of Health (R01 EB009100). We thank Dr. Bryan Pfister for his help with *in vitro* slice preparations and Dr. Tara Alvarez for commenting on the manuscript.

References

1. A. Abdo and M. Sahin, "Feasibility of neural stimulation with floating-light-activated microelectrical stimulators," *IEEE Trans. Biomed. Circ. Syst.* **5**(2), 179–188 (2011).
2. A. Abdo et al., "Floating light-activated microelectrical stimulators tested in the rat spinal cord," *J. Neural Eng.* **8**(5), 056012 (2011).
3. A. Abdo et al., "Intraspinal stimulation with light activated microstimulators," in *Proc. 5th International IEEE EMBS Conference on Neural Engineering*, pp. 546–549 (2011).
4. R. Biran, D. C. Martin, and P. A. Tresco, "The brain tissue response to implanted silicon microelectrode arrays is increased when the device is tethered to the skull," *J. Biomed. Mater. Res. A* **82**(1), 169–178 (2007).
5. J. F. Hetke et al., "Silicon ribbon cables for chronically implantable microelectrode arrays," *IEEE Trans. Biomed. Eng.* **41**(4), 314–321 (1994).
6. F. W. Koehler et al., "Near infrared spectroscopy: the practical chemical imaging solution," *Spectrosc. Eur.* **14**(3), 12, 14, 16, 18, 19 (2002).
7. K. D. Desmet et al., "Clinical and experimental applications of NIR-LED photobiomodulation," *Photomed. Laser Surg.* **24**(2), 121–128 (2006).
8. Q. Zhang et al., "Study of near infrared technology for intracranial hematoma detection," *J. Biomed. Opt.* **5**(2), 206–213 (2000).
9. Y. Lampl et al., "Infrared laser therapy for ischemic stroke: a new treatment strategy—results of the NeuroThera Effectiveness, and Safety Trial-1 (NEST-1)," *Stroke* **38**(6), 1843–1849 (2007).
10. P. Gurnani, *Near Infrared Spectroscopic Measurement of Human and Animal Brain Structures*, Master, University of Texas at Arlington, Texas (2003).
11. W. F. Cheong, S. A. Prael, and A. J. Welch, "A review of the optical properties of biological tissues," *IEEE J. Quantum Electron.* **26**(12), 2166–2185 (1990).
12. J. S. Dam et al., "Multiple polynomial regression method for determination of biomedical optical properties from integrating sphere measurements," *Appl. Opt.* **39**(7), 1202–1209 (2000).
13. J. D. Bronzino, J. D. Enderle, and S. M. Blanchard, *Introduction to Biomedical Engineering*, 2nd ed., Elsevier, Burlington, MA (2005).
14. R. W. Berg and D. Kleinfeld, "Vibrisa movement elicited by rhythmic electrical microstimulation to motor cortex in the aroused rat mimics exploratory whisking," *J. Neurophysiol.* **90**(5), 2950–2963 (2003).
15. A. N. Yaroslavsky et al., "Optical properties of selected native, and coagulated human brain tissues *in vitro* in the visible, and near infrared spectral range," *Phys. Med. Biol.* **47**(12), 2059–2073 (2002).
16. L. Zhang, A. Shi, and H. Lu, "Determination of optical coefficients of biological tissue from a single integrating-sphere," *J. Mod. Opt.* **59**(2), 121–125 (2012).
17. H. R. Eggert and V. Blazek, "Optical properties of human brain tissue, meninges, and brain tumors in the spectral range of 200 to 900 nm," *Neurosurgery* **21**(4), 459–464 (1987).
18. M. Johns et al., "Determination of reduced scattering coefficient of biological tissue from a needle-like probe," *Opt. Express* **13**(13), 4828–4842 (2005).
19. A. Abdo, B. Pfister, and M. Sahin, "Scattering coefficient and penetration depth of NIR light in the rat brain gray matter," presented at the *Neural Interface Conference*, Salt Lake City, UT (2012).
20. L. V. Wang, "Monte Carlo simulation package: modeling of photon transport in multi-layered tissues," Washington University in St. Louis, MO, <http://labs.seas.wustl.edu/bme/Wang/mc.html>.
21. G. Palumbo and R. Pratesi, *Lasers, and Current Optical Techniques in Biology*, The Royal Society of Chemistry, Cambridge, UK (2004).
22. W. Gottschalk, "Ein messverfahren zur bestimmung der optischen parameter biologischer Gevebe *in vitro*," Ph.D., Universitaet Fridriciana, Karlsruhe (1992).
23. A. Roggan et al., "Determination of optical tissue properties with double integrating sphere technique and Monte Carlo simulations," *Proc. SPIE*, **2100**, 179026 (1994).
24. E. K. Chan et al., "Effects of compression on soft tissue optical properties," *IEEE J. Sel. Top. Quantum Electron.* **2**(4), 943–950 (1996).