Errata: Lateral and axial measurement differences between spectral-domain optical coherence tomography systems

Francisco A. Folgar
Eric L. Yuan
Sina Farsiu
Cynthia A. Toth

Errata: Lateral and axial measurement differences between spectral-domain optical coherence tomography systems

Francisco A. Folgar,^a Eric L. Yuan,^a Sina Farsiu,^{a,b} and Cynthia A. Toth^{a,b}

^aDuke University, Department of Ophthalmology, Durham, North Carolina 27710

^bDuke University, Department of Biomedical Engineering, Durham, North Carolina 27710

[DOI: [10.1117/1.JBO.20.5.059802](https://doi.org/10.1117/1.JBO.20.5.059802)]

This article [*J. Biomed. Opt.* **19**(1), 016014 (2014)] was originally published online on 17 January 2014 with errors in the row and column labels of Tables 4 and 5 on p. 5. The corrected tables are reprinted below.

This article was corrected online on 18 May 2015.

Table 4 Conversion factors for lateral measurements across platforms.

	Convert from this platform				
	Lateral scaling	Zeiss Cirrus™	Heidelberg Spectralis™	Bioptigen Envisu™	Bioptigen SDOIS
Convert to this platform	Zeiss Cirrus™		0.980	0.991	0.994
	Heidelberg Spectralis™	1.020		1.011	1.014
	Bioptigen Envisu™	1.009	0.989		1.003
	Bioptigen SDOIS	1.006	0.986	0.997	

Table 5 Conversion factors for axial measurements across platforms.

	Convert from this platform				
	Axial scaling	Zeiss Cirrus™	Heidelberg Spectralis™	Bioptigen Envisu™	Bioptigen SDOIS
Convert to this platform	Zeiss Cirrus™		1.011	1.037	1.174
	Heidelberg Spectralis™	0.990		1.025	1.160
	Bioptigen Envisu™	0.964	0.975		1.132
	Bioptigen SDOIS	0.852	0.862	0.884	