Journal of Biomedical Optics

BiomedicalOptics.SPIEDigitalLibrary.org

Ilem-spFRET: improved Iem-spFRET method for robust FRET measurement

Jiang Zhang
Fangrui Lin
Liuying Chai
Lichun Wei
Tongsheng Chen

SPIE.

Jiang Zhang, Fangrui Lin, Liuying Chai, Lichun Wei, Tongsheng Chen, "Ilem-spFRET: improved Iem-spFRET method for robust FRET measurement," *J. Biomed. Opt.* **21**(10), 105003 (2016), doi: 10.1117/1.JBO.21.10.105003.

Ilem-spFRET: improved lem-spFRET method for robust FRET measurement

Jiang Zhang, Fangrui Lin, Liuying Chai, Lichun Wei, and Tongsheng Chen*

South China Normal University, College of Life Science, MOE Key Laboratory of Laser Life Science and Institute of Laser Life Science, Guangzhou 510631, China

Abstract. We recently developed a quantitative Förster resonance energy transfer (FRET) measurement method based on emission-spectral unmixing (lem-spFRET). We here developed an improved lem-spFRET method (termed as Ilem-spFRET) for more robust FRET measurement in living cells. First, two background (BG) spectral fingerprints measured from blank living cells are introduced to remove BG and autofluorescence. Second, we introduce a ρ factor denoting the ratio of two molar extinction coefficient ratios (γ) of acceptor to donor at two excitations into Ilem-spFRET for direct measurement of the γ values using a tandem construct with unknown FRET efficiency (E). We performed Ilem-spFRET on our microscope-spectrometer platform to measure the γ values of Venus (V) to Cerulean (C) and the E values of C32V, CVC, VCV, and VCVV constructs, respectively, in living Huh7 cells. For the C32V or CVC cells, the lem-spFRET and Ilem-spFRET methods measured consistent E values. However, for the cells especially with low expressing levels of VCV or VCVV, the E values measured by lem-spFRET showed large deviations and fluctuations, whereas the Ilem-spFRET method greatly improved the measured E values. Collectively, Ilem-spFRET is a powerful and robust tool for quantitatively measuring FRET signal in living cells. © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.21.10.105003]

Keywords: Förster resonance energy transfer; emission-spectral unmixing; fluorescent proteins; extinction coefficient ratio; living cells; microscopy.

Paper 160462R received Jul. 3, 2016; accepted for publication Sep. 22, 2016; published online Oct. 13, 2016.

1 Introduction

Fluorescent proteins (FPs)-based Förster resonance energy transfer (FRET) microscopy has become an important tool for the analysis of protein-protein interactions within living cells in real time.¹⁻⁶ FRET quantification, measurement of absolute FRET efficiency (E) and acceptor-to-donor concentration ratio (R'), allows quantitative analysis of the degree of molecular interaction and is necessary for the scientific communication of FRET index obtained by different research groups.^{7,8} Fluorescence lifetime imaging microscopy (FLIM), acceptor photobleaching (PbFRET), and acceptor-sensitized emission (SE-FRET) methods are commonly used for FRET quantification.^{6,7,9-13}

With improvements in spectral imaging, spectral unmixing has been widely used to quantitatively resolve E and R' , named as the spFRET method, in living cells.^{6,9,14,15} The axiom of spFRET is that the net fluorescence spectrum of a FRET sample is linear superposition of the spectra of each fluorophore contained in the FRET sample.^{16,17} Generally, the spFRET method uses uncorrected spectral fingerprints of donor and acceptor to divide the fluorescence spectrum of a FRET sample into the contributions from donor and acceptor. Therefore, the spFRET method requires at least three reference samples separately expressing donor-only, acceptor-only, and donor-acceptor tandem construct to calibrate the measurement system, and the spectra of FRET samples are generally performed under the same imaging conditions with reference samples, including the

same illumination intensity and instrumental setting.^{6,9,17} This rigorous restriction makes it very difficult to perform a complete FRET quantification within single live cells.

In order to circumvent the preparations of three reference samples separately expressing donor-only and acceptor-only as well as donor-acceptor tandem construct for every FRET measurement, Levy et al.¹⁴ proposed a novel spFRET method that utilizes the corrected spectral fingerprints of donor and acceptor to linearly resolve two corrected emission spectra of a FRET sample with two different excitations. The corrected spectral fingerprints of donor and acceptor are independent of the measurement system; thus, it is unnecessary to measure them for subsequent quantitative FRET measurement when they are predetermined by using this method to perform quantitative FRET measurement on a stable measurement system, such as wide-field fluorescence microscopy that is very stable over at least 3 months.^{11,18} Recently, we developed an independent emission-spectral unmixing-based spFRET method, named Ilem-spFRET, to directly measure the E and R' values of a FRET sample without any external references.¹² Ilem-spFRET circumvents the rigorous restriction of keeping constant measurement conditions for all FRET measurements; thus, it can be performed under the optimal imaging conditions by adjusting the illumination intensity and instrumental settings for cells expressing different FPs levels. Ilem-spFRET requires to predetermine the molar extinction coefficient ratio (γ) of acceptor-to-donor for the correction of acceptor excitation crosstalk.¹² In reality, we used a standard tandem construct with known FRET efficiency previously measured by using the PbFRET method to

*Address all correspondence to: Tongsheng Chen, E-mail: chentsh@scnu.edu.cn and chentsh126@126.com

predetermine γ values.^{12,19} Accurate measurement of the E value is crucial for the measurement of γ and subsequent FRET quantification. However, inapplicability of the PbFRET and FLIM methods for GFP-YFP pair and instability of PbFRET limit the live-cell application of the Iem-spFRET method.

In this report, we developed an improved Iem-spFRET method (termed as Ilem-spFRET) for robust FRET measurement in living cells. First, we introduce two background (BG) spectral fingerprints measured from blank cells without FPs with two different excitations into Iem-spFRET, which improves the anti-interference ability of FRET measurement. Second, we introduce a ρ factor denoting the relationship between two molar extinction coefficients ratios of acceptor to donor at two different excitations into Iem-spFRET, which circumvents predetermining the E value of a tandem construct for the measurement of γ values. We performed Iem-spFRET and Ilem-spFRET methods on our microscope-spectrometer (MSM) platform¹⁸ to measure the E and R' values of tandem constructs with different acceptor-donor stoichiometries in single living Huh7 cells, and our experimental results firmly demonstrated that Iem-spFRET method has better robustness than the Iem-spFRET method especially for the cells with low expression levels of FPs.

2 Theory

2.1 Spectral Unmixing of FRET Samples

For an FRET sample containing free donors at concentration C_d and free acceptors at concentration C_a as well as the paired donor at concentration C_D and acceptor at C_A , the emission spectra should be a linear combination of six contributions from the free donor, paired donor, free acceptor, and paired acceptor, including direct excitation and sensitized emission, as well as BG. Similar to the Iem-spFRET method we recently developed,¹² we can sort these fluorescence components according to these emission characteristics of donor and acceptor as well as BG, respectively

$$F^i(\lambda) = I^i \epsilon_D^i \varphi_D \{ [C_D(1-E) + C_d] e_D(\lambda) + [C_D E \varphi_A / \varphi_D + \gamma^i (C_A + C_a) \varphi_A / \varphi_D] e_A(\lambda) \} + F_B^i(\lambda), \quad (1)$$

where E is the FRET efficiency of the paired donor-acceptor, I^i is the excitation intensity at wavelength λ^i ($i = 1, 2$); ϵ_D^i and ϵ_A^i are the extinction coefficients of donor and acceptor at wavelength λ^i , and $\gamma^i = \epsilon_D^i / \epsilon_A^i$ is the molar extinction coefficient ratio of acceptor to donor at wavelength λ^i ;¹² φ_D and φ_A are quantum yields of donor and acceptor; $e_D(\lambda)$ and $e_A(\lambda)$ normalized to unit area are the spectral fingerprints of donor and acceptor; $F_B^i(\lambda)$ is the spectra of both BG and autofluorescence with λ^i excitation, and is normalized to the unit area as the spectral fingerprints of BG [$e_B^i(\lambda)$]. Equation (1) becomes

$$F^i(\lambda) = \alpha^i e_D(\lambda) + \beta^i e_A(\lambda) + c^i e_B^i(\lambda), \quad (2)$$

where α^i , β^i , and c^i are the weight factors of donor, acceptor, and BG, respectively. We define the weight ratio (δ^i) of donor to acceptor at λ^i excitation

$$\delta^i = \frac{\alpha^i}{\beta^i} = \frac{C_D + C_d - C_D E}{C_D E \varphi_A / \varphi_D + \gamma^i (C_A + C_a) \varphi_A / \varphi_D}. \quad (3)$$

2.2 ρ Factor

The measured emission fluorescence intensity spectra $F_D^{i,\text{ref}}(\lambda)$ and $F_A^{i,\text{ref}}(\lambda)$ for the cells exclusively expressing donors or acceptors, which is essential for predetermining the spectral fingerprints of donor and acceptor, are

$$F_D^{i,\text{ref}}(\lambda) = I^{i,\text{ref}} \epsilon_D^i C_D^{\text{ref}} \varphi_D \eta^i(\lambda) e_D(\lambda), \quad (4)$$

$$F_A^{i,\text{ref}}(\lambda) = I^{i,\text{ref}} \epsilon_A^i C_A^{\text{ref}} \varphi_A \eta^i(\lambda) e_A(\lambda), \quad (5)$$

where the superscript i ($i = 1, 2$) refers to the excitation wavelength and the subscripts D and A denote donor and acceptor, respectively; $I^{i,\text{ref}}$ is the excitation intensity, C_D^{ref} and C_A^{ref} are the concentrations of donor-only and acceptor-only samples, respectively. The functions $\eta^i(\lambda)$ are the collection efficiencies of the instrument used and may be different for different excitation wavelengths due to differences in filters.

The emission spectra of fluorophores in fluid solvents were independent of the excitation wavelength.²⁰ We can define a ρ factor to denote the relationship between two molar extinction coefficient ratios of acceptor to donor at two excitations (λ^2 and λ^1)

$$\rho = \frac{\gamma^2}{\gamma^1} = \frac{\epsilon_D^1 \epsilon_A^2}{\epsilon_D^2 \epsilon_A^1} = \frac{F_D^{1,\text{ref}}(\lambda) / \eta^1(\lambda)}{F_D^{2,\text{ref}}(\lambda) / \eta^2(\lambda)} \bigg/ \frac{F_A^{1,\text{ref}}(\lambda) / \eta^1(\lambda)}{F_A^{2,\text{ref}}(\lambda) / \eta^2(\lambda)} = \frac{\nu_D}{\nu_A}. \quad (6)$$

Here, ν_D and ν_A are the excitation absorption coefficients, which can be obtained by recording the emission intensities of donor-only or acceptor-only sample at two excitations (λ^i , $i = 1, 2$) as follows:

$$\nu_D = \frac{I^{1,\text{ref}} \epsilon_D^1}{I^{2,\text{ref}} \epsilon_D^2} = \frac{F_D^{1,\text{ref}}(\lambda) / \eta^1(\lambda)}{F_D^{2,\text{ref}}(\lambda) / \eta^2(\lambda)}, \quad (7)$$

$$\nu_A = \frac{I^{1,\text{ref}} \epsilon_A^1}{I^{2,\text{ref}} \epsilon_A^2} = \frac{F_A^{1,\text{ref}}(\lambda) / \eta^1(\lambda)}{F_A^{2,\text{ref}}(\lambda) / \eta^2(\lambda)}. \quad (8)$$

2.3 Formulas of Ilem-spFRET

Considering δ^i [Eq. (3)] at two different excitations and ρ factor [Eq. (6)], we can obtain the FRET formulas including the apparent FRET efficiency (E_f) and the ratio (R') of total acceptor concentration C_A' to total donor concentration C_D'

$$E_f = E \frac{C_D}{C_D'} = E \frac{C_D}{C_D + C_d} = \frac{\delta^2 \rho - \delta^1}{\delta^1 \delta^2 (\rho - 1) \varphi_A / \varphi_D + \delta^2 \rho - \delta^1}, \quad (9)$$

$$R' = \frac{C_A'}{C_D'} = \frac{C_A + C_a}{C_D + C_d} = \frac{(\delta^1 - \delta^2) / \gamma^1}{\delta^1 \delta^2 (\rho - 1) \varphi_A / \varphi_D + \delta^2 \rho - \delta^1}. \quad (10)$$

2.4 Predetermination of γ Factor

After ρ factor is predetermined, γ^i can be easily predetermined using a tandem construct with the same donor and acceptor fluorophores. In this case, we define the $R^i = C_A^i/C_D^i = n$. According to Eq. (10), we obtain γ^1

$$\gamma^1 = \frac{(\delta^1 - \delta^2)/n}{\delta^1 \delta^2 (\rho - 1) \varphi_A / \varphi_D + \delta^2 \rho - \delta^1}. \quad (11)$$

3 Materials and Methods

3.1 Reagents and Plasmids

Cerulean and Venus-kras plasmid constructs were purchased from Addgene Company (Cambridge, Massachusetts). The FRET-standard constructs, including C32V (Cerulean-32-Venus, Addgene plasmid 29396), CVC (Cerulean-5-Venus-5-Cerulean, Addgene plasmid 27788), VCV (Venus-5-Cerulean-5-Venus, Addgene plasmid 27788), and VCVV (Venus-5-Cerulean-5-Venus-6-Venus, Addgene plasmid 27789), were kindly provided by the Vogel Lab (National Institutes of Health, Bethesda, Maryland). Plasmid DNA of CFP-Bax and YFP-Bax was kindly provided by Dr. Prehn.⁴ Plasmid DNA of 18AA was kindly supplied by Professor Kaminski.⁷ Staurosporine (STS) was purchased from Sigma-Aldrich Co. LLC (Santa Clara).

3.2 Cell Culture and Transfection

Huh7 and HepG2 cells line were obtained from the Department of Medicine, Jinan University (Guangzhou, China). Cells were cultured in Dulbecco's modified Eagle's medium (DMEM, Gibco, Grand Island) containing 10% fetal calf serum (FCS; Sijiqing, Hangzhou, China) at 37°C in a humidified atmosphere containing 5% CO₂. For transfection, Huh7 or HepG2 cells were cultured in DMEM containing 10% FCS in 35-mm glass dish with a density 4×10^4 cells/ml. After 24 h, all plasmids were transfected into cells for 24 to 48 h when the cells reached 70% to 90%. Turbofect^{TM} (Fermentas Inc., Glen Burnie, Maryland) was used as a transfection reagent.

3.3 Microscope-Spectrometer Platform

FRET measurements were implemented on a user-built MSM platform that consists of a widefield fluorescence microscope (Axio Observer, Carl Zeiss, Oberkochen, Germany) with a $40 \times /1.3$ NA oil objective and a miniature fiber optic spectrometer (QE65 Pro, Ocean Optics, Florida).¹⁸ The MSM platform, equipped with a metal halide lamp (X-Cite 120, Carl Zeiss, Oberkochen, Germany) and AxioVision Rel. 4.8 (Carl Zeiss, Oberkochen, Germany) software for CCD operation as well as SpecSuite (Ocean Optics, Florida) software for spectra detection, has two detection channels: a microscopic imaging channel with a CCD camera (AxioCam MRm, Carl Zeiss, Oberkochen, Germany) for guiding cells and a spectral detection channel with a QE65 Pro spectrometer for collecting spectra from the guided cells. The spectra channel was accorded to the middle of CCD imaging channel. The detection area (S) of the spectrometer is $2.826 \times 10^5 \mu\text{m}^2$ that accords to a $68.99 \mu\text{m}^2$ of the sample area in this report.

The illumination intensity can be attenuated in five discrete steps ($T_1 = 0\%$, 12.5%, 25%, 50%, and 100%), and another neutral density filters controller with six discrete steps

($T_2 = 2\%$, 20%, 40%, 60%, 80%, and 100%) can be used for the same purpose. In our study, the excitation optical filters were BP405/20 (Ex405 nm) (BP 405/20× Exciter, Chroma) and BP 436/20 (Ex436 nm) (BP 436/20, Carl Zeiss, Oberkochen, Germany). 455-nm dichroic mirror (455DCLP Dichroic, Chroma) and emission long pass filter of LP455 (455DCLP Dichroic, Chroma) were used as emission control. The spectral detection range is 460 to 620 nm. Each count(λ) or $E(\lambda)$ at the emission wavelength λ is related to the photos from about a 0.761-nm wavelength range. In addition, excitation bandpass filter of BP 500/25 (Ex500 nm), primary dichroic mirror FT 515, and emission bandpass filter of BP 535/30 (Carl Zeiss, Oberkochen, Germany) were used for the acceptor photobleaching.

4 Results and Discussions

4.1 ρ Factor and Spectral Fingerprints

The lifetime of the metal halide lamp we used for excitation is much longer than 2000 h, and thus our excitation is very stable. Furthermore, spectral responsivity of our MSM platform was carefully calibrated by using a standard lamp (LS-1-CAL, Ocean Optics, Florida), just as described previously.¹⁸ Our MSM platform has the same calibration curve [$K(\lambda) = 1/\eta(\lambda)$] of spectral sensitivity at an interval of 1 month (Appendix A1), further demonstrating the stability of the MSM platform.

BG spectra were measured from 18 blank living Huh7 cells without expressing FPs with 405 and 436 nm excitations, respectively [Figs. 1(a) and 1(b)]. Figure 1(c) shows the images of cells separately expressing Cerulean and Venus with 405 nm (λ^1) and 436 nm (λ^2) excitations, respectively. Figures 1(d) and 1(e) show the spectra of the cell indicated by the red circles in Fig. 1(c) and the average BG spectra in Figs. 1(a) and 1(b). According to Eqs. (7) and (8), the excitation absorption coefficients $v_D = 0.2738$ are calculated from 460 to 560 nm wavelength range [Fig. 1(d)] and $v_A = 0.2090$ calculated from 520 to 550 nm wavelength range [Fig. 1(e)]. Statistical results from 14 cells expressing Cerulean and 13 cells expressing Venus were $v_D = 0.2773 \pm 0.0064$ and $v_A = 0.2051 \pm 0.0322$, respectively. Thus, $\rho = 1.3520$ by Eq. (6). The normalized spectra of Cerulean and Venus from above cells with 436 nm excitation and the normalized BG spectra calculated from 18 blank living Huh7 cells were used as the spectral fingerprints of Cerulean (e_D) and Venus (e_A) as well as BGs (e_B^1 and e_B^2), respectively [Fig. 1(f)].

Although careful calibration of the measurement platform is necessary for Ilem-spFRET and lem-spFRET methods, this step is not required for every FRET measurement due to the stability of our platform for at least 3 months. In addition, because of the independence of γ^i on the illumination intensity and emission-spectral transmission as well as FP concentration,^{6,12,20} predetermining the ρ factor can be a one-off event for given FPs for at least 3 months on our platform. However, most imaging systems do not have this stability, and so investigators had better not skip the calibration step.¹⁸ In addition, as the two excitations used here share the same emission channel ($\eta^1(\lambda) = \eta^2(\lambda)$), the calibration of our system is very simple and can be completed in a few seconds.¹⁸

We found that the normalized spectra of Cerulean (e_D) and Venus (e_A) in living cells with 436 nm excitation were consistent with those with 405 nm excitation.¹⁸ In fact, much of the literature assume that the spectra of FPs in living cells are

Fig. 1 Spectral fingerprints of Cerulean [$e_D(\lambda)$] and Venus [$e_A(\lambda)$] as well as BG [$e_B^i(\lambda)$, $i = 1$ or 2] inside living Huh7 cells. (a) and (b) Spectra of 18 blank cells without FPs at 405 nm (a) ($i = 1$) or 436 nm (b) ($i = 2$) excitation under the same measurement conditions. (c) Images of cells separately expressing Cerulean and Venus with 405 or 436 nm excitation. Scale bar: 10 μm . (d) and (e) Emission spectra of Cerulean (d) and Venus (e) inside the living cell indicated by the red circles in (c) and the average of spectra of 18 cells in (a) and (b), as well as the corresponding v_D and v_A . (f) Normalized spectral fingerprints of donor (e_D) and acceptor (e_A) as well as BG (e_B^1 and e_B^2) at two excitations.

independent of excitation wavelength.^{6,14,15,18,20–23} Therefore, Eq. (6) should be applicable to the FPs in living cells.

As the extinction coefficient of Venus at 405 nm is very low and 436 nm excitation can effectively excite both Cerulean and Venus,²⁴ so the measured normalized spectra of Cerulean and Venus with 436 nm excitation were used as their spectral fingerprints. Once e_D , e_A , and e_B^i are predetermined, we need not measure them for subsequent quantitative FRET measurement. In addition, acquisition range of the spectral fingerprints must overlap well with the emission spectra of donor and acceptor to insure the accurate measurement of Ilem-spFRET.¹⁴ In this study, the spectral ranges covering $\sim 96.3\%$ and $\sim 96.8\%$ of Cerulean and Venus [Fig. 1(f)] emission spectra fulfill this requirement.

Although we introduce ρ factor into Ilem-spFRET, there is no need to add any external references for determining the ρ factor. In addition, ρ factor and spectral fingerprints can be determined at the same time [Figs. 1(d) and 1(e)]. We here predetermine the ρ factor by measuring the excitation absorption coefficients (v_D and v_A), which can be obtained from the fluorescence intensities of cells exclusively expressing donor and acceptor with two excitations [Eqs. (7) and (8)]. The excitation intensity ratio ($I^{1,\text{ref}}/I^{2,\text{ref}}$) between two excitations must be constant for donor-only and acceptor-only reference, respectively. In fact, the $I^{1,\text{ref}}/I^{2,\text{ref}}$ is constant for a stable excitation light source; we thus only need to measure both v_D and v_A under the same instrumental conditions for determining the ρ factor. In addition, v_D and v_A are constant for a given donor–acceptor pair and optical system and are also independent of the emission wavelength [Eqs. (7) and (8)]. In order to improve the accuracy of measurement, we chose the spectra of Cerulean in 460 to 560 nm wavelength range to calculate v_D [Fig. 1(d)] and the

spectra of Venus in 520 to 550 nm wavelength range to calculate v_A [Fig. 1(e)] for the determination of ρ factor (1.3520). Meanwhile, we also used the spectra of both Cerulean and Venus in the same wavelength range of 520 to 550 nm [Figs. 1(d) and 1(e)] to calculate the v_D and v_A , and obtained consistent ρ factor (1.3477).

4.2 Extinction Coefficient Ratios of Venus to Cerulean in Living Huh7 Cells

To predetermine the extinction coefficient ratio (γ^i) of Venus (V) to Cerulean (C) in living Huh7 cells on our platform, which is necessary for Ilem-spFRET measurement, we transfected Huh7 cells with C32V, and the representative cell images with 405 ($i = 1$) or 436 nm ($i = 2$) excitation are shown in Fig. 2(a). The corresponding spectra and unmixing components by using Eq. (2) are shown in Figs. 2(b) and 2(c), and $\delta^1 = 1.4168$ and $\delta^2 = 1.1797$. Substituting these parameters and the predetermined $\rho = 1.3520$ into Eq. (11), we calculate $\gamma^1 = 0.3298$ and $\gamma^2 = 0.4459$, where $\varphi_D = 0.62$, $\varphi_A = 0.57$,^{25,26} and $n = 1$. The statistical γ^1 and γ^2 values from 14 live Huh7 cells by Ilem-spFRET were $\gamma^1 = 0.3055 \pm 0.0108$ and $\gamma^2 = 0.4130 \pm 0.0147$ [Table 1 and Fig. 2(d)]. Meanwhile, we also used the PbFRET method^{7,27} to measure the FRET efficiency (E) of C32V and then determined the γ^i values by using the Ilem-spFRET method [Fig. 2(d) and Appendix A].

The real values of φ_D and φ_A (quantum yield) are the biggest unknown in our calculation due to the dependence of φ_D and φ_A on the cell environment. However, the fact that the E values of FRET constructs in living cells obtained by a φ -dependent method are consistent with those obtained by a φ -independent method demonstrates that the cell environment does not

Fig. 2 γ^j values of Venus to Cerulean in living Huh7 cells. (a) Fluorescence images of representative cells expressing C32V at 405 or 436 nm excitation. Scale bar: 10 μ m. (b) and (c) Spectral unmixing of the corresponding spectra by Ilem-spFRET with 405 (b) or 436 nm (c) excitation in (a), where $\alpha^1 = 0.1315$, $\beta^1 = 0.0897$, $c^1 = 0.0032$, $\alpha^2 = 0.5300$, $\beta^2 = 0.4493$, and $c^2 = 0.0008$. (d) Statistical γ^1 and γ^2 values from at 14 living cells by using lem-spFRET and Ilem-spFRET methods, respectively.

Table 1 γ values calculated by lem-spFRET or Ilem-spFRET method.

Cell	lem-spFRET					Ilem-spFRET			
	δ^1	δ^2	E	γ^1	γ^2	δ^1	δ^2	γ^1	γ^2
1	1.4159	1.1819	0.2767	0.2790	0.3890	1.4168	1.1797	0.3298	0.4459
2	1.3409	1.1361	0.2818	0.3007	0.4057	1.3468	1.1351	0.3102	0.4194
3	1.3669	1.1694	0.2797	0.2935	0.3903	1.3794	1.1649	0.2998	0.4053
4	1.3612	1.1578	0.2848	0.2867	0.3871	1.3656	1.1559	0.2962	0.4005
5	1.3751	1.1614	0.2759	0.2968	0.4022	1.3771	1.1594	0.3079	0.4163
6	1.4232	1.2046	0.2641	0.2983	0.4004	1.4313	1.2013	0.3070	0.4151
7	1.4336	1.2146	0.2572	0.3064	0.4080	1.4471	1.2103	0.3133	0.4235
8	1.4658	1.2409	0.2585	0.2917	0.3914	1.4724	1.2357	0.3008	0.4066
9	1.3177	1.1086	0.2823	0.3102	0.4219	1.3170	1.1074	0.3214	0.4346
10	1.2224	1.0430	0.3143	0.2958	0.4008	1.2259	1.0430	0.3059	0.4135
11	1.4223	1.2124	0.2275*	0.3633	0.4656	1.4365	1.2069	0.3036	0.4105
12	1.3768	1.1791	0.2132*	0.4084*	0.5126*	1.3888	1.1760	0.2917	0.3943
13	1.2717	1.0864	0.2393*	0.4113*	0.5223*	1.2759	1.0856	0.2973	0.4020
14	1.2158	1.0424	0.2105*	0.4958*	0.6133*	1.2180	1.0424	0.2916	0.3943
Mean $\pm$ SD				0.3313 $\pm$ 0.0643	0.4365 $\pm$ 0.0679			0.3055 $\pm$ 0.0108	0.4130 $\pm$ 0.0147

Note: Significantly larger than the average value.

seriously affect the ϕ value (or the ϕ ratio) of FPs.^{6,12,14,15,18,23} Based on our previous studies,^{12,18,23} we also used the ϕ values from a reference for our calculation.^{25,26}

ρ factor not only bridges the two molar extinction coefficient ratios (γ^1 and γ^2) but also can be used to determine the γ values by using a tandem construct with unknown FRET efficiency [Eq. (11)], which is an incremental improvement for the Ilem-spFRET method. Predetermination of γ values can be considered as a one-off event for given FPs and instruments.¹² For the determination of γ values by using a standard FRET construct, the most merit of the Ilem-spFRET method is that we need not measure the FRET efficiency of this construct. Although both the Iem-spFRET and Ilem-spFRET methods obtained consistent γ values, the Ilem-spFRET method showed more stable results [Fig. 2(d) and Table 1], demonstrating the robustness of the Ilem-spFRET method. In contrast to the Iem-spFRET method, Ilem-spFRET circumvents premeasuring the E value of the tandem construct, which not only simplifies

the measurement process of γ values but also makes Ilem-spFRET very applicable to the quantitative measurement of GFP-YFP pair. Accurate measurement of the E values of the reference construct is crucial for the determination of γ by using the Iem-spFRET method (Table 1). In fact, cellular motility of living cells and the degree of acceptor photobleaching may affect the PbFRET measurement,²⁸ thus resulting in the unstable γ values measured by the Iem-spFRET method.

4.3 Implementation of Iem-spFRET and Ilem-spFRET for Cells Separately Expressing C32V, CVC, VCV, and VCVV

We next performed the Iem-spFRET and Ilem-spFRET methods on our platform to measure the E and R^I values of C32V, CVC, VCV, and VCVV inside living Huh7 cells. Figure 3(a) shows images of representative cells separately expressing C32V, CVC, VCV, and VCVV with 405 and 436 nm excitations.

Fig. 3 Implementation of both Iem-spFRET and Ilem-spFRET methods on MSM platform for FRET quantification with representative cells separately expressing C32V, CVC, VCV, and VCVV in living Huh7 cells. (a) Fluorescence images of cells with 405 (upper panels) or 436 nm (lower panels) excitations. Scale bar: 10 μ m. (b) and (e) Spectra normalized to unit area of C32V (b), CVC (c), VCV (d), and VCVV (e) constructs inside the cell indicated by red circles in (a) with 405 and 436 nm excitation, respectively. (f) Statistical E values of C32V, CVC, VCV, and VCVV tandem constructs in living Huh7 cells obtained by Iem-spFRET and Ilem-spFRET methods, respectively. (g) Statistical R^I values obtained by Iem-spFRET and Ilem-spFRET methods from Huh7 cells exclusively expressing C32V, CVC, VCV, and VCVV tandem constructs.

The corresponding normalized spectra are shown in Figs. 3(b)–3(e), and the unmixing weights by using Ilem-spFRET are listed in Table 2. Substitute these unmixing weights into Eqs. (9) and (10) to obtain the corresponding E and R^f values: 29.97% and 0.973 for C32V, 47.66% and 0.521 for CVC, 66.33% and 2.296 for VCV, and 74.86%, and 3.154 for VCVV, respectively (Table 2). The unmixing weights by using Iem-spFRET for the same cells are also listed in Table 2, and the corresponding E and R^f values were 28.36% and 0.994 for C32V, 46.70% and 0.539 for CVC, 60.87% and 2.952 for VCV, and 39.93% and 9.710 for VCVV, respectively.

Statistical E and R^f values measured by Ilem-spFRET were $28.34\% \pm 2.80\%$ and 0.99 ± 0.04 for C32V ($n = 16$), $48.00\% \pm 1.28\%$ and 0.54 ± 0.03 for CVC ($n = 15$), $67.89\% \pm 1.66\%$ and 2.09 ± 0.17 for VCV ($n = 15$), $69.02\% \pm 6.26\%$ and 3.68 ± 1.31 for VCVV ($n = 16$), respectively [Figs. 3(f) and 3(g)]. We also used the Iem-spFRET method to measure the E and R^f values of C32V, CVC, VCV, and VCVV constructs for the same cells, and the calculated E and R^f values were $26.85\% \pm 2.63\%$ and 1.00 ± 0.07 for C32V, $47.48\% \pm 1.54\%$ and 0.54 ± 0.04 for CVC, $60.47\% \pm 3.89\%$ and 3.04 ± 0.45 for VCV, $56.33\% \pm 16.4\%$ and 5.74 ± 3.19 for VCVV, respectively [Figs. 3(b) and 3(c)]. Here, the E values measured by the two methods were basically consistent with $28.67\% \pm 0.95\%$ measured by lux-FRET for C32V,¹⁸ $41.4\% \pm 3.2\%$ for CVC measured by Vogel's laboratory with FLIM method.⁹ Statistical E values of VCV by the Iem-spFRET method are consistent with $64.5 \pm 3.2\%$ measured by Vogel's laboratory with FLIM method.⁹ The R^f values calculated by the two methods are in agreement with the corresponding theoretical values of 1 for C32V, 0.5 for CVC. The statistical R^f

value of VCV by the Iem-spFRET method is also consistent with the theoretical value of 2. However, for VCV and VCVV constructs, the E and R^f values measured by the Iem-spFRET method had large deviations and fluctuations, but the Ilem-spFRET method showed greatly improved measured results [Figs. 3(f) and 3(g)].

For the cells expressing low levels of VCV or VCVV with higher FRET efficiency, the measured E and R^f values by using the Iem-spFRET method showed large deviations and fluctuations, whereas the results measured by using Ilem-spFRET exhibited significant improvements [Figs. 3(f) and 3(g)]. Although the autofluorescence intensity of cells was very low, different autofluorescence intensity of each cell might affect the FRET measurement. When the expression levels of VCV and VCVV are very low, direct subtraction of the BG signal from the measured spectra of cells expressing VCV or VCVV seriously affected the spectra in the 460 to 500 nm wavelength range, especially with 405-nm excitation [Figs. 3(d) and 3(e)]. Generally, the spectra of BG and autofluorescence should be stable; thus, they can be introduced as the BG spectral fingerprints to resolve the measured spectrum of FRET sample for removing the BG and autofluorescence. Ilem-spFRET simply subtracted the averaged BG signal (including BG and autofluorescence) obtained from many cells from the measured fluorescence signal for each cell expressing FPs, whereas Iem-spFRET used the BG spectral fingerprint as a component to linearly unmix the measured spectrum for eliminating the influences of autofluorescence of each cell. Therefore, Ilem-spFRET improves the anti-interference ability and robustness of FRET measurements in living cells [Figs. 3(f) and 3(g)].

Table 2 E and R^f values as well as unmixing parameters by Iem-spFRET or Ilem-spFRET methods.

Construct	Iem-spFRET				Ilem-spFRET					
	δ^1	δ^2	E (%)	R^f	δ^1	δ^2	c^1	c^2	E (%)	R^f
C32V	1.2717	1.0864	28.36	0.994	1.2759	1.0856	0.0049	0.0119	29.97	0.973
CVC	0.8983	0.8258	46.70	0.539	0.8955	0.8230	0.0075	0.0432	47.66	0.521
VCV	0.2682	0.2243	60.87	2.952	0.2683	0.2272	0.0035	0.0137	66.33	2.296
VCVV	0.1807	0.1409	39.93	9.710	0.1597	0.1333	0.0251	0.1571	74.86	3.154

Fig. 4 Spectra of C32V (a) and VCV (b) constructs inside the cells indicated by red circles in Fig. 3(a) as well as BG spectra with 405 and 436 nm excitation, respectively.

4.4 Good Robustness of Ilem-spFRET Method

We next evaluated the effect of fluorescence intensity (or expression level of FPs) on Ilem-spFRET and Ilem-spFRET methods. Figure 4 shows the spectra of the C32V and VCV cells in Fig. 3(a) as well as the BG spectra with 405 and 436 excitations. Compared with the spectra of C32V cells [Fig. 4(a)], the 476-nm

peak of VCV cells is very low, especially for the excitation of 405 nm [Fig. 4(b)], due to the high FRET efficiency of VCV. We chose 470 to 490 nm emission data as $I_0^i(CHO)$, and the relative intensities of the C32V cell in Fig. 4(a) with 405 and 436 excitations at CHO were $I_0^{405} = 42.1$, $I_0^{436} = 177.9$, respectively, and the corresponding E and R^f values measured by using Ilem-spFRET and Ilem-spFRET were consistent (Table 3, cell 5).

Table 3 E and R^f values of C32V and VCV constructs measured by Ilem-spFRET or Ilem-spFRET method.

Cell	C32V						VCV					
			Ilem-spFRET		Ilem-spFRET				Ilem-spFRET		Ilem-spFRET	
	I_0^{405}	I_0^{436}	E (%)	R^f	E (%)	R^f	I_0^{405}	I_0^{436}	E (%)	R^f	E (%)	R^f
1	22.4	95.7	30.49	0.92	31.57	0.92	7.4	27.4	51.41	3.88	66.65	2.02
2	23.6	96.6	25.10	1.11	27.63	1.05	8.5	32.0	52.59	4.15	66.06	2.39
3	25.2	107.0	26.31	0.98	27.85	0.97	8.7	34.4	59.71	3.15	67.48	2.15
4	33.9	143.4	28.97	1.03	30.81	1.00	10.7	42.5	61.42	3.00	69.22	1.96
5	42.1	177.9	28.36	0.99	29.97	0.97	11.1	45.6	63.93	2.78	69.38	2.04
6	50.4	217.9	29.87	0.99	31.77	0.95	12.6	52.6	64.17	2.54	68.93	1.95
Mean $\pm$ SD			28.18 $\pm$ 2.09	1.00 $\pm$ 0.06	29.93 $\pm$ 1.81	0.98 $\pm$ 0.04			58.87 $\pm$ 5.59*	3.25 $\pm$ 0.63*	67.95 $\pm$ 1.42	2.09 $\pm$ 0.17

Note: Significant deviation.

Fig. 5 Ilem-spFRET analysis on Bax oligomerization in single live HepG2 cells. (a) Fluorescence images of a cell coexpressing CFP-Bax and YFP-Bax. Scale bar, 10 μ m. (b) and (c) Spectral unmixing of the corresponding spectra with 405 (b) or 436 nm (c) excitation for the cell in (a) by Ilem-spFRET. (d) Fluorescence images of a cell coexpressing CFP-Bax and YFP-Bax after STS treatment for 4 h. Scale bar, 10 μ m. (e) and (f) Spectral unmixing of the corresponding spectra with 405 (e) or 436 nm (f) excitation for the cell in (d) by Ilem-spFRET.

The relative intensities of the VCV cell in Fig. 4(b) with 405 and 436 excitations at CH0 were $I_0^{405} = 12.6$ and $I_0^{436} = 52.6$, respectively, and Iem-spFRET and Ilem-spFRET obtained different E and R' values (Table 3, cell 6). Just as listed in Table 3, both the Iem-spFRET and Ilem-spFRET methods can obtain consistent results for C32V cells with different fluorescence intensity. However, for VCV cells with different fluorescence intensity, only Ilem-spFRET obtained consistent results, indicating the good robustness of the Ilem-spFRET method to fluorescence intensity.

The measured fluorescence intensity is proportional to the excitation intensity and the expression levels of FPs. To reduce photobleaching, the excitation intensity was controlled with $T \leq 1\%$ in our system (Appendix A3). Therefore, the fluorescence intensity in this report mainly depends on the expression levels of FPs. In addition, we can choose 430 and 450 nm wavelengths as two excitations to enhance the emission intensity of Cerulean and Venus, which may further improve the robustness of the Ilem-spFRET method.

4.5 Analysis on STS-Induced Bax Oligomerization in Living Cells

Bax is a proapoptotic protein required for the process of mitochondrial outer membrane permeabilization.⁴ We here used the Ilem-spFRET method to analyze the staurosporine (STS)-induced Bax oligomerization in living HepG2 cells just as described previously.¹² Figure 5 shows the fluorescence image of representative cells coexpressing CFP-Bax and YFP-Bax (left) and the corresponding fluorescence spectra with 405 nm (middle) or 436 nm (right) excitation. We found that Bax distributed evenly in cytoplasm in healthy cells [Fig. 5(a)] but translocated to mitochondria [Fig. 5(d)]. Ilem-spFRET analysis showed that the corresponding E value was 0.25% for control cells and 35.23% for STS-treated cells. The statistical E value was $1.15 \pm 0.79\%$ for healthy cells ($n = 15$) and $16.02 \pm 12.30\%$ for STS-treated cells ($n = 15$), indicating that STS induced Bax translocation to mitochondria and subsequent oligomerization.

5 Conclusion

We here introduce the spectra of blank cells at two excitations as BG spectral fingerprints to resolve the BG and autofluorescence from the measured spectra of cells expressing FPs, which

significantly enhances the robustness of the Iem-spFRET method. In addition, we also introduce a ρ factor that can be simultaneously obtained with the predetermination of the spectral fingerprints of donor and acceptor for predetermining the extinction coefficient ratios (γ) of acceptor to donor at excitations, which not only circumvents premeasuring the E value of a tandem reference construct but also makes the measurement of γ values more simple and stable. The Ilem-spFRET method significantly enhances the success rate of quantitative FRET measurement in living cells and will expand the application fields of FRET microscopy especially in cell biology.

Appendix: Miscellaneous Formatting Details

A1 Calibration of Our MSM Platform

MSM platform was carefully calibrated by using a standard lamp.¹⁸ Figure 6 shows the corresponding spectral sensitivity calibration curve: $K(\lambda) = E(\lambda)/\text{Count}_{\text{lamp}}(\lambda)$, where $E(\lambda)$ is the reference spectrum of halogen lamp directly measured by the calibrated QE65 Pro spectrometer and $\text{Count}_{\text{lamp}}(\lambda)$ is the spectrum of halogen lamp obtained by our MSM platform. The same $K(\lambda)$ curves were obtained at interval of 1 month [Fig. 2(b)], demonstrating the stability of the MSM platform.

A2 Implementation of Iem-spFRET for the Measurement of γ Values of Venus to Cerulean

Figure 7(a) shows the images of the cells in Fig. 2(a) with 436 or 500 nm excitations before and after partial acceptor photobleaching by using the maximum 500 nm excitation. The corresponding spectra from the cell area indicated by a red circle are shown in Fig. 7(b), and obtained $E = 27.67\%$ by using the Eq. (4.14) in the literature.⁷ By unmixing the corrected spectra of C32V from the cell area indicated by red circle in Fig. 7(b) using Iem-spFRET method, we obtained $\delta^1 = 1.4159$ and $\delta^2 = 1.1819$ [Figs. 7(c) and 7(d)], and the corresponding $\gamma^1 = 0.2790$ and $\gamma^2 = 0.3890$. Statistical γ^1 and γ^2 values from the same 14 live Huh7 cells were $\gamma^1 = 0.3313 \pm 0.0643$ and $\gamma^2 = 0.4360 \pm 0.0679$ [Table 1 and Fig. 2(d)]. Table 1 shows parameters from 14 live Huh7 cells by using Ilem-spFRET and Iem-spFRET methods, respectively.

Fig. 6 Spectral calibration of our spectrometer-microscope platform. (a) Spectral sensitivity calibration curve. (b) Spectra sensitivity calibration curves $K_1(\lambda)$, $K_2(\lambda)$, $K_3(\lambda)$, and $K_4(\lambda)$ measured at intervals of 1 month.

Fig. 7 γ^i values of Venus to Cerulean in living Huh7 cells measured by lem-spFRET method. (a) Fluorescence images of cell in Fig. 2 at 436 or 500 nm excitation before and after partial acceptor photobleaching. Scale bar: 10 μm . (b) Correction spectra from the cell area indicated by a red circle in (a) and the chose emission data of donor channel I_{DD} (CH1: 470 to 490 nm) and acceptor channel I_{AA} (CH2: 525 to 535 nm). Here, $I_{DD} = 153$, $I_{AA} = 219$, $I_{DD}^{\text{post}} = 172$, $I_{AA}^{\text{post}} = 148$, and the corresponding E value calculated by using PbFRET method was 27.67%. (c) and (d) Spectral unmixing of the corresponding spectra of the cells in Fig. 2(a) with 405 nm (c) or 436 nm (d) excitation by lem-spFRET, where $\alpha^1 = 0.1241$, $\beta^1 = 0.0876$, $\alpha^2 = 0.5198$, and $\beta^2 = 0.4398$.

Fig. 8 Dynamical spectra of cells expressing VCVV during continuous excitation with $T = 2.5\%$ (a) and (c) and 1% (b) and (d), respectively. Scale bar: 10 μm .

A3 Photobleaching Effect of Excitation

We found that among C32V, CVC, VCV, and VCVV the acceptor of VCVV was easily photobleached. We thus used cells expressing VCVV construct to evaluate the photobleaching effects of excitations on constructs (Fig. 8). We used 405 or 436 nm excitation to continuously excite cells expressing VCVV construct for 20 s and found that excitation with 2.5% of transmission ($T = 2.5\%$) resulted in a significant acceptor photobleaching [Figs. 8(a) and 8(c)], while excitation with $T = 1\%$ did not induce significant photobleaching [Figs. 8(b) and 8(d)]. Thus, $T \leq 1\%$ of illumination intensity was chosen for following experiments without indicated T value.

Acknowledgments

The authors would like to thank Prof. S. S. Vogel (NIH/NIAAA) for providing C32V, CVC, VCV, and VCVV plasmids. This work was supported by the National Natural Science Foundation of China (NSFC) (Grant Nos. 61527825 and 81471699) and the Science and Technology Plan Project of Guangdong Province (No. 2014B090901060).

References

- G. W. Gordon et al., "Quantitative fluorescence resonance energy transfer measurements using fluorescence microscopy," *Biophys. J.* **74**(5), 2702–2713 (1998).
- E. Galperin, V. V. Verkhusha, and A. Sorkin, "Three-chromophore FRET microscopy to analyze multiprotein interactions in living cells," *Nat. Methods* **1**(3), 209–217 (2004).
- R. N. Day, C. F. Booker, and A. Periasamy, "Characterization of an improved donor fluorescent protein for Förster resonance energy transfer microscopy," *J. Biomed. Opt.* **13**(3), 031203 (2008).
- H. Düssmann et al., "Single-cell quantification of Bax activation and mathematical modelling suggest pore formation on minimal mitochondrial Bax accumulation," *Cell Death Differ.* **17**(2), 278–290 (2010).
- M. Mohsin et al., "Genetically encoded FRET-based nanosensor for in vivo measurement of leucine," *Biosens. Bioelectron.* **50**, 72–77 (2013).
- J. Wlodarczyk et al., "Analysis of FRET signals in the presence of free donors and acceptors," *Biophys. J.* **94**(3), 986–1000 (2008).
- A. D. Elder et al., "A quantitative protocol for dynamic measurements of protein interactions by Förster resonance energy transfer-sensitized fluorescence emission," *J. R. Soc. Interface* **6**(S1), S59–S81 (2009).
- A. Zeug et al., "Quantitative intensity-based FRET approaches—a comparative snapshot," *Biophys. J.* **103**(9), 1821–1827 (2012).
- C. Thaler et al., "Quantitative multiphoton spectral imaging and its use for measuring resonance energy transfer," *Biophys. J.* **89**(4), 2736–2749 (2005).
- H. Wallrabe and A. Periasamy, "Imaging protein molecules using FRET and FLIM microscopy," *Curr. Opin. Biotechnol.* **16**(1), 19–27 (2005).
- T. Zal and N. R. Gascoigne, "Photobleaching-corrected FRET efficiency imaging of live cells," *Biophys. J.* **86**(6), 3923–3939 (2004).
- J. Zhang et al., "Quantitative FRET measurement using emission-spectral unmixing with independent excitation crosstalk correction," *J. Microsc.* **257**(2), 104–116 (2015).
- L. Zhang et al., "Binomial distribution-based quantitative measurement of multiple-acceptors fluorescence resonance energy transfer by partially photobleaching acceptor," *Appl. Phys. Lett.* **104**(24), 243706 (2014).
- S. Levy et al., "SpRET: highly sensitive and reliable spectral measurement of absolute FRET efficiency," *Microsc. Microanal.* **17**(2), 176–190 (2011).
- S. Mustafa et al., "Quantitative Förster resonance energy transfer efficiency measurements using simultaneous spectral unmixing of excitation and emission spectra," *J. Biomed. Opt.* **18**(2), 026024 (2013).
- T. Zimmermann, "Spectral imaging and linear unmixing in light microscopy," *Adv. Biochem. Eng. Biotechnol.* **95**, 245–265 (2005).
- A. D. Hoppe et al., "N-way FRET microscopy of multiple protein-protein interactions in live cells," *PLOS One* **8**(6), e64760 (2013).
- L. Chai et al., "Miniature fiber optic spectrometer-based quantitative fluorescence resonance energy transfer measurement in single living cells," *J. Biomed. Opt.* **20**(3), 037008 (2015).
- J. Zhang et al., "Spectral measurement of acceptor-to-donor extinction coefficient ratio in living cells," *Micron* **68**, 98–106 (2015).
- J. R. Lakowicz, *Principles of Fluorescence Spectroscopy*, 3rd ed., Springer, Singapore (2006).
- R. Y. Tsien, "The green fluorescent protein," *Annu. Rev. Biochem.* **67**, 509–544 (1998).
- M. A. Rizzo et al., "Optimization of pairings and detection conditions for measurement of FRET between cyan and yellow fluorescent proteins [J]," *Microsc. Microanal.* **12**(03), 238–254 (2006).
- M. Y. Du et al., "Wide-field microscopic FRET imaging using simultaneous spectral unmixing of excitation and emission spectra," *Opt. Express* **24**(14), 16037–16051 (2016).
- N. C. Shaner, P. A. Steinbach, and R. Y. Tsien, "A guide to choosing fluorescent proteins," *Nat. Methods* **2**(12), 905–909 (2005).
- M. A. Rizzo et al., "An improved cyan fluorescent protein variant useful for FRET," *Nat. Biotechnol.* **22**, 445–449 (2004).
- T. Nagai et al., "A variant of yellow fluorescent protein with fast and efficient maturation for cell-biological applications," *Nat. Biotechnol.* **20**, 87–90 (2002).
- H. Yu et al., "An empirical quantitative fluorescence resonance energy transfer method for multiple acceptors based on partial acceptor photobleaching," *Appl. Phys. Lett.* **100**(25), 253701 (2012).
- H. Yu et al., "Ma-PbFRET: multiple acceptors FRET measurement based on partial acceptor photobleaching," *Microsc. Microanal.* **19**(01), 171–179 (2013).

Jiang Zhang received his MS degree in optics from South China Normal University. His current research interests include quantitative FRET imaging. He is now working at the China National Infrared and Galvanothermy Product Quality Supervision and Inspection Center.

Fangrui Lin is studying for his MS degree at South China Normal University. His current research interests include fluorescence microscopy and biomedical optics.

Liuying Chai received her MS degree in optics from South China Normal University. She is studying for her PhD at Shanghai Jiao Tong University. She is the author of five journal papers.

Lichun Wei is studying for her MS degree at South China Normal University. Her current research interests include fluorescence microscopy and biomedical optics.

Tongsheng Chen received his PhD in biomedical engineering from Huazhong University of Science and Technology, China. He has 18 years of experience in fluorescence microscopy and biomedical optics. He is the author of more than 120 journal papers and is also a member of SPIE.