Journal of Biomedical Optics

BiomedicalOptics.SPIEDigitalLibrary.org

Ultrafast optical property map generation using lookup tables

Joseph Angelo
Christina R. Vargas
Bernard T. Lee
Irving J. Bigio
Sylvain Gioux

SPIE.

Joseph Angelo, Christina R. Vargas, Bernard T. Lee, Irving J. Bigio, Sylvain Gioux, "Ultrafast optical property map generation using lookup tables," *J. Biomed. Opt.* **21**(11), 110501 (2016), doi: 10.1117/1.JBO.21.11.110501.

Ultrafast optical property map generation using lookup tables

Joseph Angelo,^{a,b} Christina R. Vargas,^c
Bernard T. Lee,^c Irving J. Bigio,^{b,d} and
Sylvain Gioux^{c,e,*}

^aBeth Israel Deaconess Medical Center, Department of Medicine, 330 Brookline Avenue, Boston, Massachusetts 02215, United States

^bBoston University, Department of Biomedical Engineering, 44 Cummington Mall, Boston, Massachusetts 02215, United States

^cBeth Israel Deaconess Medical Center, Department of Surgery, 330 Brookline Avenue, Boston, Massachusetts 02215, United States

^dBoston University, Department of Electrical Engineering, 44 Cummington Mall, Boston, Massachusetts 02215, United States

^eUniversity of Strasbourg, ICube Laboratory, 300 Boulevard Sébastien Brant, 67412 Illkirch, France

Abstract. Imaging technologies working in the spatial frequency domain are becoming increasingly popular for generating wide-field maps of optical properties, enabling rapid analysis of tissue parameters. While acquisition methods have become faster and are now performing in real-time, processing methods remain slow, precluding real-time display of information. We present solutions that rapidly solve the inverse problem for extracting optical properties by use of advanced lookup tables (LUTs). We present methods and results based on a dense, linearly sampled lookup table and an analytical representation that generate maps of absorption and reduced scattering in ~ 10 ms, which is 100x faster than the standard method, with $\leq 4\%$ error compared to the Monte-Carlo simulation. Combined with real-time acquisition methods, the proposed techniques enable video-rate feedback of real-time property maps, enabling full video-rate guidance in the clinic. © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.21.11.110501]

Keywords: spatial frequency domain; lookup table; real-time; diffuse reflectance; imaging.

Paper 160565LR received Aug. 18, 2016; accepted for publication Nov. 7, 2016; published online Nov. 30, 2016.

Rapid quantitative imaging of tissue optical properties, namely absorption (μ_a) and reduced scattering (μ'_s), has long been a challenge in the field of biomedical optics. The recent introduction of spatial frequency domain imaging (SFDI) provides a transformative approach capable of measuring optical properties over a large field-of-view.¹ In principle, SFDI relies on the analysis of the tissue response to structured illumination (patterns of light) in the spatial frequency domain in a multipixel manner over an entire field-of-view at once.^{2,3} More specifically, the spatial-frequency-dependent response, called the modulation transfer function, is calibrated using a phantom with known optical properties to determine the tissue diffuse reflectance

(R_d), which is then used to extract the optical properties using a light propagation model.

While the method is rapid, SFDI has until recently involved the measurement of several images to form maps of optical properties, typically a total of six images (two spatial frequencies at three phases), although a method employing only a single spatial frequency and three phases has also been described. Recent developments in acquisition methods reduced the number of images necessary to extract optical properties from 6 to 2,⁴ and even from 6 to a single image with the method called single snapshot of optical properties (SSOP).⁵ In their most advanced implementations, such methods enable profile-corrected measurements of optical properties in a single snapshot, in turn facilitating true real-time acquisition of optical properties.⁶

Unfortunately, while acquisition methods can be performed in real time, the processing is still commonly achieved postacquisition. The most rapid processing method employs a lookup table (LUT) approach, by which solutions for diffuse reflectance are generated from a light propagation model (diffusion, Monte Carlo) or empirically with various spatial frequencies and optical properties.^{3,7,8} Such a method allows one to directly link a calibrated diffuse reflectance measurement at known spatial frequencies to a unique solution of optical properties (μ_a and μ'_s). However, because it necessitates interpolations within the LUT for each pixel in the image, this method is still time-consuming (seconds), preventing the use of SFDI for true real-time imaging of optical properties.

In this letter, we present forms of LUTs allowing rapid extraction of optical properties from the measurement of calibrated diffuse reflectance. We propose two solutions that do not necessitate interpolations and thus dramatically reduce computation time. The first method consists of a hyper-dense linearized LUT, and the second invokes an analytical representation of the LUT. These methods are described, implemented, and compared to the standard LUT method in terms of precision, accuracy, and computation time. Combined with real-time acquisition, this letter facilitates real-time quantitative optical imaging of tissue properties.

The two methods described in this letter are based on a previously developed LUT.³ This standard LUT method uses “white” Monte-Carlo simulation (WMC, with zero absorbance) to model the spatially resolved impulse response, i.e., the steady-state diffuse reflectance $R_d(\rho)$, of a collimated point-source illumination for a given set of μ_a , μ'_s , n , and g , which is then Fourier transformed to determine the spatial frequency response of the diffuse reflectance.^{9,10} To generate the standard LUT, WMC was used to simulate 10^7 photons into a homogeneous medium with index of refraction $n = 1.4$ and anisotropy factor $g = 0.9$. The model used a detector with a numerical aperture of 0.22 and the radial bins were sampled in increments of $\Delta\rho = 0.09$ mm, allowing a maximum spatial frequency of over 5 mm^{-1} . This process is repeated over several optical properties, and stored in a table that associates the diffuse reflectance at two spatial frequencies with a unique pair of optical properties. Recovering optical properties, therefore, involves measuring the diffuse reflectance of the sample and searching the table to find the corresponding optical property values (hence the name “lookup table”). With this method, it is important to understand that the optical properties are linearly sampled, so we refer to this method as the linear OP LUT. Using this method, the

*Address all correspondence to: Sylvain Gioux, E-mail: sgiox@unistra.fr

measured R_d values are used to search through the nonlinear mapping of R_d against the tissue properties in the linear OP LUT table, and then μ_a and μ'_s are acquired using cubic spline interpolation. For our linear OP LUT, the resolution of the sampled optical properties was 0.001 and 0.01 mm⁻¹ with a range of [0, 0.3] mm⁻¹ and [0.3, 3] mm⁻¹ for μ_a and μ'_s , respectively, with a table size of 81,571 points (see linear OP LUT in Fig. 1).

To avoid the time-consuming searching and interpolation steps required for the standard linear OP LUT, an LUT linear with respect to R_d (linear R_d LUT) was made by interpolating linearly sampled R_d values from 0 to 1 by 500 points within a high-resolution linear OP LUT (0.0001 and 0.001 mm⁻¹ resolution with a range of [0, 0.3] mm⁻¹ and [0.3, 3] mm⁻¹ for μ_a and μ'_s , respectively, with a table size of 8,105,701 points). The resolution of the R_d linear sampling is chosen so that measured R_d values can be found directly in the table by rounding their value to the nearest increment of 0.02, without interpolation (see linear R_d LUT in Fig. 1). Given a sample's R_d values, the LUT's corresponding indices can be calculated by parameterizing the linear R_d grid and rounding to the nearest vertex to acquire μ_a and μ'_s .

Using the linear R_d LUT, each optical property table was fit to a two-dimensional (2-D) function using a nonlinear least squares solver ("fit" using MATLAB)

$$\mu_a(\text{DC}, \text{AC}) = a_1 e^{a_2 \text{DC} + a_3 \text{AC} + a_4} + b_1 e^{b_2 \text{DC} + b_3 \text{AC} + b_4} + c_1 e^{c_2 \text{DC} + c_3} + d_1 e^{d_2 \text{AC}^3 + d_3 (\text{DC} + d_4) \text{AC}^2 + d_5 \text{DC} + d_6 \text{AC} + d_7}, \quad (1)$$

$$\mu'_s(\text{DC}, \text{AC}) = a_1 e^{a_2 \text{AC} + a_3} + b_1 e^{b_2 \text{AC} + b_3} + c_1 e^{c_2 \text{AC} + c_3} + d_1 e^{d_2 \text{AC} + d_3} + (d_1 + \text{AC}) e^{d_2 \text{DC}^2 + d_3 \text{DC} + d_4}, \quad (2)$$

where a_1 through d_4 are optimized parameters. Using these functions, optical properties can be directly deduced from the

diffuse reflectance at low spatial frequency (noted DC here) and the diffuse reflectance at high spatial frequency (noted AC here). This is referred to as the 2-D fit LUT (see 2-D fit LUT in Fig. 1).

Figure 1 summarizes the LUT formations and workflows. The white arrows describe the formation process from Monte-Carlo simulation to the linear OP LUT (arrow 1), to the linear R_d LUT (arrow 2), and finally to the 2-D fit LUT (arrow 3). The black arrows indicate the three options for workflow when using the LUTs to extract optical properties. Diffuse reflectance measurements at two spatial frequencies (low: DC, and high: AC) are used as inputs to the LUTs to extract the optical properties.

The custom imaging system and associated processing have been described extensively and results have been published, including a clinical trial.^{11,12} Briefly, the system utilizes a digital micromirror device (DMD) to project patterns of 670-nm laser illumination which is cross polarized with the collection optics to minimize specular reflections in images collected on a CCD camera.

An array of tissue-like silicone phantoms was made using India ink (Blick Art Materials, Boston, Massachusetts) and titanium dioxide (Atlantic Equipment Engineers, Bergenfield, New Jersey) for absorption and scattering, respectively. The array was made to span the optical properties $\mu_a = [0.01, 0.1]$ mm⁻¹ and $\mu'_s = [0.5, 2]$ mm⁻¹. The array was imaged using SFDI and the inversion from R_d to optical properties was then evaluated on all three LUT methods and compared to a Monte Carlo (MC) least-squares solver over square ROIs of 15 × 15 pixels.⁹

Finally, a Yorkshire pig was used as an *in vivo* model to evaluate each LUT method during realistic surgical conditions (Beth Israel Deaconess Medical Center Institutional Animal Care and Use Committee approved institutional protocol #034-2013). Abdominal skin flaps were elevated on a pair of perforator vessels, the venous pedicle was occluded, and released to restore tissue perfusion.

Fig. 1 LUT protocols: white arrows indicate the formation of each LUT, from Monte-Carlo simulation, to linear R_d sampling, and finally function fitting. Black arrows indicate three options for data flow and usage of each LUT. First, a sample's R_d (DC and AC) is measured and used as input. (a) The linear OP LUT must search and interpolate to generate optical properties, (b) the linear R_d LUT uses parametric indexing to recall optical properties, and (c) the 2-D fit LUT evaluates the function F to generate optical properties.

All computational processing and analyses were done on a 64-bit Dell Optiplex 9020 (Dell, Round Rock Texas) with an Intel Core i7-4770 CPU (Intel, Santa Clara, California). R_d inversions were repeated 100 times for each LUT type and averaged for comparison.

Figure 2 shows the results for the phantom sample array. Agreement is seen among all three LUTs for both absorption and reduced scattering maps [Fig. 2(a)]. ROIs were chosen to quantify the results for each phantom well [Fig. 2(a)] and to compare the LUT methods [Fig. 2(b)]. All methods show low relative error compared to the MC results ($\mu_a\%$, $\mu_s'\%$): linear OP LUT (1.3, 0.5), linear R_d LUT (0.2, 0.09), 2-D fit LUT (1.9, 2.8). Table 1 summarizes the processing times for each method on this sample, from a single pixel to the entire image. The linear R_d and 2-D fit LUTs exhibit better performances over the entire range. Importantly, processing shows a 100-fold improvement for an entire image (572×612 pixels), for both absorption and reduced scattering.

The *in vivo* results are shown in Fig. 3. All LUT methods show strong agreement for both absorption and reduced scattering maps [Fig. 3(a)]. A line profile was taken across the sample [dotted lines in Fig. 3(a)] to quantitatively compare the LUT results [Fig. 3(b)]. All LUTs show agreement with the MC results ($\mu_a\%$, $\mu_s'\%$): linear OP LUT (4.1, 2.1), linear R_d LUT (0.9, 0.6), 2-D fit LUT (1.5, 3.2).

Both newly developed LUT methods, the linear R_d LUT and 2-D fit LUT, decrease the processing time by a factor of $\sim 100\times$ by avoiding searching and interpolating with two different approaches. The linear R_d LUT parameterizes a linear grid using the table's sampling resolution and range and rounds to the nearest vertex that matches the sample's R_d input. The 2-D fit method avoids this discrete sampling by fitting a continuous function to each surface of μ_a and μ_s' as a function of diffuse reflectance at two spatial frequencies.

Each LUT method has its own unique relationship among inversion speed, memory space, and accuracy. The standard linear OP LUT must use interpolation for practically sized tables, which increases the required processing time. Because this table samples R_d in a nonlinear manner, increasing the sampling grid greatly increases the searching time for each input. In comparison, the linear R_d LUT is sampled with a parameterized grid, thus its speed is minimally affected by increased grid size because no searching occurs. However, because the table is linearly sampled, the density required to maximize accuracy at

Table 1 Processing speeds for LUT inversion of R_d to both μ_a and μ_s' property maps.

Number of pixels	Linear OP LUT (s)	Linear R_d LUT (s)	2-D fit LUT (s)
1	0.2814	5.207×10^{-5}	5.117×10^{-4}
10×10	0.4801	1.107×10^{-4}	5.205×10^{-4}
100×100	0.5120	5.801×10^{-4}	1.174×10^{-3}
572×672	2.492	0.0175	0.01881

high gradients in the table is used for the entire table, resulting in a large grid that requires considerable computer memory. This could potentially be mitigated by choosing a sampling function whose density increases for higher gradients of the table. The 2-D fit LUT requires less computer memory space but has variable accuracy across the table due to under-fitting, though this accuracy could potentially be weighted toward optical property values that are more likely to occur, e.g., separate LUTs for brain, liver, or skin tissues. While the presented 2-D fit LUT was fit to the entire linear R_d LUT, accuracy was weighted to prioritize optical properties reasonably expected for skin tissues ($\mu_a = [0.0035 \text{ } 0.192] \text{ mm}^{-1}$, $\mu_s' = [0.3 \text{ } 2.281] \text{ mm}^{-1}$). Within this range of optical properties, errors in comparing the 2-D fit LUT with ground truth does not exceed 0.004 mm^{-1} in absorption and 0.1072 mm^{-1} in scattering.

The new frameworks introduced can be applied to other LUT processes with possibly more dimensions. For profile corrected SFDI¹¹ or three-dimensional-SSOP,⁶ there is the potential to include height/phase dependence among the table's input parameters for optical property mapping. In addition, it is important to note that to make multispectral measurements, each wavelength is processed separately with the same LUT. Because wavelengths can be independently processed in parallel, processing time is minimally affected. Given a known set of wavelengths, one could also go directly from R_d input values to hemodynamic values such as oxygen saturation. Generating these calculations after calibrating the imaging system but before the sample acquisition makes real-time feedback for surgical guidance possible.

Fig. 2 Phantom array measurement. (a) Tissue-simulating phantom array was measured using SFDI and absorption (top row) and reduced scattering (bottom row) maps were generated using the linear OP, linear R_d , and 2-D fit LUTs. (b) Regions of interest (dotted squares) were chosen to compare each method's accuracy. Image resolution is 512×672 pixels. Scale bar represents 1 cm.

Fig. 3 *In vivo* measurement. (a) A pig skin flap vascular occlusion model was measured using SFDI and optical property maps were generated using each LUT. (b) Line profiles (dotted lines) compare each method's accuracy. Scale bar represents 1 cm.

In this letter, we introduced a new framework for LUT formation and evaluated its performance on tissue-mimicking phantoms and *in vivo* in comparison with a standard OP LUT.³ Overall, the new techniques were accurate compared to MC simulations, i.e., within 0.9% and 0.6% for linear R_d LUT μ_a and μ'_s , respectively, and within 1.9% and 3.2% for 2-D fit LUT μ_a and μ'_s , respectively, and are 100× faster than the standard linear OP LUT. These techniques help enable real-time image-guidance feedback for spatial frequency domain techniques.

The authors would like to thank Steve McClure for his helpful discussion and support. This work was supported by R01-DE-022820 (BL), K01-DK-093603, France Life Imaging, University of Strasbourg and ICube Laboratory (S.G.), and NIH/NIDDK Award Number F31-DK-105839 (J.A.).

Disclosures

No conflicts of interest, financial or otherwise, are declared by the authors.

References

1. N. Dognitz and G. Wagnieres, "Determination of tissue optical properties by steady-state spatial frequency-domain reflectometry," *Laser Med. Sci.* **13**, 55–65 (1998).
2. D. J. Cuccia et al., "Modulated imaging: quantitative analysis and tomography of turbid media in the spatial-frequency domain," *Opt. Lett.* **30**(11), 1354–1356 (2005).

3. D. J. Cuccia et al., "Quantitation and mapping of tissue optical properties using modulated imaging," *J. Biomed. Opt.* **14**(2), 024012 (2009).
4. K. P. Nadeau et al., "Advanced demodulation technique for the extraction of tissue optical properties and structural orientation contrast in the spatial frequency domain," *J. Biomed. Opt.* **19**, 056013 (2014).
5. J. Vervandier and S. Gioux, "Single snapshot imaging of optical properties," *Biomed. Opt. Express* **4**(12), 2938 (2013).
6. M. van de Giessen et al., "Real-time, profile-corrected single snapshot imaging of optical properties," *Biomed. Opt. Express* **6**(10), 4051–62 (2015).
7. G. M. Palmer and N. Ramanujam, "Monte Carlo-based inverse model for calculating tissue optical properties. Part I: theory and validation on synthetic phantoms," *Appl. Opt.* **45**(5), 1062–1071 (2006).
8. N. Rajaram et al., "Lookup table-based inverse model for determining optical properties of turbid media," *J. Biomed. Opt.* **13**(5), 050501 (2008).
9. A. Kienle and M. S. Patterson, "Determination of the optical properties of turbid media from a single Monte Carlo simulation," *Phys. Med. Biol.* **41**(10), 2221–2227 (1996).
10. J. Swartling et al., "Accelerated Monte Carlo models to simulate fluorescence spectra from layered tissues," *J. Opt. Soc. Am. A* **20**(4), 714–727 (2003).
11. S. Gioux et al., "Three-dimensional surface profile intensity correction for spatially modulated imaging," *J. Biomed. Opt.* **14**(3), 034045 (2009).
12. S. Gioux et al., "First-in-human pilot study of a spatial frequency domain oxygenation imaging system," *J. Biomed. Opt.* **16**(8), 086015 (2011).