Journal of Biomedical Optics

BiomedicalOptics.SPIEDigitalLibrary.org

Reflective Fourier ptychography

Shaun Pacheco
Guoan Zheng
Rongguang Liang

SPIE.

Reflective Fourier ptychography

Shaun Pacheco,^a Guoan Zheng,^{b,c} and Rongguang Liang^{a,*}

^aUniversity of Arizona, College of Optical Sciences, 1630 East University Boulevard, Tucson, Arizona 85721, United States

^bUniversity of Connecticut, Biomedical Engineering, 260 Glenbrook Road, Storrs, Connecticut 06269, United States

^cUniversity of Connecticut, Electrical and Computer Engineering, 371 Fairfield Way, Storrs, Connecticut 06269, United States

Abstract. The Fourier ptychography technique in reflection mode has great potential applications in tissue imaging and optical inspection, but the current configuration either has a limitation on cut-off frequency or is not practical. By placing the imaging aperture stop outside the illumination path, the illumination numerical aperture (NA) can be greater than the imaging NA of the objective lens. Thus, the cut-off frequency achieved in the proposed optical system is greater than twice the objective's NA divided by the wavelength ($2NA_{\text{obj}}/\lambda$), which is the diffraction limit for the cut-off frequency in an incoherent epi-illumination configuration. We experimentally demonstrated that the synthesized NA is increased by a factor of 4.5 using the proposed optical concept. The key advantage of the proposed system is that it can achieve high-resolution imaging over a large field of view with a simple objective. It will have a great potential for applications in endoscopy, biomedical imaging, surface metrology, and industrial inspection. © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.21.2.026010]

Keywords: microscopy; Fourier optics and signal processing; coherence imaging.

Paper 150778R received Nov. 17, 2015; accepted for publication Jan. 26, 2016; published online Feb. 18, 2016.

1 Introduction

Fourier ptychography (FP) is a recently developed imaging technique that iteratively stitches together a number of variably illuminated low-resolution (LR) images in Fourier space to recover a large field of view (FOV), high-resolution (HR) image of a complex object.¹ Essentially, FP is a coherent imaging technique that utilizes oblique plane wave illumination to shift high spatial frequency content into the passband of the imaging system, similar to the concept of synthetic aperture technique.^{2–6} FP can also be considered a phase retrieval technique that recovers the lost phase information from the irradiance measurements.^{7–10} Recently, the FP technique has been demonstrated for quantitative phase imaging,¹¹ refocusing,¹² recovering unknown pupil aberrations,¹³ and three-dimensional intensity and phase imaging.^{14,15} The acquisition speed and data requirements can be significantly improved without sacrificing image quality by utilizing a multiplexed illumination strategy,^{16,17} optimizing the sampling pattern of the light emitting diode (LED) source,¹⁸ or utilizing content adaptive illumination.¹⁹ The FP reconstruction algorithm without using an alternating projection approach has been demonstrated.^{20,21} However, in most previous demonstrations, FP was implemented in a transmission mode, which imposes a serious limitation on its applications.

To meet the practical needs for tissue imaging and surface inspection, we have been investigating reflective FP.²² However, in our previous reported system, the cut-off frequency is still limited to twice the objective's NA divided by the wavelength ($2NA_{\text{obj}}/\lambda$). We can add LEDs around the objective to increase the illumination angle, therefore increasing the cut-off frequency using FP, but this type of system is bulky and difficult to implement. Most importantly, there is potentially a large gap in illumination angle between the light from the LEDs imaged onto the aperture stop and the LEDs arranged around the objective. This

gap in illumination angles may not appropriately sample the object's Fourier spectrum for the FP reconstruction algorithm.

The key challenge in reflection mode is effectively illuminating the object using large angles since the aperture stop of the objective determines the NA of the imaging channel, while also limiting the illumination NA. We propose a solution by placing the imaging aperture stop outside the illumination path, which allows the illumination NA to be much greater than the imaging NA of the objective. Thus, the cut-off frequency achieved using FP in the proposed optical concept is greater than $2NA_{\text{obj}}/\lambda$, which is the theoretical limit for the cut-off frequency with an epi-illumination configuration. The proposed optical system will find many applications in endoscopy, biomedical imaging, surface metrology, and industrial inspection.

2 Theory

Consider an object with a complex spatial profile $o(x, y)$, which is coherently illuminated by an oblique plane wave at wavelength λ . The irradiance $I(x, y)$ is given by

$$I(x, y) = C_I |o(x, y) e^{i2\pi(\sin \theta_x/\lambda + \sin \theta_y/\lambda)} * h(x, y)|^2, \quad (1)$$

where $h(x, y)$ is the coherent point spread function, θ_x and θ_y are the angles of the illuminating plane waves in the x and y directions measured from the optical axis, and C_I is a constant. The irradiance in terms of the Fourier transforms of the complex object and coherent point spread function is

$$I(x, y) = C_I \left| \mathbf{F}_{\xi, \eta}^{-1} \left[O(\xi, \eta) H \left(\xi - \frac{\sin \theta_x}{\lambda}, \eta - \frac{\sin \theta_y}{\lambda} \right) \right] \right|^2, \quad (2)$$

where $\mathbf{F}^{-1}$ is the inverse Fourier transform operator with respect to spatial frequency coordinates ξ and η , O is the Fourier

*Address all correspondence to: Rongguang Liang, E-mail: rliang@optics.arizona.edu

Fig. 1 Illumination conditions for (a1) normal illumination, (a2) brightfield illumination, and (a3) darkfield illumination. Images of USAF target under (b1) normal illumination, (b2) brightfield illumination, and (b3) darkfield illumination. (c) Fourier spectrum of object with regions that correspond to (b1)–(b3) identified.

transform of the complex object, and H is the coherent transfer function (CTF). For a diffraction-limited system with a circular entrance pupil, the CTF is a binary filter with $H = 1$ when the radius is less than or equal to NA/λ , where NA is the numerical aperture of the imaging objective. If an object is illuminated by an oblique plane wave, the transfer function is shifted in Fourier space. Spatial frequencies that are previously attenuated by the CTF are now shifted into the passband of the imaging system. Thus, oblique plane wave illumination can be used to sample a separate region of the object's spectrum in Fourier space.

In a reflective imaging system, a convenient way to produce the oblique illuminating plane waves at the sample plane is to place an image of a point source at the back focal plane of the objective lens. Since the point source is positioned at the back focal plane of the objective, the output from the objective is an oblique collimated wave. A small LED imaged onto the back focal plane of the objective can be used to approximate a point source. This method approximates the plane waves necessary for FP. Using this set-up in air, the illumination NA_{ill} is

$$NA_{ill}(\theta) = \sin \theta_y = -y_{LED} / \sqrt{y_{LED}^2 + f^2}, \quad (3)$$

where y_{LED} is the position of the LED in the back focal plane of the objective and f is the focal length of the objective lens. Figure 1 shows the three illumination conditions and shows the position of their corresponding region in Fourier space. Figure 1(a1) shows the set-up for a normally illuminated plane wave in reflection. The image of an LED is positioned at the center of the entrance pupil of the objective. An image of a normally illuminated United States Air force (USAF) target is shown in Fig. 1(b1). Figure 1(a2) shows the illumination

conditions for brightfield illumination in a reflective imaging system. The illuminating plane wave NA_{ill} is less than the objective NA_{obj} . Figure 1(b2) shows an example of a brightfield image from a vertically tilted plane wave. Note vertical spatial frequencies in elements 4 to 6 of group 5 are now visible, which were not visible in Fig. 1(b1). Figure 1(a3) shows the illumination conditions for darkfield imaging in a reflective imaging system. The illumination NA_{ill} has to be greater than the objective NA_{obj} for darkfield imaging. If the LED is positioned at the back focal plane of the objective, the position of the LED needs to be greater than radius R_{EP} of the objective's entrance pupil. An example of a darkfield image when a USAF target is illuminated by a plane wave in the darkfield illumination condition is shown in Fig. 1(b3). The regions of Fourier space corresponding to the images in Figs. 1(b1)–1(b3) are shown in Fig. 1(c).

The FP recovery procedure utilizes a phase retrieval algorithm to estimate the complex object that creates the irradiance image captured using an oblique plane wave illumination. A more detailed description of the FP recovery procedure can be found in Refs. 1 and 17. In brief, the FP algorithm alternates between the spatial and Fourier domain for the specified region in Fourier space imposing the irradiance measurement as a constraint for that subregion in the spatial domain. The Fourier spectrum of the object is iteratively updated for N raw irradiance images, each illuminated with a unique oblique plane wave to further expand the spectrum of the object in the Fourier domain. The whole iteration scheme is repeated for M cycles until the complex reconstructed solution for the amplitude and phase of the object converges. The synthesized NA of the system after the FP recovery procedure is equal to the sum of the NA_{obj} of the objective and the maximum $NA_{ill,max}$ of the

Fig. 2 Optical system for reflective Fourier ptychography. The imaging stop is positioned outside the illumination path using a relay lens between the objective lens and tube lens.

illuminating plane waves. Therefore, the synthesized numerical aperture $NA_{syn} = NA_{obj} + NA_{ill,max}$.

An epi-illumination microscope is a convenient system for reflective FP, however, this system has a major limitation. If an epi-illumination microscope is used for FP, the LEDs are imaged onto the back focal plane of the objective, which is the location of the aperture stop in an epi-illumination microscope. Since the physical aperture stop of the objective blocks the images of the LEDs outside the aperture radius R_{EP} , the maximum illumination NA for the plane waves is the same as NA_{obj} of the objective. Thus, the maximum synthesized NA using FP with an epi-illumination microscope is $2NA_{obj}$. An epi-illumination microscope can extend the cut-off frequency of the reconstructed FP image to the incoherent limit of $2NA_{obj}/\lambda$. While there is no advantage in terms of resolution over incoherent imaging in epi-illumination mode, the reconstructed FP images are advantageous, since the transfer function is coherent-like and the contrast is higher at high spatial frequencies.²²

To achieve a cut-off frequency greater than $2NA_{obj}/\lambda$ using reflective FP, the illumination NA_{ill} must be greater than NA_{obj} . More LEDs can be arranged around the objective to obtain large angle illumination, but the system is bulky and, there is potentially a large gap in illumination angles between the light from the LEDs imaged onto the aperture stop and the LEDs arranged around the objective. The proposed optical concept for reflective FP is shown in Fig. 2, the imaging stop is placed outside the illumination path through a relay lens between the objective lens and the tube lens. The illumination path remains the same, and the illuminating light source is reflected off a beam-splitter onto the back focal plane of the objective providing the necessary plane wave illumination for FP. The imaging NA is determined by the external stop size, the illumination NA will not be limited by the stop size, but by the lens size.

With the proposed system, a simple large NA and large FOV objective can be developed for FP microscopy to obtain a HR image over a large FOV. This objective only requires diffraction-limited performance within the low imaging NA_{obj} , which is determined by the aperture placed in the conjugated plane of the aperture stop of the objective. The size of the objective lens can be increased significantly to provide large illumination NA_{ill} , since good aberration correction is not needed for the illumination path. Using this system, the synthesized NA_{syn} for reflective FP can exceed $2NA_{obj}$.

3 Experiment

For the experimental demonstration, all samples are imaged using a simple doublet lens, which is an achromatic 50-mm focal length lens (Thorlabs AC254-050-A-ML), and a scientific CMOS sensor (PCO.edge, 6.5- μm pixel pitch). This achromatic doublet is not a microscope objective, its effective imaging NA is not 0.25 as calculated from the first-order optics. It is used in this experiment to demonstrate the proposed concept because it has a small effective imaging NA 0.03 over large field of view, but with a large lens diameter of 25 mm. Because it is a single doublet, we can fully use its aperture for illumination.

Commercial objectives are not used in this experiment because all commercial objectives have internal aperture stops, we cannot illuminate the object with an angle larger than the NA. In the future, custom objectives with a large diameter, but a smaller imaging NA can be developed for specific applications. The image relay system has a magnification of -1 . A single 660-nm fiber-coupled LED with a 25-nm bandwidth and 105- μm core diameter is used as the light source. The LED is translated using an $x-y$ translation stage to move the image of the LED at the back focal plane of the objective. The LED is moved in a square array at the back focal plane of the objective to generate images of the object using 169 unique oblique illuminating plane waves. Due to the small size of the LED, the illumination is spatially coherent at the sample plane. An adjustable iris in the conjugated plane of the aperture stop of the objective is used to control the diameter of the stop. Due to the aberration of Thorlabs doublet, the remote aperture limits the imaging NA to 0.025, slightly smaller than the effective 0.03 NA.

To maintain sufficient spatial coherence in the reconstructed area, the FP image is reconstructed in patches of 150 pixels. For the 169 images, NA_{syn} is 0.113 in the ξ and η directions. First, a USAF target is imaged. The image of the USAF target using normal illumination with $NA_{obj} = 0.025$ is shown in Figs. 3(a) and 3(b), where Fig. 3(b) magnifies groups 6 to 9. With this NA, element 3 in group 5 (40.3 cycles/mm) is just resolvable. Using the proposed optical system and the FP recovery procedure with 169 unique plane wave illuminations with the proposed set-up, the synthesized NA is increased from $NA_{obj} = 0.025$ to $NA_{syn} = 0.113$. Figure 3(c) shows the irradiance image of groups 4 to 9 and 3(d) magnifies groups 6 to 9 after the FP algorithm is used. Note all elements in group 5 are now resolvable.

Fig. 3 Raw irradiance image of USAF target with $NA_{obj} = 0.025$ of (a) groups 4 to 9 and (b) groups 6 to 9. Irradiance image of USAF target after using FP recovery procedure with $NA_{syn} = 0.113$ of (c) groups 4 to 9 and (d) groups 6 to 9. Phase estimate after using FP recovery procedure of (e) groups 4 to 9 and (f) groups 6 to 9.

Furthermore, element 4 in group 7 (181 cycles/mm) is now just resolvable, as shown in Fig. 3(d). Using FP, a spatial frequency that is 4.5 times higher can be resolved with the proposed optical system. The FP recovery also estimates phase of the object, which is shown in Fig. 3(e) and magnified in Fig. 3(f).

The images reconstructed using FP are now compared with images of the USAF target when the objective has a similar NA. The NA of the objective is modified by varying the diameter of the stop, which is an adjustable iris. As shown in Fig. 4, the images acquired using an objective with an $NA = 0.04$ and the reconstructed image using FP have the same performance. The reconstructed images using FP utilize only nine brightfield images to reconstruct an image with $NA_{syn} = 0.04$. Figure 4(a) shows groups 4 to 9 and Fig. 4(b) magnifies groups 6 to 9 of the image of the USAF target reconstructed using the FP recovery procedure with the nine brightfield images. Figure 4(c) shows groups 4 to 9 and Fig. 4(d) magnifies groups 6 to 9 of the image of the USAF target acquired when diameter of the stop is increased until the NA_{obj} of the objective is 0.04. As shown in Figs. 4(a) and 4(c), all elements in group 5 are resolvable. As shown in Figs. 4(b) and 4(d), the last resolvable element is group 6 element 1 (64.0 cycles/mm). In terms of resolution, the image reconstructed using FP with nine brightfield images performs the same as an objective with $NA_{obj} = 0.04$.

Now the irradiance image of the USAF target with $NA_{syn} = 0.113$ reconstructed using all 169 images is compared with an image of the USAF target when the diameter of the stop is increased such that the objective $NA_{obj} = 0.118$. As shown in Fig. 5(a) and 5(b), group 7 element 4 (181 cycles/mm) is just resolvable, which is near the theoretical limit of 171 cycles/mm. Since the USAF target is comprised three bar targets instead of sinusoidal targets, the system is capable of resolving elements greater than the theoretical cut-off frequency at reduced contrast. Figure 5(c) shows groups 4 to 9 and 5(d) magnifies groups 6 to 9 of the USAF target when the stop is opened

such that $NA_{obj} = 0.118$. A Zemax model of the objective lens at this NA_{obj} is shown in Fig. 5(e). Note as the stop is increased to this NA_{obj} , the objective is no longer diffraction-limited. Figure 5(f) shows the spot diagram from Zemax for this objective lens with $NA_{obj} = 0.118$ when the system is optimized for minimum RMS wavefront. Note the spot diagram is larger than the Airy disk diameter, which is shown as a black circle. Thus,

Fig. 4 USAF irradiance image of (a) groups 4 to 9 and (b) groups 6 to 9 reconstructed using FP recovery procedure with nine brightfield images for a synthesized $NA_{syn} = 0.04$. USAF irradiance image of (c) groups 4 to 9 and (d) groups 6 to 9 acquired with objective with $NA_{obj} = 0.04$.

Fig. 5 USAF irradiance image reconstructed using FP recovery procedure with 169 images with synthesized $NA_{syn} = 0.113$ of (a) groups 4 to 9 and (b) groups 6-9. USAF irradiance image acquired with objective with $NA_{obj} = 0.118$ of (c) groups 4 to 9, and (d) groups 6 to 9. (e) Zemax model of objective lens when $NA_{obj} = 0.118$. (f) Spot diagram from Zemax when RMS wavefront is minimized at $NA_{obj} = 0.118$.

the aberrations in the objective when $NA_{obj} = 0.118$ degrades the performance, which is evident in Fig. 5(d), since the last resolvable element is now only group 7 element 1 (128 cycles/mm).

Due to the aberrations, this objective lens is not suitable for diffraction-limited imaging with a $NA_{obj} = 0.118$, as shown in Fig. 5(d). By using the effective NA_{obj} of 0.025 and using the FP recovery procedure with 169 images with unique oblique plane wave illumination, the reconstructed image has an equivalent

performance to a diffraction-limited objective with $NA_{syn} = 0.113$, as shown in Fig. 5(b). Therefore, using FP with the proposed reflective imaging system results in better performance than this objective can achieve, since the performance of the objective lens is degraded by aberrations.

To demonstrate the application of the proposed system for tissue imaging, a slice of tissue was cut from the chicken thigh and was sandwiched between two glass microscope slides

Fig. 6 (a) Large area of tissue from chicken thigh reconstructed with $NA_{syn} = 0.113$ using FP recovery procedure. Image of indicated subregion using (b1) objective with $NA_{obj} = 0.025$, (b2) objective with $NA_{obj} = 0.118$, and (b3) recovered using FP with $NA_{syn} = 0.113$. (c)–(g) Magnified regions indicated in (a).

to create a flat surface. Figure 6(a) shows a large area of the chicken tissue after using the FP recovery procedure using 169 raw images to obtain a NA_{syn} of 0.113. To demonstrate the improvement in the image, Fig. 6(b1) shows the raw image taken with $NA_{\text{obj}} = 0.025$ of the subregion of the image in the indicated region in the middle of 6(a), Fig. 6(b2) shows the same subregion when the diameter of the stop is increased such that NA_{obj} is 0.118, and Figure 6(b3) shows the same subregion reconstructed using FP. As shown in Fig. 6(b1), the details of the large structure are not resolvable and the smaller features particularly along the tilted white line along the bottom of the image are not resolvable. Figure 6(b2) is able to resolve the details of the large structure and show there is a small hole in the middle of it. Furthermore, it is now obvious that the large structure is surrounded by smaller circular structures that were not resolvable in Fig. 6(b1). Finally, Fig. 6(b3) resolves the finer features of the large structure as well as the smaller structures surrounding the large structure with higher contrast than Fig. 6(b2) since the image in Fig. 6(b2) is degraded by aberrations. Figure 6(c)–6(g) magnify other regions of Fig. 6(a). Note the small circular structures pervade all of these images. Therefore, the reflective FP optical system is capable of imaging reflective, planar objects with a resolution greater than the incoherent limit.

4 Conclusion

By placing the imaging stop outside the illumination path in a reflective microscope, the NA of the illumination can be much greater than the NA of the imaging system. Using this optical concept and the FP recovery procedure, a synthesized NA_{syn} of the optical system can be greater than $2NA_{\text{obj}}$, which is the theoretical limit using an epi-illumination microscope. By imaging a USAF target, it was experimentally shown that the resolution of the objective was increased from $NA_{\text{obj}} = 0.025$ to $NA_{\text{syn}} = 0.113$, which is an enhancement by a factor of 4.5. Furthermore, a slice of tissue from the thigh of a chicken sandwiched between two microscope slides was also imaged to demonstrate the system performance.

The key advantage of the proposed concept is that it can achieve HR imaging over the large FOV with a simple objective. The objective only requires diffraction-limited performance within the imaging NA, while easily increasing the lens diameter for large angle illumination without controlling the aberrations. With significantly less requirements on aberration correction (namely, spherical aberration, coma, field curvature, astigmatism, and distortion), the objective lens can be much simpler with much larger working distance and FOV.

The proposed method is not limited to a resolution improvement of 4.5. In practice, the system performance is limited by the diameter of the objective lens. Assuming the diameter of the objective lens is large enough to transmit the illuminating plane wave without vignetting, the maximum achievable shift in Fourier space is $NA_{\text{ill,max}} = 1.0$ in air. Therefore, the maximum achievable synthesized NA in air using this method is $(1 + NA_{\text{obj}})$, where NA_{obj} is the imaging NA of the objective lens.

In the future, we can also use the multiplexed illumination scheme^{16,17} or the nonuniform Fourier-space sampling scheme¹⁸

to further reduce the number of raw images for achieving the same imaging performance. We envision the proposed optical system can find applications in endoscopy, biomedical imaging, surface metrology, and industrial inspection.

References

1. G. Zheng, R. Horstmeyer, and C. Yang, "Wide-field, high-resolution Fourier ptychographic microscopy," *Nat. Photonics* **7**, 739–745 (2013).
2. V. Mico et al., "Synthetic aperture superresolution with multiple off-axis holograms," *J. Opt. Soc. Am. A* **23**, 3162–3170 (2006).
3. J. Di et al., "High resolution digital holographic microscopy with a wide field of view based on a synthetic aperture technique and use of linear CCD scanning," *Appl. Opt.* **47**, 5654–5659 (2008).
4. T. R. Hillman et al., "High-resolution, wide-field object reconstruction with synthetic aperture Fourier holographic optical microscopy," *Opt. Express* **17**, 7873–7892 (2009).
5. L. Granero et al., "Synthetic aperture superresolved microscopy in digital lensless Fourier holography by time and angular multiplexing of the object information," *Appl. Opt.* **49**, 845–857 (2010).
6. T. Gutzler et al., "Coherent aperture-synthesis, wide-field, high-resolution holographic microscopy of biological tissue," *Opt. Lett.* **35**, 1136–1138 (2010).
7. J. R. Fienup, "Phase retrieval algorithms: a comparison," *Appl. Opt.* **21**, 2758–2769 (1982).
8. V. Elser, "Phase retrieval by iterated projections," *J. Opt. Soc. Am. A* **20**, 40–55 (2003).
9. H. M. L. Faulkner and J. M. Rodenburg, "Movable aperture lensless transmission microscopy: a novel phase retrieval algorithm," *Phys. Rev. Lett.* **93**, 023903 (2004).
10. A. M. Maiden and J. M. Rodenburg, "An improved ptychographical phase retrieval algorithm for diffractive imaging," *Ultramicroscopy* **109**, 1256–1262 (2009).
11. X. Ou et al., "Quantitative phase imaging via Fourier ptychographic microscopy," *Opt. Lett.* **38**, 4845–4848 (2013).
12. S. Dong et al., "Aperture-scanning Fourier ptychography for 3D refocusing and super-resolution macroscopic imaging," *Opt. Express* **22**, 13586–13599 (2014).
13. X. Ou, G. Zheng, and C. Yang, "Embedded pupil function recovery for Fourier ptychographic microscopy," *Opt. Express* **22**, 4960–4972 (2014).
14. P. Li et al., "Separation of three-dimensional scattering effects in tilt-series Fourier ptychography," *Ultramicroscopy* **158**, 1–7 (2015).
15. L. Tian and L. Waller, "3D intensity and phase imaging from light field measurements in an LED array microscope," *Optica* **2**, 104–111 (2015).
16. S. Dong et al., "Spectral multiplexing and coherent-state decomposition in Fourier ptychographic imaging," *Biomed. Opt. Express* **5**, 1757–1767 (2014).
17. L. Tian et al., "Multiplexed coded illumination for Fourier ptychography with an LED array microscope," *Biomed. Opt. Express* **5**, 2376–2389 (2014).
18. K. Guo et al., "Optimization of sampling pattern and the design of Fourier ptychographic illuminator," *Opt. Express* **23**, 6171–6180 (2015).
19. L. Bian et al., "Content adaptive illumination for Fourier ptychography," *Opt. Lett.* **39**, 6648–6651 (2014).
20. L. Bian et al., "Fourier ptychographic reconstruction using Wirtinger flow optimization," *Opt. Express* **23**, 4856–4866 (2015).
21. R. Horstmeyer et al., "Solving ptychography with a convex relaxation," *New J. Phys.* **17**, 053044 (2015).
22. S. Pacheco et al., "Transfer function analysis in epi-illumination Fourier ptychography," *Opt. Lett.* **40**, 5343–5346 (2015).

Biographies for the authors are not available.