Journal of Biomedical Optics

BiomedicalOptics.SPIEDigitalLibrary.org

***In vivo* correlation mapping microscopy**

James McGrath
Sergey Alexandrov
Peter Owens
Hrebesh Subhash
Martin Leahy

SPIE.

James McGrath, Sergey Alexandrov, Peter Owens, Hrebesh Subhash, Martin Leahy, “*In vivo* correlation mapping microscopy,” *J. Biomed. Opt.* **21**(4), 046004 (2016), doi: 10.1117/1.JBO.21.4.046004.

In vivo correlation mapping microscopy

James McGrath,^a Sergey Alexandrov,^a Peter Owens,^b Hrebesh Subhash,^a and Martin Leahy^{a,c,*}

^aNational University of Ireland Galway, Tissue Optics and Microcirculation Imaging Group, National Biophotonics and Imaging Platform, Physics Department, University Road, Newcastle, Galway H91 TK33, Ireland

^bNational University of Ireland Galway, Centre for Microscopy and Imaging, University Road, Newcastle, Galway H91 TK33, Ireland

^cRCSI Royal College of Surgeons, 123 St. Stephens Green, Dublin 2 D02 YN77, Ireland

Abstract. To facilitate regular assessment of the microcirculation *in vivo*, noninvasive imaging techniques such as nailfold capillaroscopy are required in clinics. Recently, a correlation mapping technique has been applied to optical coherence tomography (OCT), which extends the capabilities of OCT to microcirculation morphology imaging. This technique, known as correlation mapping optical coherence tomography, has been shown to extract parameters, such as capillary density and vessel diameter, and key clinical markers associated with early changes in microvascular diseases. However, OCT has limited spatial resolution in both the transverse and depth directions. Here, we extend this correlation mapping technique to other microscopy modalities, including confocal microscopy, and take advantage of the higher spatial resolution offered by these modalities. The technique is achieved as a processing step on microscopy images and does not require any modification to the microscope hardware. Results are presented which show that this correlation mapping microscopy technique can extend the capabilities of conventional microscopy to enable mapping of vascular networks *in vivo* with high spatial resolution in both the transverse and depth directions. © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.21.4.046004]

Keywords: confocal microscopy; widefield microscopy; nailfold capillaroscopy; biomedical optical imaging; correlation mapping.

Paper 150813PRR received Dec. 12, 2015; accepted for publication Mar. 22, 2016; published online Apr. 12, 2016.

1 Introduction

Microcirculation is the term used to describe the small blood vessels ($<100\ \mu\text{m}$ in size) in the vasculature network, which are vital for tissue nutrition and shape as well as waste removal, blood pressure and temperature regulation, and oxygen supply. Structural and functional changes in the microcirculation are associated with various pathological conditions including diabetes, Raynaud's disease, psoriasis, systemic sclerosis, and hypertension.¹ A frequently used modality for noninvasive examination of the microcirculation is laser Doppler blood perfusion monitoring, which is based on measuring the Doppler shift induced by moving red blood cells.² Laser Doppler uses diffuse light from the top $\sim 1\ \text{mm}$ of tissue to provide average perfusion measurements over volumes of $\sim 100\ \mu\text{m}^3$ and is unable to provide information on individual smaller vessels.

Video capillaroscopy,³ an extension of bright widefield microscopy, is a simple and cost-effective modality for assessment of the microcirculation and is used in the diagnosis and follow-up of many pathological conditions. However, it is generally limited in application to sites, such as nailfold plexus and conjunctiva, where the blood vessels are close to the surface so that they are not obscured by scattering. Sidestream dark field imaging can extend these techniques for application to other cardiovascular diseases by improved imaging of the tongue and oral mucosa.^{4,5} More recently, reflectance-mode confocal microscopy has also been introduced to assess the human skin microcirculation.^{6–8} Optical coherence tomography (OCT) is an emerging noninvasive imaging technique, which can generate depth-resolved images of the microcirculation.^{9–11} However, OCT

itself does not directly produce these images of microcirculation. Several technologies have been developed to extend the capabilities of OCT to visualize the microcirculation, such as Doppler OCT.¹²

Figure 1 shows a plot of the typical resolution versus sampling depth for the widefield microscopy, confocal microscopy, and OCT modalities. OCT has the advantage of imaging 1 to 2 mm in tissue (typically, 1 mm in human skin), but with limited spatial resolution, typically on the order of 10 to 20 μm . The higher lateral resolution of widefield and confocal microscopy modalities makes them very suitable for imaging of the smaller cutaneous vessels noninvasively, in real time, and with a resolution comparable to that of histological examination. Longer wavelengths, for instance, 830 nm, have improved the imaging depth in skin up to 300 μm in some cases.⁶ Widefield and confocal microscopy can potentially, therefore, be used to complement OCT imaging as they can provide information on the smaller capillaries in the dermal–epidermal junction, whereas OCT can provide information on larger vessels deeper in the skin.

A correlation mapping algorithm has been previously developed and applied by our group to OCT.^{9–11} The technique, known as correlation mapping optical coherence tomography (cmOCT), utilizes the time-varying speckle effect. This effect, produced by moving scatters, provides contrast between regions of flow and static tissue, facilitating the generation of maps of microcirculation morphology. Recently, a similar correlation mapping algorithm has been implemented in real time by utilizing graphics processing units (GPUs).¹³

In this work, we extend the correlation mapping technique to widefield and confocal microscopy. Correlation mapping

*Address all correspondence to: Martin Leahy, E-mail: martin.leahy@nuigalway.ie

Fig. 1 Plot of resolution versus maximum sampling depth for confocal microscopy, widefield microscopy, and optical coherence tomography imaging modalities.

microscopy enables widefield and confocal microscopy to visualize flow, for instance, the microcirculation, with a high contrast, and takes advantage of the higher spatial resolution afforded by these modalities. This technique utilizes the light intensity variations over time caused by moving scatterers to provide contrast against a static background. Flow visualization is achieved by correlating temporally consecutive microscopy image frames to extract the flow information and does not require any modification or addition to the microscope hardware.¹⁴ The adaptation of the correlation mapping algorithm to microscopy images is demonstrated, including visualization of a capillary tube flow phantom through a scattering medium and *in vivo* human imaging of the volar forearm.

2 Processing Principles of Correlation Mapping Microscopy

2.1 Correlation Mapping Microscopy Processing Technique

The correlation mapping algorithm is applied to consecutive pairs of microscopy images. The first step of the algorithm is to suppress noise, which will cause decorrelation, by applying a “structural mask” to the images. The structural mask includes regions that contain structural information in the microscopy image. The mask is generated by performing a kernel average followed by a binary threshold on the source microscopy images. A kernel is a two-dimensional (2-D) array of image pixels, the size of which is determined by the user.

The next step is determination of the correlation between two temporally consecutive microscopy images captured at the same location. This is calculated by using a Pearson correlation to cross correlate a grid from microscope image A (I_A) to the same grid from image B (I_B):

$$cm(x, y) = \sum_{p=0}^M \sum_{q=0}^N \frac{[I_A(x+p, y+q) - I'_A(x, y)][I_B(x+p, y+q) - I'_B(x, y)]}{[I_A(x+p, y+q) - I'_A(x, y)]^2 [I_B(x+p, y+q) - I'_B(x, y)]^2} \quad (1)$$

Equation (1) is the Pearson correlation formula and is a measure of the linear correlation between two variables, giving a value

between +1 and −1 inclusive. According to this equation, a value of +1 represents total positive correlation and a value of −1 represents total negative correlation. We would expect static features to have correlation values close to +1, while regions of flow will have much lower correlation values.

The grids I_A and I_B are stepped across the whole image, and a correlation value is calculated at each pixel resulting in a 2-D correlation map. The resultant map has correlation values in the range of 0 ± 1 . This allows for immediate determination of flow regions as higher correlation represents static regions of the sample, and lower correlation values represent regions of flow within the sample. Noise present in the images generated by the system could lower the correlation value of static features, as noise would cause decorrelation. In our image processing, we assume correlation values of >0.6 are static features.

In the phantom studies, the ability of the correlation mapping algorithm to contrast the flow region is validated using the correlation signal-to-noise ratio (CSNR). Metrics such as the CSNR have previously been used to evaluate the quality of angiography algorithms.¹⁵ CSNR is defined as

$$CSNR = \frac{\overline{C_{flow}} - \overline{C_{static}}}{\sqrt{\sigma_{static}^2}}, \quad (2)$$

where $\overline{C_{flow}}$ and $\overline{C_{static}}$ are the average correlation values in flow and static regions, respectively, and σ_{static}^2 is the variance of the correlation within the static region. CSNR values are calculated in each of the following studies to illustrate the ability of the correlation mapping algorithm to contrast flow and static regions.

The optimal kernel size must be selected by the user. The kernel size used in this work was arbitrarily chosen for optimal image quality with the trade-off of processing time, spatial resolution, and CSNR. Larger kernels require greater processing times but are less susceptible to noise than smaller size kernels. However, very large kernel sizes causing “blurring” in the correlation map image, which will make smaller regions of flow undetectable from an *in vivo* perspective. If the aim is to detect smaller capillaries or to measure capillary diameter, this will be a significant issue. If a very small kernel is used, such as 3×3 , the resulting data set can be quite noisy as small changes in the structural signal can result in decorrelation and may be mistaken for regions of flow. An optimum kernel size must be determined. In this work, a kernel size of 11×11 was used as it delivered a good quality image with an acceptable signal-to-noise ratio while providing sufficient spatial resolution to avoid incurring any image blurring, which larger kernel sizes would cause.

Implementation of the correlation mapping microscopy algorithm is an “in-house” developed multithreaded code written using the Java programming language. The program can process 30 adjacent microscopy images of size 1000×1000 pixels in <1.5 s using eight threads. Further speedup could be achieved using optimized code or GPU-based techniques, enabling implementation of the algorithm in real time.

The dynamic range of velocity depends on a variety of factors, including frame rate and resolution. There must be enough time between frames so that the movement of the particles in the liquid can be detected within a given resolution. Confocal microscopy has better resolution, and because of this it should be able to operate at a higher frame rate and give the best dynamic range of velocities.

2.2 Widefield/Confocal Microscopes and Optical Coherence Tomography System

To validate the correlation mapping microscopy algorithm, a variety of microscope systems was used in this study. The widefield microscope used was an Olympus measuring microscope (STM-MJS2), using a broadband white light source with an Olympus MS Plan 10× objective lens with a numerical aperture (NA) of 0.3, giving the system a lateral resolution of 1.1 μm . Images were captured using a Lumenera Infinity 2 digital CCD camera operating at ~ 1 frame per second (fps).

Reflectance confocal microscopy images were acquired with an Andor Revolution spinning disk (Andor Technology Ltd., Belfast, UK) system operating at a wavelength of 640 nm and capturing images at 15 fps. Using a 10× objective lens of NA 0.3, the lateral resolution of this system was 0.85 μm , and the axial resolution was 9.31 μm .

In vivo images were acquired using a reflectance confocal microscope, Vivascope 1500[®] (Lucid Inc., Rochester, New York), a commercial clinical research system using an 830 nm (near-infrared) laser. This operating wavelength of 830 nm is at an “optical window” of the skin (mean free path for scattering ~ 100 μm), and thus the system facilitates investigation of the epidermis and papillary dermis to a depth of ~ 350 μm with a lateral resolution of 0.4 μm and an axial resolution of 1.9 μm in the center of the image field. The laser power is relatively low at 30 mW, therefore avoiding any tissue damage. An imaging frame rate of 9 fps is achievable with a field of view of 1000×1000 μm .

For evaluation of *in vivo* correlation mapping microscopy images, a comparison was performed with images taken from a commercial OCT system (OCM1300SS, Thorlabs Inc., Newton, New Jersey). This is a Fourier-domain system using a swept source laser (SL1325-P16 Thorlabs Inc.) operating at a center wavelength of 1325 nm and at a scanning rate of 16 kHz, providing an axial resolution of ~ 12 μm and a lateral resolution of 25 μm .

3 Phantom Imaging

3.1 Widefield Microscopy Phantom Imaging

To demonstrate the ability of correlation mapping microscopy to extract flow with low velocities, a capillary tube sample was fabricated. The sample consisted of a glass capillary tube, mounted on a piece of black card and filled with 3% intralipid solution, which was free to move under Brownian motion. This tube appears on the right-hand side in Figs. 2(a) and 2(b). A separate empty glass capillary tube was also mounted on the black card. This tube appears on the left-hand side in Figs. 2(a) and 2(b). Both capillary tubes had an internal diameter of 100 μm . The widefield microscope focus was set to the center of the capillary tubes, and a series of images were taken as shown in Figs. 2(a) and 2(b). Movie clips are provided for Figs. 2(a) and 2(b). It is very difficult to differentiate any flow/Brownian motion regions from static regions by examining these images and movie. The correlation map resulting from applying the correlation mapping microscopy algorithm to these two temporally consecutive images is shown in Fig. 2(c). A movie is also provided for Fig. 2(c). In this image and in the corresponding movie, it can be seen that the capillary tube on the right-hand side containing the intralipid solution moving under Brownian motion is highlighted by brighter colors, representing lower correlation values. To calculate the

Fig. 2 Correlation mapping microscopy applied to two temporally consecutive widefield microscopy images, (a) and (b). Each image shows two 100- μm inner diameter capillary tubes. In both images, the tube on the left is empty and the tube on the right contains intralipid solution moving under Brownian motion. The resulting correlation map is shown in (c) with higher correlation values shown as a darker color and lower correlation values as a brighter color. CSNR was calculated between the flow region within the green square and the static region within the blue square. The scale bar in parts (a)–(c) is 100 μm . The correlation map color scale is shown in (d). A movie corresponding to (a)–(c) is provided. Video 1: Example of a widefield microscopy image of phantom capillary tube sample. Video 2: Correlation mapping microscopy movie generated from widefield microscopy frames of a capillary tube flow phantom, as in Video 1. (Video 1, MOV, 530 KB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.1>]; Video 2, MOV, 529 KB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.2>]).

CSNR value, a region of interest in both a static area and an area of flow/Brownian motion needed to be analyzed. A 50×50 pixel region within both the empty capillary tube [blue square in Fig. 2(c)] and capillary tube filled with intralipid [green square in Fig. 2(c)] was selected. The resulting CSNR value was 10.89. This illustrates the ability of the correlation mapping microscopy algorithm to clearly visualize the region of flow within the image.

3.2 Widefield Microscopy Phantom Imaging Through Scattering Medium

The aim of the next experiment was to test the ability of correlation mapping microscopy to detect flow when imaging through a scattering medium. To test this, the sample used in Sec. 3.1 was covered with a layer of semitransparent scotch tape from the 3M company. This tape has a mean thickness of 60 μm and a scattering angle of about 5 deg. The widefield microscope was focused to the center of the capillary tubes, and a series of images was taken, an example of which is shown in Fig. 3(a). A movie is provided corresponding to Figs. 3(a) and 3(b). The capillary tube in the upper part of Fig. 3(a) contains 3% intralipid solution moving under Brownian motion; the lower capillary tube is empty. Figure 3(b) clearly shows the upper capillary tube containing the intralipid solution highlighted by brighter colors representing lower correlation values. To calculate the CSNR value, a region of interest in both a static area and an area of flow/Brownian motion needed to be analyzed. A 50×50 pixel region within both the empty capillary tube [blue square in Fig. 3(b)] and capillary tube filled with intralipid [green square in Fig. 3(b)] was selected. The resulting

Fig. 3 Correlation mapping microscopy applied to temporally consecutive widefield microscopy images of a capillary tube sample covered with scattering sticky tape. (a) A conventional microscope image of two 100- μm inner diameter capillary tubes; the tube on the bottom is empty, and the tube on the top contains intralipid solution moving under Brownian motion. The resulting correlation map is shown in (b). CSNR was calculated between the flow region within the green square and the static region within the blue square. The scale bar is 100 μm . The correlation map color scale is shown in (c). A movie corresponding to (a) and (b) is provided. Video 3: Example of a widefield microscopy image of phantom capillary tube sample when imaging through scattering medium (MOV, 164 KB). Video 4: Correlation mapping microscopy movie generated from widefield microscopy frames of a capillary tube flow phantom imaged through a scattering medium, as in Video 3 (Video 3, MOV, 296 KB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.3>]; Video 4, MOV, 295 KB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.4>]).

CSNR value was 5.12. This illustrates the ability of the correlation mapping microscopy algorithm to clearly visualize the region of flow within the image, even when imaging through a scattering medium.

3.3 Confocal Microscopy Phantom Imaging

We next tested the applicability of the correlation mapping microscopy algorithm to confocal microscopy. A capillary tube sample containing three tubes filled with 3% intralipid solution was prepared. The three tubes, with inner diameters 50, 100, and 200 μm , were mounted on a glass microscope slide as shown in Fig. 3(a). Imaging was performed with the Andor Revolution spinning disk confocal microscope, described in Sec. 2.1, with the imaging depth set to approximately the center of the 50- μm inner diameter capillary tube. A movie is provided corresponding to Figs. 4(a) and 4(b). From the conventional microscopy images, it is very difficult to differentiate static regions from regions of flow. The correlation mapping microscopy algorithm, described in Sec. 2.1, was applied to two temporally consecutive confocal microscope images of the sample. The resulting correlation map is shown in Fig. 4(b). Once again, the internal cavity of the capillary tubes is highlighted in brighter colors corresponding to low correlation values, differentiating these regions as containing flow/liquid moving under Brownian motion as opposed to the darker colored static regions between the tubes. To calculate the CSNR value, a region of interest in both a static area and an area of flow/Brownian motion needed to be analyzed. A 50 $\times$ 50 pixel region within the 50- μm capillary tube containing intralipid was selected to represent a region of Brownian motion [green square in Fig. 4(b)]. Additionally, a 50 $\times$ 50 pixel region between the 50- and 100- μm capillary tubes was selected to represent a static region [blue square in Fig. 4(b)]. The resulting CSNR value was 6.30. This result demonstrates possibility of

Fig. 4 Correlation mapping microscopy applied to temporally consecutive confocal microscopy images of a sample with capillary tubes of three sizes (inner diameter of 50 μm at the top, 100 μm in center, and 200 μm , the upper portion of which is visible at the bottom). Conventional confocal microscope image is shown in (a). The resulting correlation map is shown in (b). CSNR was calculated between the flow region within the green square and the static region within the blue square. The scale bar is 100 μm . The correlation map color scale is shown in (c). A movie is provided corresponding to (a) and (b). Video 5: Example of a confocal microscopy image of phantom capillary tube sample. Video 6: Correlation mapping microscopy movie generated from confocal microscopy frames of a capillary tube flow phantom, as in Video 5 (Video 5, MOV, 4.6 MB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.5>]; Video 6, MOV, 4.47 MB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.6>]).

the correlation mapping confocal microscopy to clearly visualize regions with flow.

3.4 In Vivo Imaging

To demonstrate the clinical suitability of correlation mapping microscopy and compare it to the established microcirculation imaging modalities, *in vivo* imaging was performed. Both *in vivo* confocal microscopy using the Vivascope 1500 and OCT using the Thorlabs systems described in Sec. 2.1 were performed on the volar forearm of a 27-year-old male with informed consent. The OCT image together with its derived cmOCT image of the volar forearm at a depth of ~ 250 μm is shown in Figs. 5(a) and 5(b), respectively. These images have a field of view of 3000 $\times$ 3000 μm . A zoomed-in portion of the OCT image is shown in Fig. 5(d), and of the cmOCT image is shown in Fig. 5(e). These images have a field of view of 500 $\times$ 500 μm , the same as that of images generated by the Vivascope 1500 reflectance confocal microscopy system. It is difficult to make out vessel structure and location from low-resolution OCT and cmOCT images in Figs. 5(a)–5(e).

An example image of the volar forearm generated by the Vivascope 1500 confocal microscopy system is shown in Fig. 5(f). The dermal papillae are visible in the confocal microscopy image of Fig. 5(f) as bright circular basal layers with a dark center. Inside these circles, the lumina of the capillary loops are apparent as black holes and appear in pairs, with an arterial and a venous capillary located next to one another. After applying the correlation mapping algorithm to these images, the microcirculation is clearly visualized with high contrast [as shown in Fig. 5(g)], allowing us to investigate vessels of size 10 to 20 μm located in the dermal–epidermal junction layer. The correlation maps can be averaged across a number of frames (five in this case) to give a clearer image. This averaging could be reduced or eliminated, if sample motion is reduced either

Fig. 5 *In vivo* correlation mapping microscopy compared to OCT. An OCT image of the volar forearm at depth $\sim 250 \mu\text{m}$ is shown in (a), with a cmOCT image derived from this shown in (b). Part (d) shows a zoomed-in region of (a) of size $500 \times 500 \mu\text{m}$, the same dimensions of images provided by the Vivascope 1500. (e) cmOCT image related to (d). An image of the volar forearm at depth $\sim 250 \mu\text{m}$ provided by the Vivascope is shown in (f). (g) The correlation mapping microscopy image corresponding to (f). CSNR was calculated between the flow region within the green square and the static region within the blue square in (g). Scale bars in (a) and (b) are $1000 \mu\text{m}$. Scale bars in (d) to (g) are $100 \mu\text{m}$. Correlation scale bar is shown in (c). Video 7: *In vivo* confocal microscopy of volar forearm. Video 8: *In vivo* correlation mapping microscopy movie generated from confocal microscopy frames of volar forearm, as in Video 7 (Video 7, MOV, 0.1 MB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.7>]; Video 8, MOV, 111 KB [URL: <http://dx.doi.org/10.1117/1.JBO.21.4.046004.8>]).

with improved mounts or motion correction image processing algorithms. Averaging is not performed on the OCT data; however, a motion compensation algorithm is applied to reduce any effect of axial or lateral motion. It is impossible to visualize vessels in the dermal papillae using most currently available commercial OCT systems. The observed locations of vessels in Fig. 5(g) match well with the locations of vessels one would expect from an examination of Fig. 5(f). Note that while the dermal papillae are easy to identify in the confocal reflectance microscopy image of Fig. 5(f), not all of these structures contain currently dilated vessels with blood flow. After application of the correlation mapping algorithm, the regions of blood flow are easily identified. To calculate the CSNR value, a region of interest in both a static area and an area of flow/Brownian motion needed to be analyzed. A static region was identified as a region of tissue outside the dermal papillae, where no blood vessels are present [blue square in Fig. 5(g)]. Within the dermal papillae, blood vessels are present, and a region in this area was chosen as the “flow” region. This is identifiable as the region within the green square in Fig. 5(g). A 50×50 pixel area of both these static and flow regions was selected, and the CSNR was calculated. The resulting CSNR value was 4.95.

It is noted that it is possible to calculate additional information, such as capillary loop density and vessel diameter, from Fig. 5(g). Calculation of these markers was performed by first loading the *in vivo* confocal images into ImageJ, which is a public domain image processing tool (version 1.49, available in Ref. 17). The ImageJ software was calibrated with the known pixel-distance relationship. The area within the field of view of the image could then be estimated. The capillary loop density can be calculated by manually identifying and counting the number of detected vessels within this area. The capillary density of the volar forearm has been calculated to be ~ 28 loops/ mm^2 ; this value is within the 14 to 30 loops/ mm^2 range previously

reported.¹⁶ Additionally, the average capillary loop diameter can be estimated. Capillary size measurements between two reference points were performed by tracing a horizontal line between clearly distinguishable inner borders of capillaries. The ImageJ software calculated the intracapillary distance in micrometers according to the previous calibration. We estimate the average capillary loop diameter to be $\sim 9.50 \mu\text{m}$, which is close to the value reported in previous studies.⁶ Such parameters are relatively easy to determine from correlation mapping images as the patent capillaries are immediately identified. Capillaries in the dermal papillae have diameters that are below the resolution of typical commercially available OCT systems and therefore cannot be detected by such systems. Correlation mapping confocal microscopy can complement OCT in scenarios requiring imaging of small blood vessels in the dermal-epidermal junction.

4 Summary

This work has demonstrated the application of the correlation mapping algorithm to widefield and confocal microscopy modalities, including *in vivo* confocal microscopy. The application of this algorithm to microscopy modalities allows identification of regions of fluid flow with high spatial resolution. Such regions of flow can be very difficult to distinguish in the raw microscopy images and movies. However, we showed that after application of the correlation mapping algorithm, regions of fluid flow can easily be differentiated from static regions with high contrast even when imaging through scattering media. For the flow phantoms, just two temporally consecutive microscope image frames are required to visualize flow. *In vivo* human imaging of the volar forearm using a clinical reflectance confocal microscopy system and the correlation mapping algorithm demonstrated high resolution visualization of blood vessels, which is below the resolution of the typical commercial systems of other modalities, such as OCT and laser Doppler flowmetry.

With correlation mapping, we can clearly visualize flow areas, and dynamics could potentially be estimated. These results suggest the correlation mapping microscopy is an excellent choice for imaging of different dynamic events with high resolution, for instance, visualization of blood flow in small size vessels using *in vivo* confocal microscopy. The correlation mapping algorithm does not require additional hardware modifications to the microscope system, it is implemented entirely in software, and real time processing of microscope images can be achieved.

Acknowledgments

This work was supported by the National Biophotonics Imaging Platform, Ireland, funded under the Higher Education Authority PRTL Cycle 4, cofunded by the Irish government and the European Union—Investing in your future. The authors would like to thank the Professor Steven Jacques of Oregon Health and Science University and National Center for Laser Applications, NUI Galway, for the use of their widefield microscope. The study was approved by the Research Ethics Committee of NUI Galway.

References

1. M. J. Leahy, Ed., *Microcirculation Imaging*, p. 37, Wiley-VCH, Weinheim, Germany (2012).
2. S. M. Daly and M. J. Leahy, “‘Go with the flow’: a review of methods and advancements in blood flow imaging,” *J. Biophotonics* **6**(3), 217–255 (2013).
3. F. Gallucci et al., “Indications and results of videocapillaroscopy in clinical practice,” *Adv. Med. Sci.* **53**(2), 149–157 (2008).
4. P. T. Goedhart et al., “Sidestream dark field (SDF) imaging: a novel stroboscopic LED ring-based imaging modality for clinical assessment of the microcirculation,” *Opt. Express* **15**(23), 15101–15114 (2007).
5. C. Ince, “Sidestream dark field imaging: an improved technique to observe sublingual microcirculation,” *Crit. Care* **9**(Suppl. 1), P72 (2005).
6. M. A. Altintas et al., “Insight in human skin microcirculation using *in vivo* reflectance mode confocal laser scanning microscopy,” *J. Digital Imaging* **23**(4), 475–481 (2010).
7. E. Cinotti et al., “Quantification of capillary blood cell flow using reflectance confocal microscopy,” *Skin Res. Technol.* **20**(3), 373–378 (2014).
8. M. Venturini et al., “*In vivo* reflectance confocal microscopy features of cutaneous microcirculation and epidermal and dermal changes in diffuse systemic sclerosis and correlation with histological and videocapillaroscopic findings,” *Eur. J. Dermatol.* **24**(3), 349–355 (2014).
9. J. Enfield, E. Jonathan, and M. J. Leahy, “*In vivo* imaging of the microcirculation of the volar forearm using correlation mapping optical coherence tomography (cmOCT),” *Biomed. Opt. Express* **2**(5), 1184–1193 (2011).
10. E. Jonathan, J. Enfield, and M. J. Leahy, “Correlation mapping method for generating microcirculation morphology from optical coherence tomography (OCT) intensity images,” *J. Biophotonics* **4**(9), 583–587 (2011).
11. A. Zam et al., “Feasibility of correlation mapping optical coherence tomography (cmOCT) for anti-spoof sub-surface fingerprinting,” *J. Biophotonics* **6**(9), 663–667 (2013).
12. X. Wang, T. Milner, and J. Nelson, “Characterization of fluid flow velocity by optical Doppler tomography,” *Opt. Lett.* **20**(11), 1337–1339 (1995).
13. Y. Watanabe, H. Numazawa, and D. Kamiyama, “GPU accelerated correlation mapping OCT for real-time imaging of microvasculature,” in *2013 Conf. on Lasers and Electro-Optics Pacific Rim (CLEO-PR ‘13)*, pp. 1–2 (2013).
14. J. McGrath et al., “Correlation mapping microscopy,” *Proc. SPIE* **9322**, 93220L (2015).
15. S. Gao et al., “Optimization of the split-spectrum amplitude-decorrelation angiography algorithm on a spectral optical coherence tomography system,” *Opt. Lett.* **40**(10), 2305–2308 (2015).
16. P. Agache, P. Humbert, and H. Maibach, *Measuring the Skin*, p. 439, Springer, Springer-Verlag Berlin, Heidelberg, Germany (2004).
17. W. S. Rasband, “ImageJ,” U. S. National Institutes of Health, Bethesda, Maryland, <http://imagej.nih.gov/ij/> (1997–2015).

James McGrath is a student at the National University of Ireland, Galway. He received his BSc and HDipAppSc degrees in applied physics and electronics and software design and development from the National University of Ireland, Galway, in 2010 and 2011, respectively. His current research interests include high-resolution OCT and microcirculation imaging. He is a member of SPIE.

Sergey Alexandrov is a research scientist at NUI Galway. His main area of research is the development of new optical methods and tools, such as imaging and measuring techniques (holographic, interference, polarization, microscopy, OCT, spectral, etc.) for biomedical and other applications. His contributions have resulted in the creation of new optical methods and techniques for nondestructive testing, applications to different samples, including optical elements and systems, biomedical samples, and human beings *in vivo*.

Peter Owens is a research associate at the Department of Anatomy at NUI Galway. He is running the light microscopy facility for the Centre for Microscopy and Imaging, also at NUI Galway. In conjunction, he is also coordinating the educational program for the National Biophotonics Imaging Platform of Ireland.

Hrebesh Subhash is a senior research fellow at NUI Galway. His research interests include high sensitive microcirculation imaging based on optical microangiography and correlation-mapping OCT. He has worked on development and advancement of miniature integrated optical coherence tomography instruments based on MRO™ technology with compact imaging for a range of biological and non-biological imaging applications, measurement and analysis applications from living tissue to industrial materials.

Martin Leahy is currently chair of applied physics at NUI Galway. His main research interest is in the advancement of existing technologies, as well as the development of new modalities such as cmOCT for imaging of the microcirculation. He is an adjunct professor at the Royal College of Surgeons, fellow of the Institute of Physics in Ireland, fellow of the Royal Academy of Medicine in Ireland and fellow of SPIE.