Laser-induced heating integrated with a microfluidic platform for real-time DNA replication and detection

Min-Sheng Hung
Chia-Chin Ho
Chih-Pin Chen

Laser-induced heating integrated with a microfluidic platform for real-time DNA replication and detection

Min-Sheng Hung,^{a,*} Chia-Chin Ho,^b and Chih-Pin Chen^a

^aNational Chiayi University, Department of Biomechatronic Engineering, No. 300 Syuefu Road, Chiayi 60004, Taiwan

^bChip Win Technology Co., Ltd., Biomedical and Advanced Systems Integration, No. 18 Sec. 2, Seng Yi Road, Zhubei City, Hsinchu 30261, Taiwan

Abstract. This study developed a microfluidic platform for replicating and detecting DNA in real time by integrating a laser and a microfluidic device composed of polydimethylsiloxane. The design of the microchannels consisted of a laser-heating area and a detection area. An infrared laser was used as the heating source for DNA replication, and the laser power was adjusted to heat the solutions directly. In addition, strong biotin–avidin binding was used to capture and detect the replicated products. The biotin on one end was bound to avidin and anchored to the surface of the microchannels, whereas the biotin on the other end was bound to the quantum dots (Qdots). The results showed that the fluorescent intensity of the Qdots bound to the replicated products in the detection area increased with the number of thermal cycles created by the laser. When the number of thermal cycles was ≥ 10 , the fluorescent intensity of the Qdots was directly detectable on the surface of the microchannels. The proposed method is more sensitive than detection methods entailing gel electrophoresis. © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.21.8.087003]

Keywords: microfluidics; laser; quantum dot; DNA replication.

Paper 160294RR received May 10, 2016; accepted for publication Jul. 27, 2016; published online Aug. 12, 2016.

1 Introduction

DNA is a long-chain polymer that contains genetic instructions for guiding biological development and functional operations. With rapid technological developments in recent years, biological nanotechnology has been applied extensively in medical, food, agricultural, and other industries to detect biological samples such as DNA and protein. Polymerase chain reaction (PCR) is typically used to replicate the target DNA for analysis. However, uneven heating of solution samples occur often in conventional PCR machines because of their low thermal transfer efficiency. Consequently, heating and cooling are time intensive, thereby slowing product replication.

Conduction and convection are the primary methods for thermal transfer in most biological nanotechnology-based PCR chips.^{1–3} The heat capacity of materials must be considered when selecting chip materials because it indirectly affects PCR efficiency. Reducing the sample volume can accelerate heating and cooling, thereby increasing the reaction rate. For example, integrated continuous-flow PCR and microdroplets coated with oil were used to perform PCR thermal cycling in microchannels.^{4,5}

Studies have indicated that PCR thermal cycling can be accelerated by enhancing the heat transfer efficiency of chip materials or by reducing the volume of reaction solutions. However, chip materials must be selected according to the processing procedures and cost effectiveness of chip fabrication. Moreover, solution evaporation during thermal cycling must be accounted for in reducing the solution volume. Therefore, using optical technology as the heat source for PCR processing has attracted increasing research attention. When thermal radiation is used as the heat source in PCR and chips are fabricated using

materials transparent to radiation, the radiant energy does not accumulate as heat on the chip surfaces during heating. Thus, in addition to reduce temperature hysteresis in chips, this method accelerates heating and cooling, effectively shortening the duration of PCR operations.⁶

A laser beam can be focused on a small area and with a high energy concentration, they can heat target objects quickly because of the photothermal effect. Therefore, lasers are often used as the heating source for microfluidic chips.^{7–15} Furthermore, a laser can be used to heat a specific area of the solution rapidly while preventing the surrounding components from being heated. In addition, switching off the radiant heat source can directly reduce the temperature of the sample without necessitating additional cooling devices. Using a laser-heating system integrated with microfluidic devices, Tan and Takeuchi,¹⁰ and Hung and Huang,¹⁴ have focused a laser on a heating area which generated bubbles to release cells. Subsequently, a laser was used to induce cell lysates and denature DNA. By using an infrared (IR) laser to irradiate DNA, DNA denaturation can be achieved^{16–20} and applied to target-sequence detection.¹⁹ Because lasers enable the rapid heating of substances, they can be employed to induce microdroplet PCR^{21–23} and to investigate PCR within microchannels.²⁴

After PCR, the delay before product analysis can affect the target-sequence detection rate. In conventional PCR product detection, the target sequence is detected through agarose gel electrophoresis by using a fluorescent dye (e.g., ethidium bromide; EtBr). However, colloidal production and electrophoresis are time intensive and complex. Therefore, the current methods for increasing PCR product detection speed are primarily based on optical detection technologies which are advantageous because of their ease of observation, rapid detection rate,

*Address all correspondence to: Min-Sheng Hung, E-mail: mshung@mail.nyu.edu.tw

high sensitivity, and low-detection limit. An example of an optical detection technology is quantitative PCR (qPCR), which is realized by integrating fluorescence sensing and PCR. A major characteristic of qPCR is that calibration is performed during PCR to enable quantitative product analysis: the fluorescence change in the target analyte before and after the reaction is converted into a quantitative relationship. Among the various qPCR systems, fluorescence-based detection technologies are the most advanced and can feature either an intercalating fluorescent dye (e.g., SYBE Green I) or a fluorescent probe (e.g., molecular beacon).^{25,26} However, real-time qPCR instruments are expensive and large. Therefore, developing a portable qPCR instrument is a critical research focus.^{27,28}

The use of microchannels in the development of micro-PCR chips and relevant detection technology are vital for developing microsystems that are easy to operate. Because of the specific properties of microchannels produced using polydimethylsiloxane (PDMS), such as high light transmittance, ease of production, and chemical stability, PDMS has been extensively applied in the development of biomedical microsystems.^{29–37} Although PDMS has many advantages, it is permeable to gases and organic molecules.³⁸ Many groups have investigated different forms of surface modification to prevent biosamples from being absorbed on the surface of PDMS during PCR; e.g., using compounds such as polyvinylpyrrolidone,³⁹ polybrene,⁴⁰ bovine serum albumin (BSA),⁴¹ 3-aminopropyltriethoxysilane,²⁸ and glycerol.³⁵ The present study used BSA to prevent biosamples from being absorbed on the surface of microchannels.

By integrating a laser optical system with electro-osmotic flow, this study developed a PDMS microfluidic device that can rapidly replicate and detect DNA. An IR laser was used as the heat source for thermal cycles and a DNA sample was replicated directly within the microchannels. Subsequently, avidin–biotin binding was utilized to immobilize one end of the replicated products to the surface of the microchannels. Simultaneously, the fluorescent intensity was used to detect the target sequences following replication. Electro-osmotic flow was utilized to drive solution delivery to the detection area and induce solution flow in the microfluidic device by exploiting the high fluid shear stress at the solid–liquid interface.^{18,42–47}

2 Materials and Methods

In PCR, which is based on the principle of DNA replication, a forward primer (F-primer) and reverse primer (R-primer) are separately designed for the two ends of the DNA fragments requiring amplification. Furthermore, DNA polymerases (e.g., Taq DNA polymerase) are employed to amplify particular DNA fragments by repeating DNA denaturation, primer annealing, and extension. In this study, a laser was used to heat the target DNA for replication, and the following method was employed to capture the replication products within the microchannels and to demonstrate real-time product detection (Fig. 1). First, avidin–biotin binding was utilized to immobilize one end of the products to the surface of the microchannels. Second, the fluorescent intensity of the quantum dots (Qdots) was used as the basis for product detection.

Figure 2 is a schematic depicting the experimental system used in this study. The system comprises a microfluidic device, a peristaltic pump for solution replacement (Ismatec IPC), an IR laser (1455 nm, CW, Raman fiber laser, IPG Laser GmbH), an observation system comprising a microscope (Olympus BX-51, Japan),

Fig. 1 Schematic describing the process of replication and detection of the DNA strand.

Fig. 2 Schematic of the experimental system setup.

and an electron bombardment charge-coupled device camera (C7190-23, Hamamatsu, Japan). The relationship between the laser output power and temperature was calibrated using fluorescent dye (100 μ M rhodamine B, Ex $\sim$ 550 nm/Em $\sim$ 565 nm, Echo Chemical Co., Ltd., Taiwan)^{48–52} and a heating plate with temperature control. A few groups have investigated rhodamine B diffusing into PDMS; the result may cause a systematic error in the temperature measurement.^{48,50–52} To prevent this phenomenon from occurring, the present study conducted laser temperature calibration in a microchannel comprising a coverslip and a glass plate. The temperature in the microchannels was measured using a thermocouple. Each image was analyzed by selecting and averaging the intensity values in a 30×30 pixels region (representing an area of $20 \times 20 \mu\text{m}^2$). A room-temperature intensity image was initially taken and used to normalize a subsequent intensity image of the same area after heating. Figure 3 shows the calibration results of the 12 experiments. According to Fig. 3, Table 1 lists the operating conditions under which laser heating was employed for thermal cycling.

Figure 4 shows the design of the microfluidic device. The sample solutions were injected into the laser-heating area

Fig. 3 Temperature calibrated curve of laser power in using fluorescence-based temperature measurement.

Table 1 Cycle conditions of laser-induced heating and PCR device.

	Laser cycling conditions	PCR device cycling conditions
Denature	144 mW (94°C)/3 min	94°C/3 min
Denature	144 mW (94°C)/20 s	94°C/20 s
Anneal	48 mW (55°C)/20 s	55°C/20 s
Extend	74 mW (72°C)/40 s	72°C/40 s
Final extension	74 mW (72°C)/7 min	72°C/7 min

(diameter: 200 μm) for DNA replication. Subsequently, electro-osmotic flow was utilized to drive the products into the detection area. The procedure for microfluidic device production is as follows:^{28,30,37,53} (a) lithography was performed to produce a negative photoresist master mold (SU8-2025, MicroChem). (b) PDMS mixing solution (Sylgard 184, Dow Corning) was used to fabricate a PDMS replica from the SU8 mold.

Fig. 4 Schematic of the microfluidic device.

(c) The surface of the PDMS replica was modified using oxygen plasma and bound with a coverslip. The height of the micro-channels was approximately 40 μm .

The test sample used in this study was λ DNA (48502 bp), and fluorescent dye (YO-PRO-1, Ex 491 nm/Em 509 nm) was used to stain the DNA. The positions of replication were from 24,808 to 25,237 bp on the 5' end of the λ DNA (replication length: 429 bp). The biotinylated primers with biotin on the 5' end (Mission Biotech Co., Taiwan), as designed in this study, were F-primer: 5'-GTCTTCCTGCCTCCAGTTC-3' and R-primer: 5'-TTACCTACGACAGGACACAC-3'.

Prior to the experiment, avidin was prepared in the detection area and Qdots were injected to confirm the Qdot adsorption on the surface of the microchannels. The experimental procedure is described as follows:¹⁸ (1) 100 ng/mL of avidin in a pH 7.2 buffer (10 mM Tris-HCl, 50 mM KCl) was injected into the detection area [inlet 2 $\rightarrow$ outlet 2 of Fig. 4, as shown in Fig. 5(a) (the photo was artificially brightened)] and incubated at 25°C for 50 min. (2) The buffer was used to wash away any surplus avidin (inlet 2 $\rightarrow$ outlet 2 of Fig. 4). (3) BSA solution (20 mg/mL, Sigma-Aldrich Co.) was injected into the detection area (inlet 1 $\rightarrow$ outlet 2 of Fig. 4) and incubated at 25°C for 20 min to block the adsorption of the biosamples and Qdots onto the microchannel surface. (4) A 1 nM Qdot (Qdots 655 with streptavidin, Em 653 nm, $\phi \sim 20$ nm, Invitrogen Co.) was injected into the detection area (inlet 2 $\rightarrow$ outlet 2 of Fig. 4) and incubated at 25°C for 50 min. (5) The buffer was used to rinse away any surplus Qdots (inlet 2 $\rightarrow$ outlet 2 of Fig. 4).

Subsequently, the DNA in the microchannels was replicated and detected. The procedure is described as follows: (1) The sample solution (1 ng/ μL of λ DNA, 10 μM YO-PRO-1, 1% β -methylphenethylamine, 1 μM biotinylated F-primer, 1 μM biotinylated R-primer, 0.2 mM dNTP, 0.02 U/mL of Pro Taq DNA Polymerase, 10 $\times$ buffer) was injected into the laser-heating area (inlet 1 $\rightarrow$ outlet 1 of Fig. 4). (2) Laser heating was employed to replicate the DNA. (3) Electro-osmotic flow (DC 15 V/cm) was used to deliver the products to the detection area [Fig. 5(b)] for incubation at 25°C for 40 min. (4) The buffer

Fig. 5 Visualization of flow used in fluorescent microspheres (1 μm): (a) injecting flow (photo was artificially brightened) and (b) electro-osmotic flow from the laser-heating area to the detection area.

was used to rinse surplus products (inlet 2 $\rightarrow$ outlet 2 of Fig. 4). (5) Qdots (1 nM) were injected into the detection area (inlet 2 $\rightarrow$ outlet 2 of Fig. 4) and incubated at 25°C for 50 min. (6) The buffer was used to rinse any surplus Qdots (inlet 2 $\rightarrow$ outlet 2 of Fig. 4). Blocking and washing were of particular importance to ensure that a favorable signal-to-noise ratio was achieved.

Theoretically, approximately 3.4×10^7 avidin molecules were coated in the detection area (the volume of the detection area was $39 \times 10^{-3} \mu\text{L}$ and the molecular weight of avidin was 69,000 Da), and the laser spot was approximately 20 μm in diameter. The heating area was located within the 20- μm laser spot. Therefore, approximately 2.4×10^3 λDNA molecules were replicated in the laser-heating area (the volume of the laser-heating area was $12.6 \times 10^{-4} \mu\text{L}$ and the molecular weight of λDNA is 650 Da/bp). Furthermore, the number of primers was approximately 6.02×10^{11} , and approximately 9.4×10^{11} Qdots entered the detection area.

3 Results and Discussion

BSA-coated Qdots and uncoated Qdots were injected into the microchannels and incubated for 50 min after the surface of the microchannels was coated with avidin; subsequently, the buffer was used to rinse surplus Qdots. Figure 6 shows the nonspecific adsorption of these Qdots on the surface of the microchannels and indicates that coating the surface with BSA effectively reduced the nonspecific adsorption of Qdots on the surface of the microchannels. (Please note that the photos were artificially brightened.) BSA coating on the microchannels can effectively block the adsorption of Qdots; it also can block biosamples adsorbed onto the microchannels of PDMS during the DNA replication process. As mentioned regarding previous studies,^{28,35,38–41} modifying PDMS microchannels is necessary for successful DNA replication. Our experiment revealed that

this result was also true for our DNA replication microchannels. When the sample of the λDNA mixture was injected into the laser-heating area of the PDMS microchannels without modification in our work, the fluorescence of λDNA was almost observed on the surface of the microchannels rather than in the solution (because of the λDNA adsorption on the surface of the microchannels), thus exhibiting DNA replication failure (data not shown).

A λDNA solution was injected into the laser-heating area of the microchannels for replication. To investigate whether solution evaporation occurred, this study used microchannels with and without a droplet of mineral oil at the inlet of the microfluidics during laser irradiation. In the experiment, 10, 15, 20, 25, and 30 thermal cycles were performed. No bubble was generated in the microchannels, even without mineral oil and at 30 thermal cycles through laser-induced heating. The result is consistent with that of our previous study: bubbles could be generated at a laser power greater than 157 mW¹⁴ (94°C at a laser power of 144 mW in the present work). The reason may be that the laser irradiation area was small, the temperature was lower than the boiling point of the solution, and the pressure was not acutely changed in the microchannels during the laser-induced heating process in our experiment.⁵⁴

Figures 7(a)–7(d) show the fluorescent intensity of the Qdots adsorbed on the surface of the microchannels after 10, 20, 25, and 30 thermal cycles, respectively. (Please note that the photos of Fig. 7 were artificially brightened.) Clearly, the fluorescent intensity of the Qdots adsorbed on the surface of the microchannels after replication increased with the number of thermal cycles, indicating that more products bound to avidin on the surface of the microchannels. Notably, the velocity of electroosmotic flow is critical for carrying the replicated products to the detection area for successful detection. If the time of the applied field is longer or shorter, the products flow through

Fig. 6 Fluorescent images of Qdots (photos were artificially brightened) on the microchannel surfaces: (a) without BSA and (b) coating with BSA.

Fig. 7 Fluorescent images of Qdots (photos were artificially brightened) on the microchannel surfaces after thermal cycles of (a) 10, (b) 20, (c) 25, and (d) 30 through laser-induced heating.

the detection area and toward the outlet or cannot flow into the detection area, respectively. Therefore, the velocity of electro-osmotic flow must be measured before DNA replication. In our work, the velocity of electro-osmotic flow was 2×10^{-4} m/s [Fig. 5(b)] and the time of the applied field was 15 s for carrying the products toward the detection area. In addition, the PDMS microchannels can be used only once for DNA replication and detection because the avidin on the surface of the detection area was bound and occupied by the replicated products.

The fluorescent intensity of the Qdots can be expressed as $R = (I_t - I_{bla}) / (I_{pro} - I_{bla})$, where I_t denotes the fluorescent intensity of the Qdots adsorbed to the surface of the microchannels after DNA replication, I_{pro} is the fluorescent intensity of Qdots attached to the surface of microchannels before DNA replication, and I_{bla} is the background intensity of the surface of the microchannels.⁵⁵ These fluorescence intensities [distribution of gray values (max: 256; min: 0) of the total pixels in the image] were calculated using a commercial image-processing software (ImageJ). $R > 1$ indicates a difference in the calculated and background intensities. The area of the surface of the microchannels within the detection area for the calculation was $650 \times 1500 \mu\text{m}^2$. Figure 8 shows the calculated results. R exceeded 10 when the number of thermal cycles was only 10, indicating that the fluorescent intensity of the Qdots differed considerably before and after replication. R exceeded 25 when the number of thermal cycles was 20, indicating that more postreplicated products had bound with the avidin on the surface of the microchannels. When the number of thermal cycles exceeded 20, the increase in R slowed, possibly because most of the avidin on the surface of the microchannels had already bound to the postreplicated products. Regarding the theoretical efficiency of DNA amplification, approximately 2.5×10^8 products (slightly greater than the number of avidins in the detection area) and 2.6×10^{12} (10^5 -fold that of the avidins) were formed when the numbers of thermal cycles were

20 and 30, respectively. Therefore, beyond 20 thermal cycles, the increase in the number of replicated products did not result in a significant increase in the fluorescent intensity of the Qdots, which is attributable to the saturated binding between the products and the avidins in the detection area.

To compare gel electrophoresis detection with the proposed method, we recovered the replication products in the microchannels obtained through laser-induced heating, and agarose gel electrophoresis was used to detect the products. The results are listed in Fig. 9: column 4 indicates the marker; columns 3 and 6, respectively, indicate the replicated products after 20 and 30 thermal cycles of using PCR devices (cycling conditions shown in Table 1); and columns 1, 2, and 5 indicate the

Fig. 8 Calculated results of fluorescent intensity at various thermal cycles through laser-induced heating.

Fig. 9 Results of DNA replication. (Column 1: laser heating 20 cycles, Column 2: laser heating 25 cycles, Column 3: 20 cycles by PCR device, Column 4: marker, Column 5: laser heating 30 cycles, and Column 6: 30 cycles from using the PCR device.)

replicated products obtained after 20, 25, and 30 thermal cycles, respectively, through laser-induced heating. In the gel electrophoresis method, the replicated products that involved using laser-induced heating were detected only after 30 thermal cycles (Fig. 9), implying that the DNA was insufficient. By contrast, in the proposed method, the products were detected in ≥ 10 thermal cycles. Therefore, compared with gel electrophoresis-based detection methods, the proposed method is significantly more sensitive and faster.

4 Conclusions

This study developed a real-time microfluidic platform for DNA replication and detection that exhibits the features of portability, rapid heating, cooling, and high-detection sensitivity. Microchannels were sectioned into a laser-heating area and a detection area, and an IR laser was used as the heat source for thermal cycling. The replicated products were driven to the detection area through electro-osmotic flow. Finally, the fluorescent intensity of the Qdots was used to detect the replicated products.

This study demonstrated that surface coating microchannels with BSA effectively inhibits nonspecific adsorption of Qdots. Additionally, because of the design of the biotinylated primers, biotins on one end of the replicated products bind to the avidins on the surface of the microchannels, thereby becoming immobilized. Moreover, the biotins on the other end of the products bind to the Qdots. Therefore, the fluorescent intensity of the Qdots can be used to detect the products in real time. Compared with gel electrophoresis-based detection methods, the developed method entails avidin–biotin binding for direct and more sensitive product detection in microchannels. Through this method, the products can be detected after only 10 thermal cycles of laser-induced heating, thus achieving superior detection sensitivity. Furthermore, because a laser was used as the heat source, no heater is required in the microchannels; thus, a simple microchannel design is realized. The method

enables rapid heating and cooling and thus can be applied to microanalysis systems for analyzing microliter samples such as those in microdroplet-based microfluidics.

Acknowledgments

We thank Professor Masao Washizu and Professor Hidehiro Oana of the University of Tokyo, and Professor Yu-Yuan P. Wo of National Chiayi University for valuable discussions. The authors deeply appreciate the partial financial support by the Ministry of Science and Technology, Taiwan (Grant Nos. NSC 102-2628-B-415-001-MY2 and MOST 104-2313-B-415-005).

References

1. M. Bu et al., "Design and theoretical evaluation of a novel microfluidic device to be used for PCR," *J. Micromech. Microeng.* **13**(4), S125–S130 (2003).
2. H. Gong et al., "Microfluidic handling of PCR solution and DNA amplification on a reaction chamber array biochip," *Biomed. Microdevices* **8**(2), 167–176 (2006).
3. Y.-F. Hsieh et al., "Polymerase chain reaction with phase change as intrinsic thermal control," *Appl. Phys. Lett.* **102**(17), 173701 (2013).
4. S. Park et al., "Advances in microfluidic PCR for point-of-care infectious disease diagnostics," *Biotechnol. Adv.* **29**(6), 830–839 (2011).
5. W. Wu, K.-T. Kang, and N.-Y. Lee, "Bubble-free on-chip continuous-flow polymerase chain reaction: concept and application," *Analyst* **136**(11), 2287–2293 (2011).
6. G. P. McDermott et al., "Multiplexed target detection using DNA-binding dye chemistry in droplet digital PCR," *Anal. Chem.* **85**(23), 11619–11627 (2013).
7. M. N. Silyadnev et al., "Photothermal temperature control of a chemical reaction on a microchip using an infrared diode laser," *Anal. Chem.* **73**(16), 4037–4044 (2001).
8. T. Arakawa et al., "High-speed particles and biomolecules sorting microsystem using thermosensitive hydrogel," *Meas. Sci. Technol.* **17**(12), 3141–3146 (2006).
9. M. A. Yaseen et al., "Laser-induced heating of dextran-coated mesocapsules containing indocyanine green," *Biotechnol. Prog.* **23**(6), 1431–1440 (2007).
10. W.-H. Tan and S. Takeuchi, "Dynamic microarray system with gentle retrieval mechanism for cell-encapsulating hydrogel beads," *Lab Chip* **8**(2), 259–266 (2008).
11. L. H. Thamdrup, N. B. Larsen, and A. Kristensen, "Light-induced local heating for thermophoretic manipulation of DNA in polymer micro- and nanochannels," *Nano Lett.* **10**(3), 826–832 (2010).
12. A. Q. Jian et al., "Microfluidic flow direction control using continuous-wave laser," *Sens. Actuators A* **188**, 329–334 (2012).
13. W. Lee and X. Fan, "Intracavity DNA melting analysis with optofluidic lasers," *Anal. Chem.* **84**(21), 9558–9563 (2012).
14. M.-S. Hung and Y.-T. Huang, "Laser-induced heating for cell release and cellular DNA denaturation in a microfluidics," *Biochip J.* **7**(4), 319–324 (2013).
15. N. E. Dina et al., "Structural changes induced in grapevine (*Vitis vinifera* L.) DNA by femtosecond IR laser pulses: a surface-enhanced Raman spectroscopic study," *Nanomaterials* **6**(6), 96 (2016).
16. P. Baaske, S. Dühr, and D. Braun, "Melting curve analysis in a snapshot," *Appl. Phys. Lett.* **91**(13), 133901 (2007).
17. C. B. Mast and D. Braun, "Thermal trap for DNA replication," *Phys. Rev. Lett.* **104**(18), 188102 (2010).
18. M.-S. Hung, O. Kurosawa, and M. Washizu, "Single DNA molecule denaturation using laser-induced heating," *Mol. Cell. Probes* **26**(3), 107–112 (2012).
19. E. K. Wheeler et al., "On-chip laser-induced DNA dehybridization," *Analyst* **138**(13), 3692–3696 (2013).
20. R. Zhao et al., "Laser-assisted single-molecule refolding (LASR)," *Biophys. J.* **99**(6), 1925–1931 (2010).
21. H. Terazono et al., "Development of a high-speed real-time polymerase chain reaction system using a circulating water-based rapid heat-exchange," *Jpn. J. Appl. Phys.* **49**(6s), 06GM05 (2010).

22. K. Hettiarachchi, H. Kim, and G. W. Faris, "Optical manipulation and control of real-time PCR in cell encapsulating microdroplets by IR laser," *Microfluid. Nanofluid.* **13**(6), 967–975 (2012).
23. H. Kim, S. Vishniakou, and G. W. Faris, "Petri dish PCR: laser-heated reactions in nanoliter droplet arrays," *Lab Chip* **9**(9), 1230–1235 (2009).
24. Y. Yu et al., "Quantitative polymerase chain reaction using infrared heating on a microfluidic chip," *Anal. Chem.* **84**(6), 2825–2829 (2012).
25. D. K. Yang et al., "Selection of aptamers for fluorescent detection of alpha-methylacyl-CoA racemase by single-bead SELEX," *Biosens. Bioelectron.* **62**, 106–112 (2014).
26. K.-C. Li et al., "Melting analysis on microbeads in rapid temperature-gradient inside microchannels for single nucleotide polymorphisms detection," *Biomicrofluidics* **8**(6), 064109 (2014).
27. D. Sugumar et al., "Rapid multi sample DNA amplification using rotary-linear polymerase chain reaction device (PCRDisc)," *Biomicrofluidics* **6**(1), 014119 (2012).
28. D. Khodakov et al., "DNA capture-probe based separation of double-stranded polymerase chain reaction amplification products in poly(dimethylsiloxane) microfluidic channels," *Biomicrofluidics* **6**(2), 026503 (2012).
29. S. W. Rhee et al., "Patterned cell culture inside microfluidic devices," *Lab Chip* **5**(1), 102–107 (2005).
30. K.-Y. Lien et al., "Microfluidic systems integrated with a sample pretreatment device for fast nucleic-acid amplification," *J. Microelectromech. Syst.* **17**(2), 288–301 (2008).
31. N. Ramalingam et al., "Real-time PCR-based microfluidic array chip for simultaneous detection of multiple waterborne pathogens," *Sens. Actuators B* **145**(1), 543–552 (2010).
32. J. S. Kuo and D. T. Chiu, "Disposable microfluidic substrates: transitioning from the research laboratory into the clinic," *Lab Chip* **11**(16), 2656–2665 (2011).
33. J. H. L. Beal et al., "A rapid, inexpensive surface treatment for enhanced functionality of polydimethylsiloxane microfluidic channels," *Biomicrofluidics* **6**(3), 36503 (2012).
34. C. S. Thompson and A. R. Abate, "Adhesive-based bonding technique for PDMS microfluidic devices," *Lab Chip* **13**(4), 632–635 (2013).
35. N. B. Trung et al., "Multi-chamber PCR chip with simple liquid introduction utilizing the gas permeability of polydimethylsiloxane," *Sens. Actuators B Chem.* **149**(1), 284–290 (2010).
36. A. C. Hatch et al., "Continuous flow real-time PCR device using multi-channel fluorescence excitation and detection," *Lab Chip* **14**(3), 562–568 (2014).
37. M.-S. Hung and H.-Y. Chang, "A simple microfluidics for real time plasma separation and hCG detection from whole blood," *J. Chin. Inst. Eng.* **38**(6), 685–691 (2015).
38. M. M. Kim et al., "The improved resistance of PDMS to pressure-induced deformation and chemical solvent swelling for microfluidic devices," *Microelectron. Eng.* **124**, 66–75 (2014).
39. J. A. Kim et al., "Fabrication and characterization of a PDMS-glass hybrid continuous-flow PCR chip," *Biochem. Eng. J.* **29**(1–2), 91–97 (2006).
40. D. Erickson et al., "Electrokinetically based approach for single-nucleotide polymorphism discrimination using a microfluidic device," *Anal. Chem.* **77**(13), 4000–4007 (2005).
41. T. Nakayama et al., "Circumventing air bubbles in microfluidic systems and quantitative continuous-flow PCR applications," *Anal. Bioanal. Chem.* **386**(5), 1327–1333 (2006).
42. C. Chen and J. G. Santiago, "A planar electroosmotic micropump," *J. Microelectromech. Syst.* **11**(6), 672–683 (2002).
43. M. Washizu et al., "Stretching yeast chromosomes using electroosmotic flow," *J. Electrostatics* **57**(3–4), 395–405 (2003).
44. X. Wang et al., "Electroosmotic pumps and their applications in microfluidic systems," *Microfluid. Nanofluid.* **6**(2), 145–162 (2009).
45. Y. Ai et al., "A low-voltage nano-porous electroosmotic pump," *J. Colloid Interface Sci.* **350**(2), 465–470 (2010).
46. M.-S. Hung et al., "Stretching DNA fibers out of a chromosome in solution using electroosmotic flow," *J. Chin. Soc. Mech. Eng.* **30**(4), 289–295 (2009).
47. M.-S. Hung and P.-C. Chen, "Extending DNA from a single cell using integrated system of electro-osmosis and AFM," *J. Med. Biol. Eng.* **30**(1), 29–34 (2010).
48. D. Ross et al., "Temperature measurement in microfluidic systems using a temperature-dependent fluorescent dye," *Anal. Chem.* **73**(17), 4117–4123 (2001).
49. D. Erickson, D. Sinton, and D. Li, "Joule heating and heat transfer in poly(dimethylsiloxane) microfluidic systems," *Lab Chip* **3**(3), 141–149 (2003).
50. R. Samy, T. Glawdel, and C. L. Ren, "Method for microfluidic whole-chip temperature measurement using thin-film poly(dimethylsiloxane)/rhodamine B," *Anal. Chem.* **80**(2), 369–375 (2008).
51. S. H. Lee et al., "Use of directly molded poly(methyl methacrylate) channels for microfluidic applications," *Lab Chip* **10**(23), 3300–3306 (2010).
52. H. Sasaki et al., "Parylene-coating PDMS microfluidic channels prevents the absorption of fluorescent dyes," *Sens. Actuators B* **150**(1), 478–482 (2010).
53. M.-S. Hung and Y.-T. Chang, "Single cell lysis and DNA extending using electroporation microfluidic device," *Biochip J.* **6**(1), 84–90 (2012).
54. T. Sato et al., "Evidence for hydrogen generation in laser- or spark-induced cavitation bubbles," *Appl. Phys. Lett.* **102**(7), 074105 (2013).
55. H. Du et al., "Sensitivity and specificity of metal surface-immobilized molecular beacon biosensors," *J. Am. Chem. Soc.* **127**(21), 7932–7940 (2005).

Min-Sheng Hung is a professor at the National Chiayi University of Taiwan. He received his BS and MS degrees in engineering from the National Cheng Kung University and National Taiwan University in 1993 and 1995, respectively, and his PhD in mechanical engineering from National Taiwan University in 2001. He is a reviewing associate editor of an EI Compendex (Engineering Village) journal, and has written two book chapters. His current research interests include laser and bionanotechnology.

Chia-Chin Ho is an engineer at Chip Win Technology Co., Ltd. of Taiwan. He received his BS and MS degrees in Engineering from the Feng Chia University and National Chiayi University in 2012 and 2014, respectively. His current research interests include microfluidics and microelectromechanical.

Chih-Pin Chen received his BS and MS degrees in Engineering from the National Chiayi University in 2011 and 2014, respectively. His current research interests include liquid-crystal display and nanotechnology.