Scattering correction through a space-variant blind deconvolution algorithm

Koberstein-Schwarz Benno
Omlor Lars
Schmitt-Manderbach Tobias
Mappes Timo
Ntziachristos Vasilis

Scattering correction through a space-variant blind deconvolution algorithm

Koberstein-Schwarz Benno,^{a,b,*} Omlor Lars,^a Schmitt-Manderbach Tobias,^a Mappes Timo,^c and Ntziachristos Vasilis^b

^aCarl Zeiss AG, Corporate Research and Technology, Carl-Zeiss-Promenade 10, 07745 Jena, Germany

^bInstitute for Biological and Medical Imaging, Technische Universität München and Helmholtz Zentrum, München, Ingolstädter Landstrasse 1, 85764 Neuherberg, Germany

^cCarl Zeiss Vision International GmbH, Technology and Innovation, Turnstr. 27, 73430 Aalen, Germany

Abstract. Scattering within biological samples limits the imaging depth and the resolution in microscopy. We present a prior and regularization approach for blind deconvolution algorithms to correct the influence of scattering to increase the imaging depth and resolution. The effect of the prior is demonstrated on a three-dimensional image stack of a zebrafish embryo captured with a selective plane illumination microscope. Blind deconvolution algorithms model the recorded image as a convolution between the distribution of fluorophores and a point spread function (PSF). Our prior uses image information from adjacent z-planes to estimate the unknown blur in tissue. The increased size of the PSF due to the cascading effect of scattering in deeper tissue is accounted for by a depth adaptive regularizer model. In a zebrafish sample, we were able to extend the point in depth, where scattering has a significant effect on the image quality by around 30 μm . © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.21.9.096005]

Keywords: scattering correction; three-dimensional imaging; selective plane illumination microscope; blind deconvolution.

Paper 160274LRRR received Apr. 28, 2016; accepted for publication Aug. 15, 2016; published online Sep. 12, 2016.

1 Introduction

One of the most prominent limitations for light microscopy of biological samples is the imaging depth. Scattering in tissue is the limiting factor defining the depth in which biological structures may still be resolved. Larger samples can only be imaged superficially. Deconvolution algorithms for multiview selective plane illumination microscopy (SPIM) have already been shown to increase the ability to image thick samples, provided they are optically accessible from multiple angles.¹

SPIM^{2–4} utilizes an extreme darkfield setup by illuminating only the focal plane in the sample by a light sheet. The illumination optics is usually positioned perpendicular to the detection objective. By scanning the sample through the light sheet, a three-dimensional (3-D) image stack of the sample can be acquired in a short period of time. The technique offers a high acquisition rate of entire volumes and low photo damage and bleaching, thus making it an ideal tool for *in-vivo* imaging of fast processes or developing organisms. In multiview SPIM,^{1,5,6} the sample is rotated and multiple image stacks are recorded from different angles. Alternatively, multiple excitation and detection optics allow imaging of the sample from 2 to 4 directions without rotating the sample. The different views are later used to reconstruct a single volume. The most prominent technique to combine the images are blind deconvolution algorithms.

In this work, we introduce a regularizer and a prior to improve the reconstruction and make it more robust. For demonstration purposes, we concentrate on using a single SPIM image stack rather than using multiview data (although the regularizer is applicable to multiview data). This approach offers

the advantage of using an SPIM setup without multiple detection optics and a fast acquisition without rotation of the sample. Since the SPIM image stack has to meet only some weak requirements for the reconstruction, it is also possible to improve already existing single view SPIM data.

2 Reconstruction Algorithm

The scattering influence on intensity within a sample can be treated as a convolution between the real object and an increasingly broad point spread function (PSF). The PSF describes the effect an imaging system has on a point source. For SPIM, the PSF is shaped by three main contributors: (1) the illumination function of the sample. The shape of the light sheet as well as broadening of the light sheet within the sample due to scattering determines the illuminated areas of the sample. Only from these regions we can expect fluorescence. To acquire a full 3-D image of the sample, the light sheet and the sample are moved relative to each other. The direction of movement will be denoted as z -direction, while the imaging plane will be labeled the $x-y$ plane. The function describing the illumination will be called P_{illum} . (2) Fluorescence molecules, which are illuminated in the current z -position, can be considered point sources. The light from these sources gets scattered on the way to the objective and will contribute to the PSF. This contribution will be denoted as P_{scatt} . (3) Eventually, the detection optics contributes to the PSF. This contribution P_{det} is sample independent and—for most cases considered here—smaller than the PSF broadening due to scattering. In the following, all three separate contributions are condensed into one single PSF $P_z(x, y)$

*Address all correspondence to: Koberstein-Schwarz Benno, E-mail: koberstein-schwarz@helmholtz-muenchen.de

$$I_{z_0}(x, y) = \left[\int O(x, y, z) P_{\text{illum}}(z - z_0) \times P_{\text{scatt}}(x, y, z, z_0) dx dy dz \right] \times P_{\text{det}}.$$

2.1 Spatially Invariant Algorithm

For the sake of clarity, the algorithmic details are first discussed using a spatially invariant model. In each z -plane, the detected SPIM image is treated as a convolution between the real object and the PSF. In this case, the real object is the distribution of fluorophores in the sample. An additional noise-term accounts for additive noise influences on the final image introduced by the imaging system

$$I_z = O_z \times P_z + N, \quad (1)$$

where N is the noise, I is the image, O is the real object, and P is the PSF.

If N is Gaussian noise, the original image can be estimated by minimizing the least-square error E .

$$E = \|I_z - P_z \times O_z\|_2^2. \quad (2)$$

It should be noted that image noise is typically not Gaussian but Poisson distributed. Nevertheless, the image space reconstruction algorithm (ISRA) which will be introduced below turned out to also be stable in the case of Poisson noise. This reconstruction framework is easy to implement and will be used to demonstrate the function of our regularizer and the prior we use. Equation (2) is typically ill conditioned, even if the PSF P is known. In most cases, multiple views are used to provide the necessary amount of information in order for the algorithm to converge. In our case, the acquired image stack I is the only known variable. Trying to minimize the least-square error will not necessarily converge to one solution. To obtain a converging solution, a prior is used. The main assumption for this prior is that if the step size between two consecutive z -planes is small enough so it can be assumed that the same object is illuminated in both positions, while the PSF varies. This idea of using neighboring z -planes is depicted in the illustration of the algorithm in Fig. 1(b). Under ideal conditions and under the assumption of multiple observations, P_i ($i = 1, 2, 3$), which are neighboring

z -planes, in our case, the coprimeness equation: $P_j \times I_i - P_i \times I_j = 0$, $\forall i \neq j$, can be used to determine the PSF. In the presence of noise, this equation becomes an optimization problem

$$E_{\text{coprime}} = \sum_{i \neq j} \|P_j * I_i - P_i * I_j\|_2^2.$$

The effectiveness of algorithms based on this prior has already been shown by multiple groups.⁷⁻⁹ However, they use either different views or time series of the same view with slightly varying PSF. Due to the physical properties of the imaging system, it can additionally be assumed that $PSF \geq 0$, as well as $\sum_{x,y} P(x, y) = 1$. The sum constraint is only valid if absorption in the sample can be neglected, which is reasonable for adjacent z -planes. The positive constraint is also true for the object O . Furthermore, the support of the PSF should be smaller than the image I . These constraints together with the assumptions of a constant object over neighboring z -planes [whereas the PSF P_i ($i = 1, 2, 3$) changes] lead to the following minimization problem:

$$\min_{P_i \geq 0, \sum_{P_i=1, O \geq 0}} \sum_i \|I_i - P_i \times O\|_2^2 + \sum_{i \neq j} \|P_j \times I_i - P_i \times I_j\|_2^2.$$

The last term, the coprimeness regularization mentioned above, helps to stabilize the deconvolution problem and can even lead to a unique solution.⁷

To ensure the convergence of both the PSF P and the object O , an alternating optimization scheme is used, where the optimization iterates between P and O . In this work, the optimization is done with a modified nonnegative least-square estimator, the ISRA.¹⁰⁻¹² In the first step, the PSF is updated by applying the ISRA

$$P_k(t+1) = P_k(t) \frac{O \times O^* \times I_k + \sum_{j \neq k} I_j \times I_j^* \times P_j(t) \times I_k}{(O \times O^* + \sum_{(j \neq k)} I_j \times I_j^*) \times P_k(t)}.$$

In the second iteration step, the PSF is fixed and the object is updated in the same way. The iteration scheme is illustrated in Fig. 1(b) on the right. To ensure convergence of both the PSF and the object, a proximal minimization method is applied to the

Fig. 1 (a) Schematic of the SPIM setup and (b) illustration of the algorithm with the quadratically increasing regularization (graph to the left), the coprimeness prior (to neighboring slices are used to reconstruct the red plane) and the iteration scheme to the right.

algorithm.^{13,14} Alternatively to ISRA, a Richardson–Lucy deconvolution could be used for deconvolution which should result in a slightly longer computation time.

With increasing imaging depth, the images will suffer from increased scattering. The idea in this work is to compensate for this increased scattering by adjusting the algorithm along with the imaging depth. A simple approximation for light scattering in tissue is found in Refs. 15 and 16. With increasing depth, the size of the PSF grows quadratically [see Fig. 1(b), where the increase in regularization with depth is schematically plotted]. This simple model was implemented by the use of the Tikhonov regularization.¹⁷ A higher Tikhonov regularization leads to an increased PSF size in the reconstruction. This is not an intrinsic effect of the regularization itself, but a combination effect with the constant intensity constraint $\sum P_i = 1$. The ratio between the l1-norm (sum constraint) and the l2-norm (Tikhonov regularization) is a measure of signal sparsity with higher values describing sparser signals. Since the Tikhonov term reduces the l2-norm of the solution (this is penalized in the final cost function), the regularized solution automatically has a lower sparsity, i.e., it has a larger support (broader blur).

To reflect the increase in PSF size, the regularization parameter was increased quadratically, which was empirically found to be an effective way to model the increased scattering. However, no assumptions about the exact shape of the PSF are made. As a result of this, even elliptical-shaped PSF (due to aberrations or motion) can be corrected.

2.2 Spatially Varying Algorithm

In general, the scattering will not only vary with depth, but will also depend on the $x - y$ position within the sample. Hence, Eq. (1) generalizes to

$$I_z(x, y) = O(x, y, z) \times P_z(x, y) + N.$$

Space variant deconvolution has been successfully used by other authors to remove motion artifacts.^{18,19} To account for this spatial dependency, a Gabor transform (GT_x) was added to the algorithm to divide the original image into several smaller overlapping windows. The GT_x in this algorithm²⁰ is defined as $GT_x(I) = \mathcal{F}[I \times h(x)^{0.5}]$, where $h(x)$ is the window function. In the presented algorithm, a $\sin(x)^2$ is used to model the window. Consequentially, the inverse transform is given as $I = \sum \mathcal{F}^{-1}[GT_x(I)] \cdot h(x)^{0.5}$. The original image is multiplied by overlapping window functions and the iteration of the PSF together with the deconvolution is done independently on each window. In the end, the object is put back together with the inverse transform. The size of the windows was chosen to be large enough to most closely satisfy the $\sum P_i = 1$ constraint. Since the sample is not necessarily flat and the position of the sample surface might vary from window to window, the depth regularization by the adapted Tikhonov parameter was done independently for each window. A threshold was used to locate the z -position of the sample surface and the depth was measured from there.

3 Results

We demonstrate the effect of the algorithm on a real dataset of a common model organism for biological research, a transgenic zebrafish embryo. The 3-day-old embryo of the type Fli1A had its vasculature labeled with enhanced green fluorescent protein (EGFP). Imaging was done *ex vivo* with SPIM. The

microscope setup, which is schematically shown in Fig. 1(a), was built by ZEISS, (Carl Zeiss AG, Jena, Germany) with a custom-made Apo Z 1.5 $\times$ /0.37 objective (free working distance 30 mm). Two cylindrical lenses ($f = 30$ mm) are used to achieve a double-sided illumination with a light sheet thickness of 16 μm (full width at half maximum). For detection, an sCMOS camera (pco.edge, PCO, Kelheim, Germany) is mounted on the microscope and the raw data are saved as a tiff image stack. The fish was imaged sideways with a step size of 5 μm between adjacent z -planes. Two complete 3-D image stacks of the zebrafish embryo were acquired, and the sample was rotated by 180 deg between the image stacks. While reconstruction was only performed on the first image stack (S1), the second image stack (S2) was used as a control image. Since the imaging was done from two opposed sites, structures which appear deep and scattered in S1 are more visible in S2. This circumstance is used to validate the reconstruction results from S1. To be able to use the second dataset (S2) for validation, a region from the tail of the fish with a width of about 175 μm was selected. Figures 2(a)–2(c) show image slices from three different depths: 105, 130, and 155 μm , respectively. For this sample, the depicted depths are just in the range where scattering starts to drastically reduce the image quality. In the left column, the original slice from dataset S1 is presented, the column in the middle shows the same slice after processing with the introduced algorithm, and the column to the right shows the corresponding slice from dataset S2.

The section shown was reconstructed using 25 windows. The red arrows indicate regions where the reconstructed dataset shows a clear improvement in image quality compared to the original dataset S1. The structures which can be seen in 105- μm depth agree well with the structures in the second dataset S2. The effect is still visible at a depth of 130 μm . The good agreement with the control dataset S2 can also be seen in Figs. 2(g) and 2(h), where line profiles through the slices at 105- [Fig. 2(g)] and 130- μm depth [Fig. 2(h)] are plotted. The original dataset S1 is plotted in blue, red represents our reconstruction, and green corresponds to the control dataset S2. The position of the chosen cut is indicated by a white dashed line in Figs. 2(a) and 2(b). Deeper inside the sample, the ability of the algorithm to correct the image diminishes. At a depth of 155 μm , the image improvement is no longer able to reveal detailed structures.

The adaptive regularization prevents overregularization on image planes closer to the surface. Images from all planes benefit from the used prior, since the algorithm can correct for small motion artifacts which alter the PSF shape of neighboring planes. A region of interest covering part of the head and the eye of the zebrafish is marked red in Fig. 2(d). Figure 2(f) shows a projection of the original image stack. A projection of the reconstructed image stack is depicted in Fig. 2(e). The overall thickness of the head is around 320 μm . The blur due to scattering and other effects is clearly reduced in the reconstructed image and virtually no artifacts can be seen. Two red arrows mark small details which are not/only partly visible in the original maximum intensity projection (MIP) but are clearly visible in the reconstructed MIP.

To further highlight the advantage over deconvolution algorithms without the suggested prior and regularization, two slices from the head region of the zebrafish are shown in Fig. 3. In the upper row, Figs. 3(a)–3(d) depict a plane from a depth of around 140 μm . The lower row in Figs. 3(f)–3(i) show an image plane

Fig. 2 The images show details of a zebrafish from the genetic line Fli1A 3 days post fertilization. The vasculature is labeled with EGFP. (a)–(c) show the same region from the tail of a zebrafish at different depths (105, 130, 155 μm). The column on the left shows the original dataset (S1), the column in the middle shows the reconstruction of the dataset to the left, and the column on the right side shows a second dataset (S2), which was recorded from the opposite side of the fish to validate the reconstruction results. (d) MIP of the original dataset. The red-colored box indicated the region which can be seen in (e) and (f). In (e), the MIP of the reconstructed image stack can be seen, in (f), the MIP of the original image stack. The red arrows indicate areas where the improvement can clearly be seen. (g) and (h) show cuts through the slices in (a) and (b), as indicated by the dashed lines. The blue graph corresponds to the original dataset S1, red to our reconstruction, and green represents the control dataset S2. Scale bar indicates 100 μm .

from the eye close to the surface at a depth of 5 to 10 μm . The original image slices can be seen in Figs. 3(a) and 3(f), and the result from our algorithm in Figs. 3(b) and 3(g). The two planes in Figs. 3(c) and 3(h) were reconstructed without the coprime-ness prior and with a fixed PSF. Instead of iteratively reconstructing both the PSF and the image, a fixed (simulated) Gaussian PSF was used. This kind of correction is widely used in microscopic imaging to correct the optical aberrations of the microscope. Figures 3(d) and 3(i) show a modified version of our algorithm where we kept the prior but used a constant regularization. The amount of regularization was optimized for a depth of around 140 μm . To visualize the difference between the three reconstructions, Figs. 3(e) and 3(j) show a cut through the images. The location of this cut though is indicated by a white dashed line in the original image. In Fig. 3(e), a cut through the plane of 140 μm depth is shown. The original dataset is plotted in blue, the reconstruction with fixed PSF in green, and the results from our algorithm in red. The results from the constant regularization are identical to the results from our algorithm since the regularization was optimized for the depth of 140 μm . As expected, our algorithm enhances structural contrast and is reducing the structure size to correct the scattering, while the reconstruction with the fixed PSF falls behind. In this depth, the PSF broadening due to scattering is much larger than the microscope PSF, therefore, the PSF size in the fixed reconstruction is not sufficient to yield a better result.

The effect of scattering can be neglected at a depth of only 5 μm . Here, we assume the original image Fig. 3(f) to be very close to the ground truth.

Applying excessive regularization can lead to artifacts in planes close to the surface. Figure 3(j) shows a cut through to one of these planes at a depth of 5 μm . The reconstruction with constant regularization is plotted in cyan. One artifact of overregularization could be observed in our simulation and which is also visible in Fig. 3(j) is the formation of oscillations near plateaus or structures of high intensity. The plot in Fig. 3(j) shows an example of such an oscillation at a position of around 50 μm (marked the red arrow). Our algorithm and the fixed PSF reconstruction are much closer to the original dataset at this point. In the image, the oscillations might appear as a “shadow” around bright structures and spatially overlapping objects of different intensity are artificial separated.

The selected step size of 5 μm was sufficiently small for a stable reconstruction. In the shown datasets, most observed features were at least three times larger than the step size. For samples with smaller labeled structures, a smaller step size together with a thinner light sheet would probably be beneficial.

The results from the imaged zebrafish show that the algorithm works on z-stacks of SPIM images. However, this does not mean that the algorithm is limited to this imaging modality. The algorithm should be able to work with other optical microscopy methods, as well as other imaging methods which are limited by the scattering of light in tissue. There are three requirements which must be met: a full 3-D image stack is available, the step size between two adjacent planes is small compared to the observed structures, and the scattering in the sample can sufficiently be described by the introduced model. While a single dataset was used for this demonstration, the algorithm can be extended to handle multiple datasets to form one

Fig. 3 Comparison between different deconvolution methods. The head region of the zebrafish shown in Figure 2 is chosen and two different slices from the volume are picked. The upper row (a)–(d) shows a plane from a depth of $\sim 140\ \mu\text{m}$. The lower row (f)–(i) shows a plane from the surface (of the eye). In (e) and (j), a cut through the image planes is plotted. The location of the cut is indicated by the white dotted line in the original images. In the plot, blue color corresponds to the original dataset, red to the results from our algorithm, green to the reconstruction with fixed PSF, and cyan to the reconstruction with constant regularization. Artifacts produced by the constant regularization are marked with a red arrow. Scale bar indicates $100\ \mu\text{m}$

sample, either with varying illumination or even with varying views of the sample.

All animal experiments have been carried out according to national and institutional guidelines of the Government of Upper Bavaria and the highest standards of safety have been maintained during the course of the project.

4 Conclusion

The acquired results show that the regularizer and prior presented in this work are able to partially correct for scattering within the sample and enhance the image quality and depth. The adaptive regularizer effectively avoids overregularization of the reconstruction. In the specific case of the zebrafish sample, the increase in imaging depth was about $30\ \mu\text{m}$. For this correction, only information from a single image stack was used, together with a simple model for the dispersive broadening of light in tissue. Tikhonov regularization was used efficiently as a model to account for an increased scattering, which comes along with an increased imaging depth. The advantage of this approach is that neither multiple image stacks have to be acquired nor adaptations on the hardware level are required and the algorithm can be applied to already acquired data. The algorithm only requires a volumetric image stack with a small step size between two adjacent imaging planes, thus the algorithm should also be able to work on other imaging techniques which provide a 3-D image volume.

Acknowledgments

The work has received funding from the Federal Ministry of Education and Research (BMBF), Photonic Science Germany, Tech2See-13N12623/-4.

References

1. S. Preibisch et al., "Efficient Bayesian-based multiview deconvolution," *Nat. Methods* **11**(6), 645–648 (2014).
2. J. Huisken et al., "Optical sectioning deep inside live embryos by selective plane illumination microscopy," *Science* **305**, 1007 (2004).
3. J. Huisken and D. Y. R. Stainier, "Selective plane illumination microscopy techniques in developmental biology," *Development* **136**(12), 1963–1975 (2009).
4. P. Keller et al., "Reconstruction of zebrafish early embryonic development by scanned light sheet microscopy," *Science* **322**, 1065–1069 (2008).
5. U. Krzic et al., "Multiview light-sheet microscope for rapid in toto imaging," *Nat. Methods* **9**(7), 730–733 (2012).
6. R. K. Chhetri et al., "Whole-animal functional and developmental imaging with isotropic spatial resolution," *Nat. Methods* **12**, 1171–1178 (2015).
7. G. Harikumar and Y. Bresler, "Perfect blind restoration of images blurred by multiple filters: theory and efficient algorithms," *IEEE Trans. Image Process.* **8**(2), 202–219 (1999).
8. F. Sroubek and J. Flusser, "Multichannel blind iterative image restoration," *IEEE Trans. Image Process.* **12**(9), 1094–1106 (2003).
9. F. Sroubek and J. Flusser, "Multichannel blind deconvolution of spatially misaligned images," *IEEE Trans. Image Process.* **14**(7), 874–883 (2005).
10. H. Lanteri, R. Soummer, and C. Aime, "Comparison between ISRA and RLA algorithms. Use of a Wiener filter based stopping criterion," *Astron. Astrophys. Suppl. Ser.* **140**(2), 235–246 (1999).
11. G. Archer and D. Titterton, "The iterative image space reconstruction algorithm (ISRA) as an alternative to the EM algorithm for solving positive linear inverse problems," *Stat. Sin.* **5**, 77–96 (1995).
12. A. Reader, "Generalization of the image space reconstruction algorithm," in *Nuclear Science Symp. Conf. Record*, pp. 4233–4238 (2011).
13. D. P. Bertsekas, "On penalty and multiplier methods for constrained minimization," *SIAM J. Control Optim.* **14**(2), 216–235 (1976).
14. D. P. Bertsekas and J. N. Tsitsiklis, *Parallel and Distributed Computation: Numerical Methods*, Vol. **23**, Athena Scientific (1997).
15. S. Premoze and M. Ashikhmin, "Practical rendering of multiple scattering effects in participating media," in *Eurographics Symp. on Rendering* (2004).
16. J. W. McLean, J. D. Freeman, and R. E. Walker, "Beam spread function with time dispersion," *Appl. Opt.* **37**(21), 4701–4711 (1998).
17. F. Benvenuto and M. Piana, "Regularization of multiplicative iterative algorithms with nonnegative constraint," *Inverse Probl.* **30**(3) (2014).

18. H. Hirsch et al., "Efficient filter flow for space-variant multiframe blind deconvolution," in *IEEE Conf. on Computer Vision and Pattern Recognition (CVPR)*, pp. 607–614 (2010).
19. S. Harmeling, H. Michael, and B. Schölkopf, "Space-variant single-image blind deconvolution for removing camera shake," in *Advances in Neural Information Processing Systems*, pp. 829–837 (2010).
20. M. Lamoureux et al., "A fast, discrete Gabor transform via a partition of unity," CREWES Technical Report, Vol. **15**, pp. 1067–1078 (2003).

Biographies for the authors are not available.