Recent advances toward preclinical and clinical translation of photoacoustic tomography: a review

Paul Kumar Upputuri
Manojit Pramanik

Recent advances toward preclinical and clinical translation of photoacoustic tomography: a review

Paul Kumar Upputuri and Manojit Pramanik*

Nanyang Technological University, School of Chemical and Biomedical Engineering, 62 Nanyang Drive, Singapore 637459, Singapore

Abstract. Photoacoustic imaging is an emerging hybrid imaging modality that can provide multicontrast, multiscale imaging of biological features ranging from organelles to organs. The three major embodiments of photoacoustic imaging are microscopy, endoscopy, and computed tomography. Photoacoustic tomography (PAT) or photoacoustic computed tomography allows deep-tissue imaging, and hence it is more suitable for whole body preclinical/clinical imaging applications. Due to fast-growing laser technology and ultrasound detector technology, PAT is evolving rapidly, leading to a quicker translation into clinical trials. We review the recent developments of PAT systems and their applications in preclinical and clinical practices. © 2016 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: [10.1117/1.JBO.22.4.041006](https://doi.org/10.1117/1.JBO.22.4.041006)]

Keywords: photoacoustic tomography; photoacoustic computed tomography; optoacoustic tomography; clinical translation; preclinical imaging.

Paper 160497VSSR received Jul. 18, 2016; accepted for publication Oct. 31, 2016; published online Nov. 22, 2016.

1 Introduction

Medical imaging is continuously advancing along with newer technologies. The advances in medical imaging tools [x-ray computed tomography (x-ray CT), ultrasound (US), magnetic resonance imaging (MRI), positron emission tomography (PET), and single-photon emission computed tomography etc.] have led to better diagnosis and patient care.^{1–4} In the last decade, a new imaging technology called photoacoustic imaging (PAI) has evolved and is moving rapidly from the laboratory to preclinical and clinical phases.^{5–7} PAI is a three-dimensional (3-D) hybrid imaging modality that combines the advantage of optical and acoustic imaging. In PAI, a nanosecond laser illuminates the target sample (tissue/body parts). Due to absorption of incident light by the tissue chromophores (such as melanin, red blood cells, and water), there is a local temperature rise, which in turn produces pressure waves emitted in the form of acoustics waves [also known as photoacoustic (PA) waves]. By acquiring the generated PA waves at the tissue surface, images can be formed, revealing the internal structures and function. PAI provides anatomical, functional, metabolic, molecular, and genetic contrasts of vasculature, hemodynamics, oxygen metabolism, biomarkers, and gene expression.^{7,8} These unique advantages made PAI a potential tool for many applications in vascular biology, oncology, neurology, ophthalmology, dermatology, gastroenterology, and cardiology.⁷

The merits of PAI compared with pure optical and pure ultrasonic imaging modalities are: (i) it has high intrinsic optical absorption contrast, hence no labeling is required, (ii) it can provide better penetration depth (0.1 to 12 cm)⁹ compared to optical methods, such as optical microscopy, fluorescence microscopy,¹⁰ Raman microscopy,^{11,12} multiphoton microscopy,¹³ optical coherence tomography,^{14,15} and diffuse optical tomography,¹⁶ (iii) unlike US imaging, PAI is free from speckle

artifacts, (iv) it is compact, faster, and less expensive compared with MRI, PET, and x-ray CT, (vi) it can provide both structural and functional information using multiwavelengths, (vii) it directly images optical absorption with 100% relative sensitivity and the sensitivity is two orders of magnitude greater than those of confocal microscopy and optical coherence tomography,⁷ and (viii) it can provide multiscale multicontrast images of biological structures, ranging from organelles to organs.

Several comprehensive reviews on PAI can be found in the literature.^{6,7,17–22} Three types of PAI systems have been reported: (i) photoacoustic microscopy (PAM),^{23–39} (ii) photoacoustic endoscopy (PAE),^{40–45} (iii) photoacoustic tomography (PAT) or photoacoustic computed tomography (PACT).^{46–54} In PAM, both the laser excitation and acoustic detection are focused (tightly or loosely depending on the configuration), and the dual foci are configured confocally to enhance measurement sensitivity. Depending on whether the acoustic focus or optical focus is finer, PAM is categorized into acoustic-resolution PAM^{32,37–39} or optical-resolution PAM.^{24,30,34} Subwavelength PAM was also demonstrated to achieve nanometric resolution in PAM systems.^{55–58} Like PAM, PAE is used to image internal organs, such as, the esophagus and colon. Unlike optical endoscopy, PAE can provide much deeper imaging. A 7-mm deep PAE imaging was demonstrated in the dorsal region of a rat colon *ex vivo*.⁴⁴ Both PAM and PAE are still limited to shallower imaging depth (with higher spatial resolution). For deep tissue, imaging PAT/PACT will be more suitable and hence holds a better chance to find a clinical application. This review will focus on the progress of the PAT system in recent years for translating into preclinical and clinical applications.

In PAT, typically an expanded nanosecond laser beam illuminates the tissue surface homogeneously. PA waves are acquired around the tissue boundary using a wideband ultrasound transducer (UST). Generally, circular scanning geometry in orthogonal excitation mode is preferred for deep-tissue imaging. The acquired PA signals are used to reconstruct the initial pressure

*Address all correspondence to: Manojit Pramanik, E-mail: manojit@ntu.edu.sg

Fig. 1 (a) Schematic diagram of the single-element UST-based PAT system. The PAT images of (b, c) a healthy and (d, e) a lesion tooth. (b, d) Mode-1 and (c, e) mode-2 PAT images. Reproduced with permission from Ref. 86.

rise within the tissue. Usually, a single-element UST with circular scanning or a multi-element UST array is used for PA signal detection. The progress in excitation laser technology (from Nd:YAG to LED), UST technology (from single-element to volumetric array), high-performance reconstruction algorithms made the PAT a low-cost, high-speed, portable device for biomedical applications. High-pulse energy (PE) Q-switched Nd:YAG lasers were commonly used as excitation sources for PAT. However, these lasers are expensive, bulky, and have low-repetition rate (~ 10 Hz) which limits imaging speed. Pulsed laser diode (PLD)^{29,47,59–61} or light-emitting diode (LED)⁶² can be used as an alternative to Nd:YAG laser because they are less expensive, more compact, and have high-repetition rate for fast imaging. In standard PAT, single-element UST scanning in a circular geometry was used to collect PA signals from the sample. Later nonscanning PAT systems based on linear,⁶³ semicircular,⁶⁴ circular,⁶⁵ and spherical array⁶⁶ of USTs have been developed to avoid scanning for high speed imaging. PAT was also successfully integrated with clinical US imaging systems for clinical applications including but not limited to breast cancer imaging,^{64,67} brain imaging,⁶⁸ and sentinel lymph node mapping etc.^{6,69,70}

Recent reviews on various aspects of PAT can be found in the literature. In the last couple of years, several review articles on PAT applications, such as multiscale functional and molecular imaging,⁷¹ advanced PAT multimodal imaging,⁷² PAT for biomedicine,⁷³ PAT principles and advances,⁷⁴ PAT for small animal whole body imaging,⁷⁵ PAT with imaging constant agents and recent applications,⁵ PAT for life sciences,⁷⁶ PAT for real-time multispectral imaging,⁷⁷ and clinical PAT for cancer⁷⁸ were reported. To avoid overlapping with recent review articles on PAT, in this review we focus on the progress in preclinical and clinical PAT systems from the hardware point of view, along with their applications in the last couple of years. The review is organized in the following order: Sec. 2 discusses single-element UST-based preclinical PAT systems; in Sec. 3 we discuss US array-based preclinical PAT systems; Sec. 4 is dedicated to a translational clinical US system integrated with PAI and their applications; in Sec. 5, we will briefly discuss the commercially available PAT systems; and in Sec. 6, we

conclude the review and discuss the future directions of PAT for clinical translation.

2 Single-Element Transducer-Based Preclinical PAT Systems

2.1 SET-Based PAT System Using Nd:YAG Laser

It is a widely used configuration for laboratory and preclinical studies. Here the expanded laser beam irradiates the test sample, and the SET circularly rotates around the sample to collect PA signals at different angles. SETs have been widely used in PAT owing to low-cost, wide availability, and high sensitivity. In SET-based systems, A-lines PA signal can be acquired in two different ways: (i) stop-and-go scan:^{79,80} A stepper motor moves the UST to a predefined position, collects multiple PA signals for averaging, saves the A-lines, and moves the UST to the next position, (ii) continuous scan:⁸¹ the stepper motor rotates the UST continuously at predefined speed, collects PA signals while the UST is moving, and saves the A-lines once the rotation is complete. Signal averaging can also be done later if needed. Continuous scan is faster than the stop-and-go scan. However, care should be taken while choosing the scan speed in continuous scanning to minimize the blur due to the UST motion or due to averaging. The blur should be negligible compared to the imaging resolution or object size to be imaged. Because the SET has a large active area (to increase sensitivity), it impacts the tangential resolution, especially when the sample is far from the scanning center (or closer to the UST surface). To increase the signal acceptance angle (and improve the tangential resolution), a negative lens can be added to the transducer surface.^{82,83} More recently, a modified delay-and-sum reconstruction algorithm was proposed to improve the tangential resolution without using any negative lens or virtual US detector.^{84,85}

Recently, a circular scanning-based PAT system [Fig. 1(a)] was demonstrated for the assessment of early dental lesion.⁸⁶ A Q-switched Nd:YAG laser [wavelength (λ) = 532 nm, pulse width (PW) = 8 ns, repetition rate (RR) = 10 Hz, PE = 80 mJ] was used as an excitation source. The laser energy

Fig. 2 (a) Schematic and (b) photograph of the slip-ring-based PAT system: UST, ultrasonic transducer; Amp, amplifier; APOR, acoustically penetrable optical reflector; US, US. *In vivo* animal imaging results by two cylindrically focused transducers with a 5-MHz central frequency: (a–f) cross-sectional PAT images of a nude mouse, (g) photograph of the nude mouse. Reproduced with permission from Ref. 88.

density on the surface of the sample was maintained at $\sim 11 \text{ mJ/cm}^2$ ($< \text{ANSI safety limit } 20 \text{ mJ/cm}^2$)⁸⁷ to provide an effective penetration depth of no less than 9 mm. This is deep enough to cover the crown of a tooth, where the early lesions often happen. The PA signals were collected by a UST [central frequency (f_c) $\sim 4.39 \text{ MHz}$, bandwidth (BW) $\sim 100.2\%$ at -6 dB , V310/Panametrics] while scanning around the tooth. The UST was moved around the tooth with 120 steps in a circle for one complete scan. The signals were amplified (SA-230F5, NF), and recorded by a data-acquisition (DAQ) card (PCI-5105, National Instruments) with sampling frequency of 60 MHz. The signals were averaged 40 times to reduce noise at each UST position. A linear stage was employed to control the position of the scanning plane of the tooth in the vertical direction.

The mode-1 and mode-2 images of a healthy tooth are shown in Figs. 1(b) and 1(c). The mode-1 contrast comes from the diversity of optical absorption; the mode-2 contrast is related to the microstructure and mechanical properties of the tooth. The tooth crown consists of two layers: (i) enamel is the white, protective external surface of the anatomic crown and has weaker optical absorption and (ii) dentin is the yellowish tissue underlying the enamel making up the major bulk of the inner portion, and has stronger optical absorption than the enamel. The mode-1 image clearly shows the enamel (outer dark region) and dentin (inner bright region). The mode-2 image of the same cross-section shows two-layer concentric structure with clear boundary. Figures 1(d) and 1(e) show the images of a lesion. Here, the mode-1 contrast clearly shows the shape, size, and macrostructure of the tooth cross-section, but, it fails to show significant contrast on the dental lesions. However, the mode-2 contrast clearly shows the lesions in the tooth.

2.2 Multitransducer-Based PAT with Nd:YAG Laser

In this system, an electrical slip ring was used to rotate multiple SETs around the sample. Figures 2(a) and 2(b) show the schematic and photograph of the system.⁸⁸ A slip ring is a device for transmitting electrical power through rotators, which can avoid

the cable intertwist problem of the rotating UST. This is the first time a slip ring was used in a PAI setup. A Ti:Sapphire laser ($\lambda \sim 808 \text{ nm}$, $\text{PW} \sim 12 \text{ ns}$, $\text{RR} = 10 \text{ Hz}$, LT-2211A, LOTIS TII, Minsk, Belarus) was used as an excitation source. The diverging ring-shaped beam from the conical lens was reflected by an acoustically penetrable optical reflector (APOR) to irradiate the sample surface at a horizontal plane. The laser fluence at the surface was $\sim 15 \text{ mJ/cm}^2$ (less than 20 mJ/cm^2 ANSI safety limit). The signals of each transducer were preamplified 28 dB (ZFL500LN, Mini-Circuit, New York) before passing through the slip ring, and then amplified another 32 dB (5072PR, Olympus NDT, Massachusetts) before being digitized. The amplified signals were recorded by a DAQ card with a sampling rate of 100 MHz (CSE4327, DynamicSignals, LLC, Illinois). For each cross-sectional scanning, 1800 signals were acquired and then reconstructed by the simple delay-and-sum algorithm. Two identical cylindrically focused transducers with a central frequency of 5 MHz (V307, Olympus NDT, Massachusetts) were used to collect the PA signals.

The system was tested on *in vivo* animal study on a $\sim 20\text{-g}$ BALB/c male nude mouse. During imaging, the mouse was anesthetized by an isoflurane machine (0.8% to 1.0% vaporized isoflurane with an oxygen airflow rate of 1.0 L/min). Cross-sectional images showing the spinal cord, kidneys, liver, and one major vessel are shown in Figs. 2(c)–2(h). Figure 2(i) shows the area of the animal where PA cross-sectional images were recorded. The signal-to-noise ratio (SNR) of *in vivo* images was $\sim 25 \text{ dB}$. The total imaging time was 9 min. These results promise that the system could provide *in vivo* animal whole-body imaging.

Although Nd:YAG/ OPO or dye-based lasers could provide high energy pulses for deep-tissue PAT imaging, these are expensive, bulky (even require optical table to house them), and are not suitable for high-speed imaging with SET due to the low laser pulse RR. Typically, it takes several minutes to obtain one cross-sectional image. Portable OPO lasers are commercially available but are even more expensive.^{89–91} Alternatively, PLD and LED can be promising light excitation sources which could make PAT a low-cost, high-speed, and portable system.

Fig. 3 (a) Schematic and (b) photograph of the PLD-PAT system: PLD, pulsed laser diode; LDU, laser driver unit; CSP, circular scanning plate; S, sample; MPS, motor pulley system; M, motor; DAQ, data acquisition card; R/A/F, US signal receiver, amplifier, and filter; UST, ultrasound transducer. PAT reconstructed images of horse-hair phantom acquired using 2.25-MHz UST at scan times (c) 5 s and (d) 3 s. Deep-tissue images of blood and ICG at (e,f) 1-cm depth and (g, h) 2-cm depth inside the chicken breast tissue. Reproduced with permission from Ref. 47. *In vivo* PAT imaging of rat brain in lateral plane at different scanning speeds acquired using a 2.25-MHz UST: (i) 30 s, (j) 10 s, (k) 3 s. SS, superior sagittal sinus; TS, transverse sinus; (l) SNR as a function of scan time.

2.3 SET-Based PAT with Pulsed Laser Diode

The PLD-PAT system is shown in Figs. 3(a) and 3(b).^{47,59} Here, the laser was integrated inside the scanner. The PLD ($\lambda \sim 803$ nm, PW ~ 136 ns, RR = 7 kHz, PE ~ 1.4 mJ, Quantel, France) was the excitation source. The PLD was controlled by the laser driver unit (LDU) which consists of a temperature controller (LaridTech, MTTC1410), a 12-V power supply (Votcraft, PPS-11810), a variable power supply (to change the laser output power), and a function generator (to control the laser RR). The PE and RR can be controlled independently with variable power supply (BASETech, BT-153), and function generator (FG250D, Funktionsgenerator), respectively. The function generator provides a transistor-transistor logic signal to synchronize the DAQ with the laser excitation. A flat UST (V323/SU, $f_c \sim 2.25$ MHz, BW $\sim 70\%$, Olympus NDT) was driven by a step motor (M) (Lin Engineering,

Silverpak 23C) to scan circularly around the sample. The UST detected signals were subsequently amplified, and band pass filtered by the US signal receiver (R/A/F) unit (Olympus-NDT, 5072PR), and then digitized and recorded by the PC with DAQ card (GaGe, compuscope 4227). The DAQ was done at 25 Ms/s.

High-speed PAT images obtained on a horse hair phantom are shown in Figs. 3(c) and 3(d).⁵⁹ These images were obtained in 5 s, and 3 s scan time using 2.25-MHz UST. The hair has a side-length of ~ 8 mm and a diameter of ~ 150 μ m. Deep-tissue imaging experiments were carried out on a chicken breast tissue in which two low-density polyethylene (LDPE) tubes (~ 12 -mm long and ~ 0.59 -mm inner diameter) were filled with mouse blood and indocyanine green (ICG). The radiation density on the tissue surface was ~ 0.3 mJ/cm². PAT images were acquired using 2.25 MHz for 5 s, 3 s scan time at 1 cm, and 2 cm depth, respectively, as shown in Figs. 3(e)–3(h). The SNR values of

Table 1 Comparison of various parameters between PLD-PAT and OPO-PAT. Reproduced with permission from Ref. 47.

S. No.	Parameter		PLD-PAT	OPO-PAT
1	Spatial resolution (FWHM)	2.25 MHz	384 μm	381 μm
		5 MHz	185 μm	182 μm
2	SNR	2.25 MHz	29 (3 s) and 84 (30 s)	28 (30 s)
		5 MHz	24 (3 s) and 77 (30 s)	23 (30 s)
3	Imaging speed		Acceptable image in 3 s	Acceptable image in 30 s
4	Imaging depth		~ 2 cm (SNR 10 at 30 s)	~ 3 cm (SNR 3 at 30 s) ~ 4 cm (120 s scanning time)
5	Image quality	3 s	Acceptable	Not acceptable
		30 s	Good	Good
6	Pulse duration		~ 136 ns	~ 5 ns
7	RR		7 kHz	10 Hz
8	PE per pulse		1.45 mJ at laser output	100 mJ at laser output
			0.28 mJ/cm ² at sample	10 mJ/cm ² at sample
9	Laser head dimensions		11.0 $\times$ 6.0 $\times$ 3.6 cm	77.5 $\times$ 17.8 $\times$ 19.0 cm
10	Laser head weight		~ 150 gm	~ 100 Kg
11	Portability		Yes	No (optical table needed)
12	Cost		~ 15 to 25k USD	~ 90 to 140k USD

blood, ICG measured at 1 cm are ~ 25 , ~ 32 and that measured at 2 cm are ~ 8 , ~ 14 . The performance of PLD-PAT was compared against the Nd:YAG/OPO-PAT system and the results were summarized in Table 1.⁴⁷ From Table 1, it is clear that the PLD-PAT can provide 2-cm imaging depth and acceptable images in 3 s. The low PE was slightly compensated by the higher RR of PLD. Since the PLD was integrated inside the scanner, it makes the overall system compact and portable.

The PLD-PAT system was demonstrated for *in vivo* small animal brain imaging.⁹² Figures 3(i)–3(k) show the *in vivo* brain images of a healthy rat. The superior sagittal sinus (SS) and transverse sinuses (TS) of the rat brain are clearly visible. The plot in Fig. 3(l) shows the image SNR as a function of scan time. These results promise that the PLD-PAT system could provide high-speed *in vivo* imaging for preclinical applications.

2.4 SET-Based PAT with Light-Emitting Diode

The PLDs suitable for PAT are typically available at wavelengths greater than 750 nm. The LEDs are readily available at wavelengths below 650 nm where hemoglobin absorption is significantly higher, offering high contrast for superficial vascular imaging applications. The use of LEDs for PA imaging has been explored. Figure 4(a) shows the schematic of the LED-PAT system.⁶² A high power SST-90 LED ($\lambda \sim 623$ nm, PW ~ 200 ns, RR ~ 500 Hz, PE ~ 9 μJ) from Luminus was used as an excitation source. It was driven by a driver (PCO-7120, Directed Energy, Inc.), which provided a peak current of

50 A. The LED was operated at low duty cycle 0.01% for its safety. Figures 4(b) and 4(c) show the PA images of the phantom. The phantom was comprised of three 1.4-mm-diameter poly methyl methacrylate tubes filled with human blood (35% hematocrit) immersed in 1% intralipid ($\mu_s = 1$ mm⁻¹). The three tubes were fixed in a mount that was rotated by 360 deg in steps of 0.9 deg using a stepper motor. The phantom was illuminated by LED with ~ 11.5 $\mu\text{J}/\text{cm}^2$ energy density at 623 nm. The signals were collected by a cylindrically focused UST ($f_c \sim 3.5$ MHz, V383/Panametric). The signals were amplified using a low noise voltage amplifier (60 dB, Analog Modules Inc.), averaged 5000 times and saved in a computer. Figure 4(b) shows the detected time resolved PA signals as a function of scan angle. It can be seen that PA signals corresponding to a penetration depth of up to 15 mm are clearly visible. Figure 4(c) shows the reconstructed PA image of the phantom with SNR ~ 11.5 .

This study has explored the use of low-cost visible LEDs (< 50) as a PA excitation source. At 500-Hz RR, LED-PAT could image blood tubes immersed to a depth of 15 mm in Intralipid. This suggests that *in vivo* imaging of the vasculature to depths of a few mm should be possible with LED-PAT. A four-wavelength LED was also demonstrated for spectroscopic applications. This study presents a step toward the development of compact, inexpensive visible, and near-infrared multiwavelength excitation sources for PAT.⁶²

In spite of single-element large detector area transducers providing the low-cost PAT imaging system with high sensitivity,

Fig. 4 (a) Schematic of LED-PAT system, (b) time-resolved PA signals of three 1.4-mm tubes, and (c) reconstructed PA image of the three phantoms. Reproduced with permission from Ref. 62.

it is challenging to achieve real-time imaging. The time required to collect multiple A-lines data for a cross-sectional PAT image reconstruction is too slow, since a motor driven mechanism is required to move the detector around the sample. The fastest imaging speed demonstrated so far using a single-element US transducer is 3 s per cross-sectional imaging.⁴⁷ Therefore, for faster real-time imaging a US array transducer is required. In Sec. 3, we will discuss array transducer-based PAT imaging systems.

3 Ultrasound Array Transducer-Based Preclinical PAT Systems

In traditional pulse echo US imaging, array transducers (linear array, phased array, and curved array) have replaced single-element transducers decades back. It helped US imaging devices to become real-time, portable, and handheld. Similarly, the PA community has also adopted array transducers (linear, semicircular, circular, and spherical array) to develop high-speed/real-time PA imaging.^{48,63,64–66,93–95} In the following sections, we will discuss the recent progress in PAT systems with array-based USTs that could translate the PAT from bench to bedside.

3.1 Circular Planar Array-Based PAT System for Breast Imaging

PAT is fundamentally a 3-D imaging method. Light illuminating the tissue surface creates a diffusive pattern within tissue and generates PA signals from the illuminated volume. Imaging systems must then record PA signals around the boundary of that volume so that the initial pressure distribution can be reconstructed. Several array-based PAT systems including semicylindrical cup,⁶⁴ ring-shaped transducer array,^{96,97} circular planar array,⁹⁸ semispherical array-based^{99,100} PAT systems have been developed for breast imaging.

A laser optoacoustic imaging system (LOIS) based on semicylindrical array of transducers was reported for breast cancer detection.⁶⁴ The system was designed to image a single breast slice in craniocaudal or mediolateral projection with an arc-shaped array of 64 ultrawide-band acoustic transducers. The

LOIS system was clinically tested on 27 patients, and the system could provide high contrast PAIs of breast tumors, taking advantage of the blood optical absorption in the angiogenesis-related microvasculature. A functional photoacoustic tomography (fPAT) system was developed for functional imaging of breast lesions with high resolution.^{96,97} The system consists of a tunable pulsed Ti:sapphire laser pumped with a Q-switched Nd:YAG laser (Symphotic Tii, Camarillo, California) and a ring-shaped array of 64 acoustic transducers having maximum frequency response up to 2 MHz. Functional images of breast tissue including absolute total hemoglobin concentration (HbT) and oxygen saturation ($\text{StO}_2\%$) were obtained by fPAT. HbT and $\text{StO}_2\%$ maps from pathology-confirmed cancer cases showed clear detection of tumors.

A circular planar array-based PAT, named as Twente photoacoustic mammoscope, was reported for breast imaging.⁹⁸ The schematic of the system is shown in Fig. 5(a). A Q-switched Nd:YAG laser ($\lambda \sim 1064 \text{ nm}$, $\text{PW} \sim 10 \text{ ns}$, $\text{RR} \sim 10 \text{ Hz}$, $\text{PE} \sim 350 \text{ mJ}$) from Continuum Surelite, California, was used for PA excitation. The beam was maintained at a fixed position on the breast surface, with a beam area of $\sim 35 \text{ cm}^2$ and an energy density of $\sim 10 \text{ mJ/cm}^2$ (less than 100 mJ/cm^2 ANSI limit) on the skin.^{101,102} The US array consists of 588 elements ($f_c \sim 1 \text{ MHz}$, $\text{BW} \sim 130\%$) in a circular layout with diameter of 85 mm. The elements are grouped into 10 sectors; each sector is ~ 60 elements serviced by an application-specific IC (ASIC), which buffers and amplifies the signals. Each ASIC has a single output that leads to a 10 input summing amplifier. In the upgraded mammoscope, the outputs of the 10 ASICs are made to bypass the buffers, amplifiers, and summing amplifier. An independent interface unit was developed with 10 channel buffer-amplification based on the AD797AR (Analog Devices) ultralow noise op-amp. The 10 outputs are acquired at a time using two eight-channel digitizers (National Instruments PXI 5105, 60 MS/s, 12-bit) controlled by a Labview program. With averages over 10 signals per acquisition, an imaging time of 10 min was required for the complete detector area covering an field-of-view of $90 \times 85 \text{ mm}^2$ on the breast.

Figures 5(b)–5(e) show images in the craniocaudal (CC) direction of MRI and PAT for two cases with increasing

Fig. 5 (a) Schematic of the imager of the Twente photoacoustic mammoscope, (b, d) MR, and (c, e) PA CC images of the breast for two cases. (b, c) Patient P38, having an infiltrating ductal carcinoma (IDC), grade 3 of 19 mm, (d, e) patient P55, having an IDC, grade 2 of 34 mm. The dashed box in the MR image indicates the area from which the PA image is acquired. Reproduced with permission from Ref. 98.

histopathological size. The two cases (P38 and P55) are representative for lesion sizes up to 35 mm in x-ray mammography. The results show that there is good correspondence in lesion colocalization, appearance, and shape in the PAT and MRI images. The study proved that PA imaging has potential in visualization of breast cancer. For the first time, mass-like, nonmass, and ring appearances were observed in PA image intensities. However, the system needs to be tested on a large number of patients before going for clinical practice.

3.2 Semispherical Array-Based PAT Systems

A PAT system was reported for 3-D visualization of vasculature in human breast.^{100,103} The device consists of an array of 128 elements with 5-MHz central frequency and 3 mm diameter laid out in a spiral pattern on a semispherical surface with a 100-mm radius of curvature. The system could provide maximum depth of 40 mm, and its spatial resolution $\sim 250 \mu\text{m}$ was uniform over a $64 \times 64 \times 50 \text{ mm}$ field of view. A PAT system with advanced semispherical array was demonstrated for volumetric imaging.¹⁰⁴ The imaging system is based on simultaneous acquisition of PA signals by means of a custom-made matrix ultrasonic array ($f_c \sim 3.9 \text{ MHz}$, BW $\sim 4 \text{ MHz}$, Imasonic SaS, Voray, France) consisting of 256 elements densely distributed on a spherical surface [Fig. 6(a)]. The radius of the array is 40 mm and its outer diameter is 64 mm. Acoustic coupling of the PA signals to the detectors was done by means of a transparent impermeable membrane containing water. Acoustic gel is used to guarantee coupling between the membrane and surface of the imaged region. The detection array further includes a central cylindrical cavity with a diameter of 8 mm through which optical illumination was delivered. The output of an OPO laser (680 to 900 nm, RR $\sim 10 \text{ Hz}$, PW $\sim 10 \text{ ns}$, Phocus, Opotek, California) was guided through a custom-made silica fused-end fiber bundle (CeramOptics GmbH, Bonn, Germany) consisting of 480 fibers. The illumination spot at a distance of 3 cm from the surface of the detection array has $\sim 1 \text{ cm}$ diameter. PA signals are acquired with sampling frequency of 40 MS/s using a custom-made DAQ system consisting of 16 multi-channel analog-to-digital converters (FalkensteinMikrosysteme, Taufkirchen, Germany), which were triggered by the Q-switch output of the laser. Using a graphic processing unit (GPU), the

system allows imaging at a frame rate up to 50 frames per second. The volumetric imaging rate is limited by pulse RR of the laser and processing speed of the image reconstruction method. The system was demonstrated for several applications, and the recent ones are discussed below.

3.2.1 In vivo imaging of the carotid arteries in humans

The ability to image hemoglobin makes the PAT a potential tool for imaging vascular diseases. Using the system, imaging of the carotid arteries (cca) in healthy human volunteers was demonstrated.¹⁰⁵ Figure 6(b) depicts a device that is similar in operation to medical US imaging.⁷⁷ Figures 6(c) and 6(d) show the real-time two-dimensional (2-D) multispectral optical tomography (MSOT) images of oxyhemoglobin (OHb) in human hallux. The MSOT imaging was done at a video-rate of 50 frames per second, with the ability to select a different wavelength for each frame. Such a high frame rate was enabled by a fast-tuning OPO and GPU.

3.2.2 In vivo imaging of tissue chromophores on human forearm

The system design is shown in Fig. 7(a); here high RR OPO laser (PW $\sim 5 \text{ ns}$, PE $\sim 20 \text{ mJ}$, RR $\sim 50 \text{ Hz}$) from Innolas Laser GmbH, Germany, was used.¹⁰⁶ The spectral dependence of the optical absorption of the three main tissue chromophores (melanin, Hb, and HbO_2) is shown in Fig. 7(b). The 2-D volumetric PA images collected at different wavelengths and at different time points are shown in Fig. 7(c).

3.2.3 In vivo imaging of human breast

The system design and representative volumetric frames of breast are shown in Figs. 8(a)–8(c).⁹⁹ The breast is illuminated using the fast tuning OPO laser and volumetric data were acquired at multiple wavelengths in order to recover functional blood oxygenation parameters of the breast vasculature in real time. Here, concentration of the different tissue chromophores, as resolved by the spectral unmixing procedure, is pseudocolored with red, blue, and yellow for HbO_2 , HbR, and melanin, respectively. The yellow spots correspond to melanin-rich skin

Fig. 6 (a) Projection view of the shape of the individual transducer elements. (b) Handheld volumetric PAT probe, (c) real-time PA imaging of the common carotid artery (cca) in a human volunteer, real-time multispectral PA imaging of OHb in the human hallux; (d) overlay OHb image (unmixed from 18 wavelengths acquired within 400 ms) on the PA imaging (at 700 nm) of the tip of the hallux in a healthy volunteer. Reproduced with permission from Ref. 77.

Fig. 7 (a) Schematic and the clinical handheld PA imaging scanner. Demonstration of 5-D PA imaging in a healthy volunteer *in vivo* using: (b) photograph of the human forearm, the colored frames correspond to the region of interest, (c) 2-D matrix of reconstructed volumetric PA images. Each image corresponds to data taken at a different wavelength and a different time point, thus, effectively rendering 5-D data set. (d) For each time point, the wavelength-dependent data are unmixed in order to render, in real time, distribution maps of three different intrinsic tissue chromophores. Reproduced with permission from Ref. 106.

pigmentations in the skin. Since it only takes 30 ms to acquire multispectral data at three different wavelengths, this approach is equally capable of rendering functional blood oxygenation and total hemoglobin profiles in real time. This is an especially important aspect as elevated levels of hemoglobin and hypoxic behavior were found to strongly correlate with tumor angiogenesis and breast malignancies.

3.2.4 *In vivo* imaging of lymph node micrometastases and in-transit metastases from melanoma

Recently, the MSOT system was demonstrated for detection of melanoma lymph node micrometastases and in-transit metastases undetectable with fluorine 18 fluorodeoxyglucose (FDG) PET/computed tomography (CT).¹⁰⁷ The MSOT could differentiate even small melanoma metastasis from the other

Fig. 8 (a) Layout of the handheld breast imaging approach, (b-c) functional optoacoustic breast angiography using multiwavelength DAQ simultaneously resolves the concentration of major tissue chromophores in the breast of a 40-year-old volunteer in 3-D and in real time. The breast surface is located in the upper part of the cubes. Reproduced with permission from Ref. 99.

lymphadenopathies *in vivo*. The study was performed on a mouse B16F10 melanoma limb lymph node metastasis model.

3.3 Ring-Array-Based Wearable PAT System

PAT allows *in vivo* brain imaging at micrometer spatial and millisecond temporal resolution. A 3-D wearable PAT system with $\sim 200 \mu\text{m}$ spatial and 170 ms temporal resolutions is demonstrated for brain imaging in a rat while it is walking.¹⁰⁸ This is the first time a wearable PAI device was demonstrated

for brain imaging. This could open up a lot of opportunity in neurostudies, cognitive science, and behavioral studies. The photograph of the 3-D-wPAT probe and schematic of the system are shown in Figs. 9(a) and 9(b), respectively. The pulsed laser (680 to 2550 nm, PW ~ 10 ns, RR ~ 20 Hz, Surelite OPO Plus, Continuum) was split into two beams by a beam splitter. The reflected beam was recorded by a photodiode for light energy monitoring, and the transmitted beam was coupled into the proximal end of a liquid light guide-LLG (core diameter = 8 mm, NA = 0.59, LLG77631, Newport). The system contained

Fig. 9 (a) Photographs of the fabricated 3-D-wPAT probe. (b) Schematic diagram shows the light illumination, light energy monitoring, and data amplification and acquisition system. (c) Photographs show the behaving ability of a rat while wearing 3-D-wPAT. Functional imaging using multiwavelength light excitation (710 and 840 nm). (d) Total cerebral hemoglobin wPAT images. (e) Calculated in-plane oxygen saturation map. Oxygen gas was applied to the experimental chamber at the 100th second. Reproduced with permission from Ref. 108.

192 signal amplifying channels that were three-to-one multiplexed into 64 DAQ channels. Therefore, three laser pulses are required to collect the data from all the channels. With the rotatable design of the LLG interface, the animal can walk around during imaging, as shown in Fig. 9(c).

The functional imaging capability of 3-D-wPAT was demonstrated by mapping the cerebral oxygen saturation via multiwavelength irradiation in walking hyperoxic rats as shown in Figs. 9(d) and 9(e). The animal was placed in an experimental chamber and pure oxygen was pumped into the chamber 100 s after the start of imaging. The imaging duration was 300 s, and a total of 100 images were acquired. Figures 9(d) and 9(e) present the total cerebral hemoglobin images and sO_2 maps for each acoustic detection layer, respectively. The sO_2 ratio increased throughout the brain vasculature after being supplied with oxygen. As shown in Fig. 9(e), the regions marked with white arrows underwent increase in oxygen saturation. The obtained results are consistent with the idea that hemodynamic changes initiate in medial and ventral brain regions, reinforcing the importance and significance of deep-penetration 3-D brain imaging.

All the PAT systems discussed in Sec. 3 are based on home-made/custom-made array US transducers. The challenge in building array based transducers is that they require parallel data acquisition (DAQ) systems, which are expensive and most of them are also custom-made. Such systems are also not approved by the Food and Drug Administration (FDA) in the United States at present, and hence they are not preferred in clinical use. Also these systems are not optimized to perform standard US imaging. On the other hand, clinical US imaging systems are FDA-approved devices and they are routinely used in clinics/hospitals by the clinicians. Moreover, PAI and pulse-echo US imaging share the same US-receiving mechanism. Hence, PAT integrated with a clinical US system will be more acceptable to the clinicians as they will be able to provide dual-modal imaging. They can provide traditional US imaging, Doppler blood flow imaging, and so on on top of PAI. In the following section, we will discuss the integrated clinical US systems with PAT.

4 Clinical Ultrasound Imaging Systems Integrated with PAT

4.1 Performing Sentinel Lymph Node Biopsy for Breast Cancer Patients

In staging of cancer, including breast and melanoma cancers, SLN biopsy (SLNB) is a standard procedure.^{109,110} A positive biopsy result suggests that cancer has spread to the node and probably to distant organs. The identification rate and sensitivity of SLNB technique are less than 95%, even in experienced hands.^{109–111} Moreover, the complications associated with the SLNB procedure include seroma formation, lymphedema, sensory nerve injury, and limitation in the range of motion.^{112,113} Several studies have been conducted proposing noninvasive identification of SLN using a dual-modal PAT-US platform in a small animal model.^{114–116}

Recently, a handheld PAT-US system [Fig. 10(a)] was used to identify the SLN in breast cancer patients.¹¹⁷ The system was developed by modifying a clinical US system (iU22, Philips Healthcare). The excitation source was a tunable dye laser (PW ~ 6.5 ns, RR ~ 10 Hz, PrecisionScan-P, Kaarst, Germany) pumped by a 532-nm Nd:YAG laser (QuantaRay PRO-350-10, Spectra-Physics, Santa Clara, California). Light was delivered to

the sample by a fused-end, bifurcated fiber bundle. To improve the PA detection sensitivity and handiness, the fiber bundles and UST array (L12-5 and L8-4, Philips) were integrated into a single probe as shown in Fig. 10(a). The light fluence on the skin was ~ 10 mJ/cm² (less than 20 mJ/cm² ANSI safety limit). The signals were acquired by a custom-built DAQ card. The DAQ system was synchronized with laser triggering by an FPGA-based card that also performs PAT reconstruction based upon a delay-and-sum beamforming algorithm. The system can provide a frame rate of 5 Hz for PA, US, and coregistered imaging.

For SLNB, the PAT-US system can serve as a real-time modality to first locate the SLN and then guide a fine or core needle for lymph node sampling.^{115,118} Figure 10(b) shows the coregistered PA-US image, revealing the SLN and the needle used to insert the titanium marking clip. The contrast from the needle in the PAT image is much higher than that in the corresponding US image because the US array receives more emitted PA energy than it does reflected ultrasonic energy. The system is capable of locating lymph nodes via anatomical features in US and to verify the lymph node as sentinel using PAT. In addition, these results illustrate the feasibility of PAT-US to guide needles to the SLN in breast cancer patients. SLN identification can be improved further using multiwavelengths to spectrally resolve methylene blue (MB, a routinely used contrast agent in clinical practice) from blood in nearby vessels. MB absorption is strongest at 665 nm, while its absorption is very weak at wavelengths longer than 750 nm. Figure 10(c) shows an *in vivo* imaging of MB and the ability of the system to distinguish MB from surrounding blood vessels. Five minutes after MB injection and breast massage, the PAT image acquired at 650 nm [Fig. 10(d)] shows signals from both MB (i.e., lymphatic vessels) and blood vessels, while the PAT image acquired at 1064 nm [Fig. 10(e)] shows signals from blood vessels only. The images in Figs. 10(d) and 10(e) are subtracted to highlight MB only. The results are promising; however, the clinical utility of the dual-wavelength imaging requires further investigation.

One of the challenges of integrating the PAT with a clinical US system is to obtain the raw channel radio-frequency (RF) data before beamforming is done. Once these RF data are obtained, PA reconstruction is done and displayed on the monitor. The Philips iU22 clinical US system after integrating with PAT is capable of providing the raw channel RF data before beamforming. However, this feature is not available on the commercially available iU22 US system in the market. Therefore, the researchers will have to collaborate with the Philips research team to use their system specifically modified for that purpose. This limits the accessibility of such a system to many researchers interested in doing integrated US and PA imaging. Recently, there are many other US companies coming up with research US platforms capable of extracting raw channel RF data and these systems are made available for researchers to purchase off the shelves. Below we will discuss such systems and their recent progress.

4.2 High-Frame Rate (7000 Hz) Photoacoustic Imaging

Using a handheld PLD-PAT/US probe, a real-time phantom imaging up to 15 mm depth was done at 20 fps.¹¹⁹ With the OPO-PAT/US system, *in vivo* real time imaging of blood vessels in a rat was performed at 20 fps.¹²⁰ Recently, a high-frame rate (7000 Hz) PA imaging system was demonstrated by integrating a high RR PLD (the PLD used here is same as the one used in Sec. 2.3), with a clinical US system (ECUBE 12R, Alpinion,

Fig. 10 (a) Schematic of the dual-modality PA and ultrasonography imaging system and photograph of the probe, (b) *In vivo* coregistered PAT-US image of the SLN and needle. *In vivo* dual-wavelength PAT imaging: (c) image acquired at 650 nm showing both MB dye in lymphatic vessels and hemoglobin in blood vessels, (d) image acquired at 1064 nm showing primarily hemoglobin in blood vessels while suppressing the signals from MB, (e) subtracted image highlighting the locations of MB. PA_{650} and PA_{1064} : PA amplitudes measured at 650 and 1064 nm, respectively. Reproduced with permission from Ref. 117.

South Korea).¹²¹ The PLD output was made narrower using a cylindrical lens to illuminate $15 \text{ mm} \times 5 \text{ mm}$ area with 1.87 mJ/cm^2 energy density. The PA signals generated were acquired by a linear array transducer (LAT) (L3-12 transducer, Alpinion, South Korea) consisting of 128 array elements. The probe has $f_c \sim 8.5 \text{ MHz}$ (95% BW), array element pitch 0.3 mm, and elevation height 4.5 mm. The Ecube system has a 40-MHz sampling frequency and 12-bit ADC to acquire

the data and a data transfer speed of 6 Gb/s. The clinical US system captures and displays the PAI images in real-time.

The L3-12 linear array UST can be operated in two different modes, mode-1: using all the 128 arrays elements and mode-2: using 64 channels elements only (element numbers 1-64 or 65-128). PA imaging was done for the different flow rates. Using mode-1 and mode-2, 3500 fps and 7000 fps PA images were recorded. Figures 11(a)–11(h) show the PA images of ink

Fig. 11 Set of PA images showing the ink flow at various time points with the PLD-PAT system. PA imaging was done at a frame rate of 7000 fps. (a-d) Flow rate was 3 cm/s, (e-h) flow rate was 14 cm/s. (i) Photograph of the PLD-PAT clinical US system. Reproduced with permission from Ref. 121.

flowing through an LDPE tube with an inner diameter of 1.15 mm when the transducer was operated in mode-2. Here 350 frames are being recorded in 0.05 s when operated at 7000 fps. Figure 11(i) shows the photograph of the experimental setup. Figures 11(a)–11(d) correspond to the flow rate 3 cm/s, whereas Figs. 11(e)–11(h) correspond to 14 cm/s. It was observed that the ink is flowing through the tube, i.e., a high contrast air–ink interface was moving in the PA B-scan images. A high frame rate of 7000 fps photoacoustic B-scan imaging was feasible with the PLD and Ecube clinical US system, whereas with the 10-Hz Nd:YAG/OPO laser only 10 fps (max) PA imaging was possible. With 7000-fps frame rate, the theoretical flow rate that can be measured is up to 134.4 m/s. However, future studies need to be conducted to explore the various *in vivo* applications where high frame rate PA imaging would be beneficial, such as imaging circulating tumor cells, heart valves, and blood vessel.

4.3 Slit-Enabled UST Array for High Elevational Resolution PAT Imaging

A dual-modality PA-US system is increasingly used in preclinical/clinical imaging. However, the 3-D imaging capability of a linear array is limited due to its poor elevational resolution. To overcome this problem, a metal slit was used at the array focus.¹²² The imaging system involved a 128-element UST array ($f_c = 5$ MHz, elevation focus ~ 25 mm, ATL/Philips L7-4) and an Nd:YAG laser ($\lambda \sim 532$ nm, PW ~ 10 ns, RR = 10 Hz, PE = 18 mJ/pulse, Surelite SL III-10, Continuum). PA signals received by the L7-4 array were multiplexed and digitalized by a 64-channel DAQ (Vantage, Verasonics, Redmond, Washington). A thin slit [Fig. 12(a)] was formed by two metal blades with 500 μ m thickness. The bottom blade was fixed in position while the top blade was mounted on a translation stage, which allowed easy and precise control of the slit opening. To prevent sound signal transmission directly through the plate, they glued 5-mm foam on the back surface of the blade (facing the transducer) as an acoustic absorber. The slit causes the incoming PA waves to diffract along the elevation direction and, hence, significantly improves the elevation detection aperture and resolution. Figures 12(b)–12(d) show the *in situ* PAT results on a mouse abdomen without and with slit-enabled PAT. The slit improves the elevation resolution by 10 times without compromising scanning time.

4.4 Photoacoustic Imaging Using Two Linear Array Transducers

ATs are widely used in clinical US systems, but their application in PAT is still limited due to the limited-angle of view that affects the image quality. Using two LATs, one can acquire a PAT signal from different orientations. The positions of the two transducers can be adjusted to fit the specific geometry of an imaging site. Figure 13(a) shows the schematic of the experimental setup for PAT imaging with two transducers. The laser source contains a Q-switched 532-nm Nd:YAG/OPO laser (RR = 10 Hz, PW ~ 5 ns, tunable $\lambda = 680$ to 2550 nm, Surelite OPO Plus, Continuum). The system used two identical commercial LATs controlled by a SonixMDP US system (Ultrasonix Medical Corporation, Richmond). The US imaging system was capable of enabling selected transducer channels for acoustic wave transmitting. The two LATs (L14-5/38 Linear) were positioned in the same plane. Each one has 128 channels with 7.2-MHz center frequency, minimum 70% fractional BW (at -6 dB), and 0.3-mm element pitch. The acoustic signal received by the transducer array was sent to a specialized research module, SonixDAQ for DAQ. This module acquires and digitizes pre-beamformed radio-frequency signal from all the 128 channels individually at a sampling rate of 40 MHz and 12-bit resolution. Both PA and US images were reconstructed using delay-and-sum beamforming algorithm. The lateral and axial resolutions of the systems are 0.52 and 0.44 mm, respectively.

The system was tested on phantoms and a human hand. A leaf skeleton embedded in bovine gelatin, mimicking blood vessel branches in tissue, was imaged. A photograph of the sample and the transducer positions are shown in Fig. 13(b). The relative angle between the two LATs was $\sim 90^\circ$. In both images [Figs. 13(c) and 13(d)], each linear transducer can only capture part of the leaf skeleton structure which was near parallel to the lateral axis of the transducer. By incorporating the acoustic signals acquired by both linear transducers, a more complete leaf skeleton structure was obtained, as indicated in Fig. 13(e). The system was demonstrated for *in vivo* imaging of the human hand and the results promise that the two transducer approach has potential to be implemented in clinical applications.

4.5 Full-View Photoacoustic Imaging Using Thermal Encoding

PAI of irregular objects may miss important features because PA waves propagate normal to structure boundaries and may miss

Fig. 12 *In situ* experiment of mouse abdomen: (a) 3-D drawing of the *in situ* slit-PAT setup, (b) depth-encoded image of 2-D-stack PAT, (c) depth-encoded image of 3-D focal-line PAT, (d) depth-encoded image of slit-PAT. Arrows 1 and 2 point to intestine and crossed skin vessels. (b)–(d) Reproduced with permission from Ref. 122.

Fig. 13 (a) Schematic of a broad-view PAT system based on two LATs, PAT images of a leaf skeleton. (b) Photograph of the leaf skeleton phantom and sketch diagram of the transducers. Reconstructed image from (c) transducer B, (d) transducer A, and (e) both transducers. Reproduced with permission from Ref. 123.

the UST if it has a limited angle of view. To overcome such problems, an ultrasonic heating encoded approach was demonstrated.¹²⁴ The experimental setup is illustrated in Fig. 14(a). The output of excitation 532-nm Nd:YAG laser (PW = 10 ns, RR = 20 Hz, Quantel, France) was expanded by a concave lens to illuminate the sample. A linear UST array with 256 elements ($f_c \sim 15$ MHz, BW ~ 9 MHz, MS200, VisualSonics Inc., Canada) was used to detect the PA waves. The probe was connected to an US imaging system (Vevo LAZR, VisualSonics Inc., Canada), which could generate a 2-D PA image using four laser pulses. The system can provide 257 and 113 μm lateral and axial resolutions, respectively.

A custom-made ultrasonic heating transducer ($f_c = 7.5$ MHz, focal diameter 178 μm , focal depth 445 μm) can generate a focal spot size smaller than the lateral resolution of the linear transducer array for isotropic detection. A function generator synthesized a 7.5-MHz sinusoidal signal, whose amplitude was modulated by another sinusoidal wave at a frequency of 0.5 Hz. The 7.5-MHz sinusoidal signal modulated by 0.5-Hz signal was amplified to drive the heating UST. To avoid interference between the acoustic heating and PA detection,

the heating US was turned off for 1 ms before and after each PA acquisition. At each heating spot, 50 consecutive 2-D PA images were acquired at a speed of 5 images/s and reconstructed using the filtered backprojection algorithm. As a proof of concept, full-view *in vivo* mouse vascular imaging was performed with and without thermal encoding as shown in Fig. 14(b). Only horizontal blood vessel segments [features in green in Fig. 14(b)] are visible in the original PA [green signal in Fig. 14(b)] mode and more blood vessels are visible with thermal encoding [red signal in Fig. 14(b)] due to the enlarged angle of view. Since ultrasonic heating can be focused deeply, this technique can be applied to deep-tissue imaging and is promising for full-view imaging of other features of biomedical interest, such as tumor margins.

4.6 Human Finger Joint Photoacoustic Imaging

A noninvasive 2-D PAT system based on single-element UST¹²⁵ and multiple single-element USTs¹²⁶ were reported for imaging of human peripheral joints toward diagnosis of inflammatory arthritis. Later, a 3-D quantitative PAT (3-DqPAT) system was

Fig. 14 (a) Schematic of a full-view PAI system based on acoustic thermal tagging, (b) *in vivo* mouse vascular PA imaging without (green) and with (red) ultrasonic thermal encoding. Reproduced with permission from Ref. 124.

Fig. 15 (a) Photograph of the portable US scanner and the probe integrating PLD and UST array. The overlaid PA/US images of a human PIP joint in (b) sagittal and (c) transverse planes. Reproduced with permission from Ref. 119.

reported for 3-D visualization of finger joints.¹²⁷ The 3-DqPAT system consists of a Ti:sapphire laser ($\lambda \sim 805$ nm, PW ~ 10 ns, RR = 10 Hz, LOTIS, Minsk, Belarus), a spherical scanner, a detection array consisting of eight individual 1 MHz single-element flat USTs (Valpey Fisher, Hopkinton, Massachusetts), and DAQ card (PREAMP2-D and PCIAD1650, US Ultratek, Concord, California). The system was demonstrated for *in vivo* detection of osteoarthritis in the human finger joints. Later, a high-resolution 3-D PA imaging system with lateral and axial resolutions 70 and 240 μm for a finger joint was demonstrated.¹²⁸ The cross-sectional PA images of a healthy joint clearly showed major internal structures including phalanx and tendons. The *in vivo* PAI results obtained were compared with 3.0T MRI images of the finger joint. A PA/US dual-modality system was reported for imaging human peripheral joints.¹²⁹ The system utilizes a commercial US unit (Z.ONE, ZONARE Inc.) for both US and PA imaging. A linear array (L10-5, ZONARE Inc.), with a working band of 5 to 10 MHz was used to collect signals. The lateral resolution of the PA imaging in the focal plane was ~ 300 μm at a distance ~ 20 mm from the probe surface. *In vivo* results on healthy human volunteers revealed that the system could provide both the structural and functional information of intra- and extra-articular tissues.

A handheld US probe smartly integrated with a PLD was reported for dual-modality PA-US imaging [Fig. 15(a)].^{119,130} The US detection was performed with an US array (Esaote SL3323 probe) composed of 128 elements, each with a length of 5 mm and a pitch of 0.245 mm. The array has a central frequency of 7.5 MHz and a measured -6 dB BW of around 100%. The array incorporates an acoustic lens to focus the US in the elevation plane at about 20 mm distance. The probe, connected to an US scanner from MyLab_One, ESAOTE Europe B.V, Maastricht, the Netherlands. A PLD ($\lambda = 805$ nm, PW = 130 ns, RR = 10 kHz, PE = 0.56 mJ/pulse) from Quantel (Paris, France) was integrated inside the US probe. The PLD output suffers from bad beam quality and large divergence angle. Using a fused silica diffuse optical element (efficiency 80%) composed of 400- μm diffractive cells and eight discrete phase levels, the beam is homogenized and reshaped in rectangular form onto the skin. The PLD can provide fluence up to 1.3 mJ/cm² on 18.2 $\times$ 2.3 mm² skin

area. The PA-US system allows, PA imaging at 10 images/s (after averaging 20 frames per image), and US imaging at 80 frames/s (no averaging). Figures 15(b) and 15(c) show combined PA-US images of the sagittal and transverse plane of a human proximal interphalangeal (PIP) joint. Skin, bone, and joint gap are clearly visible in the US images (gray). PA images (yellowish red) depict the skin and blood vessels. The deeper PA signals correspond to the reflection of skin PA signal on bone. This bedside device has a potential to visualize rheumatoid arthritis.

All the systems described in this section are comprised of clinical US imaging platform modified to incorporate PAI making it a dual-modal imaging system. Although the PA images obtained from LATs suffer from limited view problems and lower sensitivity (compared with large single-element US transducers), this approach holds promise to make PAI into clinical one day. This is validated by interests from several small and large companies venturing into developing a commercial PAI system. In the next section, we will briefly discuss the efforts made by the companies.

5 Commercially Available PAT Systems

Worldwide several companies are developing commercial US/PAT imaging systems: FUJIFILM VisualSonics Inc., Toronto, Canada; iThera Medical GmbH, Munich, Germany; TomoWave Laboratories Inc, Houston; Seno Medical Instruments Inc., Texas; Endra Inc, Ann Arbor, Michigan; Optosonics Inc., Oriental, North Carolina, and so on. Real-time assessment of tissue hypoxia *in vivo* was demonstrated using Vevo2100 ultrahigh frequency ($f_c \sim 75$ MHz, imaging resolution ~ 30 μm) US and PAI system from VisualSonics.¹³¹ Recently, *in vivo* liver iron quantitation¹³² and detection of lymph node micrometastases and in-transit metastases from melanoma in a mouse model¹⁰⁷ were demonstrated using iThera inVision-128 and inVision 256-TF, respectively. LOIS-3-D from TomoWave was used for demonstrating melanin nanoparticles as a PA contrast agent.¹³³ A tomographic PA scanner (Nexus 128, Endra Inc.) was used for animal imaging and contrast agent studies.¹³⁴ A dedicated 3-D PAT system from Optosonics was demonstrated for a breast cancer imaging system.¹³⁵ Among the large radiology companies, Philips Healthcare has put

significant effort to integrate PA imaging with its iU22 US system.¹³⁶ Moreover, General Electric has also shown interest in the past on integrating PAI in its US imaging platform.¹³⁷ Hopefully, in the future we will see more progress in this domain from these companies as well as other large healthcare companies.

6 Advancement in PAT Reconstruction Algorithms

The development of the PAT imaging system will be incomplete without the appropriate reconstruction algorithm development in parallel. PAT reconstruction can be considered as an inverse problem. From the boundary PA data, various reconstruction algorithms can be used for obtaining the initial pressure rise distribution inside the tissue.^{138–145} Several advancements have happened in the last couple of years in the PAT reconstruction algorithms: (a) a least-square-based algorithm for accurate reconstruction with a fewer number of transducers,¹⁴⁶ (b) basis pursuit deconvolution algorithm to retain the structural information accurately,¹⁴⁷ (c) filtered-backprojection algorithm,¹⁴⁸ (d) single-stage algorithm for high quality (compared to traditional two-stage algorithms) PA imaging,¹⁴⁹ (e) pulse decomposition algorithm in the time-domain for weak, noisy PA signal reconstruction,¹⁵⁰ (f) an algorithm integrating focal-line-based 3-D image reconstruction with coherent weight to improve the elevation resolution of commercial linear UST array-based PAT,¹⁵¹ (g) a multiview Hilbert transformation method for full-view PAT imaging using linear UST array,⁶³ (h) asymmetric DAQ optimized with compressed sensing method for full-view PAT,¹⁵² and (i) improving tangential resolution in PAT with a modified delay-and-sum algorithm.^{84,85}

7 Conclusions and Future Directions

In this review, we discussed the state-of-art PAT systems and their applications from laboratory research to clinical patient care. PAT has shown progress in many aspects: from preclinical to clinical, from bench to bedside, from static to dynamic, from 2-D to five-dimensional (5-D) volumetric imaging, from expensive to affordable. However, future undertakings are necessary to further mature the PAT technology for real-time clinical applications.

From the laser excitation point of view, the ultrafast laser technology has advanced from bulk Nd:YAG lasers to tiny LEDs. However, PLDs and LEDs demonstrated so far have low PE (less than 1.5 mJ/pulse) and high energy tunable PLDs, and LEDs are yet to be developed for deep-tissue and PA spectroscopic imaging. Along with laser technology, the US detection technology is also advancing to produce high frequency and broadband USTs. A Fabry–Perot polymer film etalon can sense US from DC to several tens of MHz.^{19,153} However, it is a contact approach and allows imaging only superficial features located within a few millimetres inside the tissue surface. The PZT-based USTs have been widely used for US detection. Single-element USTs from 1 MHz to ~1.2 GHz have been used for PAI.^{31,154} One-dimensional array-based USTs are available up to ~45 MHz, 2-D array based are up to 70 MHz.¹⁵⁵ Newer technologies for US transducers also have been explored. Capacitive micromachined ultrasonic transducers (CMUTs) have been demonstrated for PAI as well. The ease of making 2-D or 3-D arrays of transducer with this technology, and other advantages like larger bandwidth, easier integration of the CMUTs with other electronics etc. may make this technology

suitable for PAI as well.^{155,156} Further investigations are needed to mature this technology to be used both in traditional US imaging as well as in PAI.

From the image reconstruction point of view, recovering the initial pressure distribution alone is also a challenging problem for quantitative PA imaging. In general, a PA signal is a function of optical absorption coefficient and local fluence. To make the PAT a robust quantitative measurement tool, the local fluence must be compensated accurately. A quantitative PAT by stochastic search for direct recovery of the optical absorption field was reported recently.¹⁵⁷ The progress toward this study was summarized in the literature.¹⁵⁸ Extracting the absolute chromophore concentrations from PA images obtained at multiple wavelengths is a nontrivial aspect of PAT but is essential for accurate functional and molecular imaging. More information on the quantitative PAI and its challenges can be found in the review.^{159,160}

One of the key application areas of PAT is breast cancer imaging. However, penetration depth is a key obstacle. A penetration depth of ~4 cm has been achieved *in vivo* in the human breast using an excitation wavelength of 800 nm.¹⁰⁰ Twente PA mammoscope^{161,162} and volumetric handheld optoacoustic angiography⁹⁹ seem to be promising tools for breast cancer studies. However, extensive clinical studies need to be performed on this. A high-pulse-energy wavelength-cycling NIR laser system was designed and developed for 3-D PAT of the breast.⁹⁰ It could provide 5-cm deep breast imaging. The laser was designed to improve efficacy and ease of use in the clinic. Wavelength-cycling provides a pulse sequence that repeatedly alternates between two wavelengths (756 and 797 nm) that provide differential imaging. This improves coregistration of captured differential images and quantification of blood oxygen saturation. Extending the penetration depth and image contrast can be achieved by using an exogenous contrast agent with strong absorption in the NIR region. Several contrast agents based on metallic, inorganic, organic nanoparticles, quantum dots, and so on were demonstrated for deep PAT imaging.^{22,163,164–173} Research work is progressing to find a stable, nontoxic, biocompatible, and yet powerful exogenous contrast agent for PAT.^{163,174,175} Very recently, a 12-cm deep-tissue imaging was demonstrated using phosphorus phthalocyanine as an exogenous contrast agent at 1064 nm.⁹

Another application of PAT is metabolic rate of oxygen (MRO₂) imaging. MRO₂ quantifies metabolism and thus it is crucial for screening and treatment of metabolic diseases, particularly cancers and cerebral disorders.⁷¹ Fortunately, PAT allows complete assessment of MRO₂ on a single imaging modality. The PAT measurement of MRO₂ in humans is expected to have a profound impact on studies of tumor metabolism and neovascularization. Another important aspect in which PAT has shown progressing is multimodal PAT imaging. PAT has been successfully integrated with other complementary imaging techniques, e.g., fluorescence and optical coherent tomography. One may refer to reviews focused on this topic.^{71,72,176}

Overall, PAT has very promising preclinical and clinical applications. Several research groups and companies are also working toward making PAI small, compact, easy-to-use, and low-cost, and, therefore, making it possible for translating into the clinic. Improvement in terms of image quality, penetration depth, imaging speed, cost, and size of the system will further enhance the commercialization and clinical applications of this promising technology.

Acknowledgments

The authors would like to acknowledge financial support from the Tier 1 research grant funded by the Ministry of Education in Singapore (RG31/14: M4011276, RG41/14: M4011285) and Tier 2 research grant funded by Ministry of Education in Singapore (ARC2/15: M4020238). Authors have no relevant financial interests in the manuscript and no other potential conflicts of interest to disclose.

References

1. G. Lu and B. Fei, "Medical hyperspectral imaging: a review," *J. Biomed. Opt.* **19**(1), 010901 (2014).
2. M. B. Aldrich et al., "Seeing it through: translational validation of new medical imaging modalities," *Biomed. Opt. Express* **3**(12), 764–776 (2012).
3. A. R. Kherlopian et al., "A review of imaging techniques for systems biology," *BMC Syst. Biol.* **2**(74), 1–18 (2008).
4. M. Xu and L. V. Wang, "Time-domain reconstruction for thermoacoustic tomography in a spherical geometry," *IEEE Trans. Med. Imaging* **21**(7), 814–822 (2002).
5. Y. Zhou, J. Yao, and L. V. Wang, "Tutorial on photoacoustic tomography," *J. Biomed. Opt.* **21**(6), 061007 (2016).
6. L. V. Wang and L. Gao, "Photoacoustic microscopy and computed tomography: from bench to bedside," *Annu. Rev. Biomed. Eng.* **16**, 155–185 (2014).
7. L. V. Wang and S. Hu, "Photoacoustic tomography: in vivo imaging from organelles to organs," *Science* **335**(6075), 1458–1462 (2012).
8. M. Schwarz et al., "Three-dimensional multispectral photoacoustic mesoscopy reveals melanin and blood oxygenation in human skin in vivo," *J. Biophotonics* **9**(1–2), 55–60 (2016).
9. Y. Zhou et al., "A phosphorus phthalocyanine formulation with intense absorbance at 1000 nm for deep optical imaging," *Theranostics* **6**(5), 688–697 (2016).
10. P. K. Upputuri et al., "Recent developments in vascular imaging techniques in tissue engineering and regenerative medicine," *BioMed Res. Intl.* **2015**, 1–9 (2015).
11. P. K. Upputuri et al., "Super-resolution coherent anti-Stokes Raman scattering microscopy with photonic nanojets," *Opt. Express* **22**(11), 12890–12899 (2014).
12. P. K. Upputuri et al., "Circularly polarized coherent anti-Stokes Raman scattering microscopy," *Opt. Lett.* **38**(8), 1262–1264 (2013).
13. P. P. Mondal et al., "Developments in single and multi-photon fluorescence microscopy for high resolution imaging," *J. Indian Inst. Sci.* **93**(1), 15–33 (2013).
14. R. Raghunathan et al., "Optical coherence tomography for embryonic imaging: a review," *J. Biomed. Opt.* **21**(5), 050902 (2016).
15. A. M. Zysk et al., "Optical coherence tomography: a review of clinical development from bench to bedside," *J. Biomed. Opt.* **12**(5), 051403 (2007).
16. H. Yang et al., "Handheld miniature probe integrating diffuse optical tomography with photoacoustic imaging through a MEMS scanning mirror," *Biomed. Opt. Express* **4**(3), 427–432 (2013).
17. J. Yao and L. V. Wang, "Sensitivity of photoacoustic microscopy," *Photoacoustics* **2**(2), 87–101 (2014).
18. X. Cai et al., "Photoacoustic microscopy in tissue engineering," *Materials Today* **16**(3), 67–77 (2013).
19. P. Beard, "Biomedical photoacoustic imaging," *Interface Focus* **1**(4), 602–631 (2011).
20. S. Mallidi, G. P. Luke, and S. Emelianov, "Photoacoustic imaging in cancer detection, diagnosis, and treatment guidance," *Trends Biotechnol.* **29**(5), 213–221 (2011).
21. V. Ntziachristos, "Going deeper than microscopy: the optical imaging frontier in biology," *Nat. Methods* **7**(8), 603–614 (2010).
22. C. Kim, C. Favazza, and L. V. Wang, "In vivo photoacoustic tomography of chemicals: high-resolution functional and molecular optical imaging at new depths," *Chem. Rev.* **110**(5), 2756–2782 (2010).
23. J. Y. Kim et al., "Fast optical-resolution photoacoustic microscopy using a 2-axis water-proofing MEMS scanner," *Sci. Rep.* **5**, 07932 (2015).
24. S.-L. Chen, L. J. Guo, and X. Wang, "All-optical photoacoustic microscopy," *Photoacoustics* **3**(4), 143–150 (2015).
25. B. Ning et al., "Simultaneous photoacoustic microscopy of microvascular anatomy, oxygen saturation, and blood flow," *Opt. Lett.* **40**(6), 910–913 (2015).
26. L. Zeng et al., "Label-free optical-resolution photoacoustic microscopy of superficial microvasculature using a compact visible laser diode excitation," *Opt. Express* **23**(24), 31026–31033 (2015).
27. B. Dong et al., "Isometric multimodal photoacoustic microscopy based on optically transparent micro-ring ultrasonic detection," *Optica* **2**(2), 169–176 (2015).
28. W. Song et al., "Reflection-mode in vivo photoacoustic microscopy with subwavelength lateral resolution," *Biomed. Opt. Express* **5**(12), 4235–4241 (2014).
29. T. Wang et al., "A low-cost photoacoustic microscopy system with a laser diode excitation," *Biomed. Opt. Express* **5**(9), 3053–3058 (2014).
30. M.-L. Li and P.-H. Wang, "Optical resolution photoacoustic microscopy using a Blu-ray DVD pickup head," *Proc. SPIE* **8943**, 894315 (2014).
31. E. M. Strohman, E. S. Berndt, and M. C. Kolios, "High frequency label-free photoacoustic microscopy of single cells," *Photoacoustics* **1**(3–4), 49–53 (2013).
32. S. Han et al., "In vivo virtual intraoperative surgical photoacoustic microscopy," *Appl. Phys. Lett.* **103**(20), 203702 (2013).
33. C. Zhang et al., "Reflection-mode submicron-resolution in vivo photoacoustic microscopy," *J. Biomed. Opt.* **17**(2), 020501 (2012).
34. G. Ku et al., "Photoacoustic microscopy with 2-microm transverse resolution," *J. Biomed. Opt.* **15**(2), 021302 (2010).
35. S. Hu, K. Maslov, and L. V. Wang, "Second-generation optical-resolution photoacoustic microscopy with improved sensitivity and speed," *Opt. Lett.* **36**(7), 1134–1136 (2011).
36. W. Shi et al., "Optical resolution photoacoustic microscopy using novel high-repetition-rate passively Q-switched microchip and fiber lasers," *J. Biomed. Opt.* **15**(5), 056017 (2010).
37. L. Song et al., "Fast 3-D dark-field reflection-mode photoacoustic microscopy in vivo with a 30-MHz ultrasound linear array," *J. Biomed. Opt.* **13**(5), 054028 (2008).
38. H. F. Zhang et al., "Functional photoacoustic microscopy for high-resolution and noninvasive in vivo imaging," *Nat. Biotechnol.* **24**(7), 848–851 (2006).
39. K. Maslov, G. Stoica, and L. V. Wang, "In vivo dark-field reflection-mode photoacoustic microscopy," *Opt. Lett.* **30**(6), 625–627 (2005).
40. J. Xiao et al., "Photoacoustic endoscopy with hollow structured lens-focused polyvinylidene fluoride transducer," *Appl. Opt.* **55**(9), 2301–2305 (2016).
41. C. Chen et al., "Mechanical characterization of intraluminal tissue with phase-resolved photoacoustic viscoelasticity endoscopy," *Biomed. Opt. Express* **6**(12), 4975–4980 (2015).
42. J.-M. Yang et al., "Simultaneous functional photoacoustic and ultrasonic endoscopy of internal organs in vivo," *Nat. Med.* **18**(8), 1297–1302 (2012).
43. J.-M. Yang et al., "A 2.5-mm diameter probe for photoacoustic and ultrasonic endoscopy," *Opt. Express* **20**(21), 23944–23953 (2012).
44. J.-M. Yang et al., "Volumetric photoacoustic endoscopy of internal organs: a phantom and in situ study," *Proc. SPIE* **7564**, 75640D (2010).
45. J. M. Yang et al., "Photoacoustic endoscopy," *Opt. Lett.* **34**(10), 1591–1593 (2009).
46. R. Li et al., "Assessing breast tumor margin by multispectral photoacoustic tomography," *Biomed. Opt. Express* **6**(4), 1273–1281 (2015).
47. P. K. Upputuri and M. Pramanik, "Performance characterization of low-cost, high-speed, portable pulsed laser diode photoacoustic tomography (PLD-PAT) system," *Biomed. Opt. Express* **6**(10), 4118–4129 (2015).
48. X. Yang et al., "Photoacoustic tomography of small animal brain with a curved array transducer," *J. Biomed. Opt.* **14**(5), 054007 (2009).
49. X. Yang and L. V. Wang, "Monkey brain cortex imaging by photoacoustic tomography," *J. Biomed. Opt.* **13**(4), 044009 (2008).
50. M. Pramanik et al., "Design and evaluation of a novel breast cancer detection system combining both thermoacoustic (TA) and photoacoustic (PA) tomography," *Med. Phys.* **35**(6), 2218–2223 (2008).
51. G. Ku et al., "Thermoacoustic and photoacoustic tomography of thick biological tissues toward breast imaging," *Technol. Cancer Res. Treat.* **4**(5), 559–565 (2005).

52. G. Ku et al., "Imaging of tumor angiogenesis in rat brains in vivo by photoacoustic tomography," *Appl. Opt.* **44**(5), 770–775 (2005).
53. X. D. Wang et al., "Noninvasive laser-induced photoacoustic tomography for structural and functional in vivo imaging of the brain," *Nat. Biotechnol.* **21**(7), 803–806 (2003).
54. R. A. Kruger et al., "Thermoacoustic CT with radiowaves: a medical imaging paradigm," *Radiology* **211**(1), 275–278 (1999).
55. P. K. Upputuri, M. Krishnan, and M. Pramanik, "Microsphere enabled sub-diffraction limited optical resolution photoacoustic microscopy—a simulation study," *J. Biomed. Opt.* (2016) (in press).
56. P. K. Upputuri et al., "Super-resolution photoacoustic microscopy using photonic nanojets: a simulation study," *J. Biomed. Opt.* **19**(11), 116003 (2014).
57. A. Danielli et al., "Label-free photoacoustic nanoscopy," *J. Biomed. Opt.* **19**(8), 086006 (2014).
58. C. Zhang, K. Maslov, and L. V. Wang, "Subwavelength-resolution label-free photoacoustic microscopy of optical absorption in vivo," *Opt. Lett.* **35**(19), 3195–3197 (2010).
59. P. K. Upputuri and M. Pramanik, "Pulsed laser diode based optoacoustic imaging of biological tissues," *Biomed. Phys. Eng. Express* **1**(4), 045010–045017 (2015).
60. L. Zeng et al., "3D-visual laser-diode-based photoacoustic imaging," *Opt. Express* **20**(2), 1237–1246 (2012).
61. J. S. Allen and P. Beard, "Pulsed near-infrared laser diode excitation system for biomedical photoacoustic imaging," *Opt. Lett.* **31**(23), 3462–3464 (2006).
62. J. T. Allen and C. P. Beard, "High power visible light emitting diodes as pulsed excitation sources for biomedical photoacoustics," *Biomed. Opt. Express* **7**(3), 1260–1270 (2016).
63. G. Li et al., "Multiview Hilbert transformation for full-view photoacoustic computed tomography using a linear array," *J. Biomed. Opt.* **20**(6), 066010 (2015).
64. S. A. Ermilov et al., "Laser optoacoustic imaging system for detection of breast cancer," *J. Biomed. Opt.* **14**(2), 024007 (2009).
65. X. Wang, "System and method for photoacoustic tomography of joints," US Patent US 2008/0173093 A1 (2008).
66. X. L. Dean-Ben and D. Razansky, "Portable spherical array probe for volumetric real-time optoacoustic imaging at centimeter-scale depths," *Opt. Express* **21**(23), 28062–28071 (2013).
67. H. Ke et al., "Performance characterization of an integrated ultrasound, photoacoustic, and thermoacoustic imaging system," *J. Biomed. Opt.* **17**(5), 056010 (2012).
68. L. Nie et al., "Photoacoustic tomography through a whole adult human skull with a photon recycler," *J. Biomed. Opt.* **17**(11), 110506 (2012).
69. X. Cai et al., "Photoacoustic tomography of foreign bodies in soft biological tissue," *J. Biomed. Opt.* **16**(4), 046017 (2011).
70. M. Pramanik and L. V. Wang, "Thermoacoustic and photoacoustic sensing of temperature," *J. Biomed. Opt.* **14**(5), 054024 (2009).
71. J. Yao, J. Xia, and L. V. Wang, "Multiscale functional and molecular photoacoustic tomography," *Ultrason. Imaging* **38**(1), 44–62 (2016).
72. J. Xia, Y. Wang, and H. Wan, "Recent progress in multimodal photoacoustic tomography," *X-Acoust.: Imaging Sens.* **1**(1), (2015).
73. J. Yao and L. V. Wang, "Breakthroughs in photonics 2013: photoacoustic tomography in biomedicine," *IEEE Photonics J.* **6**(2), 1–6 (2014).
74. J. Xia, J. Yao, and L. V. Wang, "Photoacoustic tomography: principles and advances," *Prog. Electromagn. Res.* **147**, 1–22 (2014).
75. J. Xia and L. V. Wang, "Small-animal whole-body photoacoustic tomography: a review," *IEEE Trans. Biomed. Eng.* **61**(5), 1380–1389 (2014).
76. L. V. Wang and J. Yao, "A practical guide to photoacoustic tomography in the life sciences," *Nat. Methods* **13**(8), 627–638 (2016).
77. A. Taruttis and V. Ntziachristos, "Advances in real-time multispectral optoacoustic imaging and its applications," *Nat. Photonics* **9**(4), 219–227 (2015).
78. K. S. Valluru and J. K. Willmann, "Clinical photoacoustic imaging of cancer," *Ultrasonography* **35**(4), 267–280 (2016).
79. G. Ku et al., "Multiple-bandwidth photoacoustic tomography," *Phys. Med. Biol.* **49**(7), 1329–1338 (2004).
80. R. A. Kruger et al., "Photoacoustic ultrasound (PAUS)—reconstruction tomography," *Med. Phys.* **22**(10), 1605–1609 (1995).
81. R. Ma et al., "Multispectral optoacoustic tomography (MSOT) scanner for whole-body small animal imaging," *Opt. Express* **17**(24), 21414–21426 (2009).
82. M. Pramanik, G. Ku, and L. V. Wang, "Tangential resolution improvement in thermoacoustic and photoacoustic tomography using a negative acoustic lens," *J. Biomed. Opt.* **14**(2), 024028 (2009).
83. C. Li, G. Ku, and L. V. Wang, "Negative lens concept for photoacoustic tomography," *Phys. Rev. E* **78**(2), 021901 (2008).
84. S. K. Kalva and M. Pramanik, "Experimental validation of tangential resolution improvement in photoacoustic tomography using a modified delay-and-sum reconstruction algorithm," *J. Biomed. Opt.* **21**(8), 086011 (2016).
85. M. Pramanik, "Improving tangential resolution with a modified delay-and-sum reconstruction algorithm in photoacoustic and thermoacoustic tomography," *J. Opt. Soc. Am. A* **31**(3), 621–627 (2014).
86. R. Cheng et al., "Noninvasive assessment of early dental lesion using a dual-contrast photoacoustic tomography," *Sci. Rep.* **6**, 21798 (2016).
87. *American National Standard for Safe Use of Lasers ANSI Z136.1-2000*, American National Standards Institute, Inc., New York, NY (2000).
88. Z. Deng, W. Li, and C. Li, "Slip-ring-based multi-transducer photoacoustic tomography system," *Opt. Lett.* **41**(12), 2859–2862 (2016).
89. <http://www.opotek.com/product/phocus-mobile>.
90. A. A. Oraevsky et al., "Advanced laser system for 3D optoacoustic tomography of the breast," *Proc. SPIE* **9708**, 97085B (2016).
91. A. A. Oraevsky et al., "Quantitative assessment of photoacoustic tomography systems integrating clinical ultrasound transducers using novel tissue-simulating phantoms," *Proc. SPIE* **9323**, 932333 (2015).
92. P. K. Upputuri and M. Pramanik, "High-speed pre-clinical brain imaging using pulsed laser diode based photoacoustic tomography (PLD-PAT) system," *Proc. SPIE* **9708**, 97084R (2016).
93. C. Li et al., "Real-time photoacoustic tomography of cortical hemodynamics in small animals," *J. Biomed. Opt.* **15**(1), 010509 (2010).
94. J. Gamelin et al., "Curved array photoacoustic tomographic system for small animal imaging," *J. Biomed. Opt.* **13**(2), 024007 (2008).
95. Y. Zeng et al., "Photoacoustic and ultrasonic coimage with a linear transducer array," *Opt. Lett.* **29**(15), 1760–1762 (2004).
96. X. Li et al., "High resolution functional photoacoustic tomography of breast cancer," *Med. Phys.* **42**(9), 5321–5328 (2015).
97. L. Xi et al., "Design and evaluation of a hybrid photoacoustic tomography and diffuse optical tomography system for breast cancer detection," *Med. Phys.* **39**(5), 2584–2594 (2012).
98. M. Heijblom et al., "Photoacoustic image patterns of breast carcinoma and comparisons with magnetic resonance imaging and vascular stained histopathology," *Sci. Rep.* **5**, 11778 (2015).
99. X. L. Dean-Ben et al., "Volumetric hand-held optoacoustic angiography as a tool for real-time screening of dense breast," *J. Biophotonics* **9**(3), 253–259 (2016).
100. R. A. Kruger et al., "Photoacoustic angiography of the breast," *Med. Phys.* **37**(11), 6096–6100 (2010).
101. B. Lashkari and A. Mandelis, "Comparison between pulsed laser and frequency-domain photoacoustic modalities: signal-to-noise ratio, contrast, resolution, and maximum depth detectivity," *Rev. Sci. Instrum.* **82**(9), 094903–094914 (2011).
102. "IEC 60825-1:1993+ A1:1997+ A2:2001, in Part 1: Equipment Classification, Requirements and User's Guide Edition 1:1993 Consolidated with Amendments 1:1997 and 2:2001, International Electrotechnical Commission, Geneva, Switzerland.
103. D. R. Reinecke et al., "Co-registered photoacoustic, thermoacoustic, and ultrasound mouse imaging," *Proc. SPIE* **7564**, 756420 (2010).
104. X. L. Dean-Ben, A. Ozbek, and D. Razansky, "Volumetric real-time tracking of peripheral human vasculature with GPU-accelerated three-dimensional optoacoustic tomography," *IEEE Trans. Med. Imaging* **32**(11), 2050–2055 (2013).
105. A. Dima and V. Ntziachristos, "Non-invasive carotid imaging using optoacoustic tomography," *Opt. Express* **20**(22), 25044–25057 (2012).
106. X. L. Deán-Ben and D. Razansky, "Adding fifth dimension to optoacoustic imaging: volumetric time-resolved spectrally enriched tomography," *Light: Sci. Appl.* **3**(1), e137 (2014).
107. V. Neuschmelting et al., "Lymph node micrometastases and in-transit metastases from melanoma: in vivo detection with multispectral optoacoustic imaging in a mouse model," *Radiology* **280**(1), 137–150 (2016).

108. J. Tang et al., "Wearable 3-D photoacoustic tomography for functional brain imaging in behaving rats," *Sci. Rep.* **6**, 25470 (2016).
109. K. M. McMasters et al., "Sentinel lymph node biopsy for breast cancer: a suitable alternative to routine axillary dissection in multi-institutional practice when optimal technique is used," *J. Clin. Oncol.* **18**(13), 2560–2566 (2000).
110. D. Krag et al., "The sentinel node in breast cancer," *N. Engl. J. Med.* **339**(14), 941–946 (1998).
111. O. A. Ung et al., "Australasian experience and trials in sentinel lymph node biopsy: the RACS SNAC trial," *Asian J. Surg.* **27**(4), 284–290 (2004).
112. A. D. Purushotham et al., "Morbidity after sentinel lymph node biopsy in primary breast cancer: results from a randomized controlled trial," *J. Clin. Oncol.* **23**(19), 4312–4321 (2005).
113. K. K. Swenson et al., "Comparison of side effects between sentinel lymph node and axillary lymph node dissection for breast cancer," *Ann. Surg. Oncol.* **9**(8), 745–753 (2002).
114. G. P. Luke et al., "Sentinel lymph node biopsy revisited: ultrasound-guided photoacoustic detection of micrometastases using molecularly targeted plasmonic nanosensors," *Cancer Res.* **74**(19), 5397–5408 (2014).
115. C. Kim et al., "Handheld array-based photoacoustic probe for guiding needle biopsy of sentinel lymph nodes," *J. Biomed. Opt.* **15**(4), 046010 (2010).
116. T. N. Erpelding et al., "Sentinel lymph nodes in the rat: noninvasive photoacoustic and US imaging with a clinical US system," *Radiology* **256**(1), 102–110 (2010).
117. A. Garcia-Urbe et al., "Dual-modality photoacoustic and ultrasound imaging system for noninvasive sentinel lymph node detection in patients with breast cancer," *Sci. Rep.* **5**, 15748 (2015).
118. D. Piras et al., "Photoacoustic needle: minimally invasive guidance to biopsy," *J. Biomed. Opt.* **18**(7), 070502 (2013).
119. K. Daoudi et al., "Handheld probe integrating laser diode and ultrasound transducer array for ultrasound/photoacoustic dual modality imaging," *Opt. Express* **22**(21), 26365–26374 (2014).
120. L. G. Montilla et al., "Real-time photoacoustic and ultrasound imaging: a simple solution for clinical ultrasound systems with linear arrays," *Phys. Med. Biol.* **58**(1), N1–N12 (2013).
121. K. Sivasubramanian and M. Pramanik, "High frame rate photoacoustic imaging at 7000 frames per second using clinical ultrasound system," *Biomed. Opt. Express* **7**(2), 312–323 (2016).
122. Y. Wang et al., "Slit-enabled linear-array photoacoustic tomography with near isotropic spatial resolution in three dimensions," *Opt. Lett.* **41**(1), 127–130 (2016).
123. W. Shu et al., "Broadening the detection view of 2D photoacoustic tomography using two linear array transducers," *Opt. Express* **24**(12), 12755–12768 (2016).
124. L. Wang et al., "Ultrasonic-heating-encoded photoacoustic tomography with virtually augmented detection view," *Optica* **2**(4), 307–312 (2015).
125. X. Wang, D. L. Chamberland, and D. A. Jamadar, "Noninvasive photoacoustic tomography of human peripheral joints toward diagnosis of inflammatory arthritis," *Opt. Lett.* **32**(20), 3002–3005 (2007).
126. J. Xiao et al., "Quantitative two-dimensional photoacoustic tomography of osteoarthritis in the finger joints," *Opt. Express* **18**(14), 14359–14365 (2010).
127. Y. Sun, E. S. Sobel, and H. Jiang, "First assessment of three-dimensional quantitative photoacoustic tomography for in vivo detection of osteoarthritis in the finger joints," *Med. Phys.* **38**(7), 4009–4017 (2011).
128. L. Xi and H. Jiang, "High resolution three-dimensional photoacoustic imaging of human finger joints in vivo," *Appl. Phys. Lett.* **107**(6), 063701 (2015).
129. G. Xu et al., "Photoacoustic and ultrasound dual-modality imaging of human peripheral joints," *J. Biomed. Opt.* **18**(1), 010502 (2013).
130. K. Daoudi et al., "Handheld probe for portable high frame photoacoustic/ultrasound imaging system," *Proc. SPIE* **8581**, 85812I (2013).
131. M. Gerling et al., "Real-time assessment of tissue hypoxia in vivo with combined photoacoustics and high-frequency ultrasound," *Theranostics* **4**(6), 604–613 (2014).
132. F. Maccarinelli et al., "Photoacoustic molecular imaging for liver iron quantitation," *J. Biomed. Opt.* **21**(5), 056008 (2016).
133. A. Liopo, R. Su, and A. A. Oraevsky, "Melanin nanoparticles as a novel contrast agent for optoacoustic tomography," *Photoacoustics* **3**(1), 35–43 (2015).
134. J. V. Jokerst et al., "Cellulose nanoparticles are a biodegradable photoacoustic contrast agent for use in living mice," *Photoacoustics* **2**(3), 119–127 (2014).
135. R. A. Kruger et al., "Dedicated 3D photoacoustic breast imaging," *Med. Phys.* **40**(11), 113301 (2013).
136. J. Dean et al., "Real-time photoacoustic data acquisition with Philips iU22 ultrasound scanner," *Proc. SPIE* **6856**, 685622 (2008).
137. K. Zell et al., "First practical experiences with the optoacoustic/ultrasound system OPUS," *Proc. SPIE* **6856**, 68560S (2008).
138. C. Lutzweiler, X. L. Dean-Ben, and D. Razansky, "Expediting model-based optoacoustic reconstructions with tomographic symmetries," *Med. Phys.* **41**(1), 013302 (2014).
139. C. Huang et al., "Full-wave iterative image reconstruction in photoacoustic tomography with acoustically inhomogeneous media," *IEEE Trans. Med. Imaging* **32**(6), 1097–1110 (2013).
140. N. A. Rejesh, H. Pullagurla, and M. Pramanik, "Deconvolution-based deblurring of reconstructed images in photoacoustic/thermoacoustic tomography," *J. Opt. Soc. Am. A* **30**(10), 1994–2001 (2013).
141. M. Xu and L. V. Wang, "Universal back-projection algorithm for photoacoustic computed tomography," *Phys. Rev. E* **71**(1), 016706 (2005).
142. K. P. Köstli and P. C. Beard, "Two-dimensional photoacoustic imaging by use of Fourier-transform image reconstruction and a detector with an anisotropic response," *Appl. Opt.* **42**(10), 1899–1908 (2003).
143. M. Xu and L. V. Wang, "Pulsed-microwave-induced thermoacoustic tomography: filtered backprojection in a circular measurement configuration," *Med. Phys.* **29**(8), 1661–1669 (2002).
144. M. Xu and L. V. Wang, "Time-domain reconstruction for thermoacoustic tomography in a spherical geometry," *IEEE Trans. Med. Imaging* **21**(7), 814–822 (2002).
145. G. Palttauf et al., "Iterative reconstruction algorithm for optoacoustic imaging," *J. Acoust. Soc. Am.* **112**(4), 1536–1544 (2002).
146. C. B. Shaw et al., "Least squares QR-based decomposition provides an efficient way of computing optimal regularization parameter in photoacoustic tomography," *J. Biomed. Opt.* **18**(8), 080501 (2013).
147. J. Prakash et al., "Basis pursuit deconvolution for improving model-based reconstructed images in photoacoustic tomography," *Biomed. Opt. Express* **5**(5), 1363–1377 (2014).
148. H. Huang et al., "An adaptive filtered back-projection for photoacoustic image reconstruction," *Med. Phys.* **42**(5), 2169–2178 (2015).
149. M. Haltmeier, L. Neumann, and S. Rabanser, "Single-stage reconstruction algorithm for quantitative photoacoustic tomography," *Inverse Probl.* **31**(6), 065005 (2015).
150. L. Liu et al., "Photoacoustic tomography from weak and noisy signals by using a pulse decomposition algorithm in the time-domain," *Opt. Express* **23**(21), 26969–26977 (2015).
151. D. Wang et al., "Coherent-weighted three-dimensional image reconstruction in linear-array-based photoacoustic tomography," *Biomed. Opt. Express* **7**(5), 1957–1965 (2016).
152. M. Cao et al., "Full-view photoacoustic tomography using asymmetric distributed sensors optimized with compressed sensing method," *Biomed. Signal Process. Control* **21**, 19–25 (2015).
153. E. Zhang, J. Laufer, and P. Beard, "Backward-mode multiwavelength photoacoustic scanner using a planar Fabry-Perot polymer film ultrasound sensor for high-resolution three-dimensional imaging of biological tissues," *Appl. Opt.* **47**(4), 561–577 (2008).
154. Y. Hou et al., "Thin polymer etalon arrays for high-resolution photoacoustic imaging," *J. Biomed. Opt.* **13**(6), 064033 (2008).
155. O. Warshavski et al., "Experimental evaluation of cMUT and PZT transducers in receive only mode for photoacoustic imaging," *Proc. SPIE* **9708**, 970830 (2016).
156. S. R. Kothapalli et al., "Deep tissue photoacoustic imaging using a miniaturized 2-D capacitive micromachined ultrasonic transducer array," *IEEE Trans. Biomed. Eng.* **59**(5), 1199–1204 (2012).
157. M. Venugopal et al., "Quantitative photoacoustic tomography by stochastic search: direct recovery of the optical absorption field," *Opt. Lett.* **41**(18), 4202–4205 (2016).
158. T. Saratoon et al., "A gradient-based method for quantitative photoacoustic tomography using the radiative transfer equation," *Inverse Probl.* **29**(7), 075006 (2013).

159. B. Cox et al., "Quantitative spectroscopic photoacoustic imaging: a review," *J. Biomed. Opt.* **17**(6), 061202 (2012).
160. B. T. Cox, J. G. Laufer, and P. C. Beard, "The challenges for quantitative photoacoustic imaging," *Proc. SPIE* **7177**, 717713 (2009).
161. M. Heijblom et al., "Visualizing breast cancer using the Twente photoacoustic mammoscope: what do we learn from twelve new patient measurements?" *Opt. Express* **20**(11), 11582–11597 (2012).
162. D. Piras et al., "Photoacoustic imaging of the breast using the Twente photoacoustic mammoscope: present status and future perspectives," *IEEE J. Sel. Top. Quantum Electron.* **16**(4), 730–739 (2010).
163. S. Huang et al., "A dual-functional benzobisthiadiazole derivative as an effective theranostic agent for near-infrared photoacoustic imaging and photothermal therapy," *J. Mater. Chem. B* **4**(9), 1696–1703 (2016).
164. M. Olivo and U. S. Dinis, *Frontiers in Biophotonics for Translational Medicine*, Springer (2016).
165. J. M. Yoo, J. H. Kang, and B. H. Hong, "Graphene-based nanomaterials for versatile imaging studies," *Chem. Soc. Rev.* **44**(14), 4835–4852 (2015).
166. G. Hong et al., "Carbon nanomaterials for biological imaging and nanomedicinal therapy," *Chem. Rev.* **115**(19), 10816–10906 (2015).
167. L. Nie and X. Chen, "Structural and functional photoacoustic molecular tomography aided by emerging contrast agents," *Chem. Soc. Rev.* **43**(20), 7132–7170 (2014).
168. G. P. Luke et al., "Silica-coated gold nanoplates as stable photoacoustic contrast agents for sentinel lymph node imaging," *Nanotechnology* **24**(45), 455101 (2013).
169. Z. Zha et al., "Biocompatible polypyrrole nanoparticles as a novel organic photoacoustic contrast agent for deep tissue imaging," *Nanoscale* **5**(10), 4462–4467 (2013).
170. G. Ku et al., "Copper sulfide nanoparticles as a new class of photoacoustic contrast agent for deep tissue imaging at 1064 nm," *ACS Nano* **6**(8), 7489–7496 (2012).
171. D. Pan et al., "Recent advances in colloidal gold nanobeacons for molecular photoacoustic imaging," *Contrast Media Mol. Imaging* **6**(5), 378–388 (2011).
172. K. Homan et al., "Silver nanosystems for photoacoustic imaging and image-guided therapy," *J. Biomed. Opt.* **15**(2), 021316 (2010).
173. A. De la Zerda et al., "Carbon nanotubes as photoacoustic molecular imaging agents in living mice," *Nat. Nanotechnol.* **3**(9), 557–562 (2008).
174. Y. Shi et al., "Targeted Au-core-Ag-shell nanorods as a dual-functional contrast agent for photoacoustic imaging and photothermal therapy," *Biomed. Opt. Express* **7**(5), 1830–1841 (2016).
175. J. Xia, C. Kim, and J. F. Lovell, "Opportunities for photoacoustic-guided drug delivery," *Curr. Drug Targets* **16**(6), 571–581 (2015).
176. J. Kim et al., "Photoacoustic imaging platforms for multimodal imaging," *Ultrasonography* **34**(2), 88–97 (2015).

Paul Kumar Upputuri is currently working as a research fellow at Nanyang Technological University. He received his PhD in physics from the Indian Institute of Technology (IIT), Madras, India in 2010. He has published around 65 papers in journals and conference proceedings. His research areas are biomedical imaging, coherent Raman microscopy, ultrafast-lasers, optical metrology, and instrumentation. He is a member of SPIE, OSA, Indian Laser Association (ILA), and Optical Society of India (OSI).

Manojit Pramanik received his PhD in biomedical engineering from Washington University in St. Louis, Missouri, USA. He is an assistant professor of the School of Chemical and Biomedical Engineering, Nanyang Technological University, Singapore. His research interests include the development of photoacoustic/thermoacoustic imaging systems, image reconstruction methods, clinical application areas such as breast cancer imaging, molecular imaging, contrast agent development, Monte Carlo simulation of light propagation in biological tissue.