Deep neural network-based bandwidth enhancement of photoacoustic data

Sreedevi Gutta
Venkata Suryanarayana Kadimesetty
Sandeep Kumar Kalva
Manojit Pramanik
Sriram Ganapathy
Phaneendra K. Yalavarthy

Deep neural network-based bandwidth enhancement of photoacoustic data

Sreedevi Gutta,^a Venkata Suryanarayana Kadimesetty,^a Sandeep Kumar Kalva,^b Manojit Pramanik,^b Sriram Ganapathy,^c and Phaneendra K. Yalavarthy^{a,*}

^aIndian Institute of Science, Department of Computational and Data Sciences, Bangalore, Karnataka, India

^bNanyang Technological University, School of Chemical and Biomedical Engineering, Singapore

^cIndian Institute of Science, Department of Electrical Engineering, Bangalore, Karnataka, India

Abstract. Photoacoustic (PA) signals collected at the boundary of tissue are always band-limited. A deep neural network was proposed to enhance the bandwidth (BW) of the detected PA signal, thereby improving the quantitative accuracy of the reconstructed PA images. A least square-based deconvolution method that utilizes the Tikhonov regularization framework was used for comparison with the proposed network. The proposed method was evaluated using both numerical and experimental data. The results indicate that the proposed method was capable of enhancing the BW of the detected PA signal, which in turn improves the contrast recovery and quality of reconstructed PA images without adding any significant computational burden. © 2017 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: 10.1117/1.JBO.22.11.116001]

Keywords: photoacoustic data; deep neural network; bandwidth enhancement; reconstruction.

Paper 170494LRR received Jul. 28, 2017; accepted for publication Oct. 9, 2017; published online Nov. 2, 2017.

1 Introduction

Photoacoustic tomography (PAT) is a noninvasive and hybrid biomedical imaging modality that combines the merits of both optical and ultrasound imaging techniques.^{1–5} In this modality, a nanosecond duration short laser pulse is used to illuminate the tissue of interest. The deposited energy is absorbed by the tissue, resulting in a small temperature rise in the order of milliKelvin. Subsequently, acoustic waves are generated due to thermoelastic expansion and are collected using ultrasonic transducers placed on the boundary of the tissue. The detected photoacoustic (PA) signal is used to map the initial pressure with the help of reconstruction algorithms. The initial pressure is proportional to the absorption coefficient of the tissue, assuming homogeneous fluence distribution.

The reconstruction algorithms estimate the initial pressure distribution from the measured PA signal using either analytical or model-based iterative techniques.⁶ The performance of the reconstruction algorithms depends on the quality of PA signal. The factor that affects this performance other than signal-to-noise ratio (SNR) are limited bandwidth (BW) of acoustic detectors.^{7,8} The limited BW of the transducer results in the spatially invariant blurring of the reconstructed PA images. The signal received at the acoustic detectors is always band-limited. As it is unavoidable to use band-limited transducers, an additional step to restore the original PA signal will improve the quality of the reconstructed PA images. To have real-time imaging, it is also necessary that this additional step is computationally efficient.

The PA signal detected using the band-limited transducer can be simulated by convolving the original PA signal with the transducer impulse response.⁹ Previously, deconvolution algorithms were proposed to overcome the BW limitation of the ultrasound transducer.⁹ These algorithms attempt to restore the original PA

signal by deconvolving the detected PA signal with the transducer impulse response, which can be estimated either experimentally or can be obtained from the manufacturer. The deconvolution algorithms are in general ill-posed, necessitating the need of regularization to obtain a unique solution.¹⁰ Recently, a method to recover out-of-band frequencies was proposed for PA imaging. This method was based on inverse (Wiener) filtering, and it requires a best estimate of the frequency response of the ultrasound transducers.¹¹

Application of deep convolutional neural network for natural image deconvolution was proposed recently by Xu et al.,¹² earlier to this it was applied to astronomical image reconstruction,¹³ semantic segmentation,¹⁴ and music classification.¹⁵

In this letter, a simple deep neural network (DNN) was proposed to enhance the BW of the detected PA signal, which reconstructs all frequency components. The network was trained with band-limited signal as input and output being full BW signal. The enhanced/predicted acoustic (PA) signals were used as input to the analytical reconstruction algorithms, such as backprojection to quantitatively obtain accurate PA images. Similar framework could be applicable to ultrasound imaging (including ultrasound tomography) to enhance the BW of the recorded echo signal.

2 Materials and Methods

2.1 Photoacoustic Wave Propagation Modeling

The PA wave propagation is modeled using the following differential equation:⁶

$$\nabla^2 P(x, t) - \frac{1}{v_s^2} \frac{\partial^2 P(x, t)}{\partial t^2} = -\Gamma \frac{\partial H(x, t)}{\partial t}, \quad (1)$$

*Address all correspondence to: Phaneendra K. Yalavarthy, E-mail: yalavarthy@iisc.ac.in

where v_s represents the speed of sound in the medium, Γ denotes the Gruneisen coefficient, and $H(x, t)$ represents the absorbed energy per unit time per unit volume with x indicating the spatial location. The above equation [Eq. (1)] can be equivalently written as

$$\nabla^2 P(x, t) - \frac{1}{v_s^2} \frac{\partial^2 P(x, t)}{\partial t^2} = 0, \quad (2)$$

with initial conditions $\partial P / \partial t|_{t=0} = 0$ and $P|_{t=0} = \Gamma H(x)$. The acoustic wave equation can be discretized as a linear system of equations reducing to a matrix form as^{16,17}

$$AP_0 = b, \quad (3)$$

with the system matrix A containing the impulse responses of all the pixels in the imaging grid, P_0 indicates the initial pressure distribution, and b is the band-limited PA data collected by the transducers. Typically, Green's function approach is deployed to measure the impulse response for its computational efficiency.¹⁸ The backprojection (analytical) type image reconstruction scheme becomes^{16,17}

$$P_{bp} = A^T b, \quad (4)$$

where A^T indicates the transpose of the matrix. The acquired PA signal in the presence of noise can be written as

$$b = \bar{b} * h + \text{noise}, \quad (5)$$

where $\bar{b}$ denotes the original PA signal and h indicates the impulse response of the transducer.

2.2 Least Square Deconvolution

A least square deconvolution with Tikhonov regularization was proposed earlier to remove the effect due to finite BW of the transducer.⁹ Equation (5) can be rewritten as

$$b = H\bar{b} + n, \quad (6)$$

where H is the circulant convolution matrix formed from h . The original PA signal can be obtained by minimizing the cost function given as

$$C(\bar{b}) = \|b - H\bar{b}\|^2 + \lambda \|\bar{b}\|^2, \quad (7)$$

with λ being the regularization parameter. The minimum for the cost function is obtained by equating the gradient of Eq. (7) to zero

$$H^T b = H^T H \bar{b}_{Tik} + \lambda \bar{b}_{Tik}, \quad (8)$$

where T indicates the transpose of a matrix and $\bar{b}_{Tik}$ is a least square deconvoluted signal. Equation (8) can be written in time domain as

$$h[-n] * b[n] = h[-n] * h[n] * \bar{b}_{Tik}[n] + \lambda \bar{b}_{Tik}[n], \quad (9)$$

where n denotes the time index. In frequency domain, Eq. (9) becomes

$$\text{conj}(H)B = \text{conj}(H)H\bar{B}_{Tik} + \lambda \bar{B}_{Tik}, \quad (10)$$

where B represents the Fourier transformed version of b . $\bar{B}_{Tik}$ from Eq. (10) is evaluated as

$$\bar{B}_{Tik} = \frac{\text{conj}(H)B}{\text{conj}(H)H + \lambda}. \quad (11)$$

The time-domain signal, $\bar{b}_{Tik}$, can be obtained by applying inverse Fourier transform to $\bar{B}_{Tik}$. The corrected PA signal $\bar{b}_{Tik}$ is then used for image reconstruction using Eq. (4).

2.3 Proposed Deep Neural Network

Recently, deep learning has been proven as a promising technique for many visual recognition tasks.^{19,20} Deep learning-based improvements to medical imaging has been discussed widely and known to have huge potential, especially in enhancing the raw data.²¹ The deep feedforward neural networks extract the features through the cascade of several nonlinear activation functions (Fig. 1). As the enhancement of BW of the collected boundary data can be seen as extrapolation of a nonlinear function, the approach followed in this work is to utilize the DNN in predicting the missing frequencies. This in turn removes the effect of finite BW of transducers. The idea of

Fig. 1 The neural network architecture listing the layers utilized in the proposed DNN. The network takes band-limited signal as input and predicts full BW signal. FC, fully connected, and ReLU, rectified linear unit.

the proposed network is to formulate the deconvolution process as a nonlinear regression.

The input to the network is a band-limited signal $x = (x_1, x_2, \dots, x_N)$, and the output is a full BW signal $y = (y_1, y_2, \dots, y_N)$, where x_i and $y_i \in \mathbb{R}$. Given a training dataset $(x_i, y_i)_{i=1}^N$, our goal is to determine a nonlinear function ϕ that can effectively map between the input band-limited signal x and the output full BW signal y : $y = \phi(x)$. The network architecture was shown in Fig. 1. The proposed network contains five fully connected (FC) layers, one being the input layer and the other being the output layer and the rest three are hidden layers. Our network architecture is similar to the decoder network.²² Note that the encoder-decoder network was proven to be efficient in the deep learning literature and has been applied in many scenarios, such as predicting the missing data.²² The first and third hidden layers (h_1 and h_3 in Fig. 1) are designed with 768 $[512 + (512/2)]$ nodes, i.e., by increasing the original number of nodes by 50%. The second hidden layer (h_2 in Fig. 1) contains 256 $(512/2)$ nodes, i.e., considering half of the original nodes. The k 'th layer of the network evaluates: $h_k = f(W_k h_{k-1} + b_k)$, where h_k indicates the k 'th hidden layer, f is the activation function, W_k represents the matrix of trainable weights, and b_k is a vector of trainable biases. The activation function utilized in this network except for the last layer (as it is a regression problem) is rectified linear unit (ReLU). It should be noted that ReLUs are computationally efficient as they do not require to compute any exponentiation and division.²³

The network was trained with mean-squared error (MSE) as a loss function. The MSE can be written as

$$\text{MSE} = \frac{1}{N} \sum_{i=1}^N \|\hat{y}_i - \phi(x_i)\|^2, \quad (12)$$

where $\hat{y}_i$ represents the predictions of the model on the training set. The optimal weights for the network are determined in such a way that the MSE reduces by observing a training set (x_i, y_i) . In other words, the goal of training is to learn the model parameters weights and biases by minimizing MSE (cost function), which measures the discrepancy between the target (expected) and the predicted output. An efficient ADAM optimization algorithm²⁴ that optimizes stochastic objective functions was used to find the optimal weights. It combines the merits of AdaGrad,²⁵ which is efficient for sparse gradients, and RMSProp,²⁶ which is known to be efficient for nonstationary objectives. Note that this method only requires evaluation of first-order gradients.

2.4 Numerical and Experimental Study

A computational grid of dimension 100 mm $\times$ 100 mm containing 1000 $\times$ 1000 pixels was considered. The data were generated for the objects of size 400 $\times$ 400. Hundred point detectors of center frequency 2.25 MHz and 70% BW were placed around the tissue surface equidistantly on a circle of radius 37.02 mm. The data were recorded for a sampling frequency of 12.5 MHz with a total of 512 time steps. The medium was assumed to be homogeneous with uniform speed of sound as 1500 m/s and with no absorption and dispersion.

The proposed model is also validated using the experimental data. The schematic representation of the PAT imaging system used for performing the experiment is shown in Fig. 1(e) of

Ref. 27. A Q -switched Nd:YAG laser (Continuum, Surelite Ex) delivering laser pulses of 532-nm wavelength of 5-ns duration at 10-Hz repetition rate was used. Four right-angle uncoated prisms (PS911, Thorlabs) and one uncoated planoconcave lens (LC1715, Thorlabs) were used to incident laser energy of density ~ 9 mJ/cm² (< 20 mJ/cm²: ANSI safety limit)²⁸ on the phantom. A triangular-shaped horse hair phantom (side-length of hair: 10 mm and diameter of hair: 0.15 mm) data were collected at 100 locations around the sample using an ultrasound transducer (Olympus NDT, V306-SU) of 2.25-MHz center frequency with 13-mm-diameter active area and 70% BW. Another experimental phantom, circular in shape, made using low-density polyethylene tubes (5-mm inner diameter) filled with black Indian ink was also utilized to evaluate the proposed method. The tubes were placed at 0 and 15 mm from the scanning center and affixed at the bottom of the acrylic slab. The detected PA signals were first amplified and filtered using a pulse amplifier (Olympus NDT, 5072PR) and then recorded using a data acquisition (DAQ) card (GaGe, compuscope 4227) inside a desktop (Intel Xeon 3.7 GHz 64-bit processor, 16-GB RAM, running Windows 10 operating system). A transistor-transistor logic sync signal from the laser was used to synchronize the DAQ with laser illumination. The original data were collected at a sampling frequency of 25 MHz containing 1024 time points, and simulations were performed at a sampling frequency of 12.5 MHz as the original data were subsampled to 512 time points. Figures of merit, such as Pearson's correlation coefficient (PC), contrast-to-noise ratio (CNR), and SNR, were used to quantitatively show the improvement obtained from the proposed method and are defined in Refs. 17 and 27.

For training of the network, numerical breast phantoms²⁹ were used to obtain 286,300 (generated using 2863 phantoms, obtained via considering two-dimensional slices in the Z -direction of these breast phantoms, with 100 detector positions for each slice) samples of full BW and limited BW PA signals. The breast phantoms were chosen for training as they provide close realization of patient anatomical structures, thus enabling realistic numerical experiments. An open-source k -wave toolbox was utilized to generate the data from these phantoms.³⁰ The network was trained for a total of 100 epochs on a batch size of 100. The validation was performed on 1% of the training data. Keras software (Ref. 31), with backend as Theano (Ref. 32), which is an open-source python library was utilized for training of the proposed network. The training loss in the end of 100 epochs is 2.0826×10^{-4} . All computations were performed on a Linux workstation with a dual six-core Intel Xeon processor having a speed of 2.66 GHz with 64-GB RAM. To evaluate the degree of correlation between the target and the reconstructed image, PC was computed. It is given as¹⁷

$$\text{PC}(P_0, P_{\text{recon}}) = \frac{\text{cov}(P_0, P_{\text{recon}})}{\sigma(P_0)\sigma(P_{\text{recon}})}, \quad (13)$$

where P_0 represents the target initial pressure distribution, P_{recon} is the reconstructed initial pressure distribution, σ is the standard deviation, and cov is the covariance.

To measure the differentiability of the region of interest with respect to background, CNR was utilized. It can be defined as¹⁷

$$\text{CNR} = \frac{\mu_{\text{roi}} - \mu_{\text{back}}}{(\sigma_{\text{roi}}^2 a_{\text{roi}} + \sigma_{\text{back}}^2 a_{\text{back}})^{1/2}}, \quad (14)$$

where μ and σ correspond to the mean and standard deviation, respectively. The $a_{\text{roi}} = \frac{A_{\text{roi}}}{A_{\text{total}}}$, where A_{roi} indicates the number of pixels with nonzero initial pressure rise in the target and A_{total} is for the total number of pixels in the imaging grid, $a_{\text{back}} = \frac{A_{\text{back}}}{A_{\text{total}}}$ with A_{back} denoting the number of pixels with zero initial pressure rise in the target.

For the experimental case, where it is not possible to determine the target initial pressure distribution, SNR was computed. It is expressed as

$$\text{SNR}(\text{dB}) = 20 \times \log_{10} \left(\frac{S}{n} \right), \quad (15)$$

with S denoting the peak-to-peak signal amplitude and n corresponds to standard deviation of noise.

3 Results and Discussion

Three numerical phantoms (different from the training data) as shown in the first column of Fig. 2 were considered to evaluate the effectiveness of the proposed method: (a) blood vessel network is generally used as PA numerical phantom for imaging blood vasculature, (b) Derenzo phantom containing various size circular pressure distribution, and (c) PAT phantom to simulate sharp edges. The PA time-domain and frequency response plots for full BW, limited BW, predicted data from least square deconvolution technique, and predicted data from DNN were shown in Fig. 3. Figures 3(a)–3(b) correspond to the time-domain and frequency spectrum plot for the blood vessel phantom as shown in Fig. 2(a) using the transducer located at 9 o'clock position. The time-domain and amplitude spectrum plots for the Derenzo Fig. 2(f) and PAT phantom Fig. 2(k) for the data collected using the same transducer were given in Figs. 3(c)–3(f), respectively. The BW enhancement using the proposed DNN can be observed from these plots (especially from the frequency response plots), where frequency response

of the signal estimated using the proposed method was more close to full BW signal response.

Due to the limited BW of the transducers, the signal collected by the transducers is highly attenuated below 0.6 MHz and above 3.5 MHz. It can be noted that the removal of low frequencies has significant effect on the signal amplitude. It was evident that the proposed DNN was better able to predict the missing frequencies than the least square deconvolution method. In this case, originally collected PA signals are bandpass filtered, the high-frequency components are generally more attenuated.¹¹ As the proposed network is formulated as a deconvolution process, which essentially recovers the high-frequency components (equivalent of removing the blur), it can be noticed from Fig. 3 that the proposed DNN method was recovering higher frequency components close to expected signals compared to lower frequency components.

The reconstructed images using the full BW signal were shown in the second column of Fig. 2. The third column of Fig. 2 corresponds to the reconstructed images for limited BW signal. The reconstruction results for the data predicted using least square deconvolution method with $\lambda = 0.1$ (heuristic choice as reported in Ref. 9) were presented in the fourth column of Fig. 2. The last column of Fig. 2 corresponds to the data predicted using the proposed DNN. The quantitative metrics for these results were given in Table 1. It was clearly evident from Fig. 2 that the analytical backprojection using band-limited signal was insufficient in recovering the original object. Previously, the model-based algorithms as proposed in Ref. 17 were shown to recover the original object shape, but these methods were computationally expensive. It can be seen that the analytical reconstruction using enhanced BW PA signal from DNN was able to provide at least three times better quantitative accuracy compared to the state-of-the-art deconvolution method. The quantitative accuracy can be determined from the initial pressure rise values as shown in target phantoms. Note that the proposed network output in the training is noiseless full BW signal, the predicted signals using the trained network

Fig. 2 Numerical phantoms used in evaluating this work: (a) blood vessel network, (f) Derenzo phantom, and (k) PAT phantom. Reconstructed (backprojected) initial pressure images with 100 detectors using (b, g, l) full BW signal, (c, h, m) limited BW signal, (d, i, n) predicted signal from least square deconvolution method, and (e, j, o) predicted signal from the proposed DNN. The SNR of the data is at 40 dB.

Fig. 3 The PA signal time-domain and frequency response plot for the data collected by a single transducer. (a and b) Blood vessel phantom shown in Fig. 2(a), (c and d) Derenzo phantom shown in Fig. 2(f), and (e and f) PAT phantom shown in Fig. 2(k).

Table 1 The quantitative metrics, PC, and CNR of the reconstructed images presented in Fig. 2.

	Blood vessel		Derenzo		PAT	
	PC	CNR	PC	CNR	PC	CNR
Numerical experiments						
Full BW	0.85	5.74	0.92	5.03	0.96	9.89
Limited BW	0.30	1.14	0.26	0.05	0.22	0.01
Least square deconvolution	0.38	1.44	0.36	0.17	0.32	0.29
DNN (proposed)	0.78	4.22	0.70	1.92	0.75	2.54

will be a closer approximation to the expected noiseless signal. Thus, denoising step is inherently embedded in the proposed method.

As the network was trained with blood vessel type of phantoms, the quantitative accuracy obtained from the proposed method was superior for blood vessel phantom compared to the other numerical phantoms (Derenzo and PAT). Increasing the training set with a wide range of phantoms can further improve the reconstruction results. Training of the proposed

network took around 3 h, and it should be noted that the training was done only once for specific detection parameters. The training time could be further decreased (as low as 15 min) with utilization of general purpose graphics processing units. The computational time for reconstructing the initial pressure distribution using backprojection was 1.05 seconds. In addition, predicting the data using least square deconvolution method requires 80.58 ms, whereas the proposed DNN takes 49.48 ms. The proposed network was almost 1.63 times faster than the state-of-the-art least square deconvolution method.

The experimental results for the horse hair phantom data were shown in the first row of Fig. 4. The reconstructed image using the limited BW data was shown in Fig. 4(a). Figures 4(b) and 4(c) correspond to the reconstructed images for the data predicted using the least square deconvolution method with $\lambda = 0.1$ and the proposed DNN, respectively. Representative time-domain signal and frequency response plots for the experimental horse hair phantom data were given in Figs. 4(d) and 4(e) corresponding to the results shown in Figs. 4(a)–4(c). The results for the ink tube phantom were given in the second row of Fig. 4. The initial pressure distribution for the limited BW, predicted data from least square deconvolution, and the proposed DNN methods were shown in Figs. 4(f)–4(h), respectively. The corresponding representative time-domain signal and frequency response plots were shown

Fig. 4 Reconstructed (backprojected) initial pressure images for horse hair phantom and ink tube phantom using (a and f) limited BW signal, (b and g) predicted signal from least square deconvolution method, and (c and h) predicted signal from the proposed DNN. Corresponding estimated SNR is specified at the bottom of each image. The PA signal time-domain and frequency response plot for the transducer located at 9 o'clock position for (d and e) horse hair phantom and (i and j) ink tube phantom.

in Figs. 4(i) and 4(j). The proposed DNN provides enhanced time-domain response compared to its counterpart. As the initial pressure rise was unavailable for the experimental data, the SNR was computed and specified below each image. It can be seen that the proposed method was able to provide better SNR images compared to the state-of-the-art deconvolution method. The least square deconvolution method works well in the complete data case (close to 200 detectors for experiments similar to the ones conducted here),⁹ whereas only limited data case (with only 100 detectors) was considered in this work, due to which the performance is comparatively limited.

The performance of the deep learning model is typically constrained by the quality of the training data. Deep learning techniques perform better with large amounts of data. It was shown in the literature³³ that maintaining diversity among the training samples can have significant effect on the outcome of the training. For the proposed method to provide best performance, the training data distribution should include the numerical and experimental cases presented here. But, in here, even though this was not fulfilled for all cases (especially Derenzo, PAT, and experimental ink tube phantom), the model was still able to provide superior performance in terms of enhancing the BW of the PA signal (as evidenced by the figures of merit reported in Table 1 and Fig. 4).

4 Conclusion

The utilization of the band-limited signals without any preprocessing in the image reconstruction procedure results in loss of quantitative accuracy. As experimentally available transducers have only finite BW, the signals that are recorded will always be band-limited. The proposed DNN helps to predict the full BW signal from the band-limited signal and is proven to be effective compared to the state-of-the-art methods. It was also shown through the numerical and experimental phantom experiments that using the predicted signals from DNN in the reconstruction procedure (backprojection) results in better contrast recovery as well as improved PA image quality. In summary, as commercially available ultrasound detectors are BW limited, typically having only 70% BW, the proposed method can be seen as a promising preprocessing method to perform the BW enhancement without adding any significant computational burden. These methods are critical for making the PA imaging more appealing for real-time imaging studies. The

developed python code was provided as an open source to the enthusiastic users.³⁴

Disclosures

The authors have no financial interests or conflicts of interest to report.

Acknowledgments

This work was supported by the Department of Biotechnology (DBT) Innovative Young Biotechnologist Award (IYBA) (Grant No. BT/07/IYBA/2013-13) and DBT Bioengineering Grant (No. BT/PR7994/MED/32/284/2013). The authors would also like to acknowledge financial support from the Tier 1 grant funded by the Ministry of Education in Singapore (Nos. RG41/14: M4011285 and RG48/16: M4011617) and the Singapore National Research Foundation administered by the Singapore Ministry of Health National Medical Research Council (No. NMRC/OFIRG/0005/2016; M4062012).

References

1. L. H. V. Wang and S. Hu, "Photoacoustic tomography: in vivo imaging from organelles to organs," *Science* **335**(6075), 1458–1462 (2012).
2. L. Li et al., "Single-impulse panoramic photoacoustic computed tomography of small-animal whole-body dynamics at high spatiotemporal resolution," *Nat. Biomed. Eng.* **1**(5), 0071 (2017).
3. P. K. Upputuri and M. Pramanik, "Recent advances toward preclinical and clinical translation of photoacoustic tomography: a review," *J. Biomed. Opt.* **22**(4), 041006 (2017).
4. L. V. Wang and J. Yao, "A practical guide to photoacoustic tomography in the life sciences," *Nat. Methods* **13**(8), 627–638 (2016).
5. Y. Zhou, J. Yao, and L. V. Wang, "Tutorial on photoacoustic tomography," *J. Biomed. Opt.* **21**(6), 061007 (2016).
6. A. Rosenthal, V. Ntziachristos, and D. Razansky, "Acoustic inversion in optoacoustic tomography: a review," *Curr. Med. Imaging Rev.* **9**(4), 318–336 (2013).
7. M. Haltmeier and G. Zangerl, "Spatial resolution in photoacoustic tomography: effects of detector size and detector bandwidth," *Inverse Prob.* **26**(12), 125002 (2010).
8. M. Xu and L. H. V. Wang, "Analytic explanation of spatial resolution related to bandwidth and detector aperture size in thermoacoustic or photoacoustic reconstruction," *Phys. Rev. E* **67**(5), 056605 (2003).
9. N. A. Rejesh, H. Pullagurta, and M. Pramanik, "Deconvolution based deblurring of reconstructed images in photoacoustic/thermoacoustic tomography," *J. Opt. Soc. Am. A* **30**(10), 1994–2001 (2013).
10. S. M. Riad, "The deconvolution problem: an overview," *Proc. IEEE* **74**(1), 82–85 (1986).

11. M. Cao et al., "Spread spectrum photoacoustic tomography with image optimization," *IEEE Trans. Biomed. Circuits Syst.* **11**(2), 411–419 (2017).
12. L. Xu et al., "Deep convolutional neural network for image deconvolution," in *Advances in Neural Information Processing Systems*, pp. 1790–1798 (2017).
13. R. Flamary, "Astronomical image reconstruction with convolutional neural networks," arXiv:1612.04526 (2016).
14. H. Noh, S. Hong, and B. Han, "Learning deconvolution network for semantic segmentation," in *Proc. of the IEEE Int. Conf. on Computer Vision (ICCV)*, pp. 1520–1528 (2015).
15. K. Choi, G. Fazekas, and M. Sandler, "Explaining deep convolutional neural networks on music classification," arXiv:1607.02444 (2016).
16. C. B. Shaw et al., "Least squares QR-based decomposition provides an efficient way of computing optimal regularization parameter in photoacoustic tomography," *J. Biomed. Opt.* **18**(8), 080501 (2013).
17. J. Prakash et al., "Basis pursuit deconvolution for improving model-based reconstructed images in photoacoustic tomography," *Biomed. Opt. Express* **5**(5), 1363–1377 (2014).
18. E. Galetti, D. Halliday, and A. Curtis, "A simple and exact acoustic wavefield modeling code for data processing, imaging, and interferometry applications," *Geophysics* **78**, F17 (2013).
19. R. Girshick et al., "Rich feature hierarchies for accurate object detection and semantic segmentation," in *Proc. of the IEEE Conf. on Computer Vision and Pattern Recognition*, pp. 580–587 (2014).
20. A. Krizhevsky, I. Sutskever, and G. E. Hinton, "ImageNet classification with deep convolutional neural networks," in *Advances in Neural Information Processing Systems*, pp. 1097–1105 (2012).
21. G. Wang, "A perspective on deep imaging," *IEEE Access* **4**, 8914–8924 (2016).
22. V. Badrinarayanan, A. Kendall, and R. Cipolla, "SegNet: a deep convolutional encoder-decoder architecture for image segmentation," arXiv:1511.00561 (2015).
23. M. D. Zeiler et al., "On rectified linear units for speech processing," in *IEEE Int. Conf. on Acoustics, Speech and Signal Processing (ICASSP)*, pp. 3517–3521 (2013).
24. D. Kingma and J. Ba, "Adam: a method for stochastic optimization," arXiv:1412.6980 (2014).
25. J. Duchi, E. Hazan, and Y. Singer, "Adaptive subgradient methods for online learning and stochastic optimization," *J. Mach. Learn. Res.* **12**, 2121–2159 (2011).
26. T. Tieleman and G. Hinton, "Lecture 6.5-RMSProp, COURSE: neural networks for machine learning," University of Toronto, Technical Report (2012).
27. S. K. Kalva and M. Pramanik, "Experimental validation of tangential resolution improvement in photoacoustic tomography using modified delay-and-sum reconstruction algorithm," *J. Biomed. Opt.* **21**(8), 086011 (2016).
28. American National Standard Institute, *American National Standard for Safe Use of Lasers*, ANSI Z136.1-2000, Laser Institute of America, Orlando, Florida (2000).
29. Y. Lou et al., "Generation of anatomically realistic numerical phantoms for optoacoustic breast imaging," *Proc. SPIE* **9708**, 970840 (2016).
30. B. E. Treeby and B. T. Cox, "k-Wave: MATLAB toolbox for the simulation and reconstruction of photoacoustic wave fields," *J. Biomed. Opt.* **15**(2), 021314 (2010).
31. F. Chollet, "Keras," <https://keras.io/> (2015).
32. Theano Development Team, "Theano: a python framework for fast computation of mathematical expressions," <http://arxiv.org/abs/1605.02688> (2016).
33. L. Jiang et al., "Self-paced learning with diversity," in *Advances in Neural Information Processing Systems*, pp. 2078–2086 (2014).
34. S. Gutta et al., "Deep neural network-based bandwidth enhancement of photoacoustic data," <https://sites.google.com/site/sercmig/home/dnnbwe> (25 July 2017).

Sreedevi Gutta received her bachelor's degree from the Prasad V Potluri Siddhartha Institute of Technology, Andhra Pradesh, India, in 2014. Currently, she is a graduate student at the Department of Computational and Data Sciences, Indian Institute of Science, Bangalore, India. Her research interests include computational methods in photoacoustic imaging and inverse problems in medical imaging.

Venkata Suryanarayana Kadimesetty received his master's degree from Indian Institute of Science, Bangalore in 2009. Currently, he is a graduate student at the Department of Computational and Data Sciences, Indian Institute of Science, Bangalore, India. His research interests include medical image analysis and reconstruction.

Sandeep Kumar Kalva received his master's degree in biomedical engineering from the Indian Institute of Technology (IIT), Hyderabad, India, in 2015. He is currently a PhD student in the School of Chemical and Biomedical Engineering, Nanyang Technological University, Singapore. His research area is on functional photoacoustic imaging for various clinical applications.

Manojit Pramanik received his PhD in biomedical engineering from Washington University in St. Louis, Missouri, USA. Currently, he is an assistant professor of the School of Chemical and Biomedical Engineering, Nanyang Technological University, Singapore. His research interests include the development of photoacoustic/thermoacoustic imaging systems, image reconstruction methods, clinical application areas such as breast cancer imaging, molecular imaging, contrast agent development, and Monte Carlo simulation of light propagation in biological tissue.

Sriram Ganapathy received his master's degree in engineering from the Indian Institute of Science, Bangalore, India in 2006, and his PhD from the Center for Language and Speech Processing, Johns Hopkins University in 2011. He is an assistant professor in the Department of Electrical Engineering, Indian Institute of Science, Bangalore. His research interests are in signal processing, machine learning, deep learning and neuroscience with applications to robust speech recognition, speech enhancement, speech coding and audio analytics including biometrics.

Phaneendra K. Yalavarthy received his MSc degree in engineering from the Indian Institute of Science, Bangalore, India, and his PhD in biomedical computation from Dartmouth College, Hanover, New Hampshire, in 2007. He is an associate professor in the Department of Computational and Data Sciences, Indian Institute of Science, Bangalore. His research interests include medical image computing, medical image analysis, and biomedical optics.