Journal of Biomedical Optics

BiomedicalOptics.SPIEDigitalLibrary.org

Optimal wavelengths for subdiffuse scanning laser oximetry of the human retina

Mathi Damodaran
Arjen Amelink
Johannes F. de Boer

SPIE.

Mathi Damodaran, Arjen Amelink, Johannes F. de Boer, "Optimal wavelengths for subdiffuse scanning laser oximetry of the human retina," *J. Biomed. Opt.* **23**(8), 086003 (2018), doi: 10.1117/1.JBO.23.8.086003.

Optimal wavelengths for subdiffuse scanning laser oximetry of the human retina

Mathi Damodaran,^{a,*} Arjen Amelink,^b and Johannes F. de Boer^a

^aVrije Universiteit Amsterdam, LaserLaB, Department of Physics and Astronomy, Amsterdam, The Netherlands

^bNetherlands Organisation for Applied Scientific Research TNO, Department of Optics, Delft, The Netherlands

Abstract. Retinal blood vessel oxygenation is considered to be an important marker for numerous eye diseases. Oxygenation is typically assessed by imaging the retinal vessels at different wavelengths using multispectral imaging techniques, where the choice of wavelengths will affect the achievable measurement accuracy. Here, we present a detailed analysis of the error propagation of measurement noise in retinal oximetry, to identify optimal wavelengths that will yield the lowest uncertainty in saturation estimation for a given measurement noise level. In our analysis, we also investigate the effect of hemoglobin packing in discrete blood vessels (pigment packaging), which may result in a nonnegligible bias in saturation estimation if unaccounted for under specific geometrical conditions, such as subdiffuse sampling of smaller blood vessels located deeper within the retina. Our analyses show that using 470, 506, and 592 nm, a fairly accurate estimation of the whole oxygen saturation regime [0 1] can be realized, even in the presence of the pigment packing effect. To validate the analysis, we developed a scanning laser ophthalmoscope to produce high contrast images with a maximum pixel rate of 60 kHz and a maximum 30-deg imaging field of view. Confocal reflectance measurements were then conducted on a tissue-mimicking scattering phantom with optical properties similar to retinal tissue including narrow channels filled with absorbing dyes to mimic blood vessels. By imaging at three optimal wavelengths, the saturation of the dye combination was calculated. The experimental values show good agreement with our theoretical derivations. © The Authors. Published by SPIE under a Creative Commons Attribution 3.0 Unported License. Distribution or reproduction of this work in whole or in part requires full attribution of the original publication, including its DOI. [DOI: [10.1117/1.JBO.23.8.086003](https://doi.org/10.1117/1.JBO.23.8.086003)]

Keywords: retinal oximetry; scanning laser ophthalmoscope; hemoglobin; oxygen saturation; phantoms; blood vessel diameter.

Paper 180340R received Jun. 7, 2018; accepted for publication Aug. 2, 2018; published online Aug. 27, 2018.

1 Introduction

Hemoglobin oxygen saturation in the blood flowing through the retinal vasculature is an important physiological parameter involved in the pathophysiology of numerous retinal diseases including diabetic retinopathy,^{1,2} glaucoma,^{3–5} retinitis pigmentosa,^{6–8} age-related macular degeneration,⁹ and retinal vessel occlusions.¹⁰ Several retinal diseases result in reduced oxygen circulation in the retina causing hypoxia,¹¹ which is one of the key drivers of angiogenesis. As such, measuring the retinal blood oxygen saturation will enable monitoring of the development of hypoxia and thus will allow preventing the loss of retinal tissue due to vasoproliferation through timely therapeutic interventions. Furthermore, although damage occurring to the retina in the advanced stages of many retinal diseases associated with hypoxia has been established, oxygenation measurements could also aid in understanding the onset or presence of systemic diseases. For example, a recent study reported that elevated oxygen saturation in retinal blood vessels is an indicator of Alzheimer's disease.^{12,13}

The earliest attempts to demonstrate retinal vessel oximetry date back several decades.^{14,15} Various imaging modalities, such as fundus camera,^{16–23} hyperspectral imaging,^{24–26} scanning laser ophthalmoscope (SLO),^{27,28} adaptive optics SLO,²⁹ and optical coherence tomography (OCT)^{30,31} have since then been employed to estimate oxygen saturation in retinal blood vessels. These noninvasive techniques all involve recording

light reflected from the fundus and measuring the amount of light absorbed by the blood flowing through the retinal vessels at multiple wavelengths. Retinal oxygen saturation is typically calculated from reflection intensities recorded by a point or array detector, and the error on the measured intensities propagates to a saturation error. Using a detailed error analysis, we aim to identify optimum wavelengths that will yield minimum uncertainty on the saturation estimation for a given measurement error in the intensities.

One factor that has not been considered in retinal oximetry to date is the fact that hemoglobin is not distributed homogeneously in the tissue but confined to discrete blood vessels. This effect is particularly relevant for smaller blood vessels that occupy only a small portion of the total probed volume in a quasicontinuous/subdiffuse scheme. As a result, a fraction of the incident photons are reflected from the tissue without ever encountering a blood vessel, causing the apparent absorption coefficient of blood packed in small vessels to be different compared with when the blood would have been homogeneously distributed through the probed tissue. Such an effect is called the pigment packaging effect, and several research studies have experimentally verified this effect.^{32–34} We have included this effect in retinal oximetry algorithms to increase the accuracy of the saturation estimation, in particular for small blood vessels located deeper in the retina, probed using a (sub)diffuse measurement technique. Large and superficially located blood vessels, whose diameters are comparable with the detection aperture, are not expected to suffer from this effect as the probability that the detected photons have not traveled through the

*Address all correspondence to: Mathi Damodaran, E-mail: m.damodaran@vu.nl

Fig. 1 Two different vessel geometries in subdiffuse retinal oximetry. (a) A large blood vessel with a diameter comparable to the collection area present superficially in the retina. In this case, the probability that a detected photon has not passed through the blood vessel is negligible, and hence, the pigment packing effect is not observed. (b) A small blood vessel embedded in the scattering medium (retinal tissue) is illuminated with a narrow beam, whereas the collection aperture is much larger than the illumination aperture and the vessel diameter. In this case, the blood vessel occupies only a small portion of the total probed volume and hence the pigment packing effect should be considered.

blood vessel is negligible. In this paper, we present modified retinal oximetry equations that include the pigment packaging effect and evaluate the effect of pigment packaging on saturation estimation. We then identify optimum wavelengths for retinal oximetry using a detailed error analysis of the oximetry equations including the pigment packaging effect and identify the wavelengths that give minimum saturation error for both discussed geometries [Figs. 1(a) and 1(b)] and minimize the saturation offset in the presence of the pigment packing effect.

To validate our theoretical predictions, we have fabricated a thin, multilayer retina-mimicking phantom with narrow channels representing the blood vessels in the retina. The reduced scattering coefficient (μ'_s) of the scattering medium in the layers was chosen to be similar to the expected values for the retina. The retina-mimicking phantom was placed in a model eye. An SLO using a supercontinuum laser as the light source was used to scan the artificial retina. Two dyes of suitable concentrations were tailored to have absorption coefficients comparable with oxy- and deoxyhemoglobin with an artificial isosbestic point at 548 nm. The procedure for selecting optimum wavelengths to estimate the saturation within this artificial vessel was applied to this phantom and three optimal wavelengths (488, 548, and 612 nm) were selected based on this analysis. Using the resulting wavelengths, the experiments were conducted on the phantom. There was a good agreement between the predicted and real values of saturation of the different dye combinations using these three optimal wavelengths.

2 Theory of Retinal Oximetry and Identifying Optimum Wavelengths

2.1 Theory of Retinal Oximetry

Figure 2(a) shows the absorption spectrum of the oxy- and deoxyhemoglobin having various isosbestic points, i.e., wavelengths where the absorption of oxy and deoxyhemoglobin is equal. Figure 2(b) shows a typical fundus image of a healthy

human eye obtained with 570-nm illumination. A general method for measuring the absorbance of blood at a particular blood vessel in the retina is shown in the inset of Fig. 2(b), where the intensity profile of the blood vessel [dotted line in Fig. 2(b)] is plotted as a function of the location on the retina. The scattering properties of blood have been reported in Bosschaart et al.,³⁵ where it was shown that although blood has a high scattering coefficient, it also has a high scattering anisotropy, resulting in a reduced scattering coefficient of ≈ 1 to 2 mm^{-1} . Further, the reduced scattering coefficient of blood is close to the reduced scattering coefficient of the surrounding tissue³⁶ and about an order of magnitude smaller than the absorption coefficient of blood in the 450- to 650-nm wavelength range.

In a retinal imaging experiment, the measured intensity $I(x_o, \lambda)$ (see Fig. 1) from a spatial location x_o in the retina for a given illumination wavelength λ depends on the incident illumination intensity distribution $I_i(x_i, \lambda)$ at the point x_i , and the backscattered reflectance function of the retina $R(x_o, x_i, \lambda)$. We consider a quasiconfocal or subdiffuse scheme, where a narrow, focused beam illuminates a point x_i in the retina and the detection pinhole are collected from a larger volume around a point x_o . Let $\bar{x}_o$ be the location of an averaged volume of tissue surrounding x_o that contributes to the collected intensity, $I(x_o, \lambda)$. The extent of volume averaging for any location concerning the spatial distribution of the optical properties at that location is determined by the incident illumination profile $I_i(x_i, \lambda)$ and the detection pinhole. Under these conditions, we write the measured intensity I as

$$I(x_o, \lambda) = \int I_i(x_i, \lambda) \cdot R(x_o, x_i, \lambda) \cdot dx_i = I_i(x_o, \lambda) \cdot R(\bar{x}_o, \lambda). \quad (1)$$

The function $R(\bar{x}_o, \lambda)$ describes the reflectivity of the tissue averaged over a small tissue volume at the location x_o due to scattering and absorption. Let x_b and x_t denote the recorded

Fig. 2 (a) Absorption spectrum of oxy- and deoxyhemoglobin from 350 to 1000 nm assuming a concentration of 150 mg of hemoglobin in 1 mL of blood.³³ The inset shows the absorption spectrum from 450 to 650 nm. The absorption of these two types of hemoglobins is different for most wavelengths except for the isosbestic points (denoted by black dashed circles: 506, 522, 548, 569, and 586 nm in the inset) where the absorption depends on factors other than the oxygen saturation. (b) Fundus reflectance of the right eye of a healthy adult male at 570 nm recorded using a Fundus camera (Oxymap T1, Oxymap ehf., Iceland, fitted to a Topcon fundus camera, Topcon TRC 50-DX, Topcon corp., Japan). The inset shows a typical vessel absorption profile across a blood vessel (dotted line) in the fundus photo demonstrating a drop in the detected intensity at the blood vessel location (x_b) compared with the intensity at the tissue location (x_t).

locations at the center of a blood vessel and in the adjacent tissue, respectively. Then, the light collected from a tissue location x_t is given as

$$I(x_t, \lambda) \propto I_i(x_t, \lambda) \cdot R(\bar{x}_t, \lambda) = I_i(x_t, \lambda) \cdot R[\mu_s(\bar{x}_t, \lambda), p(\bar{x}_t, \theta, \lambda)], \quad (2)$$

where we have stated that $R(\bar{x}_t, \lambda)$ is a function of the average scattering coefficient $[\mu_s(\bar{x}_t, \lambda)]$ of the volume of tissue around x_t (denoted by $\bar{x}_t$) at the wavelength λ and of the scattering phase function $[p(\bar{x}_t, \theta, \lambda)]$ of the same tissue volume. We have assumed that there is no blood present in tissue volume $\bar{x}_t$ and that any other absorbing molecules within the tissue volume $\bar{x}_t$ in the neural retina have a negligible contribution to the reflected intensities. Similarly, the light collected from the center of a blood vessel location x_b is given as

$$I(x_b, \lambda) \propto I_i(x_b, \lambda) \cdot R(\bar{x}_b, \lambda) = I_i(x_b, \lambda) \cdot R[\mu_s(\bar{x}_b, \lambda), p(\bar{x}_b, \theta, \lambda), \mu_a(\bar{x}_b, \lambda)], \quad (3)$$

where we have stated that $R(\bar{x}_b, \lambda)$ is a function of the average scattering coefficient $\mu_s(\bar{x}_b, \lambda)$ of the volume of tissue around x_b (denoted by $\bar{x}_b$) at the wavelength λ , of the scattering phase function $p(\bar{x}_b, \theta, \lambda)$ of the same tissue volume, and the effective absorption coefficient of the tissue volume containing the blood vessel $\mu_a(\bar{x}_b, \lambda)$. We can now write the relative optical density OD_λ of the tissue at the blood vessel location compared with the surrounding tissue at a particular wavelength as the logarithm of the ratio of tissue and blood vessel reflectivity at the same wavelength, i.e.,

$$\begin{aligned} OD_\lambda &= \ln \left[\frac{I(x_t, \lambda)}{I(x_b, \lambda)} \right] \\ &= \ln \left[\frac{R(\bar{x}_t, \lambda)}{R(\bar{x}_b, \lambda)} \right] = \langle L_{\text{eff}}(\lambda) \rangle \cdot \mu_a(\bar{x}_b, \lambda) + G(\lambda). \end{aligned} \quad (4)$$

Here, we have assumed that the incident light intensity was the same at both locations and that the attenuation of the reflected light due to the blood within $\bar{x}_b$ is governed by modified Beer-Lamberts law,^{37,38} with $\langle L_{\text{eff}}(\lambda) \rangle$, the effective path length of photons traveling through the probed volume before reaching the detector. The effective path length $\langle L_{\text{eff}}(\lambda) \rangle$ depends on the absorption and scattering properties of the probed volume. $G(\lambda)$ is a factor that accounts for any apparent increase or decrease in the ODs purely due to scattering differences within the volumes $\bar{x}_b$ and $\bar{x}_t$ i.e., due to the difference in scattering properties of blood within the blood vessel compared with the surrounding tissue. As in the wavelength range 450 to 650 nm, the reduced scattering coefficient of blood is similar to that of the surrounding tissue, we do not expect G to be a significant factor, in particular for the case shown in Fig. 1(b), where the blood vessel occupies only a small fraction of the interrogated tissue volume. Equation (4) assumes that all additional reflections occurring internally in the system due to optics and the stray reflections from the cornea are accounted for, e.g., by subtracting a reference measurement. The absorption coefficient $\mu_a(\bar{x}_b, \lambda)$ of the blood is a function of saturation (S) of the blood (i.e., the fraction of the oxygenated hemoglobin concentration to the total concentration of hemoglobin) and is given as

$$\mu_a(\bar{x}_b, \lambda) = [S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda)]. \quad (5)$$

The parameters $\mu_a^{\text{HbO}_2}(\lambda)$ and $\mu_a^{\text{Hb}}(\lambda)$ are the absorption coefficients of oxy- and deoxyhemoglobin, respectively (assuming a concentration of 150 mg of hemoglobin in 1 mL of blood) and are well-known functions of wavelength. Combining Eqs. (4) and (5)

$$\text{OD}_\lambda = \langle L_{\text{eff}}(\lambda) \rangle \cdot \nu \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda)] + G(\lambda), \quad (6)$$

where ν is the fraction of interrogated tissue occupied by the blood vessel.^{33,34} We define the ratio $\rho_{\lambda_1|\lambda_2}$ of ODs at two wavelengths λ_1 and λ_2 as

$$\begin{aligned} \rho_{\lambda_1|\lambda_2} &= \frac{\text{OD}_{\lambda_1}}{\text{OD}_{\lambda_2}} \\ &= \frac{\langle L_{\text{eff}}(\lambda_1) \rangle \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda_1) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda_1)] + G(\lambda_1)}{\langle L_{\text{eff}}(\lambda_2) \rangle \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda_2) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda_2)] + G(\lambda_2)}. \end{aligned} \quad (7)$$

If the scattering and absorption coefficients at the two wavelengths are close to each other, $\langle L_{\text{eff}}(\lambda_1) \rangle \simeq \langle L_{\text{eff}}(\lambda_2) \rangle$ and $G(\lambda_1) \simeq G(\lambda_2)$, we will later discuss the validity of this assumption. Under this condition, we can write S as a function of $\rho_{\lambda_1|\lambda_2}$ as

$$S_{\lambda_1|\lambda_2} = \frac{\rho \cdot \mu_a^{\text{Hb}}(\lambda_2) - \mu_a^{\text{Hb}}(\lambda_1)}{[\mu_a^{\text{Hb}}(\lambda_2) - \mu_a^{\text{HbO}_2}(\lambda_2)] \cdot \rho + [\mu_a^{\text{HbO}_2}(\lambda_1) - \mu_a^{\text{Hb}}(\lambda_1)]} + \alpha \cdot G, \quad (8)$$

where $\alpha \cdot G$ ($\alpha = \frac{\rho-1}{\langle L_{\text{eff}} \rangle \cdot [\mu_a^{\text{HbO}_2}(\lambda_2) - \mu_a^{\text{Hb}}(\lambda_2)] \cdot \rho + [\mu_a^{\text{HbO}_2}(\lambda_1) - \mu_a^{\text{Hb}}(\lambda_1)]}$) is an offset, if any, to the saturation due to scattering. As discussed earlier, for visible wavelengths we expect $G(\lambda) \rightarrow 0$, and in the remainder of the analyses, we only focus on the first term in the right-hand side of Eq. (8).

Although not mathematically necessary, it is advantageous to choose one of the wavelengths as an isosbestic wavelength to simplify the calculations, especially for error propagation purposes. If λ_2 is taken to be the isosbestic wavelength, $\lambda_2 = \lambda_i$, $\mu_a^{\text{HbO}_2} = \mu_a^{\text{Hb}} = \mu_a^i$. In that case, Eq. (8) can be written as

$$S_{\lambda_a|\lambda_i} = \frac{\rho \cdot \mu_a^i(\lambda_i) - \mu_a^{\text{Hb}}(\lambda_a)}{\mu_a^{\text{HbO}_2}(\lambda_a) - \mu_a^{\text{Hb}}(\lambda_a)}, \quad (9)$$

where we have denoted the saturation sensitive “anisobestic” wavelength as λ_a . We will now first consider the geometry as shown in Fig. 1(a), where the blood vessel especially a large, superficial blood vessel occupies the majority of the probed volume.

2.1.1 Optimum wavelengths without pigment packing

If there is a large blood vessel with a diameter close to or greater than the size of the collection aperture directly below the illumination point, as shown in Fig. 1(a), the assumption that the blood vessel occupies only a portion of the probed volume does not hold. In such geometries, the blood vessel occupies majority of the probed volume. Monte Carlo simulations^{39,40} have suggested that in this type of geometry with a quasiconfocal aperture, the photons reaching the detector have predominantly been backscattered from within the blood vessel. In this case,

$\langle L_{\text{eff}}(\lambda) \rangle$ in Eq. (6) becomes the average backscatter path length $\langle L_{\text{bs}}(\lambda) \rangle$ from within the blood vessel (and blood volume fraction ν becomes 1)

$$\text{OD}_\lambda = \langle L_{\text{bs}}(\lambda) \rangle \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda)] + G(\lambda). \quad (10)$$

Under the condition that scattering coefficient and absorption coefficient of blood is approximately equal at both wavelengths λ_a and λ_i , $\langle L_{\text{bs}}(\lambda_a) \rangle \approx \langle L_{\text{bs}}(\lambda_i) \rangle$. The saturation in this case is given by Eq. (9).

Turning now to the error on the saturation, this error is given as

$$\Delta S = \frac{dS}{d\rho} \cdot \Delta \rho, \quad (11)$$

where $\Delta \rho_{\lambda_a|\lambda_i}$ is given by propagating the error on OD_a and OD_i in Eq. (7)

$$\Delta \rho_{\lambda_a|\lambda_i} = \frac{\text{OD}_a}{\text{OD}_i} \cdot \sqrt{\left(\frac{\Delta \text{OD}_a}{\text{OD}_a}\right)^2 + \left(\frac{\Delta \text{OD}_i}{\text{OD}_i}\right)^2}. \quad (12)$$

The error in the ODs is given by the underlying error of the reflected intensities $I(x_t, \lambda)$ and $I(x_b, \lambda)$ as

$$\Delta \text{OD}_\lambda = \sqrt{\left[\frac{\Delta I(x_t, \lambda)}{I(x_t, \lambda)}\right]^2 + \left[\frac{\Delta I(x_b, \lambda)}{I(x_b, \lambda)}\right]^2}. \quad (13)$$

Due to very low absorption at wavelengths >650 nm, there is a poor contrast between the recorded intensity at the tissue location and the blood vessel location that results in $\text{OD} \rightarrow 0$. This results in large errors in ρ [following Eq. (12)] and therefore, in saturation estimation. Absorption is very high in the Soret band (430 nm) resulting in most of the light being absorbed by the blood resulting in a very poor signal-to-noise ratio (SNR) at the blood vessel locations. Most importantly, the blue light hazard for the wavelengths <450 nm substantially limits the maximum permissible exposure of the retina for light⁴¹ of these wavelengths. Thus, we exclude the wavelengths <450 nm and <650 nm for the remainder of the analysis. From Eqs. (11)–(13), it is evident that the standard error of the mean recorded intensities by the detector ($\Delta I/I$ or standard deviation/mean) propagates to an error in the ODs and then to an error in ρ , and thus contributes to the saturation error ΔS . For example, a 1.0% error in the intensities ($\Delta I/I = 0.01$) propagates to an error in the ODs ($\Delta \text{OD} = 0.014$). If $\text{OD}_{\lambda_1} = 0.5$ and $\text{OD}_{\lambda_2} = 0.7$, this gives 3.4% error in ρ according to Eq. (12), this error in ρ propagates to saturation error in a complex manner ΔS through Eq. (11).

Figure 3 shows the calculated saturation error for different oxygen saturation levels in the retina when 506 nm was used as the isosbestic wavelength and the anisobestic wavelength varied over the 450- to 650-nm ranges, assuming a 1% measurement error in the intensity. For low saturation values [0.0 and 0.5], choosing the anisobestic wavelength in the ranges 458 to 474 nm gives a saturation error $<5\%$, while for high saturation values [0.5 and 1.0], choosing the anisobestic wavelength in the ranges 591 to 599 nm gives a saturation error $<5\%$. A similar trend was found for other isosbestic points shown in Fig. 2(a). Note, however, that the condition that the absorption coefficients

Fig. 3 Saturation error ΔS as a function of λ_a when estimating different oxygenation levels from 0% to 100%, with 1% error on the recorded intensities. A wavelength of 506 nm was chosen as the isosbestic wavelength. This graph holds for geometry described in Fig. 1(a), where the photons are mostly backscattered from a blood vessel.

are similar at $\mu_a(\lambda_a) \approx \mu_a(\lambda_i)$, so that $\langle L_{bs}(\lambda_a) \rangle \approx \langle L_{bs}(\lambda_i) \rangle$ is met best by choosing 506 nm as the isosbestic wavelength [Fig. 2(a)].

2.1.2 Optimum wavelengths including pigment packing

Figure 1(b) shows a geometry, where a small blood vessel is embedded in the scattering retina. In this case, the blood vessel occupies only a fraction of the total probed volume by the imager. As a result, Eq. (6) (which assumes a homogeneous distribution of hemoglobin throughout the probed tissue volume) does not hold as the absorbing hemoglobin molecules are not homogeneously distributed but are contained in a discrete package (the blood vessel) that occupies only part of the probed tissue volume. This packaging of absorbing molecules effectively flattens the apparent absorption spectrum for high absorption coefficients, and is called the pigment packaging effect. It has previously been shown that this effect can be modeled through the incorporation of a correction factor to the absorption spectrum.^{32–34} Equation (5) in this case is modified to

$$\mu_a(\bar{x}_b, \lambda) = \nu \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda)] \cdot C_\lambda, \quad (14)$$

where the correction factor C_λ ^{34,42} is given as

$$C_\lambda(S, \lambda, D_{\text{ves}}) = \frac{1 - \exp\{-D_{\text{ves}} \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda)]\}}{D_{\text{ves}} \cdot [S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot \mu_a^{\text{Hb}}(\lambda)]}, \quad (15)$$

where D_{ves} is the blood vessel diameter. The OD can now be written as

$$\text{OD}_\lambda = \frac{\langle L_{\text{eff}}(\lambda) \rangle \cdot \nu}{D_{\text{ves}}} \cdot \{S \cdot \mu_a^{\text{HbO}_2}(\lambda) + (1 - S) \cdot [\mu_a^{\text{Hb}}(\lambda)]\} \cdot C_\lambda + G(\lambda). \quad (16)$$

The saturation S can be written as [ignoring the term $\alpha \cdot G$ ($\alpha = \frac{\rho-1}{\langle L_{\text{eff}} \rangle \cdot \nu \cdot \{\mu_a^{\text{HbO}_2}(\lambda_a) - \mu_a^{\text{Hb}}(\lambda_a)\}} \cdot C_{\lambda_a}$) as was done for Eq. (9)]

$$S_{\lambda_a|\lambda_i} = \frac{(C_{\lambda_i}/C_{\lambda_a}) \cdot \rho \cdot \mu_a^i(\lambda_i) - \mu_a^{\text{Hb}}(\lambda_a)}{\mu_a^{\text{HbO}_2}(\lambda_a) - \mu_a^{\text{Hb}}(\lambda_a)}, \quad (17)$$

where S is now also a function of the blood vessel diameter D_{ves} , due to the addition of the diameter-dependent correction factor.

A detailed error analysis of Eq. (13) is necessary to find the optimal wavelengths. The error propagation equations gives the error in saturation ΔS as

$$\Delta S = \sqrt{\left(\frac{\partial S}{\partial \rho}\right)^2 \cdot (\Delta \rho)^2 + \left(\frac{\partial S}{\partial D_{\text{ves}}}\right)^2 \cdot (\Delta D_{\text{ves}})^2}. \quad (18)$$

The error ΔS is a function of S , D_{ves} , and λ_a for a chosen λ_i . Now we proceed to analyze each term in Eq. (18) and its contribution to the saturation error ΔS . For an image-based estimation of the blood vessel diameter D_{ves} (explained in Sec. 3.3) using an SLO technique, we chose a constant error on the estimation of the diameter (ΔD_{ves}) as 12 μm based on spatial resolution of the system (see Table 1). The error on $\rho_{\lambda_a|\lambda_i}$ is given by error propagation in Eq. (12).

Figures 4 and 5 show the color-coded graph of the calculated errors in saturation as a function of D_{ves} and λ_a considering a 1% error ($\Delta I/I$) in the intensities when 506 nm (lowest μ_a^i in the 450- to 650-nm interval) and 548 nm (highest μ_a^i in the 450- to 650-nm interval) were used as isosbestic points, respectively. For simplicity, all the saturation error values (ΔS) > 0.10 were set to 0.10 to emphasize the combinations with accuracy better than 0.1. From Fig. 4, we see that when 506 nm is used as an isosbestic wavelength, there are two spectral bands optimal for oximetry. For low saturation values (0.00 to 0.25), either of the bands, 460 to 480 nm or 590 to 600 nm could be used. But for higher saturation levels (0.75 and 1.00), only the 589 to 600 nm gives $\Delta S < 0.1$. When 548 nm is used as an isosbestic wavelength (Fig. 5), 589 to 600 nm is not ideal for detecting the low saturation values and 460 to 480 nm is not suitable for high saturation values.

Figure 6 shows the saturation error for a 50- μm [Figs. 6(a)–6(c)] and a 100- μm [Figs. 6(d)–6(f)] blood vessel for different saturation levels. From these graphs, we see that that all the five isosbestic wavelengths perform similarly for estimating fully oxygenated blood ($S = 1.00$) when the 590- to 600-nm wavelength range is used as an anisosbestic wavelength. However, for estimating low saturation values, 522, 586, and 506 nm perform marginally better than 548 and 569 nm when the 460- to 480-nm wavelength range is used as an anisosbestic wavelength.

Figure 7 shows ΔS values as a function of λ_a ($\lambda_i = 506$ nm) for a fully oxygenated 50- and 100- μm blood vessel for different

Table 1 System parameters of the SLO.

System parameter	Value
Spot size in the retinal plane (at 548 nm)	8.9 μm (Airy disc diameter)
Maximum field of view (FOV)	30 deg in air
Pixel rate	30 KHz
Digital resolution	11.7 μm

Fig. 4 Saturation error ΔS for different λ_a when 506 nm is chosen as the isosbestic wavelength. The standard error on the mean recorded intensity values ($\frac{\Delta I}{I}$) was assumed to be 1%. All values of $\Delta S > 0.10$ were made as 0.10.

Fig. 5 Saturation error ΔS for different λ_a when 548 nm is chosen as the isosbestic wavelength. The standard error on the mean recorded intensity values ($\frac{\Delta I}{I}$) was assumed to be 1%. All values of $\Delta S > 0.10$ were made as 0.10.

levels of the standard error of the mean ($\Delta I/I$). It can be seen in this figure that to reach saturation accuracy levels better than 5% for 100% oxygenation when measuring near 594 nm, the required accuracy level of the mean intensity was 1% for both the 50- and 100- μm vessels.

We note that the optimal wavelength bands calculated in the presence, as well as in the absence of pigment packaging, are very similar. In fact, it is observed that if the ratio of correction factors $C_{\lambda_i}/C_{\lambda_a}$ in Eq. (17) is 1, Eq. (17) reduces to Eq. (9). Thus, choosing wavelengths within the [460 and 480] and

Fig. 6 Saturation error ΔS as a function of λ_a for different saturation levels. [(a and d): $S = 0.00$; (b and e): $S = 0.50$; (c) $S = 1.00$] for a $50\text{ }\mu\text{m}$ (a, b, and c) blood vessel and a $100\text{ }\mu\text{m}$ (d, e, and f) blood vessel when different isosbestic wavelengths in the 500- to 600-nm range were chosen as the λ_i .

Fig. 7 Saturation error ΔS as a function of λ_a when estimating full oxygenation ($S = 1.00$) for a $50\text{-}\mu\text{m}$ blood vessel (a) and $100\text{-}\mu\text{m}$ blood vessel (b) for different levels of % error in the intensities. The isosbestic point chosen was 506 nm.

[590 and 600] bands, where $C_{506}/C_{\lambda_a} \approx 1$ would serve the purpose of minimizing the saturation error independent of the effect of pigment packing. For such wavelength combinations, Eq. (9) can be used even in the presence of pigment packing without applying any correction.

Figures 8(a) and 8(d) show the difference in saturation estimated with [Eq. (17)] and without [Eq. (9)] pigment packing effect for a $50\text{-}\mu\text{m}$ blood vessel displayed as an offset ($S_{\text{offset}} = S_{\text{Eq.17}} - S_{\text{Eq.9}}$), at an isosbestic wavelength of $\lambda_i = 506\text{ nm}$.

The dark blue region in the color map in Figs. 8(a) and 8(d) denotes the situation when $C_{506}/C_{\lambda_a} \approx 1$, i.e., when pigment packing has no effect. Unfortunately, this condition is satisfied for different wavelengths in the [460 and 480] and [590 and 600] wavelength bands for different saturation levels. Figures 8(b)

and 8(e) show the saturation error [Eq. (11)] in the [460 and 480] and [590 and 600] wavelength bands, respectively, assuming a 1% measurement error of intensities and when 506 nm was used as an isosbestic wavelength. Figures 8(c) and 8(f) show the combined effect of saturation error from a 1% measurement error and ignoring pigment packing, indicating that using the 470- to 506-nm pair, the combined error is minimized for saturation in the range [0.00 and 0.55], and using the 592- and 506-nm pair, the combined error is minimized for saturation in the ranges [0.40 and 1.00]. Thus, the triad of wavelengths 470, 506, and 592 nm appear to be optimal for retinal oximetry. At these wavelengths, (i) the blood absorption coefficients are approximately equal, and as a consequence, $C_{506}/C_{\lambda_a} \approx 1$, which minimizes the effect of pigment packaging on saturation

Fig. 8 (a and d) Calculated saturation difference with and without pigment packing effect shown as an offset ($S_{\text{offset}} = S_{\text{Eq.17}} - S_{\text{Eq.9}}$). (b and e) Saturation error calculated from Eq. (11) assuming a 1% measurement error of intensities. (c and f) Addition of the saturation offset and the saturation error for wavelengths [460 and 480] and [590 and 600], respectively.

Fig. 9 Schematic of the multiwavelength SLO used for imaging: Supercontinuum (SC) laser source (NKT Photonics A/S, Denmark) was used as the light source. An acoustic-optic tunable filter, (AOTF) (Select, NKT Photonics A/S, Denmark) coupled to the supercontinuum source provided a spectrally tunable illumination source for the multiple wavelength measurements. The polychromatic light beam from the AOTF was coupled into a single-mode fiber and was relayed to a model eye. The model eye was constructed with an uncoated lens L4 ($f = 15.4$ mm) having a curvature very similar to the cornea of the human eye. The sample was placed in the focus of L4. The light was reflected and backscattered by the sample and 90% of the reflected light passed through the beam splitter in the return path and split using dichroic mirrors into multimode fibers with a core diameter of $100 \mu\text{m}$.

estimation. Similar absorption and scattering coefficients at these wavelengths satisfy the assumption that $\langle L_{\text{bs}}(\lambda_i) \rangle \simeq \langle L_{\text{bs}}(\lambda_a) \rangle$ and $\langle L_{\text{eff}}(\lambda_a) \rangle \simeq \langle L_{\text{eff}}(\lambda_i) \rangle$, which are conditions for our algorithms to hold, (ii) the blood absorption coefficients are sufficiently high to yield a good vascular contrast in the images and, finally, (iii) the saturation error is low because the error in measured intensities has a minimized propagation of the error to OD and ρ .

3 Experimental Validation

To validate our theoretical predictions, we have fabricated a thin, multilayer retina-mimicking phantom with narrow channels representing blood vessels in the retina. The retina-mimicking phantom was placed in a model eye. A scanning laser ophthalmoscope (SLO) built in-house using a supercontinuum laser as the light source was used to scan the artificial retina at different wavelengths, to experimentally validate our theory.

3.1 Scanning Laser Ophthalmoscope—Description of the System

A multiwavelength SLO⁴³⁻⁴⁶ was developed for the continuous acquisition of *en face* images of a phantom retina as shown in Fig. 9. The pixel rate of the system for imaging the phantoms was 30 kHz, although the maximum pixel rate of the system was 60 kHz. The size of a single image frame was 512×512 pixels. Thus, each image took 8.7 s to record. The essentially static nature of the phantoms precluded motion artifacts in the recorded image. The key system parameters are shown in Table 1. Using a relatively large core diameter of $100 \mu\text{m}$ in the detection compared with a single-mode core diameter of $2.5 \mu\text{m}$ in illumination, eliminated the speckle noise in the final image at the expense of spatial resolution due to reduced confocality.

3.2 Measurements in a Model Eye Using a Retina Mimicking Phantom

Uniformly scattering phantoms were manufactured using a method previously described by de Bruin et al.⁴⁷ Using mixtures with different wt. % of TiO₂, layered phantoms were made in a Petri dish, each layer cured in a thermal oven at 80°C for 60 min. Thin metallic wires of different diameter thicknesses (90 and 140 μm) were embedded in the topmost scattering layer of the phantom while still in the viscous state and then cured. Then, a transparent layer of silicone without any TiO₂ particles was poured finally to provide mechanical stability to the phantom. After solidifying completely, the metallic wires were removed from the phantom. For our experiments, we used three-layer phantoms [Fig. 10(a)] with a transparent top layer, a thin 200-μm layer with $\mu'_s = 1 \text{ mm}^{-1}$ with channels in the middle, and a bottom layer with $\mu'_s = 5 \text{ mm}^{-1}$, respectively.

Figure 10(b) shows an OCT B-scan (cross section) of the phantom used for the measurements in the model eye. Two water-soluble dyes—Evans Blue (Sigma Aldrich B.V., The Netherlands) and red food colorant (Lebenmittelfarbe, Birkmanns Voedekleurstof, Germany) were used to mimic the deoxyhemoglobin and oxyhemoglobin absorptions, respectively. Using a suitable concentration of these dyes, an artificial isosbestic point was created at 548 nm as shown in Fig. 10(c). Using a syringe, different calibrated proportions of the dyes

were injected into the channels. If [EB] is the concentration of Evans Blue and [FC] is the concentration of the red food colorant, we define a parameter ς as follows:

$$\varsigma = \frac{[\text{FC}]}{[\text{FC}] + [\text{EB}]} \quad (19)$$

An analysis similar to the analysis in Sec. 2 was performed to find the wavelengths resulting in low errors when estimating the parameter ς . The optimum wavelengths for the dye combination were found to be 500, 548 (isosbestic), and 610 nm, respectively. Due to the choice of the dichroic mirrors (DM1 and DM2 in Fig. 9) in the detection channel, 488 and 612 nm were chosen as the two anisobestic wavelengths to measure the saturation. Using a suitable ratio of [FC] and [EB], solutions with five different ς (0.00, 0.25, 0.50, 0.75, and 1.00) were made and injected into the channels using a syringe. The phantoms were imaged simultaneously using the three different wavelengths, 488-nm [bandwidth, full width at half maximum (FWHM): 2 nm], 612-nm (bandwidth, FWHM: 3 nm), and 548-nm (isosbestic) (bandwidth, FWHM: 2 nm), and *en face* face images were created for each wavelength. The number of reflected photons from TiO₂ and the dye-filled channel was estimated from the image and the parameter ς was then calculated after estimating the ODs, ρ and D_C (channel diameter) values from the images, respectively.

Fig. 10 (a) Layered structure of the phantoms: the phantoms were made to mimic the scattering properties of the retina and two cylindrical channels (red circles) with a diameters of 90 and 140 μm were embedded in the scattering medium to mimic the blood vessels in the retina. (b) OCT B-scan (in house 1310-nm OCT, average of 15 consecutive B scans) of the phantom with the 140-μm hollow (empty) channel (1) and 90-μm hollow (empty) channel (2). Scale bar is 300 μm in each direction. (c) Comparison between absorption spectrum of the dyes chosen for phantom measurements and blood.

Fig. 11 Typical phantom blood vessel intensity profile: the blood vessel absorption creates a dip in the intensity profile with a local minimum at point b . The OD is calculated as a ratio of the intensity of point t (the linear approximation of the tissue intensity in the absence of the blood vessel) and point b . The points v_1 and v_2 are the points with a maximum slope along the vessel walls, and t_1 and t_2 are the tissue locations used for the linear approximation of the tissue intensities.

3.3 Estimating Vessel Diameter from the Images

Figure 11 shows the typical intensity profile of a phantom blood vessel. The phantom blood vessel diameter was determined by the distance (in pixels) between the points with maximum slope magnitude along the blood vessel profiles v_1 and v_2 , respectively. The intensity at the blood vessel location was directly estimated by the minima of the vessel profile.

The intensity at the tissue location t was estimated using the points t_1 and t_2 by a linear approximation. The points t_1 and t_2 represent the locations where the light has not interacted with the blood and are at a fixed number of pixels from v_1 and v_2 , respectively, such that the locations t and b are affected by similar factors except the absorption by hemoglobin. In our analysis, we took the points t_1 and t_2 , at a distance of $2 \times$ diameter from on each side of v_1 and v_2 , respectively. To increase the accuracy of the estimated intensities (points t and b), the intensities were averaged along multiple locations in the tissue and in the middle of the blood vessel as shown in Fig. 11.

The intensity $I(\bar{x}_b, \lambda)$ was obtained by averaging the intensities of the point b along the valley (along with the y -dimension that is denoted by dotted lines). The point t and thereby $I(\bar{x}_t, \lambda)$ is obtained by averaging the points in an area as shown in Fig. 11.

3.4 Experimental Results with Retinal Phantoms

Figures 12(a)–12(c) show the reflectance measurements for 100% food coloring ($\zeta = 1.00$) showing high absorption at 488 nm [Fig. 12(a)]. Figures 12(d)–12(f) show the reflectance measurements for 0% food coloring ($\zeta = 0.00$). Measurement points were chosen along the center of the channel and in the tissue (TiO₂ layer in this case) to estimate the reflected light intensity (I_b and I_t). To bring the standard error on the mean intensity in both the tissue and along the blood vessels to acceptable values of $\leq 1\%$, up to 400 points were averaged in the tissue location and 40 points were averaged in the blood vessel location. The intensity profiles along the lines

indicated in Figs. 12(d)–12(f) are shown in Fig. 12(g). The dip in the intensity profile due to the absorption of blood is obvious in Fig. 12(g). The channels in the phantom are slightly curved due to the process of manufacturing the phantoms as described in Sec. 3.2. The phantom was slightly angled during measurements to reduce specular reflections. This resulted in a slight change in the nonuniformity of the collected intensities within the imaging field. The bright spots in Figs. 12(a) and 12(d) are due to specular reflections from the surface of the sample and could be caused by dust on the surface.

The comparison between the theoretical (with and without pigment packing correction) and experimentally obtained average ρ values in the phantom channel is shown in Fig. 13, for both the 90- and 140- μm channels used in the phantom. Our phantom geometry resembles the geometry in Fig. 1(a) here the pigment packing effect is not expected to play a large role. Indeed, Fig. 13 clearly shows that the estimated values closely follow the theoretical values without correction for pigment packaging. The estimated diameter of the channels from the images for 140 and 90 μm were $144 \pm 6 \mu\text{m}$ and $92 \pm 7 \mu\text{m}$, respectively.

In Fig. 13, both the fitted curves with and without pigment packing describe the measurements reasonably well as the difference between the uncorrected and corrected calculated ρ values is small. This illustrates that by picking optimum wavelengths that fulfill the conditions stated in Sec. 2, the deviation between the ρ values with and without pigment packing could be minimized. For nonoptimal wavelengths, the difference in ρ values with and without correction is much higher (data not shown) and hence, using the simplified model presented without pigment packing would result in large offsets to the saturation estimation for smaller blood vessels embedded deeper in tissue.

Figure 14 shows the pseudocolor map of the saturation overlaid on the corresponding reflectance image from the SLO for the 140- μm blood vessel for five different calibrated values. The estimated values show the overall mean and standard deviation of the saturation in the channel. For the estimation of ζ , the 488- to 548-nm pair was used for $\zeta = 0.00, 0.25$, and 0.50 , whereas the 548- to 612-nm pair was used for $\zeta = 0.75$ and 1.00 , respectively.

4 Discussion

SLO-based retinal oximetry has advantages over conventional fundus camera-based method regarding resolution and contrast. A critical factor for clinically relevant retinal oximetry is able to determine blood oxygen saturation in small retinal vessels—capillaries, venules, and arterioles. It is in these microvessels that the oxygen saturation is expected to decrease in response to increased metabolic demand or decreased oxygen delivery capacity. The larger retinal vessels ($>100 \mu\text{m}$) are expected to be less sensitive to changes in tissue metabolic demand or microvascular dysfunction and are therefore not ideal as early hypoxia markers. The clinical value of fundus camera-based oximeters is most likely limited due to the low spatial resolution, even if the oxygen saturation estimates would be accurate and robust taking into account our optimized wavelengths and our proposed algorithms that include pigment packaging. SLO-based oximeters have been investigated before, but their performance was limited due to suboptimal choices for the wavelengths used, combined with oximetry algorithms that did not take pigment packaging into account, which is likely a nonnegligible factor for smaller blood vessels located deeper in the retina.

Fig. 12 (a)–(f) Reflectance measurements of the phantoms in the model eye at three different wavelengths (λ , columns) and two different saturation levels (ζ , rows). The intensity profiles along the lines indicated in (d)–(f) is plotted in (g): 488-nm (blue), 548-nm (green), and 612-nm (red). Scale bar is 300 μm .

In this paper, for the first time, we have addressed factors that affect the accuracy of SLO-based oximeters, i.e., measurement noise and pigment packaging when measuring small, embedded blood vessels. Although we have used relatively large diameter blood vessels in our validation experiments, this was a limitation of the phantom fabrication technique yielding unstable phantoms at low channel diameters and is not a limitation of the imaging technique as a current SLO technology can image blood vessels $<50 \mu\text{m}$.

In our SLO design, we used a single-mode fiber illumination and multimode fiber detection. Hence the imaging is not truly confocal, but quasiconfocal or subdiffuse. We believe that quasiconfocal, subdiffuse detection reduces the central light reflex, a bright specular reflection from the center of a blood vessel that is common to many fundus imaging modalities that do not use cross polarizers. When the central light reflex is present in an image, the intensity in the center of the vessel could be extracted

by reconstructing the blood vessel profile using analytical models, such as a fourth-order polynomial.^{39,48} This adds additional error to the blood vessel OD estimation and thereby, the saturation estimation. Longitudinal chromatic aberration (LCA) is another potential source of error in oximetry due to the dispersion in the human eye.⁴⁹ LCA causes the spatial location of each pixel in the multispectral image to be in a slightly different axial plane for different wavelengths. This effect can lead up to a 250- μm longitudinal focal shift between the 470- and 592-nm wavelengths; however, this effect can be corrected as has been done previously by others^{50,51} using an achromatizing lens in the beam path. The optimum wavelength analysis was made with a 1-nm bandwidth around each central wavelength. The filter bandwidth choice can be incorporated into the oximetry analysis by multiplying the absorption spectrum of hemoglobin and the filter transmission spectrum and obtaining an average absorption coefficient for that spectral band, and then

Fig. 13 Comparison between the theoretical and experimental ρ values for the different ratio of dye concentrations in (a and b) the 90- μm channel, (c and d) 140- μm channel, respectively. The average values in the channels for each value are plotted as a box plot, with “+” values representing the outliers. Left panels (blue lines, panels a and c) use the 488- to 548-nm wavelength pair for ρ calculations, right panels (red lines, panels b and d) use the 548- to 612-nm wavelength pair for ρ calculations.

Fig. 14 Pseudocolor images of the saturation s superimposed on the reflectance image for the 140- μm blood vessel. Scale bar is 300 μm . The estimated values with and without pigment packing are presented as mean (standard deviation) of all values of estimated s in each channel.

using this averaged absorption coefficients in Eqs. (9) and (17) to obtain a saturation estimation, and therefore the saturation error.

As explained previously, the noise in the images propagates into the saturation error. To achieve acceptable levels of saturation error (we aim for <5%, but it depends on the clinical application), a large number of points along the blood vessel and the

tissue location have to be averaged. Although this is relatively easy in our experiments due to the homogeneous nature of the phantom, an *in-vivo* image contains structural information and choosing many points might lead to an incorrect estimation of the intensity in the absence of a blood vessel. A smaller number of points around the blood vessel should be averaged to overcome this problem, and this requires multiple images to be

produced within a short amount of time. Fundus cameras are based on snapshot imaging, and it might not always be possible to acquire multiple images continuously without causing discomfort to the subject or exceeding the safety limit for radiant exposure. Although our current SLO images at relatively slow speeds, significant improvements in speed can be achieved using a resonant scanner, with a reduced SNR. Inhomogeneities in absorption within a vessel due to the distribution of red blood cells is less relevant in the final saturation estimation due to the required averaging and resulting image acquisition speeds to achieve desired SNR. Current commercial SLO technology has an imaging speed of about 30 Hz, which facilitates the averaging of multiple consecutive images to accomplish noise reduction.

The effective path length ($\langle L_{\text{eff}} \rangle$) that photons travel through a tissue volume before being collected depends on the illumination and collection geometry and is also a function of the optical properties of the tissue and is, therefore, wavelength dependent. We have assumed that the effective path length for the two wavelengths used to calculate the factor ρ [Eq. (12)] is almost the same. In the wavelength range 450 to 600 nm, absorption remains in the same order and scattering varies slowly with wavelength (roughly as $1/\lambda$). However, if there is a slight mismatch in the path lengths, this results in an offset in the saturation estimation. We have calculated that a 20% mismatch in path lengths ($\langle L_{\lambda_p} \rangle / \langle L_{\lambda_i} \rangle = 1.2$) could lead up to a 30% offset in the saturation estimates. For the phantoms that we fabricated, scattering was a slowly varying function of wavelength, similar to tissue, and the absorption coefficients of the dyes in the channels were of the same order of magnitude as the absorption coefficients of oxy- and deoxyhemoglobin. For our phantoms, we have shown the ability to accurately recover the saturation in the channels, suggesting that there were no significant changes in the path length due to scattering and absorption in a tissue-mimicking phantom with relevant scattering and absorbing properties. A limiting factor of our phantom was that the dyes that mimic the oxy- and deoxyhemoglobin were nonscattering, which is different from real blood.

Additionally, in a subdiffuse approach, pigments that are present in the retina, especially in the retinal pigment epithelium (RPE) and absorption background due to choroid blood, can affect the recorded intensity. The effect of the melanin on backscattered light varies with concentration,⁵² and an *a priori* knowledge of the pigment concentration in RPE melanin and its contribution to the backscattered light can aid in removing the influence of RPE in the backscattered light. OCT-based oximetry methods might have an advantage here that layers of the retina free from the influence of RPE can be used for extracting the intensities. But, OCT being a coherent detection method suffers from speckles, and a significant amount of averaging has to be performed to improve the image quality and reduce the error on intensity down to 1%. We must emphasize that the insufficient removal of the system reflections from the recorded intensities will result in a wrong estimation of the ODs and hence the saturation. Therefore, in our experiments, all the internal system reflections were subtracted by taking a reference measurement without any sample.

There have been several attempts to choose the optimal wavelengths for retinal oximetry, but none of them have accounted for the hemoglobin packaging effect. Most of these attempts were aimed at two wavelength oximetries^{15,53,54} using one isosbestic wavelength and another nonisosbestic

wavelength, where $|\mu_a^{\text{HbO}_2}(\lambda) - \mu_a^{\text{Hb}}(\lambda)|$ is maximum, the popular choice being 570 (isosbestic) and 600 nm, respectively. Although this wavelength combination might give sufficiently accurate saturation estimation in arteries, our analysis shows that they are not reliable for low saturation values of $S < 60\%$. Schweitzer et al.⁵⁵ reported a physiologically relevant retinal oxygenation range of 57% to 92%. However, lower saturations may be expected due to various pathological conditions. Smith et al.⁵⁶ chose 803 (isosbestic) and 670 nm to estimate the saturation. The infrared wavelengths have weak blood absorption and therefore give a poor estimate of the blood absorption leading to a large error in the saturation estimations according to our calculations. In 1988, Delori¹⁴ used three wavelengths—558, 570 (isosbestic), and 548 nm (isosbestic)—to scan a small area of the retina with blood vessel tracking to reduce eye motion artefacts. This wavelength choice, similar to the 570- to 600-nm wavelength pair, results in poor accuracy for low saturation values, $S < 60\%$. Our assessment of the modified oximetry equations suggests that the 470, 506 (isosbestic), and 592 nm triad is the best combination for oximetry. Although this choice spans a relatively broad wavelength range ≈ 130 nm, the absorption coefficients and scattering coefficients in this wavelength range are similar to those we have used in our phantoms, suggesting that the effective path length differences within this wavelength range are of minor importance.

The accuracy of saturation estimation can be potentially improved by including more wavelengths. To this end, hyperspectral imaging^{17,23–25,57} approaches toward *in vivo* retinal oximetry have also been undertaken by various groups. However, the challenge of acquiring images with sufficient SNR from multiple wavelengths at high spatial resolution, in a large field of view, with high speed and with low power levels remains.

5 Conclusions

We have presented a method to select optimal wavelengths for oximetry measurements in the retina through an error analysis on a modified oximetry equation that includes the effect of pigment packaging. The result of the analysis yielded 470, 506 (isosbestic), and 592 nm as three suitable wavelengths for accurate estimation of oxygenation in the retinal blood vessels. We have developed a quasiconfocal, multispectral SLO capable of simultaneously imaging the retina at all these wavelengths and demonstrated the validity of our approach in tissue-mimicking phantoms. *In vivo* human experiments are pending for ethical committee approval.

Disclosures

MD (P: VU Amsterdam), AA (P: VU Amsterdam), JFDB (P: VU Amsterdam)

Acknowledgments

This research was funded by the Netherlands Organisation for Scientific Research (NWO) with a Vici (JFdB) (Grant No. 91810628), the Netherlands Organisation for Health Research and Development ZonMW (Grant No. 91212061), and STW (Grant No. 12822), which are both a part of NWO, and Heidelberg Engineering GmbH. The authors would like to thank F. Feroldi, M.G.O. Grafe, L. Bartolini, and Dr. B. K. P. Lochocki for assistance with experiments; J. J. A. Weda for fabrication of the phantoms.

References

1. E. D. Cole et al., "Contemporary retinal imaging techniques in diabetic retinopathy: a review," *Clin. Exp. Ophthalmol.* **44**(4), 289–299 (2016).
2. R. Klein et al., "Diabetic retinopathy as detected using ophthalmoscopy, a nonmyriatic camera and a standard fundus camera," *Ophthalmology* **92**, 485–491 (1985).
3. G. Michelson, M. Scibor, and D. Schweitzer, "Intravascular oxygen saturation in retinal vessels in glaucoma," *Invest. Ophthalmol. Visual Sci.* **43**, U445–U445 (2002).
4. E. Nitta et al., "Retinal oxygen saturation before and after glaucoma surgery," *Acta Ophthalmol.* (October), 2015–2018 (2016).
5. O. B. Olafsdottir et al., "Retinal oximetry in primary open-angle glaucoma," *Invest. Ophthalmol. Visual Sci.* **52**(9), 6409–6413 (2011).
6. T. Eysteinson et al., "Retinal vessel oxygen saturation and vessel diameter in retinitis pigmentosa," *Acta Ophthalmol.* **92**(5), 449–453 (2014).
7. C. Türksever et al., "Retinal vessel oxygen saturation and its correlation with structural changes in retinitis pigmentosa," *Acta Ophthalmol.* **92**(5), 454–460 (2014).
8. T. Ueda-Consolvo et al., "Analysis of retinal vessels in eyes with retinitis pigmentosa by retinal oximeter," *Acta Ophthalmol.* **93**(6), e446–e450 (2015).
9. A. Geirsdottir et al., "Retinal oxygen metabolism in exudative age-related macular degeneration," *Acta Ophthalmol.* **92**(1), 27–33 (2014).
10. S. Yoneya et al., "Retinal oxygen saturation levels in patients with central retinal vein occlusion," *Ophthalmology* **109**, 1521–1526 (2002).
11. M. I. Uddin et al., "In vivo imaging of retinal hypoxia in a model of oxygen-induced retinopathy," *Sci. Rep.* **6**(1), 31011 (2016).
12. A. B. Einarsson et al., "Retinal oximetry imaging in Alzheimer's disease," *J. Alzheimer's Dis.* **49**(1), 79–83 (2015).
13. N. J. Hart et al., "Ocular indicators of Alzheimer's: exploring disease in the retina," *Acta Neuropathol.* **132**(6), 767–787 (2016).
14. F. C. Delori, "Noninvasive technique for oximetry of blood in retinal vessels," *Appl. Opt.* **27**(6), 1113–1125 (1988).
15. J. B. Hickam, H. O. Sieker, and R. Frayser, "Studies of retinal circulation and A-V oxygen difference in man," *Trans. Am. Clin. Climatol. Assoc.* **71**(657), 34–44 (1960).
16. E. DeHoog and J. Schwiegerling, "Optimal parameters for retinal illumination and imaging in fundus cameras," *Appl. Opt.* **47**(36), 6769–6777 (2008).
17. J. C. Ramella-Roman et al., "Measurement of oxygen saturation in the retina with a spectroscopic sensitive multi aperture camera," *Opt. Express* **16**(9), 6170–6182 (2008).
18. M. Hammer and D. Schweitzer, "Quantitative reflection spectroscopy at the human ocular fundus," *Phys. Med. Biol.* **47**(2), 179–191 (2002).
19. M. Hammer et al., "Retinal vessel oximetry-calibration, compensation for vessel diameter and fundus pigmentation, and reproducibility," *J. Biomed. Opt.* **13**(5), 054015 (2008).
20. T. Shiba, K. Maruo, and T. Akahoshi, "Development of a multi-field fundus photographing system using a non-myriatic camera for diabetic retinopathy," *Diabetes Res. Clin. Pract.* **45**(1), 1–8 (1999).
21. A. Geirsdottir et al., "Retinal vessel oxygen saturation in healthy individuals," *Invest. Ophthalmol. Visual Sci.* **53**(9), 5433–5442 (2012).
22. S. H. Hardarson et al., "Automatic retinal oximetry," *Invest. Ophthalmol. Visual Sci.* **47**(11), 5011–5016 (2006).
23. D. J. Mordant et al., "Spectral imaging of the retina," *Eye* **25**(3), 309–320 (2011).
24. W. R. Johnson et al., "Snapshot hyperspectral imaging in ophthalmology," *J. Biomed. Opt.* **12**(1), 014036 (2014).
25. B. Khoobehi, J. M. Beach, and H. Kawano, "Hyperspectral imaging for measurement of oxygen saturation in the optic nerve head," *Invest. Ophthalmol. Visual Sci.* **45**(5), 1464–1472 (2004).
26. A. H. Kashani et al., "Hyperspectral computed tomographic imaging spectroscopy of vascular oxygen gradients in the rabbit retina in vivo," *PLoS ONE* **6**(9), e24482–11 (2011).
27. R. A. Ashman, F. Reinholz, and R. H. Eikelboom, "Oximetry with a multiple wavelength SLO," *Int. Ophthalmol.* **23**(4–6), 343–346 (2001).
28. J. V. Kristjansdottir et al., "Retinal oximetry with a scanning laser ophthalmoscope," *Invest. Ophthalmol. Visual Sci.* **55**(5), 3120–3126 (2014).
29. H. Li et al., "Measurement of oxygen saturation in small retinal vessels with adaptive optics confocal scanning laser ophthalmoscope," *J. Biomed. Opt.* **16**(11), 110504 (2011).
30. S. P. Chong et al., "Structural and functional human retinal imaging with a fiber-based visible light OCT ophthalmoscope," *Biomed. Opt. Express* **8**(1), 323–337 (2017).
31. J. Yi et al., "Visible light optical coherence tomography measures retinal oxygen metabolic response to systemic oxygenation," *Light Sci. Appl.* **4**(9), e334 (2015).
32. J. C. Finlay and T. H. Foster, "Effect of pigment packaging on diffuse reflectance spectroscopy of samples containing red blood cells," *Opt. Lett.* **29**(9), 965–967 (2004).
33. A. Amelink, T. Christiaan, and H. J. C. M. Sterenborg, "Effect of hemoglobin extinction spectra on optical spectroscopic measurements of blood oxygen saturation," *Opt. Lett.* **34**(10), 1525–1527 (2009).
34. N. Rajaram et al., "Experimental validation of the effects of microvasculature pigment packaging on in vivo diffuse reflectance spectroscopy," *Lasers Surg. Med.* **42**(7), 680–688 (2010).
35. N. Bosschaart et al., "A literature review and novel theoretical approach on the optical properties of whole blood," *Lasers Med. Sci.* **29**(2), 453–479 (2014).
36. M. Hammer et al., "Optical properties of ocular fundus tissues—an in vitro study using the double-integrating-sphere technique and inverse Monte Carlo simulation," *Phys. Med. Biol.* **40**, 963–978 (1995).
37. A. Sassaroli and S. Fantini, "Comment on the modified Beer–Lambert law for scattering media," *Phys. Med. Biol.* **49**(14), (2004).
38. L. Kocsis, P. Herman, and A. Eke, "The modified Beer–Lambert law revisited," *Phys. Med. Biol.* **51**(5), (2006).
39. M. Hammer et al., "Light paths in retinal vessel oximetry," *IEEE Trans. Biomed. Eng.* **48**(5), 592–598 (2001).
40. P. I. Rodmell et al., "Light path-length distributions within the retina," *J. Biomed. Opt.* **19**(3), 036008 (2014).
41. International Electrotechnical Commission, "International electrotechnical commission, safety of laser products part 1: equipment classification and requirements, (Geneva, Switzerland), IEC-60825-1 (2014)," **122** (2014).
42. W. Verkrusse et al., "Modelling light distributions of homogeneous versus discrete absorbers in light irradiated turbid media," *Phys. Med. Biol.* **42**(1), 51–65 (1997).
43. R. H. Webb, G. W. Hughes, and F. C. Delori, "Confocal scanning laser ophthalmoscope," *Appl. Opt.* **26**(8), 1492–1499 (1987).
44. F. LaRocca et al., "Optimization of confocal scanning laser ophthalmoscope design," *J. Biomed. Opt.* **18**(7), 076015 (2013).
45. D. X. Hammer et al., "Line-scanning laser ophthalmoscope," *J. Biomed. Opt.* **11**(4), 041126 (2014).
46. M. Damodaran et al., "Digital micromirror device based ophthalmoscope with concentric circle scanning," *Biomed. Opt. Express* **8**(5), 2766–2780 (2017).
47. D. M. de Bruin et al., "Optical phantoms of varying geometry based on thin building blocks with controlled optical properties," *J. Biomed. Opt.* **15**(2), 025001 (2010).
48. A. Lompadó, L. W. Hillniaif, and K. R. Denninghoff, "Multi-spectral confocal scanning laser ophthalmoscope for retinal vessel oximetry," *Proc. SPIE* **3920** (X), 67–73 (2000).
49. M. Vinas et al., "Longitudinal chromatic aberration of the human eye in the visible and near infrared from wavefront sensing, double-pass and psychophysics," *Biomed. Opt. Express* **6**(3), 948–962 (2015).
50. S. P. Chong et al., "Ultrahigh resolution retinal imaging by visible light OCT with longitudinal achromatization," *Biomed. Opt. Express* **9**(4), 1477 (2018).
51. F. LaRocca et al., "True color scanning laser ophthalmoscopy and optical coherence tomography handheld probe," *Biomed. Opt. Express* **5**(9), 3204 (2014).
52. W. Liu, S. Jiao, and H. F. Zhang, "Accuracy of retinal oximetry: a Monte Carlo investigation," *J. Biomed. Opt.* **18**(6), 066003 (2013).
53. A. J. Cohen and R. A. Laing, "Multiple scattering analysis of retinal blood oximetry," *IEEE Trans. Biomed. Eng.* **BME-23**(5), 391–400 (1976).
54. J. S. Tiedeman et al., "Retinal oxygen consumption during hyperglycemia in patients with diabetes without retinopathy," *Ophthalmology* **105**(1), 31–36 (1998).
55. D. Schweitzer et al., "In vivo measurement of the oxygen saturation of retinal vessels in healthy volunteers," *IEEE Trans. Biomed. Eng.* **46**(12), 1454–1465 (1999).

56. M. H. Smith et al., "Effect of multiple light paths on retinal vessel oximetry," *Appl. Opt.* **39**(7), 1183 (2000).
57. J. Beach, J. Ning, and B. Khoobehi, "Oxygen saturation in optic nerve head structures by hyperspectral image analysis," *Curr. Eye Res.* **32**(2), 161–170 (2007).

Mathi Damodaran is a PhD candidate in the biophotonics and medical imaging group, Vrije Universiteit Amsterdam. He received his bachelor of engineering degree in electronics and communication from Anna University, Chennai (India). He received his MSc degree in photonics from the University of St. Andrews, Scotland, in a joint program with Universiteit Gent, Belgium. The focus of his research is developing optical imaging techniques for diagnosis and followup of the retinal diseases. He is a member of SPIE.

Arjen Amelink received his master's degree in experimental physics from the University of Groningen in 1995, and his PhD in atomic

physics from the University of Utrecht in 2000. After one year at Philips Research, he started in 2001 as a postdoc at the Center for Optical Diagnostics and Therapy (CODT) at the Erasmus Medical Center in Rotterdam. From 2005 to 2013, he was an assistant professor at the CODT, where his research focused on clinical applications of optical spectroscopic technologies. Since 2014, he has been a senior scientist at the Medical Photonics research program within TNO.

Johannes F. de Boer is a full professor at the Department of Physics and Astronomy, Vrije Universiteit, Amsterdam, and a former director of LaserLaB Amsterdam. He has been an active researcher in biomedical optics for over 20 years since obtaining his PhD from the University of Amsterdam. He was an assistant professor at the Beckman Laser Institute, UC Irvine, California, and an associate professor at Harvard Medical School, Boston, Massachusetts. He is a member of SPIE and an OSA fellow.