Improving the characterization of *ex vivo* human brain optical properties using high numerical aperture optical coherence tomography by spatially constraining the confocal parameters

Jiarui Yang^a,[✉] Ichun Anderson Chen^a,[✉] Shuaibin Chang,^b
Jianbo Tang,^a Blaire Lee,^a Kivılcım Kılıç^a,[✉] Smrithi Sunil,^a
Hui Wang^b,[✉] Divya Varadarajan,^c Caroline Magnain^b,[✉]
Shih-Chi Chen^b,[✉] Irene Costantini^b,[✉] Francesco Pavone^b,[✉]
Bruce Fischl,^{c,g} and David A. Boas^{a,*}

^aBoston University, Department of Biomedical Engineering, Boston, United States

^bBoston University, Department of Electrical and Computer Engineering, Boston, United States

^cMassachusetts General Hospital, A.A. Martinos Center for Biomedical Imaging, Department of Radiology, Boston, United States

^dThe Chinese University of Hong Kong, Department of Mechanical Engineering, Hong Kong Special Administrative Region, China

^eUniversity of Florence, European Laboratory for Non-Linear Spectroscopy, Sesto Fiorentino, Florence, Italy

^fNational Research Council, National Institute of Optics, Italy

^gHealth Science and Technology/Computer Science & Artificial Intelligence Laboratory, Massachusetts Institute of Technology, Cambridge, Massachusetts, United States

Abstract

Significance: The optical properties of biological samples provide information about the structural characteristics of the tissue and any changes arising from pathological conditions. Optical coherence tomography (OCT) has proven to be capable of extracting tissue's optical properties using a model that combines the exponential decay due to tissue scattering and the axial point spread function that arises from the confocal nature of the detection system, particularly for higher numerical aperture (NA) measurements. A weakness in estimating the optical properties is the inter-parameter cross-talk between tissue scattering and the confocal parameters defined by the Rayleigh range and the focus depth.

Aim: In this study, we develop a systematic method to improve the characterization of optical properties with high-NA OCT.

Approach: We developed a method that spatially parameterizes the confocal parameters in a previously established model for estimating the optical properties from the depth profiles of high-NA OCT.

Results: The proposed parametrization model was first evaluated on a set of intralipid phantoms and then validated using a low-NA objective in which cross-talk from the confocal parameters is negligible. We then utilize our spatially parameterized model to characterize optical property changes introduced by a tissue index matching process using a simple immersion agent, 2,2'-thiodiethanol.

Conclusions: Our approach improves the confidence of parameter estimation by reducing the degrees of freedom in the non-linear fitting model.

© The Authors. Published by SPIE under a Creative Commons Attribution 4.0 Unported License. Distribution or reproduction of this work in whole or in part requires full attribution of the original publication, including its DOI. [DOI: [10.1117/1.NPh.7.4.045005](https://doi.org/10.1117/1.NPh.7.4.045005)]

*Address all correspondence to David A. Boas, dboas@bu.edu

Keywords: optical coherence tomography; human brain tissue; index matching.

Paper 20059R received Jul. 23, 2020; accepted for publication Sep. 30, 2020; published online Oct. 21, 2020.

1 Introduction

Optical coherence tomography (OCT) is a volumetric imaging technique that has been widely used to image the microstructure of biological samples.¹ The OCT signal originates from back-scattered photons and provides image contrast that reflects intrinsic differences in tissue properties within the sample. The depth profile of the OCT signal is determined by five factors: the refraction index of the sample, the optical attenuation coefficient, the back-scattering coefficient, the numerical aperture (NA) of focusing optics, and the depth of the focus in the sample.² The attenuation coefficient, i.e., the sum of the tissue scattering and absorption coefficients, is a spatially varying intrinsic property of the sample and is independent of the optics and the incident power of the system. For low-NA OCT systems, where the lateral resolution is typically larger than 10 μm and the confocal parameter (i.e., two times the Rayleigh range) is comparable to or larger than the imaging depth, a single-scattering exponential model is appropriate to describe OCT signal attenuation versus depth.^{3,4} Previous studies have shown that extracting the attenuation coefficient from low-NA OCT measurements based on Beer's law is useful in quantifying the optical properties of the sample for applications, including detecting cancerous tissue in skin, bladder, and brain,^{5–7} monitoring blood glucose concentration,⁸ characterizing atherosclerosis plaques^{3,9} and correlating collagen content with histological staining.¹⁰ However, for high-NA OCT systems where the lateral resolution reaches a few microns or less, the shape of the axial point spread function (PSF) plays an important role in modulating the OCT signal depth profile, and thus the effect of the confocal parameter and focus depth cannot be neglected when estimating the optical properties of the sample.^{4,11} We have previously shown that OCT with a 3.5- μm lateral resolution is able to differentiate laminar structures in the neocortex of human brain samples using the average intensity projection (AIP) over the effective Rayleigh range.^{2,12} Moreover, we showed that estimating the intrinsic optical properties of the sample could provide additional information to differentiate brain structures, albeit with high inter-parameter correlations that resulted in large variance.²

Extensive mapping of brain structures and neuronal connections in postmortem human brain requires high-throughput microscopic imaging over large volumes. One major limitation is penetration depth due to tissue light scattering.¹³ Different approaches have been developed to overcome this challenge, among which optical clearing agents are suitable for various applications and imaging modalities.¹⁴ Water-based optical clearing reagents, such as Scale,¹⁵ See Deep Brain,¹⁶ ClearT,¹⁷ and Clear Unobstructed Brain/Body Imaging Cocktail and Computational analysis,¹⁸ are designed to increase the penetration depth in fixed brain samples while allowing for improved preservation of fluorescence.^{19,20} Most of them, however, present other nonnegligible limitations, such as long incubation time, severe structural alteration, and incompatibility with immunostaining. A glycol derivative, 2,2'-thiodiethanol (TDE), was previously reported to be a versatile, rapid, and inexpensive water-soluble clearing agent that overcomes these limitations.^{21,22} However, the impact of TDE optical clearing via index matching on the optical property contrast that permits imaging of brain structures by OCT has not yet been well studied.

In this work, we develop a systematic approach to spatially parametrize the confocal parameters across a three-dimensional (3-D) OCT image to constrain and reduce the degrees of freedom in the nonlinear coefficient fitting problem, resulting in improved confidence in the estimated optical properties of the sample. We first evaluate our spatially parametrized model on Intralipid phantoms with varied scattering levels. We compare the estimated coefficients using a high-NA objective with a low-NA objective to validate our parametrized model. We then quantify the improvements afforded by our model compared with the previously established model that does not employ spatial parameterization.² Finally, we utilize our fitting procedure to characterize the optical property changes of a human brain motor cortex sample under index matching in TDE. We quantify the reduction in the scattering coefficients of both gray matter (GM) and white matter (WM) with index matching, and show differential responses of GM versus WM with the concentration of TDE.

2 Method

2.1 Theoretical Model

Based on Schmitt et al.⁴ and Izatt et al.,¹¹ the derived single scattering model of the OCT signal can be written as

$$A(z) = \frac{\eta e \tau_i P_0}{2 h \nu} \sqrt{R(z)}, \quad (1)$$

where P_0 is the incident power, η is the quantum efficiency of the camera sensor, e is the electron charge, τ_i is the integration time of the line-scan camera, and $h \nu$ is the photon energy. $R(z)$ is the reflectance from the sample at the depth of z , and $z = 0$ is the tissue surface. The reflectance is determined by two factors: the back-scattering coefficient (μ_b) and the total attenuation coefficient (μ_t), which is the sum of the scattering coefficient (μ_s) and the absorption coefficient (μ_a). In the near-infrared range, light attenuation within the tissue is dominated by scattering and light absorption is negligible. Therefore, we use μ_s to represent the total attenuation coefficient in this paper. Further, due to the shallow depth range of the OCT measurement, we can generally neglect the effects of multiply scattered light.²³ Both μ_b and μ_s are assumed constant over depth in one OCT A-line (an assumption we hope to remove in the future). The OCT reflectance signal versus depth can thus be expressed as

$$R(z) = \mu_b \cdot \exp(-2\mu_s z) \cdot h(z), \quad (2)$$

where $h(z)$ is the axial PSF, which is dependent on the refractive index of the medium (n), the focus depth (z_f), and the Rayleigh range (z_R). In a single-mode fiber-based OCT system, the light beam in the sample arm approximately follows a Gaussian distribution.²⁴ Since we measure our tissue samples under liquid immersion, a simplified PSF model that neglects the effect of the refractive index difference at the tissue-air interface is given as

$$h(z) = \frac{1}{1 + \left(\frac{z - z_f}{n \cdot z_R}\right)^2} = \frac{1}{1 + \left(\frac{z - z_f}{z_r}\right)^2}, \quad (3)$$

where $z_r = n \cdot z_R$ is the effective Rayleigh range. The Rayleigh range is calculated given the optics of the system as

$$z_R = \frac{\pi \omega_0^2}{\lambda_0} \approx \frac{\omega_0}{NA}, \quad (4)$$

where ω_0 is the beam waist at the focus, λ_0 is the center wavelength of the laser, and NA is the numerical aperture of the optics. However, in a highly scattering medium measured with a high-NA objective, the effective Rayleigh range is observed to deviate from the theoretical value as a function of z_f and μ_s .²⁵

$$z'_r = z_r \cdot \sqrt{1 + \frac{\mu_s z_f}{4} \left(\frac{NA}{n}\right)^2}. \quad (5)$$

This occurs because the oblique rays with longer path lengths are attenuated more than the on-axis rays, which effectively reduces the NA of the focusing optics. In our set up with NA = 0.28, this effect is negligible. In our fitting model, z_r is assumed to be constant over the entire image.

The focus depth, on the other hand, is not constant over the image because of two confounding effects. First, we must correct for the field curvature effect that arises from the different path lengths that light travels to the image focus for different positions in the image. We calibrated this field curvature correction for our system by imaging a mirror placed at the image focus.²⁶ The second confounding effect is the likelihood that the surface of the brain sample being imaged is not parallel with the image plane. To correct for this effect, we parameterize z_f within the imaging field of view (FOV) as

$$z_f(x, y) = A + Bx + Cy, \quad (6)$$

where x and y correspond to positions in the image as this accounts for the possibility that the image plane is not parallel to the surface of the sample.

Finally, the back-scattering coefficient μ_b can be quantified with known η and P_0 . However, this would require knowledge of optical losses of all optical components in the OCT system. A relative back-scattering coefficient was previously introduced as

$$\mu'_b = \left(\frac{\eta e \tau_i P_0}{2h\nu} \right)^2 \mu_b. \quad (7)$$

This relative back-scattering coefficient is system dependent and could be used to estimate the relative spatial variations in μ_b .²

2.2 Spectral-Domain OCT System

We used a commercial spectral-domain OCT system (Telesto TEL320C1, Thorlabs, New Jersey) to measure the optical properties of human brain tissue. The light source is a broadband superluminescent diode with center wavelength of 1300 nm and a full width half maximum bandwidth of 150 nm, yielding an axial resolution of 4.2 μm in tissue. The spectrometer has a 2048-pixel InGaAs line scan camera operating at an A-line rate of 76 kHz. The total imaging depth is 2.6 mm in tissue. A 10 $\times$ air objective (Mitutoyo) was used in the sample arm, which yields a lateral resolution of 3.5 μm with a theoretical Rayleigh range of 40 μm in a nonscattering medium. The maximum sensitivity of the system is 109 dB.

In spectral-domain OCT, the sensitivity drops with imaging depth due to the spectrometer resolution and discrete spectral sampling by the line-scan camera. The sensitivity roll-off function is given as²⁷

$$H(z) = \left(\frac{\sin \xi}{\xi} \right)^2 \cdot \exp \left(\frac{-\omega^2}{2 \ln 2} \xi^2 \right), \quad (8)$$

where $\xi = z/z_{\text{max}}$ is the relative depth normalized to maximum imaging depth and $\omega = \delta\lambda/\Delta\lambda$ represents the ratio of the spectrometer resolution ($\delta\lambda$) and the wavelength spacing between camera pixels ($\Delta\lambda$). A silver mirror was translated at a series of imaging depths and the measured peak reflectance was fit with the above function [Eq. (8)] to confirm the system's sensitivity roll-off.

2.3 Model Coefficient Fitting

A nonlinear least-squares solver was used for estimating the μ_s and μ'_b parameters in the nonlinear theoretical model [Eqs. (1)–(7)]. The “trusted region reflective” algorithm was used for the optimization process. The trusted region reflective algorithm is an algorithm used in solving large-scale nonlinear equations. Specifically, this algorithm is a subspace adaptation of the Coleman-Li trust region and interior method.²⁸ Trust region methods form a respected class of algorithms for solving unconstrained minimization problems mostly due to their strong convergence properties. The algorithm first defines a trust region around the current best solution, then forces the size of the region to be updated iteratively only based on a sufficient decrease in the cost function. All the curve fitting was performed in MATLAB[®]. Before estimating μ_s and μ'_b for each position in an image, we first estimate the values of the constant z_r and the z_f tilt parameterized by Eqs. (2) and (3) by performing a simultaneous fit of OCT depth profiles from multiple positions within a region of the sample where the optical properties are known to be constant. This condition is easily met in phantoms with spatially uniform optical properties. For brain samples where the optical properties are spatially varying, we utilize the agarose material in which our brain tissue is embedded for creating the condition of spatially uniform optical properties for estimating z_r and the parameterization of z_f . The value of z_r is then taken to be constant through the entire brain sample, while the value of z_f will vary linearly with x and y

as prescribed by Eq. (6). Specifically, we fit the depth profiles from multiple locations within the agarose material around the brain tissue simultaneously to extract the value of A , B , and C in Eq. (6). The linear coefficients (B and C) are always constant for different spatial locations due to flattened surface, whereas the constant term (A) is spatially varying due to the angle of the vibrotome blade varying from exactly horizontal. Since it is a linear variation, we interpolate the constant term along the direction of surface flattening to obtain the value of A for each image tile.

For comparison with the full model, we fit the four parameters (μ_s , μ'_b , z_f , and z_r) independently using Eqs. (2) and (3) for each position within a region. Due to the nonlinearity of the model and the high inter-parameter dependency, a two-step procedure was used as previously described in Ref. 2. We fit the four-parameter model to the averaged experimental data and obtained an average z_r for each brain sample. We repeat the fitting with z_r fixed to this estimated value and then estimate the spatially varying optical properties and z_f .

2.4 Phantom Experiment Set Up

A set of eight liquid optical phantoms was prepared by diluting different volume concentrations (4%, 8%, 12%, 16%, 20%, 40%, 60%, and 80%) of Intralipid 20% (batch 10LI4282, Fresenius Kabi, Germany) in deionized water to create phantoms with eight different scattering coefficients. Taking into account the properties of the different Intralipid components, pure Intralipid contains 22.7% (v/v) scattering particles (soybean oil and egg lecithin),^{29–32} the concentration-dependent scattering coefficient can be calculated based on the Twersky quotation and corrected by packing factor³³

$$\mu_s(\lambda, \phi_p) = \frac{\mu_s^{indep}(\lambda)}{0.227} \phi_p \left\{ \frac{(1 - \phi_p)^{p(\lambda)+1}}{[1 + \phi_p(p(\lambda) - 1)]^{p(\lambda)-1}} \right\}, \quad (9)$$

where λ is the wavelength, ϕ_p is the volume concentration of scattering particles in the sample, p is the scattering phase function, and μ_s^{indep} is the independent scattering coefficient. An accurate equation for the independent scattering coefficient of pure intralipid in the 500- to 2250-nm range, was derived in Ref. 34, with μ_s in mm^{-1} and λ in nm.

$$\mu_s^{indep}(\lambda) = 1.868 \times 10^9 \times \lambda^{-2.59}. \quad (10)$$

The eight liquid phantoms were imaged with a 10× air objective. Each phantom was measured at 13 different focus depths, starting with $z_f = 0$ and incrementing by 50 μm up to a depth of 600 μm in the phantom. The three shallow depths ($z_f = 0, 50, 100 \mu\text{m}$) were discarded from further analysis due to the specular reflection from the phantom surface. The deep depths ($z_f > 300 \mu\text{m}$) were also discarded due to the poor signal-to-noise ratio causing a constant noise offset in the signal, systematically deviating the experimental data from the theoretical model. Each measurement produces an image that spans over a $1 \times 1 \times 2.5 \text{ mm}^3$ volume with a 2.5- μm isotropic pixel size, but with a lateral resolution of 4 μm and an axial resolution of 3.5 μm . During acquisition, A-lines were repeated 10 times to improve the signal to noise ratio. The total acquisition time for each phantom was 13.4 s.

The raw data were resampled in k -space, compensated for dispersion, and Fourier transformed into the spatial domain following standard procedures.³⁵ Then the volume data were corrected for sensitivity roll off as described in Ref. 2. Finally, the noise floor, defined as the average intensity at an imaging depth between 2.4 and 2.6 mm where no signal is expected, was subtracted.

2.5 Tissue Preparation and Imaging

A human brain tissue was obtained from the Massachusetts General Hospital Autopsy Suite. The brain was from a neurologically normal subject without previous diagnosis of neurological deficits. The sample was fixed with 10% formalin for at least two months. The postmortem interval

did not exceed 24 h. The brain was cut into smaller blocks. Then the blocks of interest were embedded in oxidized agarose and covalently cross-linked with the agarose using borohydride borate solution. Embedded human brain blocks were washed for one month in phosphate buffer saline solution (PBS) 0.01 M at room temperature (RT) while gently shaking. Index matching was performed with serial incubations in 100 ml of 20%, 40%, and 60% (v/v) TDE in 0.01 M PBS (TDE/PBS) each for 24 h at RT while gently shaking.

One human brain block from the primary motor cortex was imaged with a 10× air objective (Mitutoyo) and a z-spacer. The surface of the embedded sample was flattened using a vibratome slicer before OCT imaging.³⁶ The volumetric imaging spanned a FOV of $1 \times 1 \times 2.5 \text{ mm}^3$, with a voxel size of 2.5- μm isotropic. The focus of the light beam was located at roughly 150 μm below the surface. Motorized *xyz* stages were integrated under the sample arm to automatically translate the sample between image tiles. The $12 \times 12 \text{ mm}^2$ section of the tissue block was imaged with a total of 576 tiles (24×24) that were acquired with 50% overlap between adjacent tiles to ensure reliable registration and to uniformly reduce speckle noise. Fiji software was used to stitch the image tiles and create the mosaic images.³⁷

To further reduce the effect of noise on fitting depth-profiles and estimating the optical properties, the volumetric images were averaged over 25 μm in *x* and *y* (i.e., 10×10 pixels) and down-sampled in *x* and *y* to a 2.5- μm voxel size. The location of the tissue surface was determined on cross-section images using an image processing edge detector.³⁸ A total depth range of 250 μm was selected for depth-profile fitting, starting from 35 μm below the tissue surface to reduce the impact of specular reflection from the surface. To compare with conventional OCT contrast, the AIP and maximum intensity projection maps within the same depth range were computed as well.

To further validate our parametrized model, the quantified scattering coefficient of the human brain sample was compared with a conventional low-NA measurement with a negligible axial PSF effect where the OCT signal is modeled as

$$R(z) \propto \exp(-2\mu_s z). \quad (11)$$

We imaged the same section of the brain sample with a 2× air objective (Thorlabs) and a z spacer, which yields a lateral resolution of 15 μm and a theoretical Rayleigh range of 280 μm in tissue. The volumetric image tiles have a FOV of $2 \times 2 \text{ mm}^2$ with a $5 \times 5 \times 2.5 \text{ mm}^3$ voxel size and 50% overlap between adjacent tiles. A linear function was fit to the logarithmic OCT signal depth profile and the slope was used to extract the μ_s value.

3 Results

3.1 Optical Properties Characterization of Liquid Phantoms

We first fit our parametrized model to the OCT signals acquired from the eight Intralipid phantoms measured at different focus depths. Example fits of the OCT signal depth dependence are shown in Figs. 1(a) and 1(b). In Fig. 1(a), we show example fits of the OCT signal depth profile of liquid phantoms with different volume concentrations (0.8%, 2.4%, 4%, and 8% v/v Intralipid). In Fig. 1(b), we show example fits of the OCT signal depth profile of 0.8% v/v Intralipid phantoms with different focus depths ($z_f = 150, 250, 350$, and $450 \text{ }\mu\text{m}$). In Fig. 1(c), the estimated scattering coefficient (μ_s) values are presented as a function of the volume concentration of Intralipid measured at two different focus depths (z_f). Error bars indicate the standard deviation of the estimated μ_s across the entire FOV. The error bars are rather small, indicating the high repeatability of the measurements and the coefficient estimation procedure. At high concentration, μ_s tends to be underestimated compared with a prior model of the concentration dependence of μ_s for Intralipid,³² probably due to the impact of multiply scattered photons becoming important for higher scattering coefficients.

In Fig. 1(d), the estimated Rayleigh range z_r is plotted as a function of both the volume concentration of Intralipid and the focus depth. We found that the estimated z_r differed by $< 1 \text{ }\mu\text{m}$ across different focus depths for each volume concentration. This indicated that we could

Fig. 1 Optical properties of Intralipid liquid phantoms. (a) Representative OCT signal depth profiles and model fitted curves of Intralipid phantoms with different volume concentrations (0.8%, 2.4%, 4%, and 8% v/v) at focus depth of 300 μm . (b) Representative OCT signal depth profiles and model fitted curves of 0.8% v/v Intralipid phantom at different focus depths (150, 250, 350, and 450 μm). (c) Estimation of scattering coefficient (μ_s) of Intralipid phantoms with respect to Intralipid volume concentration at two different focus depths. Error bars: standard deviation. The solid line indicates a previous theoretical prediction for the scattering coefficient. (d) Effective Rayleigh range (z_r) as a function of depth of focus (z_f) with respect to Intralipid volume concentration. Deep depths of focus were discarded due to the non-negligible effect of multiple scattering. (e) Shape of z_f of 0.8% v/v Intralipid phantom within a FOV with fixed z_r in model fitting. (f) Shape of z_f of 0.8% v/v Intralipid phantom within a FOV with freely estimated z_r in model fitting.

assume z_r to be constant across the FOV. We next evaluated the spatial dependence of z_f within the FOV both with z_r set as a constant and as a freely fitted parameter that was allowed to vary over the FOV, as shown in Figs. 1(e) and 1(f), respectively. We found that the spatial dependence of z_f was almost identical with z_r set as a constant or as a freely fitted parameter. Moreover, when z_r was set as a freely fitted parameter, the coefficient of correlation for the estimated z_f and z_r was calculated to be -0.043 , indicating a strong independence of these two parameters.

To quantify the improvement in fitting quality using the parametrized model compared with the full fitting model, we calculated the coefficient of variance (CV), which is defined as the standard deviation divided by the mean of extracted μ_s and μ'_b over the imaging field, as shown in Fig. 2. We first estimated both μ_s and μ'_b in an 0.8% volume concentration Intralipid phantom with varying focus depths (z_f), as shown in Figs. 2(a) and 2(b). We found that our parametrized model showed a smaller CV in both μ_s and μ'_b for all focus depths. We then estimated both μ_s and μ'_b in a set of liquid phantoms with different concentrations (0.8%, 1.6%, 2.4%, 3.2%, 4%, and 8%) with a fixed focus depth of 150 μm below the liquid surface, as shown in Figs. 2(c) and 2(d). We found that our parametrized model showed a smaller CV in both μ_s and μ'_b for all scattering levels.

3.2 Optical Properties Characterization of Human Brain

3.2.1 Validation with low-NA model

To validate the quantitative estimates of the optical properties of the brain tissue derived from the proposed parametrized model with the conventional method using a single exponential model, the same section of a human brain sample was imaged with both a high NA (10 $\times$, Mitutoyo) and a low-NA (2 $\times$, Thorlabs) objective. The impact of axial PSF was negligible for the low-NA objective since the Rayleigh range was longer than the depth range analyzed. We first plot the back-scattering (μ'_b) and scattering (μ_s) maps along with AIPs for both the low-NA and high-NA

Fig. 2 Coefficient of variance (CV) of estimated parameters of liquid phantoms shows improvement of fitting quality. (a) CV of estimated μ_s of 0.8% volume concentration Intralipid phantom over imaging field with various focus depths of both parametrized model and full model. (b) CV of estimated μ_b of 0.8% volume concentration Intralipid phantom over imaging field with various focus depths of both parametrized model and full model. (c) CV of estimated μ_s of Intralipid phantoms at 150- μm focus depth over imaging field with various scattering levels of both parametrized model and full model. (d) CV of estimated μ_b of Intralipid phantoms at 150- μm focus depth over imaging field with various scattering levels of both parametrized model and full model.

objectives, as shown in Fig. 3. Moreover, for the high-NA objective, we plot the μ_b' and μ_s maps using both the parametrized model [Fig. 3(b)] and the full model [Fig. 3(c)]. To demonstrate the ability of our parametrized model to quantitatively extract the optical properties of human brain tissue, two regions of interest (ROIs) were manually selected, including a $2 \times 2 \text{ mm}^2$ GM region and a $2 \times 2 \text{ mm}^2$ WM region, as shown in Figs. 3(a) and 3(b). The averaged scattering coefficient (μ_s) within these two ROIs are plotted in Fig. 3(d). The estimated μ_s of WM were $6.43 \pm 0.96 \text{ mm}^{-1}$ and $6.12 \pm 0.99 \text{ mm}^{-1}$ for the low-NA and high-NA objectives, respectively. The estimated μ_s of GM were $3.32 \pm 0.89 \text{ mm}^{-1}$ and $3.58 \pm 0.64 \text{ mm}^{-1}$ for the low-NA and high-NA objectives, respectively. We then performed a pairwise t-test on the extracted μ_s from both ROIs using the full and the parametrized models to determine if there is a significant difference in the mean. We found no significant difference in the mean of the estimated μ_s ($p = 0.49$ for the GM ROI and $p = 0.25$ for the WM ROI). We further performed a pairwise t-test on the extracted μ_s from both ROIs for the low-NA and high-NA objectives to determine if there is a significant difference in the mean. We found no significant difference in the mean of the estimated μ_s ($p = 0.26$ for the GM ROI and $p = 0.09$ for the WM ROI). Although the scattering coefficient is comparable for the two objectives, as expected, we see that structures such as vessels and axonal fiber tracts are more detailed in the high-NA images. As shown in Figs. 3(b) and 3(c), the full four-parameter model overestimated the scattering coefficient of GM and had higher standard deviation within both ROIs compared to the parametrized and the low-NA exponential model. Recall that we take the low-NA measurement as the ground truth result since the model fitting is a simple exponential with no dependence on the confocal parameters.

Fig. 3 Comparing the scattering coefficient map of primary motor cortex imaged by a low-NA 2× and higher-NA 10× objective. (a) The relative back-scattering (left), scattering coefficient map (middle), and AIP (right) of the brain section with a 2× objective. Some but not all vessels and fiber tracts are labeled with blue arrows. Scale bar: 1 mm. (b) The relative back-scattering (left), scattering coefficient map (middle), and AIP (right) of the brain section with a 10× objective using the parametrized model. Some but not all vessels and fiber tracts are labeled with blue arrows. Scale bar: 1 mm. (c) The relative back-scattering (left) and scattering map (middle) of the brain section with a 10× objective using the full model. Some but not all vessels and fiber tracts are labeled with blue arrows. Scale bar: 1 mm. (d) Averaged μ_s within selected ROIs for both high NA and low-NA objectives. Both parametrized model and full model were used to extract μ_s under high-NA objective.

3.2.2 Comparison with the full fitting model

We compare the results from our parametrized model with results from the previously established full four-parameter estimation model by imaging the same section of a human motor cortex sample. The same depth range data was used to fit both the full four-parameter model and our proposed parametrized model. To show the improvement for our parametrized model, we generated the CV images for both the back-scattering (μ_b') and scattering (μ_s) maps for both the parametrized model and the full model using a 5×5 pixel sliding window, as shown in Figs. 4(a) and 4(b). The value of each pixel in the CV images is defined as the standard deviation of values over the sliding window divided by the mean of values over the sliding window. We observed from the CV images that our proposed parametrized model showed a significantly smaller CV compared to the full model. The histograms of the CV values are shown in Fig. 4(d). The median of the CV for μ_s was 0.28 for the full model and was reduced to 0.13 for the parametrized model. Similarly, the median CV of μ_b' was 0.57 for the full model and was reduced to 0.26 for the parametrized model. This confirms that our proposed

Fig. 4 Comparison between the full model and the proposed parametrized model. (a) CV image of the extracted μ_s using the parametrized model. (b) CV image of the extracted μ_s using the full model. (c) CV image of the extracted μ_b using the parametrized model. (d) CV image of the extracted μ_b using the full model. (e) Histograms of CV for the estimated optical properties. Left: CV of the extracted μ_s using both the parametrized and full models. Right: CV of the extracted μ_b using both the parametrized and full models.

parametrized model is effective at reducing inter-parameter covariation and can provide more precise estimates of the optical properties while maintaining accuracy.

3.2.3 Optical properties of human brain before and after index matching

A section of a superior frontal cortex brain sample was first surface flattened using a vibratome and imaged with the high-NA objective (10 $\times$, Mitutoyo). Then the sample underwent incremental index matching with 20%, 40%, and 60% v/v TDE/PBS. The tissue section was imaged with the high-NA objective at each TDE concentration after it reached equilibrium. Due to nonuniform distortion introduced by index matching, the surface of the tissue block was reflattened at each TDE volume concentration before imaging with the high-NA objective. As shown in Figs. 5(a)–5(d) and 5(j), the AIP indicates a reduced GM/WM contrast after index matching. The GM/WM contrast is defined as the mean value of the WM intensity divided by the mean value of the GM intensity. Prior to index matching, the averaged intensity of WM over the depth range is smaller compared to GM due to the higher scattering coefficient of WM. After index matching, we see that the WM AIP becomes brighter as the ratio between WM and GM intensity reached 0.95 at 60% v/v TDE/PBS equilibrium, suggesting that the index

Fig. 5 Estimating optical property change during index matching. (a) AIP of the brain section before index matching. (b) AIP of the brain section after equilibrium in 20% v/v TDE/PBS. (c) AIP of the brain section after equilibrium in 40% v/v TDE/PBS. (d) AIP of the brain section after equilibrium in 60% v/v TDE/PBS. (e) Scattering map of the brain section before index matching. (f) Scattering map of the brain section after equilibrium in 20% v/v TDE/PBS. (g) Scattering map of the brain section after equilibrium in 40% v/v TDE/PBS. (h) Scattering map of the brain section after equilibrium in 60% v/v TDE/PBS. (i) Contrast between WM and GM during each TDE concentration. (j) Extracted scattering coefficient from WM and GM ROI during each TDE concentration. Error bar: standard deviation.

matching process reduced the scattering coefficient of WM more than GM and resulted in a more homogenous scattering level across the sample, as indicated in the extracted μ_s map, as shown in Figs. 5 (e)–5(h).

Two $3.75 \times 3.75 \text{ mm}^2$ (150×150 pixel) ROIs were selected, one from GM and one from WM, as shown in Figs. 5(e)–5(h). The average μ_s of each ROI was calculated to quantify the scattering coefficient change with index matching at each TDE concentration, as shown in Fig. 5(k). We found that after index matching to 60% v/v TDE/PBS, μ_s of GM decreased from $2.29 \pm 0.79 \text{ mm}^{-1}$ to $0.45 \pm 0.27 \text{ mm}^{-1}$ and that μ_s of WM decreased from $5.61 \pm 0.86 \text{ mm}^{-1}$ to $0.77 \pm 0.65 \text{ mm}^{-1}$. We found that the scattering coefficient of WM decreased to $3.06 \pm 0.73 \text{ mm}^{-1}$, $1.97 \pm 0.89 \text{ mm}^{-1}$, and $0.77 \pm 0.65 \text{ mm}^{-1}$ at 20%, 40%, and 60% v/v TDE/PBS equilibrium, respectively, whereas the scattering coefficient of GM decreased to $0.25 \pm 0.17 \text{ mm}^{-1}$ and $0.57 \pm 0.37 \text{ mm}^{-1}$ at 20% and 40% v/v TDE/PBS equilibrium, respectively, and then reached $0.45 \pm 0.27 \text{ mm}^{-1}$ at 60% v/v TDE/PBS equilibrium. The different behavior of GM and WM under different TDE concentration index matching is a result of the expected lower average index of refraction for GM relative to WM.³⁹ To better justify the use of the parametrized model, we also fitted the scattering coefficient at each TDE concentration with the full model, as shown in Fig. 5(k). We found that, in general, both models agreed with each other on the extracted scattering coefficient whereas the standard deviation of estimated μ_s was smaller for the parametrized model at each TDE concentration.

4 Discussion

Previous studies investigating the optical properties of biological tissues with OCT primarily adopted a simple exponential model based on Beer's law.⁴⁰ For high-resolution OCT, models including the impact of confocal parameters have been proposed, but with high inter-parameter correlation, which resulted in large variance in the extracted coefficients.² Here, we extended this prior model by developing procedures to spatially parametrize the confocal parameters to reduce the degrees of freedom of the fitting model and produce estimated values with less variance compared with the prior full four-parameter model. We have detailed a straight-forward approach of utilizing field curvature correction, estimating the constant Rayleigh range, and parameterizing the focus depth over the FOV to achieve more precision in the estimated optical properties of the sample while maintaining accuracy. We anticipate this parameterized model being used routinely for estimating tissue optical properties with high-NA OCT measurements.

There are a few remaining limitations in this revised fitting model. The refractive index difference at the liquid-tissue interface was not considered in the axial PSF model. Both Schmitt et al.⁴ and Izatt et al.¹¹ have derived PSF models that include this effect. However, these models would require more parameters and could introduce inter-parameter dependency, which could potentially cause greater bias in the prediction. However, it was reported previously that the estimated μ_s of microsphere suspensions using the Izatt model and our model had no significant difference,² suggesting that this effect is negligible. Further support for this comes from agreement in our estimated values using low-NA measurements, which are not impacted by the index of refraction mismatch, and the high-NA measurements. An additional potential limitation is that we did not consider differences in the index of refraction for different tissue types, which could cause variations in z_f and z_r within the FOV. Our current model assumed a single index of refraction in the tissue and constrained its effect on the axial PSF. Neglecting this effect introduces some uncertainty in the estimate of the optical properties. In our OCT reflectance model, the refractive index (n) plays a role in two ways: the effective Rayleigh range ($z_r = n \cdot z_R$) and the axial step size. Uncertainty in each of these will introduce uncertainty in the estimated scattering coefficient. In our proposed approach, the effective Rayleigh range was estimated directly from the depth profile of the experimental data, and thus would not be affected by uncertainty in the refractive index. But uncertainty in the refractive index will produce uncertainty in the axial step size. To quantify the impact of this uncertainty, we compared the scattering coefficient extracted using the refractive index of WM ($n = 1.45$) and that of water ($n = 1.33$), which are extreme possible values for the refractive index. We selected an ROI of 1 mm² from within the red ROI shown on the human WM sample shown in Fig. 3 and performed coefficient estimation with our parametrized method. We found that the estimated scattering coefficient varied by ~8%, which is slightly <9% change in the refractive index. It makes sense that the uncertainty in the estimated scattering coefficient scales directly with the systematic error in the index of refraction as this changes the axial length scale on which the scattering coefficient is estimated. Overall, this uncertainty is smaller than the standard deviation we observed in selected ROIs, as shown in Fig. 3(d), which varied from 17.9% to 31.1%. A third limitation is that we assumed negligible contributions from multiple scattering of light within the sample. Given that we are constraining our estimates to relatively shallow depth ranges, we are comfortable with this assumption. We do note that the effect of multiple light scattering likely revealed itself in our phantom measurements at high scattering coefficients.

A future goal for advancing this fitting model is to remove the assumption of constant optical properties over the fitting depth range and expand the model to estimate a depth-dependent scattering coefficient. Achieving this would provide 3-D maps of the tissue optical properties with high-NA OCT and greatly improve the impact for varied applications. Given that depth-resolved models have been developed for low-NA OCT and have been implemented in retinal, colon, and bladder imaging,^{41,42} we anticipate future efforts will succeed in doing the same with high NA OCT.

With this new spatially parameterized fitting model, we performed a preliminary analysis of the impact of a tissue index matching procedure on the quantitative scattering coefficient of the tissue. Our preliminary result quantified the reduction in scattering in GM and WM caused by

increasing the concentration of the index matching agent TDE. It was previously reported that index matching with TDE increased the transmittance in fixed brain slices in a concentration-dependent manner, leading to significant enhancement of the penetration depth.⁴³ We found that 60% TDE increased the scattering length in the GM/WM by 370.6%/294.4% compared to no index matching. In the case of no index matching, we note that our scattering coefficient was less than previously reported,² but this can be explained by our washing samples with PBS versus the prior work where the tissue was emersed in fixative, which is known to enhance optical scattering.⁴⁴ Interestingly, we observed that a minimum in the GM scattering coefficient was obtained with 20% to 40% TDE and that GM scattering increased at 60% TDE while WM scattering continued to decrease. This is consistent with reports that the GM index of refraction is less than that of WM.³⁹ Specifically, the index of refraction for GM has been reported to be 1.367, which matches a TDE concentration of about 30% while the index of refraction for WM has been reported to be 1.467, which matches a TDE concentration of about 65%.²¹ Future work performing a more in-depth analysis expanding on our preliminary result will help identify the optimal concentration of the index matching agent to use for maximizing image features and contrast of interest.

Disclosures

The authors have nothing to disclose.

Acknowledgments

This work was supported by NIH (Grant No. U01MH117023).

Code, Data, and Materials Availability

The code associated with the manuscript is available and will be uploaded to Github repository. The data and materials associated with the manuscript are also available per request.

References

1. J. M. Schmitt, "Optical coherence tomography (OCT): a review," *IEEE J. Sel. Top. Quantum Electron.* **5**(4), 1205–1215 (1999).
2. H. Wang et al., "Characterizing the optical properties of human brain tissue with high numerical aperture optical coherence tomography," *Biomed. Opt. Express* **8**(12), 5617–5636 (2017).
3. C. Xu et al., "Characterization of atherosclerosis plaques by measuring both backscattering and attenuation coefficients in optical coherence tomography," *J. Biomed. Opt.* **13**(3), 034003 (2008).
4. J. M. Schmitt et al., "Optical characterization of dense tissues using low-coherence interferometry," *Proc. SPIE* **1889**, 197–211 (1993).
5. A. J. Coleman et al., "Histological correlates of optical coherence tomography in non-melanoma skin cancer," *Skin Res. Technol.* **19**(1), e10–e19 (2013).
6. E. C. Cauberg et al., "Quantitative measurement of attenuation coefficients of bladder biopsies using optical coherence tomography for grading urothelial carcinoma of the bladder," *J. Biomed. Opt.* **15**(6), 066013 (2010).
7. C. Kut et al., "Detection of human brain cancer infiltration *ex vivo* and *in vivo* using quantitative optical coherence tomography," *Sci. Transl. Med.* **7**(292), 292ra100 (2015).
8. K. V. Larin et al., "Noninvasive blood glucose monitoring with optical coherence tomography: a pilot study in human subjects," *Diabetes Care* **25**(12), 2263–2267 (2002).
9. F. J. van der Meer et al., "Localized measurement of optical attenuation coefficients of atherosclerotic plaque constituents by quantitative optical coherence tomography," *IEEE Trans. Med. Imaging* **24**(10), 1369–1376 (2005).

10. Y. Yang et al., "Optical scattering coefficient estimated by optical coherence tomography correlates with collagen content in ovarian tissue," *J. Biomed. Opt.* **16**(9), 090504 (2011).
11. J. A. Izatt et al., "Optical coherence microscopy in scattering media," *Opt. Lett.* **19**(8), 590–592 (1994).
12. C. Magnain et al., "Blockface histology with optical coherence tomography: a comparison with Nissl staining," *NeuroImage* **84**, 524–533 (2014).
13. D. Kobat et al., "Deep tissue multiphoton microscopy using longer wavelength excitation," *Opt. Express* **17**(16), 13354–13364 (2009).
14. I. Costantini et al., "In-vivo and ex-vivo optical clearing methods for biological tissues," *Biomed. Opt. Express* **10**(10), 5251–5267 (2019).
15. H. Hama et al., "Scale: a chemical approach for fluorescence imaging and reconstruction of transparent mouse brain," *Nat. Neurosci.* **14**(11), 1481 (2011).
16. M.-T. Ke, S. Fujimoto, and T. Imai, "SeeDB: a simple and morphology-preserving optical clearing agent for neuronal circuit reconstruction," *Nat. Neurosci.* **16**(8), 1154 (2013).
17. T. Kuwajima et al., "ClearT: a detergent-and solvent-free clearing method for neuronal and non-neuronal tissue," *Development* **140**(6), 1364–1368 (2013).
18. E. A. Susaki et al., "Whole-brain imaging with single-cell resolution using chemical cocktails and computational analysis," *Cell* **157**(3), 726–739 (2014).
19. B. Yang et al., "Single-cell phenotyping within transparent intact tissue through whole-body clearing," *Cell* **158**(4), 945–958 (2014).
20. A. Ertürk et al., "Three-dimensional imaging of solvent-cleared organs using 3DISCO," *Nat. Protoc.* **7**(11), 1983 (2012).
21. T. Staudt et al., "2, 2'-thiodiethanol: a new water soluble mounting medium for high resolution optical microscopy," *Microsc. Res. Tech.* **70**(1), 1–9 (2007).
22. I. Costantini et al., "A versatile clearing agent for multi-modal brain imaging," *Sci. Rep.* **5**, 9808 (2015).
23. K. K. Bizheva, A. M. Siegel, and D. A. Boas, "Path-length-resolved dynamic light scattering in highly scattering random media: the transition to diffusing wave spectroscopy," *Phys. Rev. E* **58**(6), 7664 (1998).
24. T. G. van Leeuwen, D. J. Faber, and M. C. Aalders, "Measurement of the axial point spread function in scattering media using single-mode fiber-based optical coherence tomography," *IEEE J. Sel. Top. Quantum Electron.* **9**(2), 227–233 (2003).
25. P. Theer and W. Denk, "On the fundamental imaging-depth limit in two-photon microscopy," *J. Opt. Soc. Am. A* **23**(12), 3139–3149 (2006).
26. B. W. Graf, S. G. Adie, and S. A. Boppart, "Correction of coherence gate curvature in high numerical aperture optical coherence imaging," *Opt. Lett.* **35**(18), 3120–3122 (2010).
27. S. Yun et al., "High-speed spectral-domain optical coherence tomography at 1.3 μm wavelength," *Opt. Express* **11**(26), 3598–3604 (2003).
28. T. F. Coleman and Y. Li, "An interior trust region approach for nonlinear minimization subject to bounds," *SIAM J. Optim.* **6**(2), 418–445 (1996).
29. H. J. Van Staveren et al., "Light scattering in Intralipid-10% in the wavelength range of 400–1100 nm," *Appl. Opt.* **30**(31), 4507–4514 (1991).
30. X. Wen et al., "Controlling the scattering of intralipid by using optical clearing agents," *Phys. Med. Biol.* **54**(22), 6917 (2009).
31. G. Zaccanti, S. Del Bianco, and F. Martelli, "Measurements of optical properties of high-density media," *Appl. Opt.* **42**(19), 4023–4030 (2003).
32. R. Michels, F. Foschum, and A. Kienle, "Optical properties of fat emulsions," *Opt. Express* **16**(8), 5907–5925 (2008).
33. B. Aernouts et al., "Dependent scattering in Intralipid® phantoms in the 600–1850 nm range," *Opt. Express* **22**(5), 6086–6098 (2014).
34. B. Aernouts et al., "Supercontinuum laser based optical characterization of Intralipid® phantoms in the 500–2250 nm range," *Opt. Express* **21**(26), 32450–32467 (2013).
35. J. F. De Boer et al., "Improved signal-to-noise ratio in spectral-domain compared with time-domain optical coherence tomography," *Opt. Lett.* **28**(21), 2067–2069 (2003).
36. T. Ragan et al., "Serial two-photon tomography for automated *ex vivo* mouse brain imaging," *Nat. Methods* **9**(3), 255 (2012).

37. S. Preibisch, S. Saalfeld, and P. Tomancak, "Globally optimal stitching of tiled 3D microscopic image acquisitions," *Bioinformatics* **25**(11), 1463–1465 (2009).
38. J. Canny, "A computational approach to edge detection," *IEEE Trans. Pattern Anal. Mach. Intell.* **PAMI-8**, 679–698 (1986).
39. R. Bacallao, S. Sohrab, and C. Phillips, "Guiding principles of specimen preservation for confocal fluorescence microscopy," in *Handbook of Biological Confocal Microscopy*, J. B. Pawley, Ed., pp. 368–380, Springer US, Boston, MA (2006).
40. D. Levitz et al., "Determination of optical scattering properties of highly-scattering media in optical coherence tomography images," *Opt. Express* **12**(2), 249–259 (2004).
41. K. Vermeer et al., "Depth-resolved model-based reconstruction of attenuation coefficients in optical coherence tomography," *Biomed. Opt. Express* **5**(1), 322–337 (2014).
42. G. T. Smith et al., "Automated, depth-resolved estimation of the attenuation coefficient from optical coherence tomography data," *IEEE Trans. Med. Imaging* **34**(12), 2592–2602 (2015).
43. Y. Aoyagi et al., "A rapid optical clearing protocol using 2, 2'-thiodiethanol for microscopic observation of fixed mouse brain," *PLoS One* **10**(1), e0116280 (2015).
44. S. Anand et al., "Effects of formalin fixation on tissue optical properties of in-vitro brain samples," *Proc. SPIE* **9321**, 93210Z (2015).

Jiarui Yang is a PhD candidate in the Biomedical Engineering Department at Boston University. He received his BS degree in control science and engineering from Zhejiang University in June 2016. A PhD candidate in the BOAS lab, he will be applying serial sectioning optical coherence tomography to characterize postmortem human brain.

Kivilem Kılıç received her medical degree and PhD in neuroscience from Hacettepe University, Ankara, Turkey. She worked as a postdoctoral fellow at the University of California San Diego (UCSD) Neuroscience Department working with Anna Devor on studies of neurovascular coupling using two photon microscopy, optogenetics, electrophysiology, and awake animal models. She then took a senior research scientist position at Boston University (BU) Biomedical Engineering Department after serving as the lab manager for the Neurophotonics Center.

Smrithi Sunil is a PhD candidate in the Biomedical Engineering Department at Boston University. She received her BS degree in biomedical engineering with a minor in electrical engineering from Case Western Reserve University in May 2015. She is a PhD candidate in the BOAS lab, she will be applying various optical techniques to characterize changes in cerebral vascular structure, hemodynamics, and neurovascular coupling after stroke. The goal of her project is to study functional recovery after stroke.

Shih-Chi Chen is a professor in the Department of Mechanical and Automation Engineering, Chinese University of Hong Kong. He received his BS degree in mechanical engineering from the National Tsing Hua University, Taiwan, in 1999 and his SM and PhD degrees in mechanical engineering from the Massachusetts Institute of Technology, Cambridge, in 2003 and 2007, respectively. His research interests include biomedical optics and precision engineering. He received two R&D 100 Awards in 2003 and 2018, respectively.

Irene Costantini obtained her International Doctorate in Atomic and Molecular Photonics at the European Laboratory for Non-Linear Spectroscopy, University of Florence, in 2016. She then continued her work as a postdoc at the Physics Department of the University of Florence in Francesco S. Pavone's Biophysics Lab. In 2018 she obtained a research fellowship at the National Institute of Optics, National Research Council, and in 2020 a position at the Department of Biology at the University of Florence.

Francesco Pavone is developing new microscopy techniques for high resolution and high sensitivity imaging, and for laser manipulation purposes, mainly applied in the field of tissue imaging, where nonlinear optical techniques have been applied for tissue pathology detection or for neural or cardiac imaging. In this area, many contributions have been realized with new techniques based on imaging and spectroscopic content to detect neural and cardiac activity together with whole organ imaging to connect structure and functionality.

Bruce Fischl received his BA degree in mathematics at Wesleyan University, then worked as a software engineer working on neural networks. This inspired him to join Boston University's doctoral program in Cognitive and Neural Systems. At Boston University, he worked on neural networks, diffusion image enhancement and nonlocal filtering. He is currently a professor at MGH's Martinos Center where he started as a postdoc, with Anders Dale and Marty Sereno, who jointly developed the neuroimaging package FreeSurfer.

David Boas is the director of the Neurophotonics Center and a professor of Biomedical Engineering at Boston University. He received his BS degree in physics at Rensselaer Polytechnic Institute and PhD in physics at the University of Pennsylvania. He was awarded the Britton Chance Award in Biomedical Optics in 2016 for his development of several novel, high-impact biomedical optical technologies in the neurosciences, as well as following through with impactful application studies, and fostering the widespread adoption of these technologies.

Biographies of the other authors are not available.